

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: UNIO02

Lab Code: CHEM

SAS No.: P4663

SDG NO.: P4663

Lab File ID: BE101388.D

DFTPP Injection Date: 10/28/2024

Instrument ID: BNA_E

DFTPP Injection Time: 10:51

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	17.8
68	Less than 2.0% of mass 69	0.3 (1.5) 1
69	Mass 69 relative abundance	18.3
70	Less than 2.0% of mass 69	0.0 (0.1) 1
127	10.0 - 80.0% of mass 198	27.1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	4
275	10.0 - 60.0% of mass 198	20
365	Greater than 1% of mass 198	3.2
441	Present, but less than mass 443	16.5
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	20.1 (20.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BE101389.D	10/28/2024	11:44
SSTDICC005	SSTDICC005	BE101390.D	10/28/2024	12:23
SSTDICC010	SSTDICC010	BE101391.D	10/28/2024	12:59
SSTDICC020	SSTDICC020	BE101392.D	10/28/2024	13:35
SSTDICCC040	SSTDICCC040	BE101393.D	10/28/2024	14:11
SSTDICC050	SSTDICC050	BE101394.D	10/28/2024	14:49
SSTDICC060	SSTDICC060	BE101395.D	10/28/2024	15:25
SSTDICC080	SSTDICC080	BE101396.D	10/28/2024	16:01