

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: UNIO02

Lab Code: CHEM

SAS No.: P4663 SDG NO.: P4663

Lab File ID: BF140199.D

DFTPP Injection Date: 11/01/2024

Instrument ID: BNA_F

DFTPP Injection Time: 08:56

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	42.9
68	Less than 2.0% of mass 69	0.7 (1.7) 1
69	Mass 69 relative abundance	37.8
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	47
197	Less than 2.0% of mass 198	0.7
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
275	10.0 - 60.0% of mass 198	28.1
365	Greater than 1% of mass 198	3.5
441	Present, but less than mass 443	15.6
442	Greater than 50% of mass 198	98.8
443	15.0 - 24.0% of mass 442	19.3 (19.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF140200.D	11/01/2024	09:22
RES-BUILDING-PAD-NO-2	P4663-08	BF140208.D	11/01/2024	14:00
RES-BUILDING-PAD-NO-2MS	P4663-08MS	BF140209.D	11/01/2024	14:27
RES-BUILDING-PAD-NO-2MSD	P4663-08MSD	BF140210.D	11/01/2024	14:53
RES-BUILDING-PAD-NO-11	P4663-04	BF140212.D	11/01/2024	15:47