

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: UNIO02
 Lab Code: CHEM Case No.: P4663 SAS No.: P4663 SDG NO.: P4663
 Lab File ID: VY020109.D Date Analyzed: 11/01/2024
 Instrument ID: MSVOA_Y Time Analyzed: 09:02
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	249316	7.71	428486	8.62	369253	11.42
UPPER LIMIT	498632	8.213	856972	9.115	738506	11.92
LOWER LIMIT	124658	7.213	214243	8.115	184627	10.92
EPA SAMPLE NO.						
TENNANT-PAD-2-3-STOCKPILE	159567	7.71	338625	8.62	324781	11.41
RES-BUILDING-PAD-NO-11	164371	7.71	344945	8.61	309515	11.41
RES-BUILDING-PAD-NO-3	159663	7.71	335103	8.62	302177	11.41
RES-BUILDING-PAD-NO-2	168827	7.71	353605	8.62	335031	11.41
VY1101SBL01	193272	7.71	403931	8.62	365454	11.41
VY1101SBS01	224649	7.71	413315	8.62	362345	11.41
VY1101SBSD01	251864	7.71	454799	8.62	392733	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.