

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: UNIO02
 Lab Code: CHEM Case No.: P4663 SAS No.: P4663 SDG NO.: P4663
 Lab File ID: VY020109.D Date Analyzed: 11/01/2024
 Instrument ID: MSVOA_Y Time Analyzed: 09:02
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	173771	13.346				
UPPER LIMIT	347542	13.846				
LOWER LIMIT	86885.5	12.846				
EPA SAMPLE NO.						
TENNANT-PAD-2-3-STOCKPILE	118518	13.35				
RES-BUILDING-PAD-NO-11	105144	13.35				
RES-BUILDING-PAD-NO-3	103950	13.35				
RES-BUILDING-PAD-NO-2	125105	13.35				
VY1101SBL01	125829	13.35				
VY1101SBS01	169895	13.35				
VY1101SBSD01	182188	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.