ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092 NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

## GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEM	TECH PROJECT NUMBER: P4665 MA	TRIX: Water			
METH	OD: 8260D				
			NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified. (Field samples and	Method Blanks)	NA	NO	TES ✓
2.	GC/MS Tuning Specifications BFB Meet Criteria (NOTE THAT THERE ARE DIFFERENT CRIT ASP CLP, CLP AND NJ)	ERIA FOR NY			√
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series 8000 Series.	and 12 hours for			✓
4.	GC/MS Calibration - Initial Calibration performed before sample and continuing calibration performed within 24 hours of sample analysis 12 hours for 8000 series.				✓
5.	GC/MS Calibration Requirements.				$\checkmark$
	The %RSD is greater than 20% in the Initial Calibration method (821 for Acetone, Chloroethane, Chloromethane these compounds are p Linear Regression while, 1,4-Dichlorobenzene this compound is pass Quadratic Regression.	bassing on			
	The Continuous Calibration met the requirements .				
6.	Blank Contamination - If yes, list compounds and concentrations in e	each blank:		$\checkmark$	
7.	Surrogate Recoveries Meet Criteria				$\checkmark$
	If not met, list those compounds and their recoveries which fall outsid ranges.	de the acceptable			
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria				$\checkmark$
	If not met, list those compounds and their recoveries which fall outside range.	de the acceptable			
	The Blank Spike met requirements for all samples .				
	The Blank Spike Duplicate met requirements for all samples .				
9.	Internal Standard Area/Retention Time Shift Meet Criteria				$\checkmark$
	Comments:				
10.	Analysis Holding Time Met				$\checkmark$
	If not met, list number of days exceeded for each sample:				

## ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092 NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

## GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

NA NO YES

## ADDITIONAL COMMENTS:

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is)." Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

**QA REVIEW** 

Date