

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110524\
 Data File : BN034867.D
 Acq On : 05 Nov 2024 12:46
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 05 13:57:40 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN103024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 05 00:03:28 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2	1,4-Dioxane	0.439	0.427	2.7	111	0.00
3	n-Nitrosodimethylamine	0.494	0.619	-25.3#	150	0.00
4 S	2-Fluorophenol	1.212	0.999	17.6	96	0.00
5 S	Phenol-d6	1.574	1.370	13.0	103	0.00
6	bis(2-Chloroethyl)ether	1.149	1.170	-1.8	117	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	109	0.00
8 S	Nitrobenzene-d5	0.314	0.275	12.4	108	-0.01
9	Naphthalene	1.106	1.003	9.3	110	0.00
10	Hexachlorobutadiene	0.182	0.161	11.5	106	0.00
11 SURR	2-Methylnaphthalene-d10	0.579	0.483	16.6	102	0.00
12	2-Methylnaphthalene	0.723	0.606	16.2	102	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.00
14 S	2,4,6-Tribromophenol	0.198	0.099	50.0#	60	0.00
15 S	2-Fluorobiphenyl	1.575	1.390	11.7	98	0.00
16	Acenaphthylene	2.007	1.573	21.6	90	0.00
17	Acenaphthene	1.337	1.110	17.0	94	0.00
18	Fluorene	1.667	1.401	16.0	93	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.00
20	4,6-Dinitro-2-methylphenol	0.059	0.031	47.5#	66	0.00
21	4-Bromophenyl-phenylether	0.231	0.182	21.2	81	0.00
22	Hexachlorobenzene	0.259	0.230	11.2	90	0.00
23	Atrazine	0.199	0.128	35.7#	67	0.00
24	Pentachlorophenol	0.112	0.063	43.8#	65	0.00
25	Phenanthrene	1.176	1.092	7.1	95	0.00
26	Anthracene	1.101	0.908	17.5	86	0.00
27 SURR	Fluoranthene-d10	0.938	0.785	16.3	86	0.00
28	Fluoranthene	1.297	1.130	12.9	90	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	91	0.00
30	Pyrene	2.091	1.921	8.1	91	0.00
31 S	Terphenyl-d14	0.859	0.680	20.8	78	0.00
32	Benzo(a)anthracene	1.592	1.365	14.3	88	0.00
33	Chrysene	1.559	1.480	5.1	96	0.00
34	Bis(2-ethylhexyl)phthalate	1.331	0.881	33.8#	62	0.00
35 I	Perylene-d12	1.000	1.000	0.0	83	0.00
36	Indeno(1,2,3-cd)pyrene	1.471	1.414	3.9	88	0.00
37	Benzo(b)fluoranthene	1.610	1.594	1.0	91	0.00
38	Benzo(k)fluoranthene	1.571	1.620	-3.1	97	0.00
39 C	Benzo(a)pyrene	1.329	1.161	12.6	82	0.00
40	Dibenzo(a,h)anthracene	1.147	1.080	5.8	87	0.00
41	Benzo(g,h,i)perylene	1.240	1.165	6.0	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4