


Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN110424\  
Data File : BN034849.D  
Acq On : 04 Nov 2024 20:54  
Operator : RC/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270-SIM-BN103024.M  
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
Last Update : Tue Nov 05 00:03:28 2024






AutoFind: Scans 2474, 2475, 2476; Background Corrected with Scan 2467

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 53.1         | 129505     | PASS                |
| 68             | 69              | 0.00            | 2               | 1.6          | 1938       | PASS                |
| 69             | 198             | 0.00            | 100             | 49.0         | 119523     | PASS                |
| 70             | 69              | 0.00            | 2               | 0.6          | 686        | PASS                |
| 127            | 198             | 10              | 80              | 61.5         | 149968     | PASS                |
| 197            | 198             | 0.00            | 2               | 0.6          | 1346       | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 244032     | PASS                |
| 199            | 198             | 5               | 9               | 7.1          | 17358      | PASS                |
| 275            | 198             | 10              | 60              | 24.1         | 58707      | PASS                |
| 365            | 198             | 1               | 100             | 3.0          | 7442       | PASS                |
| 441            | 198             | 0.01            | 100             | 8.8          | 21576      | PASS                |
| 442            | 442             | 50              | 100             | 100.0        | 142067     | PASS                |
| 443            | 442             | 15              | 24              | 19.1         | 27149      | PASS                |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN110424\  
Data File : BN034849.D  
Acq On : 04 Nov 2024 20:54  
Operator : RC/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP

Quant Time: Nov 04 23:56:11 2024  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Nov 04 23:56:03 2024  
Response via : Initial Calibration



TIC: BN034849.D\data.ms

**(70) Pentachlorophenol (C)**

16.622min ( 0.000) 21272.94 ng




response 271619

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 62.20  | 57.32  |
| 264.00 | 61.60  | 58.20  |
| 0.00   | 0.00   | 0.00   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN110424\  
Data File : BN034849.D  
Acq On : 04 Nov 2024 20:54  
Operator : RC/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP

Quant Time: Nov 04 23:56:11 2024  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Nov 04 23:56:03 2024  
Response via : Initial Calibration



TIC: BN034849.D\data.ms

**(77) Benzidine**

19.227min ( 0.000) 0.00 ng

response 1662756

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 15.50  | 14.88  |
| 183.00 | 13.20  | 11.04  |
| 0.00   | 0.00   | 0.00   |

### DDT Breakdown

| Date          | Instrument Name  | DFTPP Data File    |
|---------------|------------------|--------------------|
| 11/4/2024     | BNA_N            | <u>BN034849.D</u>  |
|               |                  |                    |
|               |                  |                    |
| Compound Name | Response         | Retention Time     |
|               |                  |                    |
| DDT           | 864034           | 20.469             |
| DDD           | 21444            | 20.027             |
| DDE           | 934              | 19.522             |
|               |                  |                    |
|               |                  |                    |
| SUM(DDD+DDE)  | SUM(DDT+DDD+DDE) | % Breakdown Of DDT |
|               |                  |                    |
| 22378         | 886412           | 2.52               |
|               |                  |                    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP