

Surrogate Summary

SW-846

SDG No.: P4665

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P4665-02	VPB190-HYD-20241030	2-Methylnaphthalene-d10	0.4	0	0	*	30	150
		Fluoranthene-d10	0.4	0.022	5	*	30	150
		Nitrobenzene-d5	0.4	0.39	97		55	111
		2-Fluorobiphenyl	0.4	0.41	102		53	106
		Terphenyl-d14	0.4	0.47	118		58	132
P4665-02RE	VPB190-HYD-20241030RE	2-Methylnaphthalene-d10	0.4	0	0	*	30	150
		Fluoranthene-d10	0.4	0.023	6	*	30	150
		Nitrobenzene-d5	0.4	0.40	99		55	111
		2-Fluorobiphenyl	0.4	0.41	103		53	106
		Terphenyl-d14	0.4	0.42	105		58	132
P4665-05	BP-VPB-190-GW-558-560	2-Methylnaphthalene-d10	0.4	0.28	69		30	150
		Fluoranthene-d10	0.4	0.34	85		30	150
		Nitrobenzene-d5	0.4	0.28	71		55	111
		2-Fluorobiphenyl	0.4	0.30	74		53	106
		Terphenyl-d14	0.4	0.35	87		58	132
P4665-07	BP-VPB-190-GW-598-600	2-Methylnaphthalene-d10	0.4	0.38	94		30	150
		Fluoranthene-d10	0.4	0.46	114		30	150
		Nitrobenzene-d5	0.4	0.38	95		55	111
		2-Fluorobiphenyl	0.4	0.41	103		53	106
		Terphenyl-d14	0.4	0.52	129		58	132
PB164594BL	PB164594BL	2-Methylnaphthalene-d10	0.4	0.35	88		30	150
		Fluoranthene-d10	0.4	0.36	90		30	150
		Nitrobenzene-d5	0.4	0.38	95		55	111
		2-Fluorobiphenyl	0.4	0.41	102		53	106
		Terphenyl-d14	0.4	0.46	114		58	132
PB164594BS	PB164594BS	2-Methylnaphthalene-d10	0.4	0.49	123		30	150
		Fluoranthene-d10	0.4	0.32	80		30	150
		Nitrobenzene-d5	0.4	0.38	95		55	111
		2-Fluorobiphenyl	0.4	0.42	104		53	106
		Terphenyl-d14	0.4	0.52	129		58	132
PB164594BSD	PB164594BSD	2-Methylnaphthalene-d10	0.4	0.48	121		30	150
		Fluoranthene-d10	0.4	0.32	79		30	150
		Nitrobenzene-d5	0.4	0.39	97		55	111
		2-Fluorobiphenyl	0.4	0.44	109	*	53	106
		Terphenyl-d14	0.4	0.52	131		58	132