

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: P4665 SAS No.: P4665 SDG NO.: P4665
Lab File ID: VN084628.D Date Analyzed: 11/01/2024
Instrument ID: MSVOA_N Time Analyzed: 10:22
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	136355	13.794				
UPPER LIMIT	272710	14.294				
LOWER LIMIT	68177.5	13.294				
EPA SAMPLE NO.						
BP-VPB-190-TB-20241029	117747	13.79				
VPB190-HYD-20241030	113670	13.79				
BP-VPB-190-EB-20241030	115295	13.79				
BP-VPB-190-GW-538-540	124509	13.79				
BP-VPB-190-GW-558-560	121785	13.79				
BP-VPB-190-GW-583-585	124365	13.79				
BP-VPB-190-GW-598-600	115441	13.79				
VN1101WBL01	125664	13.79				
VN1101WBS01	147052	13.79				
VN1101WBSD01	139639	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.