

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB133262

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/03/2024	09:30
2	CAL2	1	Water	NA	NA	20.2	7.00	11/03/2024	09:31
3	CAL3	1	Water	NA	NA	20.2	10.02	11/03/2024	09:32
4	ICV	1	Water	NA	NA	20.2	7.01	11/03/2024	09:35
5	CCV1	1	Water	NA	NA	20.2	2.01	11/03/2024	09:40
6	P4660-02	1	Solid	20.02	20	24.2	11.87	11/03/2024	09:47
7	P4660-06	1	Solid	20.04	20	24.3	10.96	11/03/2024	09:52
8	P4660-10	1	Solid	20.03	20	24.3	8.81	11/03/2024	10:00
9	P4672-01	1	Solid	20.02	20	24.7	6.08	11/03/2024	10:10
10	P4672-02	1	Solid	20.03	20	24.6	6.60	11/03/2024	10:15
11	P4672-03	1	Solid	20.02	20	24.6	6.79	11/03/2024	10:22
12	P4672-04	1	Solid	20.03	20	24.2	6.73	11/03/2024	10:30
13	P4672-05	1	Solid	20.02	20	24.5	6.53	11/03/2024	10:37
14	P4672-05DUP	1	Solid	20.03	20	24.6	6.54	11/03/2024	10:38
15	CCV2	1	Water	NA	NA	20.2	12.02	11/03/2024	10:40

WORKLIST(Hardcopy Internal Chain)

18133262

WorkList Name : ph p4672 WorkList ID : 185049 Department : Wet-Chemistry Date : 11-03-2024 09:17:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4660-02	WC-TA2-01-C	Solid	pH	Cool 4 deg C	ENTA05	K41	10/30/2024	9045D
P4660-06	WC-WOOD-01-C	Solid	pH	Cool 4 deg C	ENTA05	K41	10/31/2024	9045D
P4660-10	WC-CONCRETE-01-C	Solid	pH	Cool 4 deg C	ENTA05	K41	10/31/2024	9045D
P4672-01	TAPLPR-SED06-103024-00-T2	Solid	pH	Cool 4 deg C	WEST04	K41	10/30/2024	9045D
P4672-02	TAPLPR-SED07-103024-00-T2	Solid	pH	Cool 4 deg C	WEST04	K41	10/30/2024	9045D
P4672-03	TAPLPR-SED03-103024-00-T2	Solid	pH	Cool 4 deg C	WEST04	K41	10/30/2024	9045D
P4672-04	TAPLPR-SED04-103024-00-T2	Solid	pH	Cool 4 deg C	WEST04	K41	10/30/2024	9045D
P4672-05	TAPLPR-SED05-103024-00-T2	Solid	pH	Cool 4 deg C	WEST04	K41	10/30/2024	9045D

Date/Time 11/03/24 09:25
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 11/03/24 09:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]