

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: P4672

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

RunNo.: LB133245

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 TOC	mg/L	1010	1000	101	90-110	08/07/2024
Sample ID: CCV1 TOC	mg/L	1030	1000	103	90-110	11/01/2024
Sample ID: CCV2 TOC	mg/L	978	1000	98	90-110	11/01/2024
Sample ID: CCV3 TOC	mg/L	1020	1000	102	90-110	11/01/2024

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: P4672

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

RunNo.: LB133262

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.01	7	100	90-110	11/03/2024
Sample ID:	CCV1						
pH		pH	2.01	2.00	101	90-110	11/03/2024
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	11/03/2024