

Duplicate Sample Summary

Client:	Weston Solutions	SDG No.:	P4672
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Sample ID:	P4653-02
Client ID:	TAPLPR-SED09-103024-00-T2MSD	Percent Solids for Spike Sample:	81.2

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TOC	mg/Kg	+/-20	2470		2510		1	2		11/01/2024

Duplicate Sample Summary

Client:	Weston Solutions	SDG No.:	P4672
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Sample ID:	P4672-05
Client ID:	TAPLPR-SED05-103124-00-T2DUP	Percent Solids for Spike Sample:	83.9

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
pH	pH	+/-20	6.53		6.54		1	0.15		11/03/2024