



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	RSB Environmental				SDG No.:	P4674			
Contract:	RSBE01	Lab Code:	CHEM		Case No.:	P4674	SAS No.:	P4674	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB69	Mercury	0.20	+/-0.20	U	0.20	CV	11/05/2024	12:00	LB133297

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>RSB Environmental</u>		SDG No.: <u>P4674</u>							
Contract: <u>RSBE01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4674</u>	SAS No.: <u>P4674</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB26	Mercury	0.20	+/-0.20	U	0.20	CV	11/05/2024	12:05	LB133297
CCB27	Mercury	0.20	+/-0.20	U	0.20	CV	11/05/2024	12:32	LB133297
CCB28	Mercury	0.20	+/-0.20	U	0.20	CV	11/05/2024	12:59	LB133297
CCB29	Mercury	0.20	+/-0.20	U	0.20	CV	11/05/2024	13:34	LB133297
CCB30	Mercury	0.20	+/-0.20	U	0.20	CV	11/05/2024	14:11	LB133297
CCB31	Mercury	0.20	+/-0.20	U	0.20	CV	11/05/2024	14:31	LB133297
CCB32	Mercury	0.20	+/-0.20	U	0.20	CV	11/05/2024	14:45	LB133297

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>RSB Environmental</u>		SDG No.: <u>P4674</u>							
Contract: <u>RSBE01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4674</u>	SAS No.: <u>P4674</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	11/06/2024	17:43	LB133323
	Arsenic	20.0	+/-20.0	U	20.0	P	11/06/2024	17:43	LB133323
	Barium	100	+/-100	U	100	P	11/06/2024	17:43	LB133323
	Cadmium	6.00	+/-6.00	U	6.00	P	11/06/2024	17:43	LB133323
	Chromium	10.0	+/-10.0	U	10.0	P	11/06/2024	17:43	LB133323
	Copper	20.0	+/-20.0	U	20.0	P	11/06/2024	17:43	LB133323
	Iron	100	+/-100	U	100	P	11/06/2024	17:43	LB133323
	Lead	12.0	+/-12.0	U	12.0	P	11/06/2024	17:43	LB133323
	Manganese	20.0	+/-20.0	U	20.0	P	11/06/2024	17:43	LB133323
	Nickel	40.0	+/-40.0	U	40.0	P	11/06/2024	17:43	LB133323
	Selenium	20.0	+/-20.0	U	20.0	P	11/06/2024	17:43	LB133323
	Silver	10.0	+/-10.0	U	10.0	P	11/06/2024	17:43	LB133323
	Zinc	40.0	+/-40.0	U	40.0	P	11/06/2024	17:43	LB133323

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>RSB Environmental</u>		SDG No.: <u>P4674</u>							
Contract: <u>RSBE01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4674</u>	SAS No.: <u>P4674</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	11/06/2024	18:27	LB133323
	Arsenic	20.0	+/-20.0	U	20.0	P	11/06/2024	18:27	LB133323
	Barium	100	+/-100	U	100	P	11/06/2024	18:27	LB133323
	Cadmium	0.21	+/-6.00	J	6.00	P	11/06/2024	18:27	LB133323
	Chromium	10.0	+/-10.0	U	10.0	P	11/06/2024	18:27	LB133323
	Copper	20.0	+/-20.0	U	20.0	P	11/06/2024	18:27	LB133323
	Iron	100	+/-100	U	100	P	11/06/2024	18:27	LB133323
	Lead	12.0	+/-12.0	U	12.0	P	11/06/2024	18:27	LB133323
	Manganese	20.0	+/-20.0	U	20.0	P	11/06/2024	18:27	LB133323
	Nickel	40.0	+/-40.0	U	40.0	P	11/06/2024	18:27	LB133323
	Selenium	20.0	+/-20.0	U	20.0	P	11/06/2024	18:27	LB133323
	Silver	10.0	+/-10.0	U	10.0	P	11/06/2024	18:27	LB133323
	Zinc	40.0	+/-40.0	U	40.0	P	11/06/2024	18:27	LB133323
CCB02	Aluminum	100	+/-100	U	100	P	11/06/2024	18:44	LB133323
	Arsenic	20.0	+/-20.0	U	20.0	P	11/06/2024	18:44	LB133323
	Barium	100	+/-100	U	100	P	11/06/2024	18:44	LB133323
	Cadmium	6.00	+/-6.00	U	6.00	P	11/06/2024	18:44	LB133323
	Chromium	10.0	+/-10.0	U	10.0	P	11/06/2024	18:44	LB133323
	Copper	20.0	+/-20.0	U	20.0	P	11/06/2024	18:44	LB133323
	Iron	100	+/-100	U	100	P	11/06/2024	18:44	LB133323
	Lead	12.0	+/-12.0	U	12.0	P	11/06/2024	18:44	LB133323
	Manganese	20.0	+/-20.0	U	20.0	P	11/06/2024	18:44	LB133323
	Nickel	40.0	+/-40.0	U	40.0	P	11/06/2024	18:44	LB133323
	Selenium	20.0	+/-20.0	U	20.0	P	11/06/2024	18:44	LB133323
	Silver	10.0	+/-10.0	U	10.0	P	11/06/2024	18:44	LB133323
	Zinc	40.0	+/-40.0	U	40.0	P	11/06/2024	18:44	LB133323
CCB03	Aluminum	100	+/-100	U	100	P	11/06/2024	19:36	LB133323
	Arsenic	20.0	+/-20.0	U	20.0	P	11/06/2024	19:36	LB133323
	Barium	100	+/-100	U	100	P	11/06/2024	19:36	LB133323
	Cadmium	6.00	+/-6.00	U	6.00	P	11/06/2024	19:36	LB133323
	Chromium	10.0	+/-10.0	U	10.0	P	11/06/2024	19:36	LB133323
	Copper	20.0	+/-20.0	U	20.0	P	11/06/2024	19:36	LB133323
	Iron	100	+/-100	U	100	P	11/06/2024	19:36	LB133323
	Lead	12.0	+/-12.0	U	12.0	P	11/06/2024	19:36	LB133323
	Manganese	20.0	+/-20.0	U	20.0	P	11/06/2024	19:36	LB133323
	Nickel	40.0	+/-40.0	U	40.0	P	11/06/2024	19:36	LB133323
	Selenium	20.0	+/-20.0	U	20.0	P	11/06/2024	19:36	LB133323
	Silver	10.0	+/-10.0	U	10.0	P	11/06/2024	19:36	LB133323
	Zinc	23.7	+/-40.0	J	40.0	P	11/06/2024	19:36	LB133323

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>RSB Environmental</u>		SDG No.: <u>P4674</u>							
Contract: <u>RSBE01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4674</u>	SAS No.: <u>P4674</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Aluminum	100	+/-100	U	100	P	11/06/2024	20:26	LB133323
	Arsenic	20.0	+/-20.0	U	20.0	P	11/06/2024	20:26	LB133323
	Barium	100	+/-100	U	100	P	11/06/2024	20:26	LB133323
	Cadmium	6.00	+/-6.00	U	6.00	P	11/06/2024	20:26	LB133323
	Chromium	10.0	+/-10.0	U	10.0	P	11/06/2024	20:26	LB133323
	Copper	20.0	+/-20.0	U	20.0	P	11/06/2024	20:26	LB133323
	Iron	100	+/-100	U	100	P	11/06/2024	20:26	LB133323
	Lead	12.0	+/-12.0	U	12.0	P	11/06/2024	20:26	LB133323
	Manganese	20.0	+/-20.0	U	20.0	P	11/06/2024	20:26	LB133323
	Nickel	40.0	+/-40.0	U	40.0	P	11/06/2024	20:26	LB133323
	Selenium	20.0	+/-20.0	U	20.0	P	11/06/2024	20:26	LB133323
	Silver	10.0	+/-10.0	U	10.0	P	11/06/2024	20:26	LB133323
	Zinc	40.0	+/-40.0	U	40.0	P	11/06/2024	20:26	LB133323
CCB05	Aluminum	100	+/-100	U	100	P	11/06/2024	21:17	LB133323
	Arsenic	20.0	+/-20.0	U	20.0	P	11/06/2024	21:17	LB133323
	Barium	100	+/-100	U	100	P	11/06/2024	21:17	LB133323
	Cadmium	6.00	+/-6.00	U	6.00	P	11/06/2024	21:17	LB133323
	Chromium	10.0	+/-10.0	U	10.0	P	11/06/2024	21:17	LB133323
	Copper	20.0	+/-20.0	U	20.0	P	11/06/2024	21:17	LB133323
	Iron	100	+/-100	U	100	P	11/06/2024	21:17	LB133323
	Lead	12.0	+/-12.0	U	12.0	P	11/06/2024	21:17	LB133323
	Manganese	20.0	+/-20.0	U	20.0	P	11/06/2024	21:17	LB133323
	Nickel	40.0	+/-40.0	U	40.0	P	11/06/2024	21:17	LB133323
	Selenium	20.0	+/-20.0	U	20.0	P	11/06/2024	21:17	LB133323
	Silver	10.0	+/-10.0	U	10.0	P	11/06/2024	21:17	LB133323
	Zinc	40.0	+/-40.0	U	40.0	P	11/06/2024	21:17	LB133323
CCB06	Aluminum	100	+/-100	U	100	P	11/06/2024	22:15	LB133323
	Arsenic	20.0	+/-20.0	U	20.0	P	11/06/2024	22:15	LB133323
	Barium	100	+/-100	U	100	P	11/06/2024	22:15	LB133323
	Cadmium	6.00	+/-6.00	U	6.00	P	11/06/2024	22:15	LB133323
	Chromium	10.0	+/-10.0	U	10.0	P	11/06/2024	22:15	LB133323
	Copper	20.0	+/-20.0	U	20.0	P	11/06/2024	22:15	LB133323
	Iron	100	+/-100	U	100	P	11/06/2024	22:15	LB133323
	Lead	12.0	+/-12.0	U	12.0	P	11/06/2024	22:15	LB133323
	Manganese	20.0	+/-20.0	U	20.0	P	11/06/2024	22:15	LB133323
	Nickel	40.0	+/-40.0	U	40.0	P	11/06/2024	22:15	LB133323
	Selenium	20.0	+/-20.0	U	20.0	P	11/06/2024	22:15	LB133323
	Silver	10.0	+/-10.0	U	10.0	P	11/06/2024	22:15	LB133323
	Zinc	40.0	+/-40.0	U	40.0	P	11/06/2024	22:15	LB133323

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: RSB Environmental		SDG No.: P4674							
Contract: RSBE01	Lab Code: CHEM	Case No.: P4674	SAS No.: P4674						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB07	Aluminum	100	+/-100	U	100	P	11/06/2024	23:07	LB133323
	Arsenic	20.0	+/-20.0	U	20.0	P	11/06/2024	23:07	LB133323
	Barium	100	+/-100	U	100	P	11/06/2024	23:07	LB133323
	Cadmium	6.00	+/-6.00	U	6.00	P	11/06/2024	23:07	LB133323
	Chromium	10.0	+/-10.0	U	10.0	P	11/06/2024	23:07	LB133323
	Copper	20.0	+/-20.0	U	20.0	P	11/06/2024	23:07	LB133323
	Iron	100	+/-100	U	100	P	11/06/2024	23:07	LB133323
	Lead	12.0	+/-12.0	U	12.0	P	11/06/2024	23:07	LB133323
	Manganese	20.0	+/-20.0	U	20.0	P	11/06/2024	23:07	LB133323
	Nickel	40.0	+/-40.0	U	40.0	P	11/06/2024	23:07	LB133323
	Selenium	20.0	+/-20.0	U	20.0	P	11/06/2024	23:07	LB133323
	Silver	10.0	+/-10.0	U	10.0	P	11/06/2024	23:07	LB133323
	Zinc	40.0	+/-40.0	U	40.0	P	11/06/2024	23:07	LB133323
CCB08	Aluminum	100	+/-100	U	100	P	11/07/2024	00:03	LB133323
	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	00:03	LB133323
	Barium	100	+/-100	U	100	P	11/07/2024	00:03	LB133323
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	00:03	LB133323
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	00:03	LB133323
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	00:03	LB133323
	Iron	100	+/-100	U	100	P	11/07/2024	00:03	LB133323
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	00:03	LB133323
	Manganese	20.0	+/-20.0	U	20.0	P	11/07/2024	00:03	LB133323
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	00:03	LB133323
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	00:03	LB133323
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	00:03	LB133323
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	00:03	LB133323
CCB09	Aluminum	100	+/-100	U	100	P	11/07/2024	00:57	LB133323
	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	00:57	LB133323
	Barium	100	+/-100	U	100	P	11/07/2024	00:57	LB133323
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	00:57	LB133323
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	00:57	LB133323
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	00:57	LB133323
	Iron	100	+/-100	U	100	P	11/07/2024	00:57	LB133323
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	00:57	LB133323
	Manganese	20.0	+/-20.0	U	20.0	P	11/07/2024	00:57	LB133323
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	00:57	LB133323
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	00:57	LB133323
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	00:57	LB133323
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	00:57	LB133323

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>RSB Environmental</u>		SDG No.: <u>P4674</u>							
Contract: <u>RSBE01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4674</u>	SAS No.: <u>P4674</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB10	Aluminum	100	+/-100	U	100	P	11/07/2024	01:53	LB133323
	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	01:53	LB133323
	Barium	100	+/-100	U	100	P	11/07/2024	01:53	LB133323
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	01:53	LB133323
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	01:53	LB133323
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	01:53	LB133323
	Iron	100	+/-100	U	100	P	11/07/2024	01:53	LB133323
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	01:53	LB133323
	Manganese	20.0	+/-20.0	U	20.0	P	11/07/2024	01:53	LB133323
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	01:53	LB133323
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	01:53	LB133323
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	01:53	LB133323
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	01:53	LB133323

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>RSB Environmental</u>		SDG No.: <u>P4674</u>							
Contract: <u>RSBE01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4674</u>	SAS No.: <u>P4674</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	11/08/2024	14:16	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	14:16	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	14:16	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	14:16	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	14:16	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	14:16	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	14:16	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	14:16	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	14:16	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	14:16	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	14:16	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	14:16	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	14:16	LB133365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>RSB Environmental</u>		SDG No.: <u>P4674</u>							
Contract: <u>RSBE01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4674</u>	SAS No.: <u>P4674</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	11/08/2024	14:50	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	14:50	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	14:50	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	14:50	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	14:50	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	14:50	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	14:50	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	14:50	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	14:50	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	14:50	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	14:50	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	14:50	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	14:50	LB133365
CCB02	Aluminum	100	+/-100	U	100	P	11/08/2024	15:33	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	15:33	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	15:33	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	15:33	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	15:33	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	15:33	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	15:33	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	15:33	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	15:33	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	15:33	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	15:33	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	15:33	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	15:33	LB133365
CCB03	Aluminum	100	+/-100	U	100	P	11/08/2024	16:28	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	16:28	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	16:28	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	16:28	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	16:28	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	16:28	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	16:28	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	16:28	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	16:28	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	16:28	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	16:28	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	16:28	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	16:28	LB133365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>RSB Environmental</u>		SDG No.: <u>P4674</u>							
Contract: <u>RSBE01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4674</u>	SAS No.: <u>P4674</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Aluminum	100	+/-100	U	100	P	11/08/2024	17:22	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	17:22	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	17:22	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	17:22	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	17:22	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	17:22	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	17:22	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	17:22	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	17:22	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	17:22	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	17:22	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	17:22	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	17:22	LB133365
CCB05	Aluminum	100	+/-100	U	100	P	11/08/2024	18:08	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	18:08	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	18:08	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	18:08	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	18:08	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	18:08	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	18:08	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	18:08	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	18:08	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	18:08	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	18:08	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	18:08	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	18:08	LB133365
CCB06	Aluminum	100	+/-100	U	100	P	11/08/2024	18:59	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	18:59	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	18:59	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	18:59	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	18:59	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	18:59	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	18:59	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	18:59	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	18:59	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	18:59	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	18:59	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	18:59	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	18:59	LB133365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>RSB Environmental</u>		SDG No.: <u>P4674</u>							
Contract: <u>RSBE01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4674</u>	SAS No.: <u>P4674</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB07	Aluminum	100	+/-100	U	100	P	11/08/2024	19:51	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	19:51	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	19:51	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	19:51	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	19:51	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	19:51	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	19:51	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	19:51	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	19:51	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	19:51	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	19:51	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	19:51	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	19:51	LB133365
CCB08	Aluminum	100	+/-100	U	100	P	11/08/2024	20:57	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	20:57	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	20:57	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	20:57	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	20:57	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	20:57	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	20:57	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	20:57	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	20:57	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	20:57	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	20:57	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	20:57	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	20:57	LB133365
CCB09	Aluminum	100	+/-100	U	100	P	11/08/2024	21:55	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	21:55	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	21:55	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	21:55	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	21:55	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	21:55	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	21:55	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	21:55	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	21:55	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	21:55	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	21:55	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	21:55	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	21:55	LB133365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>RSB Environmental</u>		SDG No.: <u>P4674</u>							
Contract: <u>RSBE01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4674</u>	SAS No.: <u>P4674</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB10	Aluminum	100	+/-100	U	100	P	11/08/2024	23:35	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	23:35	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	23:35	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	23:35	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	23:35	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	23:35	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	23:35	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	23:35	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	23:35	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	23:35	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	23:35	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	23:35	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	23:35	LB133365