

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: RSB Environmental

SDG No.: P4674

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB164687BL		WATER		Batch Number:	PB164687		Prep Date:	11/05/2024	
	Mercury	0.20	<0.20	U	0.20	CV	11/05/2024	13:15	LB133297

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: RSB Environmental

SDG No.: P4674

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB164645BL	WATER			Batch Number:	PB164645		Prep Date:	11/04/2024	
	Aluminum	50.0	<50.0	U	50.0	P	11/08/2024	21:10	LB133365
	Arsenic	10.0	<10.0	U	10.0	P	11/08/2024	21:10	LB133365
	Barium	50.0	<50.0	U	50.0	P	11/08/2024	21:10	LB133365
	Cadmium	3.00	<3.00	U	3.00	P	11/08/2024	21:10	LB133365
	Chromium	5.00	<5.00	U	5.00	P	11/08/2024	21:10	LB133365
	Copper	10.0	<10.0	U	10.0	P	11/08/2024	21:10	LB133365
	Iron	50.0	<50.0	U	50.0	P	11/08/2024	21:10	LB133365
	Lead	6.00	<6.00	U	6.00	P	11/08/2024	21:10	LB133365
	Manganese	10.0	<10.0	U	10.0	P	11/08/2024	21:10	LB133365
	Nickel	20.0	<20.0	U	20.0	P	11/08/2024	21:10	LB133365
	Selenium	10.0	<10.0	U	10.0	P	11/08/2024	21:10	LB133365
	Silver	5.00	<5.00	U	5.00	P	11/08/2024	21:10	LB133365
	Zinc	20.0	<20.0	U	20.0	P	11/08/2024	21:10	LB133365