

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client:	<u>RSB Environmental</u>	SDG No.:	<u>P4674</u>
Contract:	<u>RSBE01</u>	Lab Code:	<u>CHEM</u>
ICS Source:	<u>EPA</u>	Case No.:	<u>P4674</u>
		SAS No.:	<u>P4674</u>
		Instrument ID:	<u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	257000	255000	101	216000	294000	11/06/2024	18:08	LB133323
	Arsenic	2.46			-20	20	11/06/2024	18:08	LB133323
	Barium	5.21	6.0	87	-94	106	11/06/2024	18:08	LB133323
	Cadmium	6.38	1.0	638	-5	7	11/06/2024	18:08	LB133323
	Chromium	56.4	52.0	108	42	62	11/06/2024	18:08	LB133323
	Copper	0.49	2.0	24	-18	22	11/06/2024	18:08	LB133323
	Iron	98800	101000	98	85600	116500	11/06/2024	18:08	LB133323
	Lead	9.93			-12	12	11/06/2024	18:08	LB133323
	Manganese	4.23	7.0	60	-13	27	11/06/2024	18:08	LB133323
	Nickel	3.00	2.0	150	-38	42	11/06/2024	18:08	LB133323
	Selenium	-15.6			-20	20	11/06/2024	18:08	LB133323
	Silver	1.23			-10	10	11/06/2024	18:08	LB133323
	Zinc	8.08			-40	40	11/06/2024	18:08	LB133323
ICSAB01	Aluminum	261000	247000	106	209000	285000	11/06/2024	18:13	LB133323
	Arsenic	114	104	110	88.4	120	11/06/2024	18:13	LB133323
	Barium	529	537	98	437	637	11/06/2024	18:13	LB133323
	Cadmium	1020	972	105	826	1120	11/06/2024	18:13	LB133323
	Chromium	562	542	104	460	624	11/06/2024	18:13	LB133323
	Copper	488	511	96	434	588	11/06/2024	18:13	LB133323
	Iron	100000	99300	101	84400	114500	11/06/2024	18:13	LB133323
	Lead	57.7	49.0	118	37	61	11/06/2024	18:13	LB133323
	Manganese	516	507	102	430	584	11/06/2024	18:13	LB133323
	Nickel	1010	954	106	810	1100	11/06/2024	18:13	LB133323
	Selenium	36.0	46.0	78	26	66	11/06/2024	18:13	LB133323
	Silver	203	201	101	170	232	11/06/2024	18:13	LB133323
	Zinc	1070	952	112	809	1095	11/06/2024	18:13	LB133323
ICSA01	Aluminum	264000	255000	104	216000	294000	11/08/2024	14:24	LB133365
	Arsenic	5.52			-20	20	11/08/2024	14:24	LB133365
	Barium	1.84	6.0	31	-94	106	11/08/2024	14:24	LB133365
	Cadmium	6.11	1.0	611	-5	7	11/08/2024	14:24	LB133365
	Chromium	54.3	52.0	104	42	62	11/08/2024	14:24	LB133365
	Copper	14.3	2.0	715	-18	22	11/08/2024	14:24	LB133365
	Iron	98600	101000	98	85600	116500	11/08/2024	14:24	LB133365
	Lead	9.97			-12	12	11/08/2024	14:24	LB133365
	Manganese	3.65	7.0	52	-13	27	11/08/2024	14:24	LB133365
	Nickel	2.63	2.0	132	-38	42	11/08/2024	14:24	LB133365
	Selenium	-14.4			-20	20	11/08/2024	14:24	LB133365
	Silver	-0.087			-10	10	11/08/2024	14:24	LB133365
	Zinc	7.04			-40	40	11/08/2024	14:24	LB133365
ICSAB01	Aluminum	260000	247000	105	209000	285000	11/08/2024	14:30	LB133365
	Arsenic	115	104	111	88.4	120	11/08/2024	14:30	LB133365
	Barium	496	537	92	437	637	11/08/2024	14:30	LB133365

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: RSB Environmental **SDG No.:** P4674
Contract: RSBE01 **Lab Code:** CHEM **Case No.:** P4674 **SAS No.:** P4674
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Cadmium	979	972	101	826	1120	11/08/2024	14:30	LB133365
	Chromium	521	542	96	460	624	11/08/2024	14:30	LB133365
	Copper	499	511	98	434	588	11/08/2024	14:30	LB133365
	Iron	95700	99300	96	84400	114500	11/08/2024	14:30	LB133365
	Lead	57.3	49.0	117	37	61	11/08/2024	14:30	LB133365
	Manganese	480	507	95	430	584	11/08/2024	14:30	LB133365
	Nickel	981	954	103	810	1100	11/08/2024	14:30	LB133365
	Selenium	31.2	46.0	68	26	66	11/08/2024	14:30	LB133365
	Silver	198	201	98	170	232	11/08/2024	14:30	LB133365
	Zinc	1090	952	114	809	1095	11/08/2024	14:30	LB133365