

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: RSB Environmental **SDG No.:** P4674
Contract: RSBE01 **Lab Code:** CHEM **Case No.:** P4674 **SAS No.:** P4674

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164645BS							
Aluminum	ug/L	1000	955		96	80 - 120	P
Arsenic	ug/L	400	370		92	80 - 120	P
Barium	ug/L	100	86.7		87	80 - 120	P
Cadmium	ug/L	100	91.6		92	80 - 120	P
Chromium	ug/L	200	182		91	80 - 120	P
Copper	ug/L	150	143		95	80 - 120	P
Iron	ug/L	1500	1300		87	80 - 120	P
Lead	ug/L	500	470		94	80 - 120	P
Manganese	ug/L	100	94.7		95	80 - 120	P
Nickel	ug/L	250	235		94	80 - 120	P
Selenium	ug/L	1000	913		91	80 - 120	P
Silver	ug/L	37.5	35.9		96	80 - 120	P
Zinc	ug/L	100	99.1		99	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: RSB Environmental

SDG No.: P4674

Contract: RSBE01

Lab Code: CHEM

Case No.: P4674

SAS No.: P4674

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164687BS Mercury	ug/L	4.0	3.57		89	82 - 119	CV