

Instrument ID: BNA_E

Daily Analysis Runlog For Sequence/QC Batch ID # BE102824

Review By	yogesh	Review On	10/29/2024 4:57:44 AM
Supervise By	mohammad	Supervise On	10/30/2024 2:48:38 AM
SubDirectory	BE102824	HP Acquire Method	BNA_E
		HP Processing Method	be102824

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12323,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BE101388.D	28 Oct 2024 10:51		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BE101389.D	28 Oct 2024 11:44		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BE101390.D	28 Oct 2024 12:23	Compound #32,54,56,65 removed from 5ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BE101391.D	28 Oct 2024 12:59		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BE101392.D	28 Oct 2024 13:35	Compound #32,54 Kept on LR	RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BE101393.D	28 Oct 2024 14:11	The Calibration is Good For 8270 DOD Except com#77 and Not good for 625.1 Method	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BE101394.D	28 Oct 2024 14:49	Com#77(Benzidine) Failed in the calibration for both DOD and NON-DOD	RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BE101395.D	28 Oct 2024 15:25	Compound #69 removed from 60ppm	RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BE101396.D	28 Oct 2024 16:01	Compound #9,69,79 removed from 80ppm	RC/JU	Ok,M
10	SSTDICV040	ICVBE102824	BE101397.D	28 Oct 2024 16:37	ICV Fail for Com#77 for Both DOD and NON-DOD	RC/JU	Ok,M
11	PB164444BL	PB164444BL	BE101398.D	28 Oct 2024 17:16		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_E

Daily Analysis Runlog For Sequence/QC Batch ID # BE110424

Review By	Jagrut	Review On	11/5/2024 2:11:37 PM
Supervise By	mohammad	Supervise On	11/6/2024 3:14:23 AM
SubDirectory	BE110424	HP Acquire Method	BNA_E
		HP Processing Method	be102824

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	SP6573 SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	SP6624 S12324,10ul/1000ul sample SP6559

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BE101452.D	4 Nov 2024 12:13		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BE101453.D	4 Nov 2024 12:49		RC/JU	Ok,M
3	PB164593BL	PB164593BL	BE101454.D	4 Nov 2024 13:25		RC/JU	Ok
4	PB164593BS	PB164593BS	BE101455.D	4 Nov 2024 14:04		RC/JU	Ok,M
5	P4667-13	BP-F-7	BE101456.D	4 Nov 2024 15:42		RC/JU	Ok
6	P4680-05	BP-F25	BE101457.D	4 Nov 2024 16:18		RC/JU	Ok
7	P4680-01	BP-F26	BE101458.D	4 Nov 2024 16:51	Internal Standard Fail	RC/JU	ReRun
8	P4675-04	COMP-4	BE101459.D	4 Nov 2024 17:27	Internal Standard Fail	RC/JU	Ok
9	P4675-04MS	COMP-4MS	BE101460.D	4 Nov 2024 18:02	Internal Standard Fail	RC/JU	Ok,M
10	P4675-04MSD	COMP-4MSD	BE101461.D	4 Nov 2024 18:38	Internal Standard Fail	RC/JU	Ok,M
11	P4675-03	COMP-3	BE101462.D	4 Nov 2024 19:14		RC/JU	Ok
12	P4675-01	COMP-1	BE101463.D	4 Nov 2024 19:50		RC/JU	Ok
13	P4675-02	COMP-2	BE101464.D	4 Nov 2024 20:26	Internal Standard Fail	RC/JU	ReRun
14	P4675-05	COMP-5	BE101465.D	4 Nov 2024 21:01	Internal Standard Fail	RC/JU	ReRun
15	P4675-06	COMP-6	BE101466.D	4 Nov 2024 21:37	Internal Standard Fail	RC/JU	ReRun
16	P4667-05	BP-F-5	BE101467.D	4 Nov 2024 22:13	Internal Standard Fail	RC/JU	ReRun
17	P4667-01	BP-F-6	BE101468.D	4 Nov 2024 22:49		RC/JU	Ok
18	P4667-09	BP-F-10	BE101469.D	4 Nov 2024 23:25		RC/JU	Ok

Instrument ID: BNA_E

Daily Analysis Runlog For Sequence/QC Batch ID # BE110424

Review By	Jagrut	Review On	11/5/2024 2:11:37 PM				
Supervise By	mohammad	Supervise On	11/6/2024 3:14:23 AM				
SubDirectory	BE110424	HP Acquire Method	BNA_E	HP Processing Method	be102824		
STD. NAME		STD REF.#					
Tune/Reschk		SP6573					
Initial Calibration Stds		SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621					
CCC		SP6624					
Internal Standard/PEM		S12324,10ul/1000ul sample					
ICV/I.BLK		SP6559					
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

19	P4679-01	MH-1	BE101470.D	5 Nov 2024 00:01		RC/JU	Ok,M
----	----------	------	------------	------------------	--	-------	------

M : Manual Integration

Instrument ID: BNA_E

Daily Analysis Runlog For Sequence/QC Batch ID # BE110524

Review By	yogesh	Review On	11/6/2024 3:08:53 AM
Supervise By	mohammad	Supervise On	11/6/2024 3:16:26 AM
SubDirectory	BE110524	HP Acquire Method	BNA_E
		HP Processing Method	be102824

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12324,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BE101471.D	5 Nov 2024 10:12		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BE101472.D	5 Nov 2024 10:48		RC/JU	Ok,M
3	PB164639BL	PB164639BL	BE101473.D	5 Nov 2024 11:33		RC/JU	Ok
4	PB164639BS	PB164639BS	BE101474.D	5 Nov 2024 12:09		RC/JU	Ok,M
5	P4680-01	BP-F26	BE101475.D	5 Nov 2024 12:44		RC/JU	Ok
6	P4675-02	COMP-2	BE101476.D	5 Nov 2024 13:20		RC/JU	Ok,M
7	P4675-05	COMP-5	BE101477.D	5 Nov 2024 13:56		RC/JU	Ok
8	P4675-06	COMP-6	BE101478.D	5 Nov 2024 14:32		RC/JU	Ok
9	P4667-05	BP-F-5	BE101479.D	5 Nov 2024 15:08		RC/JU	Ok
10	P4640-04	MH-3	BE101480.D	5 Nov 2024 15:43		RC/JU	Ok
11	P4640-04MS	MH-3MS	BE101481.D	5 Nov 2024 16:16		RC/JU	Ok,M
12	P4640-04MSD	MH-3MSD	BE101482.D	5 Nov 2024 16:52		RC/JU	Ok,M
13	P4643-04	BP-F9-ADDITIONAL	BE101483.D	5 Nov 2024 17:28		RC/JU	Ok
14	P4643-08	BP-F8	BE101484.D	5 Nov 2024 18:04		RC/JU	Ok
15	P4643-12	TP-9	BE101485.D	5 Nov 2024 18:39		RC/JU	Ok
16	P4682-02	BELL-SHOP-RAGS	BE101486.D	5 Nov 2024 19:15		RC/JU	Ok
17	P4645-04	Z-02-WC	BE101487.D	5 Nov 2024 19:51		RC/JU	Ok
18	P4660-03	WC-TA2-01-C	BE101488.D	5 Nov 2024 20:27		RC/JU	Ok

Instrument ID: BNA_E

Daily Analysis Runlog For Sequence/QC Batch ID # BE110524

Review By	yogesh	Review On	11/6/2024 3:08:53 AM		
Supervise By	mohammad	Supervise On	11/6/2024 3:16:26 AM		
SubDirectory	BE110524	HP Acquire Method	BNA_E	HP Processing Method	be102824

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12324,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	P4682-01	BELL-SHOP-RAGS	BE101489.D	5 Nov 2024 21:03	Internal Standrad Fail	RC/JU	Ok,M
20	P4685-01	OK-01-11012024	BE101490.D	5 Nov 2024 21:38		RC/JU	Ok,M

M : Manual Integration