

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4675

Client: Kleinfelder

Analytical Method: 8270E DataFile: BE101474.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	RPD		Low	Limits	
							Qual	Qual		High	RPD
PB164639BS	Naphthalene	1700	1600	ug/Kg	94				62	100	
	Fluorene	1700	1600	ug/Kg	94				61	101	
	Phenanthrene	1700	1800	ug/Kg	106		*		59	103	
	Anthracene	1700	1900	ug/Kg	112		*		61	105	
	Pyrene	1700	1700	ug/Kg	100				59	103	
	Benzo(a)anthracene	1700	1900	ug/Kg	112		*		60	102	
	Chrysene	1700	1800	ug/Kg	106		*		59	101	
	Benzo(b)fluoranthene	1700	1700	ug/Kg	100				62	109	
	Benzo(a)pyrene	1700	1900	ug/Kg	112		*		63	103	
	Indeno(1,2,3-cd)pyrene	1700	1800	ug/Kg	106		*		63	101	
	Benzo(g,h,i)perylene	1700	1600	ug/Kg	94				70	108	