

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: POWE02

Lab Code: CHEM Case No.: P4675 SAS No.: P4675 SDG No.: P4675

Instrument ID: MSVOA_Y Calibration Date/Time: 11/04/2024 09:42

Lab File ID: VY020133.D Init. Calib. Date(s): 10/30/2024 10/30/2024

Heated Purge: (Y/N) Y Init. Calib. Time(s): 12:39 14:52

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
cis-1,2-Dichloroethene	0.700	0.796		13.71	20
1,1,1-Trichloroethane	1.013	1.146		13.13	20
Benzene	1.433	1.616		12.77	20
Trichloroethene	0.332	0.373		12.35	20
Toluene	0.884	1.023		15.72	20
Ethyl Benzene	1.992	2.300		15.46	20
m/p-Xylenes	0.729	0.853		17.01	20
o-Xylene	0.690	0.798		15.65	20
Isopropylbenzene	3.994	4.804		20.28	20
1,2-Dichloroethane-d4	0.624	0.639		2.4	20
Dibromofluoromethane	0.318	0.337		5.97	20
Toluene-d8	1.265	1.328		4.98	20
4-Bromofluorobenzene	0.411	0.423		2.92	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.