

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4676 SAS No.: P4676 SDG NO.: P4676

Continuing Calib Date: 11/06/2024 Initial Calibration Date(s): 10/29/2024 10/29/2024

Continuing Calib Time: 00:09 Initial Calibration Time(s): 04:56 14:43

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	5.77	5.77	5.70	5.84	0.00
Aroclor-1254-2	(2)	6.01	6.01	5.94	6.08	0.00
Aroclor-1254-3	(3)	6.40	6.40	6.33	6.47	0.00
Aroclor-1254-4	(4)	6.71	6.71	6.64	6.78	0.00
Aroclor-1254-5	(5)	7.16	7.17	7.10	7.24	0.01
TCX		3.64	3.65	3.60	3.70	0.01
DCB		9.53	9.54	9.44	9.64	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4676 SAS No.: P4676 SDG NO.: P4676

Continuing Calib Date: 11/06/2024 Initial Calibration Date(s): 10/29/2024 10/29/2024

Continuing Calib Time: 00:09 Initial Calibration Time(s): 04:56 14:43

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	4.94	4.94	4.87	5.01	0.00
Aroclor-1254-2	(2)	5.10	5.10	5.03	5.17	0.00
Aroclor-1254-3	(3)	5.52	5.52	5.45	5.59	0.00
Aroclor-1254-4	(4)	5.77	5.77	5.70	5.84	0.00
Aroclor-1254-5	(5)	6.22	6.22	6.15	6.29	0.00
TCX		2.97	2.98	2.93	3.03	0.01
DCB		8.26	8.26	8.16	8.36	0.00