

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4676 SAS No.: P4676 SDG NO.: P4676

Continuing Calib Date: 11/06/2024 Initial Calibration Date(s): 10/29/2024 10/29/2024

Continuing Calib Time: 04:34 Initial Calibration Time(s): 04:56 14:43

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.86	4.87	4.80	4.94	0.01
Aroclor-1016-2	(2)	4.89	4.89	4.82	4.96	0.00
Aroclor-1016-3	(3)	4.95	4.95	4.88	5.02	0.00
Aroclor-1016-4	(4)	5.05	5.06	4.99	5.13	0.01
Aroclor-1016-5	(5)	5.37	5.37	5.30	5.44	0.00
Aroclor-1260-1	(1)	6.58	6.58	6.51	6.65	0.00
Aroclor-1260-2	(2)	6.86	6.86	6.79	6.93	0.00
Aroclor-1260-3	(3)	7.25	7.25	7.18	7.32	0.00
Aroclor-1260-4	(4)	7.49	7.50	7.43	7.57	0.01
Aroclor-1260-5	(5)	7.83	7.83	7.76	7.90	0.00
TCX		3.64	3.65	3.60	3.70	0.01
DCB		9.53	9.53	9.43	9.63	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4676 SAS No.: P4676 SDG NO.: P4676

Continuing Calib Date: 11/06/2024 Initial Calibration Date(s): 10/29/2024 10/29/2024

Continuing Calib Time: 04:34 Initial Calibration Time(s): 04:56 14:43

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.10	4.10	4.03	4.17	0.00
Aroclor-1016-2	(2)	4.11	4.12	4.05	4.19	0.01
Aroclor-1016-3	(3)	4.29	4.30	4.23	4.37	0.01
Aroclor-1016-4	(4)	4.35	4.35	4.28	4.42	0.01
Aroclor-1016-5	(5)	4.56	4.57	4.50	4.64	0.01
Aroclor-1260-1	(1)	5.66	5.67	5.60	5.74	0.01
Aroclor-1260-2	(2)	5.87	5.87	5.80	5.94	0.00
Aroclor-1260-3	(3)	6.03	6.03	5.96	6.10	0.00
Aroclor-1260-4	(4)	6.54	6.54	6.47	6.61	0.01
Aroclor-1260-5	(5)	6.80	6.80	6.73	6.87	0.00
TCX		2.97	2.98	2.93	3.03	0.01
DCB		8.26	8.26	8.16	8.36	0.00