

Surrogate Summary

SDG No.: P4676

Client: Chemtech Consulting Group

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P4676-01	Storage-Blank-SOIL-REF-6	1,2-Dichloroethane-d4	50	39.8	80	74	125
		Dibromofluoromethane	50	46.5	93	75	124
		Toluene-d8	50	50.2	100	86	113
		4-Bromofluorobenzene	50	46.1	92	77	121
P4676-02	Storage-Blank-WATER-REF-5	1,2-Dichloroethane-d4	50	41.4	83	74	125
		Dibromofluoromethane	50	46.3	93	75	124
		Toluene-d8	50	50.0	100	86	113
		4-Bromofluorobenzene	50	43.8	88	77	121
P4676-03	Storage-Blank-WATER-REF-4	1,2-Dichloroethane-d4	50	41.9	84	74	125
		Dibromofluoromethane	50	46.4	93	75	124
		Toluene-d8	50	50.4	101	86	113
		4-Bromofluorobenzene	50	45.9	92	77	121
P4676-04	Storage-Blank-SAMPLE-MANAGEMENT	1,2-Dichloroethane-d4	50	42.8	86	74	125
		Dibromofluoromethane	50	46.1	92	75	124
		Toluene-d8	50	51.1	102	86	113
		4-Bromofluorobenzene	50	48.0	96	77	121
VX1101WBL01	VX1101WBL01	1,2-Dichloroethane-d4	50	41.0	82	74	125
		Dibromofluoromethane	50	45.0	90	75	124
		Toluene-d8	50	48.5	97	86	113
		4-Bromofluorobenzene	50	45.1	90	77	121
VX1101WBS01	VX1101WBS01	1,2-Dichloroethane-d4	50	39.5	79	74	125
		Dibromofluoromethane	50	46.9	94	75	124
		Toluene-d8	50	49.5	99	86	113
		4-Bromofluorobenzene	50	49.1	98	77	121
VX1101WBSD0	VX1101WBSD01	1,2-Dichloroethane-d4	50	40.9	82	74	125
		Dibromofluoromethane	50	47.8	96	75	124
		Toluene-d8	50	50.9	102	86	113
		4-Bromofluorobenzene	50	49.6	99	77	121