

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4676

Client: Chemtech Consulting Group

Analytical Method: SW8260-Low

Datafile : VX043657.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1101WBS01	Dichlorodifluoromethane	20	17.2	ug/L	86			69	116	
	Chloromethane	20	16.2	ug/L	81			65	116	
	Vinyl chloride	20	17.7	ug/L	89			65	117	
	Ethyl Acetate	20	19.9	ug/L	100			75	115	
	Bromomethane	20	17.8	ug/L	89			58	125	
	Chloroethane	20	24.3	ug/L	121			56	128	
	Trichlorofluoromethane	20	20.3	ug/L	102			73	115	
	1,1,2-Trichlorotrifluoroethane	20	21.4	ug/L	107			80	112	
	Tert butyl alcohol	100	7.60	ug/L	8		*	73	124	
	1,1-Dichloroethene	20	19.3	ug/L	97			74	110	
	Acrolein	100	74.0	ug/L	74			51	143	
	Acrylonitrile	100	92.4	ug/L	92			73	113	
	Acetone	100	85.3	ug/L	85			60	125	
	Carbon disulfide	20	16.9	ug/L	85			64	112	
	Methyl tert-butyl Ether	20	18.4	ug/L	92			78	114	
	Methyl Acetate	20	17.8	ug/L	89			67	125	
	Methylene Chloride	20	17.9	ug/L	90			72	114	
	trans-1,2-Dichloroethene	20	19.4	ug/L	97			75	108	
	Vinyl Acetate	100	91.4	ug/L	91			76	115	
	1,1-Dichloroethane	20	18.7	ug/L	94			78	112	
	Cyclohexane	20	19.6	ug/L	98			75	110	
	2-Butanone	100	90.8	ug/L	91			65	122	
	Carbon Tetrachloride	20	20.6	ug/L	103			77	113	
	2,2-Dichloropropane	20	19.2	ug/L	96			75	116	
	cis-1,2-Dichloroethene	20	19.4	ug/L	97			77	110	
	Bromochloromethane	20	17.4	ug/L	87			70	124	
	Chloroform	20	18.7	ug/L	94			79	113	
	1,1,1-Trichloroethane	20	19.4	ug/L	97			80	108	
	Methylcyclohexane	20	22.0	ug/L	110			72	115	
	1,1-Dichloropropene	20	20.4	ug/L	102			79	110	
	Benzene	20	20.1	ug/L	101			82	109	
	1,2-Dichloroethane	20	18.6	ug/L	93			80	115	
	Trichloroethene	20	21.4	ug/L	107			77	113	
	1,2-Dichloropropane	20	19.6	ug/L	98			83	111	
	Dibromomethane	20	19.6	ug/L	98			82	110	
	Bromodichloromethane	20	19.8	ug/L	99			83	110	
	4-Methyl-2-Pentanone	100	100	ug/L	100			74	118	
	Toluene	20	21.1	ug/L	106			82	110	
	t-1,3-Dichloropropene	20	19.3	ug/L	97			79	110	
	cis-1,3-Dichloropropene	20	20.4	ug/L	102			82	110	
	1,1,2-Trichloroethane	20	20.7	ug/L	104			83	112	
	1,3-Dichloropropane	20	20.6	ug/L	103			83	111	
	2-Chloroethyl vinyl ether	100	96.5	ug/L	97			75	124	
	2-Hexanone	100	99.9	ug/L	100			73	117	
	Dibromochloromethane	20	20.5	ug/L	103			82	110	
	1,2-Dibromoethane	20	20.1	ug/L	101			81	110	
	Tetrachloroethene	20	22.4	ug/L	112			67	123	
	Chlorobenzene	20	21.1	ug/L	106			82	109	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4676

Client: Chemtech Consulting Group

Analytical Method: SW8260-Low

Datafile : VX043657.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1101WBS01	1,1,1,2-Tetrachloroethane	20	20.8	ug/L	104			84	111	
	Hexachloroethane	20	21.4	ug/L	107			80	113	
	Ethyl Benzene	20	21.5	ug/L	108			83	109	
	m/p-Xylenes	40	43.6	ug/L	109			82	110	
	o-Xylene	20	21.9	ug/L	110		*	83	109	
	Styrene	20	22.0	ug/L	110			80	111	
	Bromoform	20	19.7	ug/L	99			79	109	
	Isopropylbenzene	20	22.3	ug/L	112			83	112	
	1,1,2,2-Tetrachloroethane	20	20.2	ug/L	101			76	118	
	1,2,3-Trichloropropane	20	19.7	ug/L	99			75	112	
	Bromobenzene	20	21.6	ug/L	108			77	116	
	N-propylbenzene	20	22.6	ug/L	113		*	83	112	
	2-Chlorotoluene	20	21.1	ug/L	106			83	110	
	1,3,5-Trimethylbenzene	20	22.0	ug/L	110			85	112	
	4-Chlorotoluene	20	21.1	ug/L	106			82	110	
	tert-Butylbenzene	20	22.6	ug/L	113		*	83	112	
	1,2,4-Trimethylbenzene	20	22.0	ug/L	110			85	111	
	Sec-butylbenzene	20	23.3	ug/L	117		*	81	114	
	p-Isopropyltoluene	20	23.1	ug/L	116			78	116	
	1,3-Dichlorobenzene	20	21.6	ug/L	108			82	108	
	1,4-Dichlorobenzene	20	21.2	ug/L	106			82	107	
	n-Butylbenzene	20	23.4	ug/L	117		*	75	115	
	1,2-Dichlorobenzene	20	21.6	ug/L	108			82	109	
	1,2-Dibromo-3-Chloropropane	20	18.5	ug/L	93			68	112	
	1,2,4-Trichlorobenzene	20	21.2	ug/L	106			75	113	
	Hexachlorobutadiene	20	23.3	ug/L	117			81	121	
	Naphthalene	20	20.6	ug/L	103			78	119	
	1,2,3-Trichlorobenzene	20	20.8	ug/L	104			76	114	