

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: P4676 SAS No.: P4676 SDG NO.: P4676
 Lab File ID: VX043654.D Date Analyzed: 11/01/2024
 Instrument ID: MSVOA_X Time Analyzed: 10:12
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	131267	12.024				
UPPER LIMIT	262534	12.524				
LOWER LIMIT	65633.5	11.524				
EPA SAMPLE NO.						
Storage-Blank-SOIL-REF-6	95140	12.02				
Storage-Blank-WATER-REF-5	86030	12.02				
Storage-Blank-WATER-REF-4	87785	12.02				
Storage-Blank-SAMPLE-MANAGEMENT	90064	12.02				
VX1101WBL01	102414	12.02				
VX1101WBS01	133771	12.02				
VX1101WBSD01	124591	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.