

Duplicate Sample Summary

Client:	METEM A GE POWER Business	SDG No.:	P4681
Project:	Parsippany Wastewater Quarterly 2024	Sample ID:	P4681-02
Client ID:	PHDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Field pH	pH	+/-20	7.09		7.10		1	0.14		11/01/2024

Duplicate Sample Summary

Client:	METEM A GE POWER Business	SDG No.:	P4681
Project:	Parsippany Wastewater Quarterly 2024	Sample ID:	P4762-01
Client ID:	EFF-WWDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/L	+/-20	0.0027	J	0.0028	J	1	4		11/11/2024

Duplicate Sample Summary

Client: METEM A GE POWER Business	SDG No.: P4681
Project: Parsippany Wastewater Quarterly 2024	Sample ID: P4762-01
Client ID: EFF-WWMSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Cyanide	mg/L	+/-20	0.047		0.047		1	0		11/11/2024