



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: METEM A GE POWER Business

SDG No.: P4681

Contract: METE01

Lab Code: CHEM

Case No.: P4681

SAS No.: P4681

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	14:16	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	14:16	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	14:16	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	14:16	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	14:16	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	14:16	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	14:16	LB133365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: METEM A GE POWER Business		SDG No.: P4681							
Contract: METE01	Lab Code: CHEM	Case No.: P4681	SAS No.: P4681						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	14:50	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	14:50	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	14:50	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	14:50	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	14:50	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	14:50	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	14:50	LB133365
CCB02	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	15:33	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	15:33	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	15:33	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	15:33	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	15:33	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	15:33	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	15:33	LB133365
CCB03	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	16:28	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	16:28	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	16:28	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	16:28	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	16:28	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	16:28	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	16:28	LB133365
CCB04	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	17:22	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	17:22	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	17:22	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	17:22	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	17:22	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	17:22	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	17:22	LB133365
CCB05	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	18:08	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	18:08	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	18:08	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	18:08	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	18:08	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	18:08	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	18:08	LB133365
CCB06	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	18:59	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	18:59	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	18:59	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	18:59	LB133365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: METEM A GE POWER Business **SDG No.:** P4681
Contract: METE01 **Lab Code:** CHEM **Case No.:** P4681 **SAS No.:** P4681

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	18:59	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	18:59	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	18:59	LB133365
CCB07	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	19:51	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	19:51	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	19:51	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	19:51	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	19:51	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	19:51	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	19:51	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	20:57	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	20:57	LB133365
CCB08	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	20:57	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	20:57	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	20:57	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	20:57	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	20:57	LB133365
CCB09	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	21:55	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	21:55	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	21:55	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	21:55	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	21:55	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	21:55	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	21:55	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	23:35	LB133365
CCB10	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	23:35	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	23:35	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	23:35	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	23:35	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	23:35	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	23:35	LB133365