

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: METEM A GE POWER Business **SDG No.:** P4681
Contract: METE01 **Lab Code:** CHEM **Case No.:** P4681 **SAS No.:** P4681
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	6.11	1.0	611	-5	7	11/08/2024	14:24	LB133365
	Chromium	54.3	52.0	104	42	62	11/08/2024	14:24	LB133365
	Copper	14.3	2.0	715	-18	22	11/08/2024	14:24	LB133365
	Lead	9.97			-12	12	11/08/2024	14:24	LB133365
	Nickel	2.63	2.0	132	-38	42	11/08/2024	14:24	LB133365
	Silver	-0.087			-10	10	11/08/2024	14:24	LB133365
	Zinc	7.04			-40	40	11/08/2024	14:24	LB133365
ICSAB01	Cadmium	979	972	101	826	1120	11/08/2024	14:30	LB133365
	Chromium	521	542	96	460	624	11/08/2024	14:30	LB133365
	Copper	499	511	98	434	588	11/08/2024	14:30	LB133365
	Lead	57.3	49.0	117	37	61	11/08/2024	14:30	LB133365
	Nickel	981	954	103	810	1100	11/08/2024	14:30	LB133365
	Silver	198	201	98	170	232	11/08/2024	14:30	LB133365
	Zinc	1090	952	114	809	1095	11/08/2024	14:30	LB133365
ICSA	Cadmium	7.97	1.0	797	-5	7	11/08/2024	14:35	LB133365
	Chromium	52.1	52.0	100	42	62	11/08/2024	14:35	LB133365
	Copper	12.0	2.0	600	-18	22	11/08/2024	14:35	LB133365
	Lead	-25.8			-12	12	11/08/2024	14:35	LB133365
	Nickel	6.25	2.0	312	-38	42	11/08/2024	14:35	LB133365
	Silver	12.8			-10	10	11/08/2024	14:35	LB133365
	Zinc	17.8			-40	40	11/08/2024	14:35	LB133365
ICSAB	Cadmium	985	972	101	826	1120	11/08/2024	14:40	LB133365
	Chromium	612	542	113	460	624	11/08/2024	14:40	LB133365
	Copper	538	511	105	434	588	11/08/2024	14:40	LB133365
	Lead	33.1	49.0	68	37	61	11/08/2024	14:40	LB133365
	Nickel	1010	954	106	810	1100	11/08/2024	14:40	LB133365
	Silver	197	201	98	170	232	11/08/2024	14:40	LB133365
	Zinc	1110	952	117	809	1095	11/08/2024	14:40	LB133365