

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: METEM A GE POWER Business

SDG No.: P4681

Contract: METE01

Lab Code: CHEM

Case No.: P4681

SAS No.: P4681

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164646BS							
Cadmium	ug/L	100	90.9		91	85 - 115	P
Chromium	ug/L	200	184		92	85 - 115	P
Copper	ug/L	150	141		94	85 - 115	P
Lead	ug/L	500	466		93	85 - 115	P
Nickel	ug/L	250	232		93	85 - 115	P
Silver	ug/L	37.5	35.6		95	85 - 115	P
Zinc	ug/L	100	96.8		97	85 - 115	P