



SDG NARRATIVE

LAB NAME: Alliance Technical Group, LLC

CASE: 51802 SDG: C0CF4

CONTRACT: 68HERH20D0011

LAB CODE: ACE

LAB ORDER ID: P4690

MODIFICATION REF. NUMBER: NA

Sample ID	EPA Sample ID	pН
P4690-01	C0CF4	1.0
P4690-02	C0CF5	1.0
P4690-02DL	C0CF5DL	1.0
P4690-05	C0CF6	1.0
P4690-06	C0CF7	1.0
P4690-07	C0CF8	1.0
P4690-08	C0CF9	1.0
P4690-09	C0CG0	1.0
P4690-10	C0CG1	1.0
P4690-11	C0CG2	1.0
P4690-12	C0CG3	1.0
P4690-13	C0CG4	1.0
P4690-14	C0CH2	1.0
P4690-15	C0CH3	1.0
P4690-16	C0CN6	1.0
P4690-17	C0CN8	1.0
P4690-18	C0CP0	1.0
P4690-19	C0CP1	1.0
P4690-20	C0CP2	1.0
P4690-21	C0CP5	1.0
P4690-22	C0CP6	1.0
P4690-22RE	C0CP6RE	1.0

20 Water samples were delivered to the laboratory intact on 11/02/2024.

Test requested on the Chain of Custody was Trace volatile Organic by Method SFAM01.1.

The temperature of the samples was measured using an I R Gun. The samples temperature was 2.5 degree Celsius for the samples received on 11/02/2024.





Trace Volatiles:

The analysis performed on instrument MSVOA_U were done using GC column DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UI

The analysis of VOC-SFAM was based on method SFAM01.1_Trace.

The Surrogate recoveries met the acceptable criteria except for,

C0CP6 [1,1-Dichloroethene-d2 - 133%, 2-Butanone-d5 - 176%, Chloroform-d - 127%, Toluene-d8 - 67%],

COCP6RE [1,1-Dichloroethene-d2 - 194%, 2-Butanone-d5 - 185%, Chloroethane-d5 - 147% and Chloroform-d - 154%],

Sample C0CP6 failed for more than three surrogates, as corrective action this sample was reanalyzed and both the run are Reported.

The Internal Standards Areas met the acceptable requirements.

Instrument Performance Check met requirements.

The Retention Times met requirements.

The Tuning criteria met requirements.

The initial Calibration criteria met requirements.

The Continuing Calibration criteria met requirements.

The Blank analysis did not indicate the presence of lab contamination.

The storage blank analysis did not indicate the presence of lab contamination.

Sample COCF5 was diluted due to high concentration.

See **Manual Integration report** for the manual integration information at the end of the case narrative.

Calculation:

Low/Med Water Level Calculation

Concentration in ug/L =
$$(Ax) (Is) (DF)$$

(Ais) (RRF) (Vo)

Where,

Ax = Area of the characteristic ion (EICP) for the compound to be measured.

Ais = Area of the characteristic ion (EICP) for the internal standard.

Amount of internal standard added in ng.

RRF = Mean Relative Response Factor from the initial calibration standard.

Vo = Total volume of water purged, in mL.

DF = Dilution Factor





Example calculation of C00	CF6 for Tetrachloroethene:
-----------------------------------	----------------------------

Ax = 26060Is = 125 RRF = 0.294DF = 1 Ais = 157314Vo. = 25 $Concentration in ug/L = \frac{(26060) (125) (1)}{(157314)(0.294)(25)}$ Reported Result = 2.82 ug/LFinal Reported Result = 2.8 ug/L

Relative Response Factor = **Dichlorodifluoromethane**: RUN **VU102324** for **0.5** ppb

RRF= <u>Area of compound</u> X <u>Conc. of Internal Standard</u>
Area of Internal Standard Conc. of Compound

RRF= <u>5894</u> X <u>5.0</u> 198718 0.5

RRF= 0.297

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature	Name: Nimisha Pandya.
Date:	Title: Document Control Officer.