

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110724\
 Data File : VL041679.D
 Acq On : 7 Nov 2024 13:31
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 11/09/2024
 Supervised By :Semsettin Yesilyurt 11/09/2024

Quant Time: Nov 08 00:12:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL110724AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LFri Aug 2
 QLast Update : Fri Nov 08 00:07:44 2024
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) Bromochloromethane		5.192	49	1085923	10.000	ppbv	0.00
33) 1,4-Difluorobenzene		6.652	114	3041887	10.000	ppbv	0.00
55) Chlorobenzene-d5		11.465	117	2745730	10.000	ppbv	0.00
System Monitoring Compounds							
68) 1-Bromo-4-Fluorobenzene		13.777	95	2034307	10.018	ppbv	0.00
Spiked Amount		10.000	Range	65 - 135	Recovery	= 100.200%	
Target Compounds							
							Qvalue
5) Vinyl Chloride		2.922	62	2094m	0.031	ppbv	
63) 1,1,2,2-Tetrachloroethane		13.047	83	6025m	0.032	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110724\
Data File : VL041679.D
Acq On : 7 Nov 2024 13:31
Operator : SY/MD
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Quant Time: Nov 08 00:12:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL110724AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Nov 08 00:07:44 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Mahesh Dadoda 11/09/2024
Supervised By : Semsettin Yesilyurt 11/09/2024

