

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: EarthEfficient LLC **SDG No.:** P4699
Contract: EART12 **Lab Code:** CHEM **Case No.:** P4699 **SAS No.:** P4699
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Lead	8.91			-12	12	11/07/2024	14:47	LB133344
ICSAB01	Lead	56.6	49.0	116	37	61	11/07/2024	14:51	LB133344