

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: EarthEfficient LLC

SDG No.: P4699

Contract: EART12

Lab Code: CHEM

Case No.: P4699

SAS No.: P4699

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164708BS Lead	mg/Kg	43.9	41.2		94	80 - 120	P