

### CALIBRATION VERIFICATION SUMMARY

Contract: EART12

Lab Code: CHEM Case No.: P4699 SAS No.: P4699 SDG NO.: P4699

Continuing Calib Date: 11/05/2024 Initial Calibration Date(s): 10/15/2024 10/16/2024

Continuing Calib Time: 13:35 Initial Calibration Time(s): 18:27 02:36

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|------------|-----------|-----------|-------|------------|
|                      |            |           | FROM      | TO    |            |
| Aroclor-1248-1 (1)   | 5.52       | 5.52      | 5.42      | 5.62  | 0.00       |
| Aroclor-1248-2 (2)   | 5.80       | 5.80      | 5.70      | 5.90  | 0.00       |
| Aroclor-1248-3 (3)   | 6.00       | 6.00      | 5.90      | 6.10  | 0.00       |
| Aroclor-1248-4 (4)   | 6.40       | 6.40      | 6.30      | 6.50  | 0.00       |
| Aroclor-1248-5 (5)   | 6.44       | 6.44      | 6.34      | 6.54  | 0.00       |
| Tetrachloro-m-xylene | 4.37       | 4.37      | 4.27      | 4.47  | 0.00       |
| Decachlorobiphenyl   | 10.07      | 10.06     | 9.96      | 10.16 | -0.01      |

### CALIBRATION VERIFICATION SUMMARY

Contract: EART12

Lab Code: CHEM Case No.: P4699 SAS No.: P4699 SDG NO.: P4699

Continuing Calib Date: 11/05/2024 Initial Calibration Date(s): 10/15/2024 10/16/2024

Continuing Calib Time: 13:35 Initial Calibration Time(s): 18:27 02:36

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1248-1 (1)   | 4.72       | 4.73      | 4.63      | 4.83 | 0.01       |
| Aroclor-1248-2 (2)   | 4.96       | 4.96      | 4.86      | 5.06 | 0.00       |
| Aroclor-1248-3 (3)   | 5.00       | 5.00      | 4.90      | 5.10 | 0.00       |
| Aroclor-1248-4 (4)   | 5.17       | 5.18      | 5.08      | 5.28 | 0.01       |
| Aroclor-1248-5 (5)   | 5.56       | 5.57      | 5.47      | 5.67 | 0.01       |
| Tetrachloro-m-xylene | 3.64       | 3.64      | 3.54      | 3.74 | 0.00       |
| Decachlorobiphenyl   | 8.64       | 8.64      | 8.54      | 8.74 | 0.00       |