

CALIBRATION VERIFICATION SUMMARY

Contract: EART12

Lab Code: CHEM Case No.: P4699 SAS No.: P4699 SDG NO.: P4699

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL01 Date Analyzed: 11/05/2024

Lab Sample No.: AR1660CCC500 Data File : PO107682.D Time Analyzed: 13:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.522	5.423	5.623	496.940	500.000	-0.6
Aroclor-1016-2	5.544	5.445	5.645	490.020	500.000	-2.0
Aroclor-1016-3	5.607	5.507	5.707	482.070	500.000	-3.6
Aroclor-1016-4	5.704	5.604	5.804	497.190	500.000	-0.6
Aroclor-1016-5	5.999	5.899	6.099	491.380	500.000	-1.7
Aroclor-1260-1	7.128	7.026	7.226	479.190	500.000	-4.2
Aroclor-1260-2	7.384	7.282	7.482	525.610	500.000	5.1
Aroclor-1260-3	7.748	7.643	7.843	518.430	500.000	3.7
Aroclor-1260-4	7.972	7.868	8.068	541.150	500.000	8.2
Aroclor-1260-5	8.287	8.181	8.381	533.350	500.000	6.7
Decachlorobiphenyl	10.071	9.958	10.158	48.700	50.000	-2.6
Tetrachloro-m-xylene	4.371	4.274	4.474	49.030	50.000	-1.9

CALIBRATION VERIFICATION SUMMARY

Contract: EART12

Lab Code: CHEM Case No.: P4699 SAS No.: P4699 SDG NO.: P4699

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL01 Date Analyzed: 11/05/2024

Lab Sample No.: AR1660CCC500 Data File : PO107682.D Time Analyzed: 13:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.723	4.626	4.826	536.740	500.000	7.3
Aroclor-1016-2	4.742	4.645	4.845	546.450	500.000	9.3
Aroclor-1016-3	4.918	4.821	5.021	521.460	500.000	4.3
Aroclor-1016-4	4.959	4.862	5.062	474.530	500.000	-5.1
Aroclor-1016-5	5.173	5.076	5.276	525.150	500.000	5.0
Aroclor-1260-1	6.204	6.107	6.307	492.350	500.000	-1.5
Aroclor-1260-2	6.391	6.294	6.494	507.430	500.000	1.5
Aroclor-1260-3	6.544	6.448	6.648	492.680	500.000	-1.5
Aroclor-1260-4	7.015	6.918	7.118	465.830	500.000	-6.8
Aroclor-1260-5	7.256	7.159	7.359	492.690	500.000	-1.5
Decachlorobiphenyl	8.636	8.539	8.739	47.340	50.000	-5.3
Tetrachloro-m-xylene	3.642	3.545	3.745	55.510	50.000	11.0