

CALIBRATION VERIFICATION SUMMARY

Contract: EART12

Lab Code: CHEM Case No.: P4699 SAS No.: P4699 SDG NO.: P4699

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL02 Date Analyzed: 11/05/2024

Lab Sample No.: AR1660CCC500 Data File : PO107697.D Time Analyzed: 18:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.521	5.423	5.623	522.470	500.000	4.5
Aroclor-1016-2	5.543	5.445	5.645	507.430	500.000	1.5
Aroclor-1016-3	5.606	5.507	5.707	503.820	500.000	0.8
Aroclor-1016-4	5.703	5.604	5.804	525.210	500.000	5.0
Aroclor-1016-5	5.998	5.899	6.099	515.260	500.000	3.1
Aroclor-1260-1	7.126	7.026	7.226	513.720	500.000	2.7
Aroclor-1260-2	7.384	7.282	7.482	520.670	500.000	4.1
Aroclor-1260-3	7.746	7.643	7.843	528.800	500.000	5.8
Aroclor-1260-4	7.971	7.868	8.068	552.130	500.000	10.4
Aroclor-1260-5	8.285	8.181	8.381	543.660	500.000	8.7
Decachlorobiphenyl	10.070	9.958	10.158	48.260	50.000	-3.5
Tetrachloro-m-xylene	4.371	4.274	4.474	55.650	50.000	11.3

CALIBRATION VERIFICATION SUMMARY

Contract: EART12

Lab Code: CHEM Case No.: P4699 SAS No.: P4699 SDG NO.: P4699

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL02 Date Analyzed: 11/05/2024

Lab Sample No.: AR1660CCC500 Data File : PO107697.D Time Analyzed: 18:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.722	4.626	4.826	575.760	500.000	15.2
Aroclor-1016-2	4.742	4.645	4.845	589.770	500.000	18.0
Aroclor-1016-3	4.918	4.821	5.021	567.830	500.000	13.6
Aroclor-1016-4	4.959	4.862	5.062	515.120	500.000	3.0
Aroclor-1016-5	5.173	5.076	5.276	560.050	500.000	12.0
Aroclor-1260-1	6.203	6.107	6.307	556.200	500.000	11.2
Aroclor-1260-2	6.391	6.294	6.494	575.150	500.000	15.0
Aroclor-1260-3	6.545	6.448	6.648	536.310	500.000	7.3
Aroclor-1260-4	7.015	6.918	7.118	525.040	500.000	5.0
Aroclor-1260-5	7.256	7.159	7.359	545.390	500.000	9.1
Decachlorobiphenyl	8.636	8.539	8.739	48.340	50.000	-3.3
Tetrachloro-m-xylene	3.642	3.545	3.745	57.920	50.000	15.8