

CALIBRATION VERIFICATION SUMMARY

Contract: EART12

Lab Code: CHEM Case No.: P4699 SAS No.: P4699 SDG NO.: P4699

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL03 Date Analyzed: 11/05/2024

Lab Sample No.: AR1660CCC500 Data File : PO107711.D Time Analyzed: 22:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.522	5.423	5.623	540.930	500.000	8.2
Aroclor-1016-2	5.544	5.445	5.645	536.510	500.000	7.3
Aroclor-1016-3	5.607	5.507	5.707	524.620	500.000	4.9
Aroclor-1016-4	5.704	5.604	5.804	546.140	500.000	9.2
Aroclor-1016-5	5.998	5.899	6.099	539.380	500.000	7.9
Aroclor-1260-1	7.128	7.026	7.226	530.520	500.000	6.1
Aroclor-1260-2	7.385	7.282	7.482	550.560	500.000	10.1
Aroclor-1260-3	7.747	7.643	7.843	552.230	500.000	10.4
Aroclor-1260-4	7.973	7.868	8.068	570.140	500.000	14.0
Aroclor-1260-5	8.286	8.181	8.381	566.790	500.000	13.4
Decachlorobiphenyl	10.071	9.958	10.158	50.590	50.000	1.2
Tetrachloro-m-xylene	4.371	4.274	4.474	55.110	50.000	10.2

CALIBRATION VERIFICATION SUMMARY

Contract: EART12

Lab Code: CHEM Case No.: P4699 SAS No.: P4699 SDG NO.: P4699

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL03 Date Analyzed: 11/05/2024

Lab Sample No.: AR1660CCC500 Data File : PO107711.D Time Analyzed: 22:34

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.723	4.626	4.826	571.620	500.000	14.3
Aroclor-1016-2	4.742	4.645	4.845	590.130	500.000	18.0
Aroclor-1016-3	4.918	4.821	5.021	564.280	500.000	12.9
Aroclor-1016-4	4.959	4.862	5.062	505.500	500.000	1.1
Aroclor-1016-5	5.172	5.076	5.276	557.810	500.000	11.6
Aroclor-1260-1	6.203	6.107	6.307	555.170	500.000	11.0
Aroclor-1260-2	6.391	6.294	6.494	582.330	500.000	16.5
Aroclor-1260-3	6.545	6.448	6.648	547.050	500.000	9.4
Aroclor-1260-4	7.015	6.918	7.118	535.770	500.000	7.2
Aroclor-1260-5	7.257	7.159	7.359	569.080	500.000	13.8
Decachlorobiphenyl	8.637	8.539	8.739	54.680	50.000	9.4
Tetrachloro-m-xylene	3.642	3.545	3.745	58.830	50.000	17.7