

Hit Summary Sheet
SW-846

SDG No.: P4718

Order ID: P4718

Client: Portal Partners Tri-Venture

Project ID: Amtrak Sawtooth Bridges 2024

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : WB-307-SB01								
P4718-01	WB-307-SB01	SOIL	Aluminum	4030		2.69	5.57	mg/Kg
P4718-01	WB-307-SB01	SOIL	Arsenic	1.77		0.32	1.11	mg/Kg
P4718-01	WB-307-SB01	SOIL	Barium	54.2		0.71	5.57	mg/Kg
P4718-01	WB-307-SB01	SOIL	Beryllium	0.40		0.013	0.33	mg/Kg
P4718-01	WB-307-SB01	SOIL	Cadmium	1.18		0.018	0.33	mg/Kg
P4718-01	WB-307-SB01	SOIL	Calcium	10400		3.12	111	mg/Kg
P4718-01	WB-307-SB01	SOIL	Chromium	15.5		0.060	0.56	mg/Kg
P4718-01	WB-307-SB01	SOIL	Cobalt	4.91		0.065	1.67	mg/Kg
P4718-01	WB-307-SB01	SOIL	Copper	14.4		0.52	1.11	mg/Kg
P4718-01	WB-307-SB01	SOIL	Iron	9340		3.00	5.57	mg/Kg
P4718-01	WB-307-SB01	SOIL	Lead	9.59		0.17	0.67	mg/Kg
P4718-01	WB-307-SB01	SOIL	Magnesium	2140		3.82	111	mg/Kg
P4718-01	WB-307-SB01	SOIL	Manganese	239		0.079	1.11	mg/Kg
P4718-01	WB-307-SB01	SOIL	Mercury	0.0080	J	0.0070	0.015	mg/Kg
P4718-01	WB-307-SB01	SOIL	Nickel	11.1		0.10	2.23	mg/Kg
P4718-01	WB-307-SB01	SOIL	Potassium	801		32.0	111	mg/Kg
P4718-01	WB-307-SB01	SOIL	Silver	0.23	J	0.058	0.56	mg/Kg
P4718-01	WB-307-SB01	SOIL	Sodium	1200		40.2	111	mg/Kg
P4718-01	WB-307-SB01	SOIL	Vanadium	18.4		0.30	2.23	mg/Kg
P4718-01	WB-307-SB01	SOIL	Zinc	31.1		0.12	2.23	mg/Kg
Client ID : WB-307-SB02								
P4718-02	WB-307-SB02	SOIL	Aluminum	1990		2.55	5.28	mg/Kg
P4718-02	WB-307-SB02	SOIL	Arsenic	1.00	J	0.31	1.06	mg/Kg
P4718-02	WB-307-SB02	SOIL	Barium	11.7		0.68	5.28	mg/Kg
P4718-02	WB-307-SB02	SOIL	Beryllium	0.21	J	0.013	0.32	mg/Kg
P4718-02	WB-307-SB02	SOIL	Cadmium	0.62		0.017	0.32	mg/Kg
P4718-02	WB-307-SB02	SOIL	Calcium	3290		2.96	106	mg/Kg
P4718-02	WB-307-SB02	SOIL	Chromium	5.65		0.057	0.53	mg/Kg
P4718-02	WB-307-SB02	SOIL	Cobalt	2.57		0.061	1.58	mg/Kg
P4718-02	WB-307-SB02	SOIL	Copper	4.19		0.50	1.06	mg/Kg
P4718-02	WB-307-SB02	SOIL	Iron	6880		2.84	5.28	mg/Kg
P4718-02	WB-307-SB02	SOIL	Lead	3.08		0.16	0.63	mg/Kg
P4718-02	WB-307-SB02	SOIL	Magnesium	1440		3.62	106	mg/Kg
P4718-02	WB-307-SB02	SOIL	Manganese	81.0		0.075	1.06	mg/Kg
P4718-02	WB-307-SB02	SOIL	Mercury	0.018		0.0070	0.016	mg/Kg
P4718-02	WB-307-SB02	SOIL	Nickel	5.02		0.095	2.11	mg/Kg
P4718-02	WB-307-SB02	SOIL	Potassium	466		30.3	106	mg/Kg

Hit Summary Sheet
SW-846

SDG No.: P4718

Order ID: P4718

Client: Portal Partners Tri-Venture

Project ID: Amtrak Sawtooth Bridges 2024

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
P4718-02	WB-307-SB02	SOIL	Silver	0.15	J	0.055	0.53	mg/Kg
P4718-02	WB-307-SB02	SOIL	Sodium	1250		38.1	106	mg/Kg
P4718-02	WB-307-SB02	SOIL	Vanadium	8.23		0.28	2.11	mg/Kg
P4718-02	WB-307-SB02	SOIL	Zinc	11.9		0.12	2.11	mg/Kg
Client ID : WB-307-SW								
P4718-04	WB-307-SW	Water	Aluminum	389		28.3	50.0	ug/L
P4718-04	WB-307-SW	Water	Barium	34.4	J	6.28	50.0	ug/L
P4718-04	WB-307-SW	Water	Cadmium	0.19	J	0.094	3.00	ug/L
P4718-04	WB-307-SW	Water	Calcium	206000		33.0	1000	ug/L
P4718-04	WB-307-SW	Water	Chromium	1.78	J	0.66	5.00	ug/L
P4718-04	WB-307-SW	Water	Iron	521		18.5	50.0	ug/L
P4718-04	WB-307-SW	Water	Magnesium	628000		39.4	1000	ug/L
P4718-04	WB-307-SW	Water	Manganese	69.7		1.46	10.0	ug/L
P4718-04	WB-307-SW	Water	Potassium	425000		685	1000	ug/L
P4718-04	WB-307-SW	Water	Sodium	5110000	D	2370	10000	ug/L
P4718-04	WB-307-SW	Water	Vanadium	3.65	J	3.06	20.0	ug/L
P4718-04	WB-307-SW	Water	Zinc	19.8	J	1.75	20.0	ug/L



SAMPLE DATA

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	11/04/24
Project:	Amtrak Sawtooth Bridges 2024	Date Received:	11/05/24
Client Sample ID:	WB-307-SB01	SDG No.:	P4718
Lab Sample ID:	P4718-01	Matrix:	SOIL
Level (low/med):	low	% Solid:	78

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units(Dry Weight)	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	4030		1	2.69	5.57	mg/Kg	11/06/24 10:00	11/07/24 23:26	SW6010	SW3050
7440-36-0	Antimony	0.17	U	1	0.17	2.79	mg/Kg	11/06/24 10:00	11/07/24 23:26	SW6010	SW3050
7440-38-2	Arsenic	1.77		1	0.32	1.11	mg/Kg	11/06/24 10:00	11/07/24 23:26	SW6010	SW3050
7440-39-3	Barium	54.2		1	0.71	5.57	mg/Kg	11/06/24 10:00	11/07/24 23:26	SW6010	SW3050
7440-41-7	Beryllium	0.40		1	0.013	0.33	mg/Kg	11/06/24 10:00	11/07/24 23:26	SW6010	SW3050
7440-43-9	Cadmium	1.18		1	0.018	0.33	mg/Kg	11/06/24 10:00	11/07/24 23:26	SW6010	SW3050
7440-70-2	Calcium	10400		1	3.12	111	mg/Kg	11/06/24 10:00	11/07/24 23:26	SW6010	SW3050
7440-47-3	Chromium	15.5		1	0.060	0.56	mg/Kg	11/06/24 10:00	11/07/24 23:26	SW6010	SW3050
7440-48-4	Cobalt	4.91		1	0.065	1.67	mg/Kg	11/06/24 10:00	11/07/24 23:26	SW6010	SW3050
7440-50-8	Copper	14.4		1	0.52	1.11	mg/Kg	11/06/24 10:00	11/07/24 23:26	SW6010	SW3050
7439-89-6	Iron	9340		1	3.00	5.57	mg/Kg	11/06/24 10:00	11/07/24 23:26	SW6010	SW3050
7439-92-1	Lead	9.59		1	0.17	0.67	mg/Kg	11/06/24 10:00	11/07/24 23:26	SW6010	SW3050
7439-95-4	Magnesium	2140		1	3.82	111	mg/Kg	11/06/24 10:00	11/07/24 23:26	SW6010	SW3050
7439-96-5	Manganese	239		1	0.079	1.11	mg/Kg	11/06/24 10:00	11/07/24 23:26	SW6010	SW3050
7439-97-6	Mercury	0.0080	J	1	0.0070	0.015	mg/Kg	11/07/24 10:15	11/07/24 15:22	SW7471B	
7440-02-0	Nickel	11.1		1	0.10	2.23	mg/Kg	11/06/24 10:00	11/07/24 23:26	SW6010	SW3050
7440-09-7	Potassium	801	N	1	32.0	111	mg/Kg	11/06/24 10:00	11/07/24 23:26	SW6010	SW3050
7782-49-2	Selenium	0.37	U	1	0.37	1.11	mg/Kg	11/06/24 10:00	11/07/24 23:26	SW6010	SW3050
7440-22-4	Silver	0.23	J	1	0.058	0.56	mg/Kg	11/06/24 10:00	11/07/24 23:26	SW6010	SW3050
7440-23-5	Sodium	1200	N	1	40.2	111	mg/Kg	11/06/24 10:00	11/07/24 23:26	SW6010	SW3050
7440-28-0	Thallium	0.49	U	1	0.49	2.23	mg/Kg	11/06/24 10:00	11/07/24 23:26	SW6010	SW3050
7440-62-2	Vanadium	18.4		1	0.30	2.23	mg/Kg	11/06/24 10:00	11/07/24 23:26	SW6010	SW3050
7440-66-6	Zinc	31.1		1	0.12	2.23	mg/Kg	11/06/24 10:00	11/07/24 23:26	SW6010	SW3050

Color Before:	Brown	Clarity Before:	Texture:	Medium
Color After:	Yellow	Clarity After:	Artifacts:	N/A
Comments:	METALS-TAL			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N = Spiked sample recovery not within control limits

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	11/04/24
Project:	Amtrak Sawtooth Bridges 2024	Date Received:	11/05/24
Client Sample ID:	WB-307-SB02	SDG No.:	P4718
Lab Sample ID:	P4718-02	Matrix:	SOIL
Level (low/med):	low	% Solid:	77.6

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units(Dry Weight)	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	1990		1	2.55	5.28	mg/Kg	11/06/24 10:00	11/07/24 23:30	SW6010	SW3050
7440-36-0	Antimony	0.16	U	1	0.16	2.64	mg/Kg	11/06/24 10:00	11/07/24 23:30	SW6010	SW3050
7440-38-2	Arsenic	1.00	J	1	0.31	1.06	mg/Kg	11/06/24 10:00	11/07/24 23:30	SW6010	SW3050
7440-39-3	Barium	11.7		1	0.68	5.28	mg/Kg	11/06/24 10:00	11/07/24 23:30	SW6010	SW3050
7440-41-7	Beryllium	0.21	J	1	0.013	0.32	mg/Kg	11/06/24 10:00	11/07/24 23:30	SW6010	SW3050
7440-43-9	Cadmium	0.62		1	0.017	0.32	mg/Kg	11/06/24 10:00	11/07/24 23:30	SW6010	SW3050
7440-70-2	Calcium	3290		1	2.96	106	mg/Kg	11/06/24 10:00	11/07/24 23:30	SW6010	SW3050
7440-47-3	Chromium	5.65		1	0.057	0.53	mg/Kg	11/06/24 10:00	11/07/24 23:30	SW6010	SW3050
7440-48-4	Cobalt	2.57		1	0.061	1.58	mg/Kg	11/06/24 10:00	11/07/24 23:30	SW6010	SW3050
7440-50-8	Copper	4.19		1	0.50	1.06	mg/Kg	11/06/24 10:00	11/07/24 23:30	SW6010	SW3050
7439-89-6	Iron	6880		1	2.84	5.28	mg/Kg	11/06/24 10:00	11/07/24 23:30	SW6010	SW3050
7439-92-1	Lead	3.08		1	0.16	0.63	mg/Kg	11/06/24 10:00	11/07/24 23:30	SW6010	SW3050
7439-95-4	Magnesium	1440		1	3.62	106	mg/Kg	11/06/24 10:00	11/07/24 23:30	SW6010	SW3050
7439-96-5	Manganese	81.0		1	0.075	1.06	mg/Kg	11/06/24 10:00	11/07/24 23:30	SW6010	SW3050
7439-97-6	Mercury	0.018		1	0.0070	0.016	mg/Kg	11/07/24 10:15	11/07/24 15:24	SW7471B	
7440-02-0	Nickel	5.02		1	0.095	2.11	mg/Kg	11/06/24 10:00	11/07/24 23:30	SW6010	SW3050
7440-09-7	Potassium	466	N	1	30.3	106	mg/Kg	11/06/24 10:00	11/07/24 23:30	SW6010	SW3050
7782-49-2	Selenium	0.35	U	1	0.35	1.06	mg/Kg	11/06/24 10:00	11/07/24 23:30	SW6010	SW3050
7440-22-4	Silver	0.15	J	1	0.055	0.53	mg/Kg	11/06/24 10:00	11/07/24 23:30	SW6010	SW3050
7440-23-5	Sodium	1250	N	1	38.1	106	mg/Kg	11/06/24 10:00	11/07/24 23:30	SW6010	SW3050
7440-28-0	Thallium	0.47	U	1	0.47	2.11	mg/Kg	11/06/24 10:00	11/07/24 23:30	SW6010	SW3050
7440-62-2	Vanadium	8.23		1	0.28	2.11	mg/Kg	11/06/24 10:00	11/07/24 23:30	SW6010	SW3050
7440-66-6	Zinc	11.9		1	0.12	2.11	mg/Kg	11/06/24 10:00	11/07/24 23:30	SW6010	SW3050

Color Before:	Brown	Clarity Before:	Texture:	Medium
Color After:	Yellow	Clarity After:	Artifacts:	N/A
Comments:	METALS-TAL			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	11/04/24
Project:	Amtrak Sawtooth Bridges 2024	Date Received:	11/05/24
Client Sample ID:	WB-307-SW	SDG No.:	P4718
Lab Sample ID:	P4718-04	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	389	N	1	28.3	50.0	ug/L	11/07/24 12:30	11/08/24 19:42	SW6010	SW3010
7440-36-0	Antimony	2.06	UN	1	2.06	25.0	ug/L	11/07/24 12:30	11/08/24 19:42	SW6010	SW3010
7440-38-2	Arsenic	3.48	UN	1	3.48	10.0	ug/L	11/07/24 12:30	11/08/24 19:42	SW6010	SW3010
7440-39-3	Barium	34.4	J	1	6.28	50.0	ug/L	11/07/24 12:30	11/08/24 19:42	SW6010	SW3010
7440-41-7	Beryllium	0.13	U	1	0.13	3.00	ug/L	11/07/24 12:30	11/08/24 19:42	SW6010	SW3010
7440-43-9	Cadmium	0.19	J	1	0.094	3.00	ug/L	11/07/24 12:30	11/08/24 19:42	SW6010	SW3010
7440-70-2	Calcium	206000		1	33.0	1000	ug/L	11/07/24 12:30	11/08/24 19:42	SW6010	SW3010
7440-47-3	Chromium	1.78	J	1	0.66	5.00	ug/L	11/07/24 12:30	11/08/24 19:42	SW6010	SW3010
7440-48-4	Cobalt	0.50	U	1	0.50	15.0	ug/L	11/07/24 12:30	11/08/24 19:42	SW6010	SW3010
7440-50-8	Copper	7.07	U	1	7.07	10.0	ug/L	11/07/24 12:30	11/08/24 19:42	SW6010	SW3010
7439-89-6	Iron	521		1	18.5	50.0	ug/L	11/07/24 12:30	11/08/24 19:42	SW6010	SW3010
7439-92-1	Lead	3.51	U	1	3.51	6.00	ug/L	11/07/24 12:30	11/08/24 19:42	SW6010	SW3010
7439-95-4	Magnesium	628000		1	39.4	1000	ug/L	11/07/24 12:30	11/08/24 19:42	SW6010	SW3010
7439-96-5	Manganese	69.7		1	1.46	10.0	ug/L	11/07/24 12:30	11/08/24 19:42	SW6010	SW3010
7439-97-6	Mercury	0.081	U	1	0.081	0.20	ug/L	11/11/24 10:15	11/11/24 14:24	SW7470A	
7440-02-0	Nickel	0.85	U	1	0.85	20.0	ug/L	11/07/24 12:30	11/08/24 19:42	SW6010	SW3010
7440-09-7	Potassium	425000		1	685	1000	ug/L	11/07/24 12:30	11/08/24 19:42	SW6010	SW3010
7782-49-2	Selenium	5.88	UN	1	5.88	10.0	ug/L	11/07/24 12:30	11/08/24 19:42	SW6010	SW3010
7440-22-4	Silver	0.58	UN	1	0.58	5.00	ug/L	11/07/24 12:30	11/08/24 19:42	SW6010	SW3010
7440-23-5	Sodium	5110000	D	10	2370	10000	ug/L	11/07/24 12:30	11/08/24 21:18	SW6010	SW3010
7440-28-0	Thallium	2.32	U	1	2.32	20.0	ug/L	11/07/24 12:30	11/08/24 19:42	SW6010	SW3010
7440-62-2	Vanadium	3.65	J	1	3.06	20.0	ug/L	11/07/24 12:30	11/08/24 19:42	SW6010	SW3010
7440-66-6	Zinc	19.8	JN	1	1.75	20.0	ug/L	11/07/24 12:30	11/08/24 19:42	SW6010	SW3010

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	METALS-TAL			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N = Spiked sample recovery not within control limits



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Portal Partners Tri-Venture				SDG No.:	P4718				
Contract:	PORT06	Lab Code:	CHEM		Case No.:	P4718	SAS No.:	P4718		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB74	Mercury	0.20	+/-0.20	U	0.20	CV	11/07/2024	14:43	LB133334	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: P4718

Contract: PORT06

Lab Code: CHEM

Case No.: P4718

SAS No.: P4718

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB47	Mercury	0.20	+/-0.20	U	0.20	CV	11/07/2024	14:48	LB133334
CCB48	Mercury	0.20	+/-0.20	U	0.20	CV	11/07/2024	15:15	LB133334
CCB49	Mercury	0.20	+/-0.20	U	0.20	CV	11/07/2024	15:43	LB133334
CCB50	Mercury	0.20	+/-0.20	U	0.20	CV	11/07/2024	16:10	LB133334

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: P4718

Contract: PORT06

Lab Code: CHEM

Case No.: P4718

SAS No.: P4718

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB80	Mercury	0.20	+/-0.20	U	0.20	CV	11/11/2024	14:01	LB133392

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: P4718

Contract: PORT06

Lab Code: CHEM

Case No.: P4718

SAS No.: P4718

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB64	Mercury	0.20	+/-0.20	U	0.20	CV	11/11/2024	14:05	LB133392
CCB65	Mercury	0.20	+/-0.20	U	0.20	CV	11/11/2024	14:41	LB133392
CCB66	Mercury	0.20	+/-0.20	U	0.20	CV	11/11/2024	14:57	LB133392

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	11/07/2024	14:38	LB133344
	Antimony	50.0	+/-50.0	U	50.0	P	11/07/2024	14:38	LB133344
	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	14:38	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	14:38	LB133344
	Beryllium	6.00	+/-6.00	U	6.00	P	11/07/2024	14:38	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	14:38	LB133344
	Calcium	2000	+/-2000	U	2000	P	11/07/2024	14:38	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	14:38	LB133344
	Cobalt	30.0	+/-30.0	U	30.0	P	11/07/2024	14:38	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	14:38	LB133344
	Iron	100	+/-100	U	100	P	11/07/2024	14:38	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	14:38	LB133344
	Magnesium	2000	+/-2000	U	2000	P	11/07/2024	14:38	LB133344
	Manganese	20.0	+/-20.0	U	20.0	P	11/07/2024	14:38	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	14:38	LB133344
	Potassium	2000	+/-2000	U	2000	P	11/07/2024	14:38	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	14:38	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	14:38	LB133344
	Sodium	2000	+/-2000	U	2000	P	11/07/2024	14:38	LB133344
	Thallium	40.0	+/-40.0	U	40.0	P	11/07/2024	14:38	LB133344
	Vanadium	40.0	+/-40.0	U	40.0	P	11/07/2024	14:38	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	14:38	LB133344

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	11/07/2024	15:05	LB133344
	Antimony	50.0	+/-50.0	U	50.0	P	11/07/2024	15:05	LB133344
	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	15:05	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	15:05	LB133344
	Beryllium	6.00	+/-6.00	U	6.00	P	11/07/2024	15:05	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	15:05	LB133344
	Calcium	2000	+/-2000	U	2000	P	11/07/2024	15:05	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	15:05	LB133344
	Cobalt	30.0	+/-30.0	U	30.0	P	11/07/2024	15:05	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	15:05	LB133344
	Iron	100	+/-100	U	100	P	11/07/2024	15:05	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	15:05	LB133344
	Magnesium	2000	+/-2000	U	2000	P	11/07/2024	15:05	LB133344
	Manganese	20.0	+/-20.0	U	20.0	P	11/07/2024	15:05	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	15:05	LB133344
	Potassium	2000	+/-2000	U	2000	P	11/07/2024	15:05	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	15:05	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	15:05	LB133344
	Sodium	2000	+/-2000	U	2000	P	11/07/2024	15:05	LB133344
	Thallium	40.0	+/-40.0	U	40.0	P	11/07/2024	15:05	LB133344
	Vanadium	40.0	+/-40.0	U	40.0	P	11/07/2024	15:05	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	15:05	LB133344
CCB02	Aluminum	100	+/-100	U	100	P	11/07/2024	15:58	LB133344
	Antimony	50.0	+/-50.0	U	50.0	P	11/07/2024	15:58	LB133344
	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	15:58	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	15:58	LB133344
	Beryllium	6.00	+/-6.00	U	6.00	P	11/07/2024	15:58	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	15:58	LB133344
	Calcium	2000	+/-2000	U	2000	P	11/07/2024	15:58	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	15:58	LB133344
	Cobalt	30.0	+/-30.0	U	30.0	P	11/07/2024	15:58	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	15:58	LB133344
	Iron	100	+/-100	U	100	P	11/07/2024	15:58	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	15:58	LB133344
	Magnesium	2000	+/-2000	U	2000	P	11/07/2024	15:58	LB133344
	Manganese	20.0	+/-20.0	U	20.0	P	11/07/2024	15:58	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	15:58	LB133344
	Potassium	2000	+/-2000	U	2000	P	11/07/2024	15:58	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	15:58	LB133344

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	15:58	LB133344
	Sodium	2000	+/-2000	U	2000	P	11/07/2024	15:58	LB133344
	Thallium	40.0	+/-40.0	U	40.0	P	11/07/2024	15:58	LB133344
	Vanadium	40.0	+/-40.0	U	40.0	P	11/07/2024	15:58	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	15:58	LB133344
CCB03	Aluminum	100	+/-100	U	100	P	11/07/2024	16:49	LB133344
	Antimony	50.0	+/-50.0	U	50.0	P	11/07/2024	16:49	LB133344
	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	16:49	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	16:49	LB133344
	Beryllium	6.00	+/-6.00	U	6.00	P	11/07/2024	16:49	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	16:49	LB133344
	Calcium	2000	+/-2000	U	2000	P	11/07/2024	16:49	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	16:49	LB133344
	Cobalt	30.0	+/-30.0	U	30.0	P	11/07/2024	16:49	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	16:49	LB133344
	Iron	100	+/-100	U	100	P	11/07/2024	16:49	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	16:49	LB133344
	Magnesium	2000	+/-2000	U	2000	P	11/07/2024	16:49	LB133344
	Manganese	20.0	+/-20.0	U	20.0	P	11/07/2024	16:49	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	16:49	LB133344
	Potassium	2000	+/-2000	U	2000	P	11/07/2024	16:49	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	16:49	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	16:49	LB133344
	Sodium	2000	+/-2000	U	2000	P	11/07/2024	16:49	LB133344
	Thallium	40.0	+/-40.0	U	40.0	P	11/07/2024	16:49	LB133344
	Vanadium	40.0	+/-40.0	U	40.0	P	11/07/2024	16:49	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	16:49	LB133344
CCB04	Aluminum	100	+/-100	U	100	P	11/07/2024	18:04	LB133344
	Antimony	50.0	+/-50.0	U	50.0	P	11/07/2024	18:04	LB133344
	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	18:04	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	18:04	LB133344
	Beryllium	6.00	+/-6.00	U	6.00	P	11/07/2024	18:04	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	18:04	LB133344
	Calcium	2000	+/-2000	U	2000	P	11/07/2024	18:04	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	18:04	LB133344
	Cobalt	30.0	+/-30.0	U	30.0	P	11/07/2024	18:04	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	18:04	LB133344
	Iron	100	+/-100	U	100	P	11/07/2024	18:04	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	18:04	LB133344

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	11/07/2024	18:04	LB133344
	Manganese	20.0	+/-20.0	U	20.0	P	11/07/2024	18:04	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	18:04	LB133344
	Potassium	2000	+/-2000	U	2000	P	11/07/2024	18:04	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	18:04	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	18:04	LB133344
	Sodium	2000	+/-2000	U	2000	P	11/07/2024	18:04	LB133344
	Thallium	40.0	+/-40.0	U	40.0	P	11/07/2024	18:04	LB133344
	Vanadium	40.0	+/-40.0	U	40.0	P	11/07/2024	18:04	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	18:04	LB133344
CCB05	Aluminum	100	+/-100	U	100	P	11/07/2024	18:37	LB133344
	Antimony	50.0	+/-50.0	U	50.0	P	11/07/2024	18:37	LB133344
	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	18:37	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	18:37	LB133344
	Beryllium	6.00	+/-6.00	U	6.00	P	11/07/2024	18:37	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	18:37	LB133344
	Calcium	2000	+/-2000	U	2000	P	11/07/2024	18:37	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	18:37	LB133344
	Cobalt	30.0	+/-30.0	U	30.0	P	11/07/2024	18:37	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	18:37	LB133344
	Iron	100	+/-100	U	100	P	11/07/2024	18:37	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	18:37	LB133344
	Magnesium	2000	+/-2000	U	2000	P	11/07/2024	18:37	LB133344
	Manganese	20.0	+/-20.0	U	20.0	P	11/07/2024	18:37	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	18:37	LB133344
	Potassium	2000	+/-2000	U	2000	P	11/07/2024	18:37	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	18:37	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	18:37	LB133344
	Sodium	2000	+/-2000	U	2000	P	11/07/2024	18:37	LB133344
	Thallium	40.0	+/-40.0	U	40.0	P	11/07/2024	18:37	LB133344
	Vanadium	40.0	+/-40.0	U	40.0	P	11/07/2024	18:37	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	18:37	LB133344
CCB06	Aluminum	100	+/-100	U	100	P	11/07/2024	19:12	LB133344
	Antimony	50.0	+/-50.0	U	50.0	P	11/07/2024	19:12	LB133344
	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	19:12	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	19:12	LB133344
	Beryllium	6.00	+/-6.00	U	6.00	P	11/07/2024	19:12	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	19:12	LB133344
	Calcium	2000	+/-2000	U	2000	P	11/07/2024	19:12	LB133344

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	19:12	LB133344
	Cobalt	30.0	+/-30.0	U	30.0	P	11/07/2024	19:12	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	19:12	LB133344
	Iron	100	+/-100	U	100	P	11/07/2024	19:12	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	19:12	LB133344
	Magnesium	2000	+/-2000	U	2000	P	11/07/2024	19:12	LB133344
	Manganese	20.0	+/-20.0	U	20.0	P	11/07/2024	19:12	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	19:12	LB133344
	Potassium	2000	+/-2000	U	2000	P	11/07/2024	19:12	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	19:12	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	19:12	LB133344
	Sodium	2000	+/-2000	U	2000	P	11/07/2024	19:12	LB133344
	Thallium	40.0	+/-40.0	U	40.0	P	11/07/2024	19:12	LB133344
	Vanadium	40.0	+/-40.0	U	40.0	P	11/07/2024	19:12	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	19:12	LB133344
CCB07	Aluminum	100	+/-100	U	100	P	11/07/2024	20:02	LB133344
	Antimony	50.0	+/-50.0	U	50.0	P	11/07/2024	20:02	LB133344
	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	20:02	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	20:02	LB133344
	Beryllium	6.00	+/-6.00	U	6.00	P	11/07/2024	20:02	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	20:02	LB133344
	Calcium	2000	+/-2000	U	2000	P	11/07/2024	20:02	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	20:02	LB133344
	Cobalt	30.0	+/-30.0	U	30.0	P	11/07/2024	20:02	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	20:02	LB133344
	Iron	100	+/-100	U	100	P	11/07/2024	20:02	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	20:02	LB133344
	Magnesium	2000	+/-2000	U	2000	P	11/07/2024	20:02	LB133344
	Manganese	20.0	+/-20.0	U	20.0	P	11/07/2024	20:02	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	20:02	LB133344
	Potassium	2000	+/-2000	U	2000	P	11/07/2024	20:02	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	20:02	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	20:02	LB133344
	Sodium	2000	+/-2000	U	2000	P	11/07/2024	20:02	LB133344
	Thallium	40.0	+/-40.0	U	40.0	P	11/07/2024	20:02	LB133344
	Vanadium	40.0	+/-40.0	U	40.0	P	11/07/2024	20:02	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	20:02	LB133344
CCB08	Aluminum	100	+/-100	U	100	P	11/07/2024	21:02	LB133344
	Antimony	50.0	+/-50.0	U	50.0	P	11/07/2024	21:02	LB133344

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	21:02	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	21:02	LB133344
	Beryllium	6.00	+/-6.00	U	6.00	P	11/07/2024	21:02	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	21:02	LB133344
	Calcium	2000	+/-2000	U	2000	P	11/07/2024	21:02	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	21:02	LB133344
	Cobalt	30.0	+/-30.0	U	30.0	P	11/07/2024	21:02	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	21:02	LB133344
	Iron	100	+/-100	U	100	P	11/07/2024	21:02	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	21:02	LB133344
	Magnesium	2000	+/-2000	U	2000	P	11/07/2024	21:02	LB133344
	Manganese	20.0	+/-20.0	U	20.0	P	11/07/2024	21:02	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	21:02	LB133344
	Potassium	2000	+/-2000	U	2000	P	11/07/2024	21:02	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	21:02	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	21:02	LB133344
	Sodium	2000	+/-2000	U	2000	P	11/07/2024	21:02	LB133344
	Thallium	40.0	+/-40.0	U	40.0	P	11/07/2024	21:02	LB133344
	Vanadium	40.0	+/-40.0	U	40.0	P	11/07/2024	21:02	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	21:02	LB133344
CCB09	Aluminum	100	+/-100	U	100	P	11/07/2024	21:55	LB133344
	Antimony	50.0	+/-50.0	U	50.0	P	11/07/2024	21:55	LB133344
	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	21:55	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	21:55	LB133344
	Beryllium	6.00	+/-6.00	U	6.00	P	11/07/2024	21:55	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	21:55	LB133344
	Calcium	2000	+/-2000	U	2000	P	11/07/2024	21:55	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	21:55	LB133344
	Cobalt	30.0	+/-30.0	U	30.0	P	11/07/2024	21:55	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	21:55	LB133344
	Iron	100	+/-100	U	100	P	11/07/2024	21:55	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	21:55	LB133344
	Magnesium	2000	+/-2000	U	2000	P	11/07/2024	21:55	LB133344
	Manganese	20.0	+/-20.0	U	20.0	P	11/07/2024	21:55	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	21:55	LB133344
	Potassium	2000	+/-2000	U	2000	P	11/07/2024	21:55	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	21:55	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	21:55	LB133344
	Sodium	2000	+/-2000	U	2000	P	11/07/2024	21:55	LB133344

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Thallium	40.0	+/-40.0	U	40.0	P	11/07/2024	21:55	LB133344
	Vanadium	40.0	+/-40.0	U	40.0	P	11/07/2024	21:55	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	21:55	LB133344
CCB10	Aluminum	100	+/-100	U	100	P	11/07/2024	22:56	LB133344
	Antimony	50.0	+/-50.0	U	50.0	P	11/07/2024	22:56	LB133344
	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	22:56	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	22:56	LB133344
	Beryllium	6.00	+/-6.00	U	6.00	P	11/07/2024	22:56	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	22:56	LB133344
	Calcium	2000	+/-2000	U	2000	P	11/07/2024	22:56	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	22:56	LB133344
	Cobalt	30.0	+/-30.0	U	30.0	P	11/07/2024	22:56	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	22:56	LB133344
	Iron	100	+/-100	U	100	P	11/07/2024	22:56	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	22:56	LB133344
	Magnesium	2000	+/-2000	U	2000	P	11/07/2024	22:56	LB133344
	Manganese	20.0	+/-20.0	U	20.0	P	11/07/2024	22:56	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	22:56	LB133344
	Potassium	2000	+/-2000	U	2000	P	11/07/2024	22:56	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	22:56	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	22:56	LB133344
	Sodium	2000	+/-2000	U	2000	P	11/07/2024	22:56	LB133344
	Thallium	40.0	+/-40.0	U	40.0	P	11/07/2024	22:56	LB133344
	Vanadium	40.0	+/-40.0	U	40.0	P	11/07/2024	22:56	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	22:56	LB133344
CCB11	Aluminum	100	+/-100	U	100	P	11/07/2024	23:47	LB133344
	Antimony	50.0	+/-50.0	U	50.0	P	11/07/2024	23:47	LB133344
	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	23:47	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	23:47	LB133344
	Beryllium	6.00	+/-6.00	U	6.00	P	11/07/2024	23:47	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	23:47	LB133344
	Calcium	2000	+/-2000	U	2000	P	11/07/2024	23:47	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	23:47	LB133344
	Cobalt	30.0	+/-30.0	U	30.0	P	11/07/2024	23:47	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	23:47	LB133344
	Iron	100	+/-100	U	100	P	11/07/2024	23:47	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	23:47	LB133344
	Magnesium	2000	+/-2000	U	2000	P	11/07/2024	23:47	LB133344
	Manganese	20.0	+/-20.0	U	20.0	P	11/07/2024	23:47	LB133344

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB11	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	23:47	LB133344
	Potassium	2000	+/-2000	U	2000	P	11/07/2024	23:47	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	23:47	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	23:47	LB133344
	Sodium	2000	+/-2000	U	2000	P	11/07/2024	23:47	LB133344
	Thallium	40.0	+/-40.0	U	40.0	P	11/07/2024	23:47	LB133344
	Vanadium	40.0	+/-40.0	U	40.0	P	11/07/2024	23:47	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	23:47	LB133344
CCB12	Aluminum	100	+/-100	U	100	P	11/08/2024	00:43	LB133344
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	00:43	LB133344
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	00:43	LB133344
	Barium	100	+/-100	U	100	P	11/08/2024	00:43	LB133344
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	00:43	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	00:43	LB133344
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	00:43	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	00:43	LB133344
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	00:43	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	00:43	LB133344
	Iron	100	+/-100	U	100	P	11/08/2024	00:43	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	00:43	LB133344
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	00:43	LB133344
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	00:43	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	00:43	LB133344
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	00:43	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	00:43	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	00:43	LB133344
	Sodium	2000	+/-2000	U	2000	P	11/08/2024	00:43	LB133344
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	00:43	LB133344
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	00:43	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	00:43	LB133344
CCB13	Aluminum	100	+/-100	U	100	P	11/08/2024	01:01	LB133344
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	01:01	LB133344
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	01:01	LB133344
	Barium	100	+/-100	U	100	P	11/08/2024	01:01	LB133344
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	01:01	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	01:01	LB133344
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	01:01	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	01:01	LB133344
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	01:01	LB133344

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB13	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	01:01	LB133344
	Iron	100	+/-100	U	100	P	11/08/2024	01:01	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	01:01	LB133344
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	01:01	LB133344
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	01:01	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	01:01	LB133344
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	01:01	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	01:01	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	01:01	LB133344
	Sodium	2000	+/-2000	U	2000	P	11/08/2024	01:01	LB133344
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	01:01	LB133344
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	01:01	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	01:01	LB133344

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	11/08/2024	14:16	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	14:16	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	14:16	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	14:16	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	14:16	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	14:16	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	14:16	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	14:16	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	14:16	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	14:16	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	14:16	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	14:16	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	14:16	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	14:16	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	14:16	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	14:16	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	14:16	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	14:16	LB133365
	Sodium	2000	+/-2000	U	2000	P	11/08/2024	14:16	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	14:16	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	14:16	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	14:16	LB133365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	11/08/2024	14:50	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	14:50	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	14:50	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	14:50	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	14:50	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	14:50	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	14:50	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	14:50	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	14:50	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	14:50	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	14:50	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	14:50	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	14:50	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	14:50	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	14:50	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	14:50	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	14:50	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	14:50	LB133365
	Sodium	2000	+/-2000	U	2000	P	11/08/2024	14:50	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	14:50	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	14:50	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	14:50	LB133365
CCB02	Aluminum	100	+/-100	U	100	P	11/08/2024	15:33	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	15:33	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	15:33	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	15:33	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	15:33	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	15:33	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	15:33	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	15:33	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	15:33	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	15:33	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	15:33	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	15:33	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	15:33	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	15:33	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	15:33	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	15:33	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	15:33	LB133365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	15:33	LB133365
	Sodium	2000	+/-2000	U	2000	P	11/08/2024	15:33	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	15:33	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	15:33	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	15:33	LB133365
CCB03	Aluminum	100	+/-100	U	100	P	11/08/2024	16:28	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	16:28	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	16:28	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	16:28	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	16:28	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	16:28	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	16:28	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	16:28	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	16:28	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	16:28	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	16:28	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	16:28	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	16:28	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	16:28	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	16:28	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	16:28	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	16:28	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	16:28	LB133365
	Sodium	2000	+/-2000	U	2000	P	11/08/2024	16:28	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	16:28	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	16:28	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	16:28	LB133365
CCB04	Aluminum	100	+/-100	U	100	P	11/08/2024	17:22	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	17:22	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	17:22	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	17:22	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	17:22	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	17:22	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	17:22	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	17:22	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	17:22	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	17:22	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	17:22	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	17:22	LB133365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	17:22	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	17:22	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	17:22	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	17:22	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	17:22	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	17:22	LB133365
	Sodium	2000	+/-2000	U	2000	P	11/08/2024	17:22	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	17:22	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	17:22	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	17:22	LB133365
CCB05	Aluminum	100	+/-100	U	100	P	11/08/2024	18:08	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	18:08	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	18:08	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	18:08	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	18:08	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	18:08	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	18:08	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	18:08	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	18:08	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	18:08	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	18:08	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	18:08	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	18:08	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	18:08	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	18:08	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	18:08	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	18:08	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	18:08	LB133365
	Sodium	2000	+/-2000	U	2000	P	11/08/2024	18:08	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	18:08	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	18:08	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	18:08	LB133365
CCB06	Aluminum	100	+/-100	U	100	P	11/08/2024	18:59	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	18:59	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	18:59	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	18:59	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	18:59	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	18:59	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	18:59	LB133365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	18:59	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	18:59	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	18:59	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	18:59	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	18:59	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	18:59	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	18:59	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	18:59	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	18:59	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	18:59	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	18:59	LB133365
	Sodium	2000	+/-2000	U	2000	P	11/08/2024	18:59	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	18:59	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	18:59	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	18:59	LB133365
CCB07	Aluminum	100	+/-100	U	100	P	11/08/2024	19:51	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	19:51	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	19:51	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	19:51	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	19:51	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	19:51	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	19:51	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	19:51	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	19:51	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	19:51	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	19:51	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	19:51	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	19:51	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	19:51	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	19:51	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	19:51	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	19:51	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	19:51	LB133365
	Sodium	941	+/-2000	J	2000	P	11/08/2024	19:51	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	19:51	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	19:51	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	19:51	LB133365
CCB08	Aluminum	100	+/-100	U	100	P	11/08/2024	20:57	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	20:57	LB133365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	20:57	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	20:57	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	20:57	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	20:57	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	20:57	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	20:57	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	20:57	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	20:57	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	20:57	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	20:57	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	20:57	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	20:57	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	20:57	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	20:57	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	20:57	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	20:57	LB133365
	Sodium	720	+/-2000	J	2000	P	11/08/2024	20:57	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	20:57	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	20:57	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	20:57	LB133365
CCB09	Aluminum	100	+/-100	U	100	P	11/08/2024	21:55	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	21:55	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	21:55	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	21:55	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	21:55	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	21:55	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	21:55	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	21:55	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	21:55	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	21:55	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	21:55	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	21:55	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	21:55	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	21:55	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	21:55	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	21:55	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	21:55	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	21:55	LB133365
	Sodium	2000	+/-2000	U	2000	P	11/08/2024	21:55	LB133365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	21:55	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	21:55	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	21:55	LB133365
CCB10	Aluminum	100	+/-100	U	100	P	11/08/2024	23:35	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	23:35	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	23:35	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	23:35	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	23:35	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	23:35	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	23:35	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	23:35	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	23:35	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	23:35	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	23:35	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	23:35	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	23:35	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	23:35	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	23:35	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	23:35	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	23:35	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	23:35	LB133365
	Sodium	742	+/-2000	J	2000	P	11/08/2024	23:35	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	23:35	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	23:35	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	23:35	LB133365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	11/19/2024	11:38	LB133524
	Antimony	50.0	+/-50.0	U	50.0	P	11/19/2024	11:38	LB133524
	Arsenic	20.0	+/-20.0	U	20.0	P	11/19/2024	11:38	LB133524
	Barium	100	+/-100	U	100	P	11/19/2024	11:38	LB133524
	Beryllium	6.00	+/-6.00	U	6.00	P	11/19/2024	11:38	LB133524
	Cadmium	6.00	+/-6.00	U	6.00	P	11/19/2024	11:38	LB133524
	Calcium	2000	+/-2000	U	2000	P	11/19/2024	11:38	LB133524
	Chromium	10.0	+/-10.0	U	10.0	P	11/19/2024	11:38	LB133524
	Cobalt	30.0	+/-30.0	U	30.0	P	11/19/2024	11:38	LB133524
	Copper	20.0	+/-20.0	U	20.0	P	11/19/2024	11:38	LB133524
	Iron	100	+/-100	U	100	P	11/19/2024	11:38	LB133524
	Lead	12.0	+/-12.0	U	12.0	P	11/19/2024	11:38	LB133524
	Magnesium	2000	+/-2000	U	2000	P	11/19/2024	11:38	LB133524
	Manganese	20.0	+/-20.0	U	20.0	P	11/19/2024	11:38	LB133524
	Nickel	40.0	+/-40.0	U	40.0	P	11/19/2024	11:38	LB133524
	Potassium	2000	+/-2000	U	2000	P	11/19/2024	11:38	LB133524
	Selenium	20.0	+/-20.0	U	20.0	P	11/19/2024	11:38	LB133524
	Silver	10.0	+/-10.0	U	10.0	P	11/19/2024	11:38	LB133524
	Sodium	2000	+/-2000	U	2000	P	11/19/2024	11:38	LB133524
	Thallium	40.0	+/-40.0	U	40.0	P	11/19/2024	11:38	LB133524
	Vanadium	40.0	+/-40.0	U	40.0	P	11/19/2024	11:38	LB133524
	Zinc	40.0	+/-40.0	U	40.0	P	11/19/2024	11:38	LB133524

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	11/19/2024	12:48	LB133524
	Antimony	50.0	+/-50.0	U	50.0	P	11/19/2024	12:48	LB133524
	Arsenic	20.0	+/-20.0	U	20.0	P	11/19/2024	12:48	LB133524
	Barium	100	+/-100	U	100	P	11/19/2024	12:48	LB133524
	Beryllium	6.00	+/-6.00	U	6.00	P	11/19/2024	12:48	LB133524
	Cadmium	6.00	+/-6.00	U	6.00	P	11/19/2024	12:48	LB133524
	Calcium	2000	+/-2000	U	2000	P	11/19/2024	12:48	LB133524
	Chromium	10.0	+/-10.0	U	10.0	P	11/19/2024	12:48	LB133524
	Cobalt	30.0	+/-30.0	U	30.0	P	11/19/2024	12:48	LB133524
	Copper	20.0	+/-20.0	U	20.0	P	11/19/2024	12:48	LB133524
	Iron	100	+/-100	U	100	P	11/19/2024	12:48	LB133524
	Lead	12.0	+/-12.0	U	12.0	P	11/19/2024	12:48	LB133524
	Magnesium	2000	+/-2000	U	2000	P	11/19/2024	12:48	LB133524
	Manganese	20.0	+/-20.0	U	20.0	P	11/19/2024	12:48	LB133524
	Nickel	40.0	+/-40.0	U	40.0	P	11/19/2024	12:48	LB133524
	Potassium	2000	+/-2000	U	2000	P	11/19/2024	12:48	LB133524
	Selenium	20.0	+/-20.0	U	20.0	P	11/19/2024	12:48	LB133524
	Silver	10.0	+/-10.0	U	10.0	P	11/19/2024	12:48	LB133524
	Sodium	2000	+/-2000	U	2000	P	11/19/2024	12:48	LB133524
	Thallium	40.0	+/-40.0	U	40.0	P	11/19/2024	12:48	LB133524
	Vanadium	40.0	+/-40.0	U	40.0	P	11/19/2024	12:48	LB133524
	Zinc	40.0	+/-40.0	U	40.0	P	11/19/2024	12:48	LB133524
CCB02	Aluminum	100	+/-100	U	100	P	11/19/2024	13:38	LB133524
	Antimony	50.0	+/-50.0	U	50.0	P	11/19/2024	13:38	LB133524
	Arsenic	20.0	+/-20.0	U	20.0	P	11/19/2024	13:38	LB133524
	Barium	100	+/-100	U	100	P	11/19/2024	13:38	LB133524
	Beryllium	6.00	+/-6.00	U	6.00	P	11/19/2024	13:38	LB133524
	Cadmium	6.00	+/-6.00	U	6.00	P	11/19/2024	13:38	LB133524
	Calcium	2000	+/-2000	U	2000	P	11/19/2024	13:38	LB133524
	Chromium	10.0	+/-10.0	U	10.0	P	11/19/2024	13:38	LB133524
	Cobalt	30.0	+/-30.0	U	30.0	P	11/19/2024	13:38	LB133524
	Copper	20.0	+/-20.0	U	20.0	P	11/19/2024	13:38	LB133524
	Iron	100	+/-100	U	100	P	11/19/2024	13:38	LB133524
	Lead	12.0	+/-12.0	U	12.0	P	11/19/2024	13:38	LB133524
	Magnesium	2000	+/-2000	U	2000	P	11/19/2024	13:38	LB133524
	Manganese	20.0	+/-20.0	U	20.0	P	11/19/2024	13:38	LB133524
	Nickel	40.0	+/-40.0	U	40.0	P	11/19/2024	13:38	LB133524
	Potassium	2000	+/-2000	U	2000	P	11/19/2024	13:38	LB133524
	Selenium	20.0	+/-20.0	U	20.0	P	11/19/2024	13:38	LB133524

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	11/19/2024	13:38	LB133524
	Sodium	2000	+/-2000	U	2000	P	11/19/2024	13:38	LB133524
	Thallium	40.0	+/-40.0	U	40.0	P	11/19/2024	13:38	LB133524
	Vanadium	40.0	+/-40.0	U	40.0	P	11/19/2024	13:38	LB133524
	Zinc	40.0	+/-40.0	U	40.0	P	11/19/2024	13:38	LB133524
CCB03	Aluminum	100	+/-100	U	100	P	11/19/2024	14:28	LB133524
	Antimony	50.0	+/-50.0	U	50.0	P	11/19/2024	14:28	LB133524
	Arsenic	20.0	+/-20.0	U	20.0	P	11/19/2024	14:28	LB133524
	Barium	100	+/-100	U	100	P	11/19/2024	14:28	LB133524
	Beryllium	6.00	+/-6.00	U	6.00	P	11/19/2024	14:28	LB133524
	Cadmium	6.00	+/-6.00	U	6.00	P	11/19/2024	14:28	LB133524
	Calcium	2000	+/-2000	U	2000	P	11/19/2024	14:28	LB133524
	Chromium	10.0	+/-10.0	U	10.0	P	11/19/2024	14:28	LB133524
	Cobalt	30.0	+/-30.0	U	30.0	P	11/19/2024	14:28	LB133524
	Copper	20.0	+/-20.0	U	20.0	P	11/19/2024	14:28	LB133524
	Iron	100	+/-100	U	100	P	11/19/2024	14:28	LB133524
	Lead	12.0	+/-12.0	U	12.0	P	11/19/2024	14:28	LB133524
	Magnesium	2000	+/-2000	U	2000	P	11/19/2024	14:28	LB133524
	Manganese	20.0	+/-20.0	U	20.0	P	11/19/2024	14:28	LB133524
	Nickel	40.0	+/-40.0	U	40.0	P	11/19/2024	14:28	LB133524
	Potassium	2000	+/-2000	U	2000	P	11/19/2024	14:28	LB133524
	Selenium	20.0	+/-20.0	U	20.0	P	11/19/2024	14:28	LB133524
	Silver	10.0	+/-10.0	U	10.0	P	11/19/2024	14:28	LB133524
	Sodium	2000	+/-2000	U	2000	P	11/19/2024	14:28	LB133524
	Thallium	40.0	+/-40.0	U	40.0	P	11/19/2024	14:28	LB133524
	Vanadium	40.0	+/-40.0	U	40.0	P	11/19/2024	14:28	LB133524
	Zinc	40.0	+/-40.0	U	40.0	P	11/19/2024	14:28	LB133524
CCB04	Aluminum	100	+/-100	U	100	P	11/19/2024	15:36	LB133524
	Antimony	50.0	+/-50.0	U	50.0	P	11/19/2024	15:36	LB133524
	Arsenic	20.0	+/-20.0	U	20.0	P	11/19/2024	15:36	LB133524
	Barium	100	+/-100	U	100	P	11/19/2024	15:36	LB133524
	Beryllium	6.00	+/-6.00	U	6.00	P	11/19/2024	15:36	LB133524
	Cadmium	6.00	+/-6.00	U	6.00	P	11/19/2024	15:36	LB133524
	Calcium	2000	+/-2000	U	2000	P	11/19/2024	15:36	LB133524
	Chromium	10.0	+/-10.0	U	10.0	P	11/19/2024	15:36	LB133524
	Cobalt	30.0	+/-30.0	U	30.0	P	11/19/2024	15:36	LB133524
	Copper	20.0	+/-20.0	U	20.0	P	11/19/2024	15:36	LB133524
	Iron	100	+/-100	U	100	P	11/19/2024	15:36	LB133524
	Lead	12.0	+/-12.0	U	12.0	P	11/19/2024	15:36	LB133524

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	11/19/2024	15:36	LB133524
	Manganese	20.0	+/-20.0	U	20.0	P	11/19/2024	15:36	LB133524
	Nickel	40.0	+/-40.0	U	40.0	P	11/19/2024	15:36	LB133524
	Potassium	2000	+/-2000	U	2000	P	11/19/2024	15:36	LB133524
	Selenium	20.0	+/-20.0	U	20.0	P	11/19/2024	15:36	LB133524
	Silver	10.0	+/-10.0	U	10.0	P	11/19/2024	15:36	LB133524
	Sodium	2000	+/-2000	U	2000	P	11/19/2024	15:36	LB133524
	Thallium	40.0	+/-40.0	U	40.0	P	11/19/2024	15:36	LB133524
	Vanadium	40.0	+/-40.0	U	40.0	P	11/19/2024	15:36	LB133524
	Zinc	40.0	+/-40.0	U	40.0	P	11/19/2024	15:36	LB133524
CCB05	Aluminum	100	+/-100	U	100	P	11/19/2024	16:22	LB133524
	Antimony	50.0	+/-50.0	U	50.0	P	11/19/2024	16:22	LB133524
	Arsenic	20.0	+/-20.0	U	20.0	P	11/19/2024	16:22	LB133524
	Barium	100	+/-100	U	100	P	11/19/2024	16:22	LB133524
	Beryllium	6.00	+/-6.00	U	6.00	P	11/19/2024	16:22	LB133524
	Cadmium	6.00	+/-6.00	U	6.00	P	11/19/2024	16:22	LB133524
	Calcium	2000	+/-2000	U	2000	P	11/19/2024	16:22	LB133524
	Chromium	10.0	+/-10.0	U	10.0	P	11/19/2024	16:22	LB133524
	Cobalt	30.0	+/-30.0	U	30.0	P	11/19/2024	16:22	LB133524
	Copper	20.0	+/-20.0	U	20.0	P	11/19/2024	16:22	LB133524
	Iron	100	+/-100	U	100	P	11/19/2024	16:22	LB133524
	Lead	12.0	+/-12.0	U	12.0	P	11/19/2024	16:22	LB133524
	Magnesium	2000	+/-2000	U	2000	P	11/19/2024	16:22	LB133524
	Manganese	20.0	+/-20.0	U	20.0	P	11/19/2024	16:22	LB133524
	Nickel	40.0	+/-40.0	U	40.0	P	11/19/2024	16:22	LB133524
	Potassium	2000	+/-2000	U	2000	P	11/19/2024	16:22	LB133524
	Selenium	20.0	+/-20.0	U	20.0	P	11/19/2024	16:22	LB133524
	Silver	10.0	+/-10.0	U	10.0	P	11/19/2024	16:22	LB133524
	Sodium	2000	+/-2000	U	2000	P	11/19/2024	16:22	LB133524
	Thallium	40.0	+/-40.0	U	40.0	P	11/19/2024	16:22	LB133524
	Vanadium	40.0	+/-40.0	U	40.0	P	11/19/2024	16:22	LB133524
	Zinc	40.0	+/-40.0	U	40.0	P	11/19/2024	16:22	LB133524
CCB06	Aluminum	100	+/-100	U	100	P	11/19/2024	17:24	LB133524
	Antimony	50.0	+/-50.0	U	50.0	P	11/19/2024	17:24	LB133524
	Arsenic	20.0	+/-20.0	U	20.0	P	11/19/2024	17:24	LB133524
	Barium	22.0	+/-100	J	100	P	11/19/2024	17:24	LB133524
	Beryllium	6.00	+/-6.00	U	6.00	P	11/19/2024	17:24	LB133524
	Cadmium	6.00	+/-6.00	U	6.00	P	11/19/2024	17:24	LB133524
	Calcium	2000	+/-2000	U	2000	P	11/19/2024	17:24	LB133524

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	11/19/2024	17:24	LB133524
	Cobalt	30.0	+/-30.0	U	30.0	P	11/19/2024	17:24	LB133524
	Copper	20.0	+/-20.0	U	20.0	P	11/19/2024	17:24	LB133524
	Iron	100	+/-100	U	100	P	11/19/2024	17:24	LB133524
	Lead	12.0	+/-12.0	U	12.0	P	11/19/2024	17:24	LB133524
	Magnesium	2000	+/-2000	U	2000	P	11/19/2024	17:24	LB133524
	Manganese	6.51	+/-20.0	J	20.0	P	11/19/2024	17:24	LB133524
	Nickel	40.0	+/-40.0	U	40.0	P	11/19/2024	17:24	LB133524
	Potassium	2000	+/-2000	U	2000	P	11/19/2024	17:24	LB133524
	Selenium	20.0	+/-20.0	U	20.0	P	11/19/2024	17:24	LB133524
	Silver	10.0	+/-10.0	U	10.0	P	11/19/2024	17:24	LB133524
	Sodium	2000	+/-2000	U	2000	P	11/19/2024	17:24	LB133524
	Thallium	40.0	+/-40.0	U	40.0	P	11/19/2024	17:24	LB133524
	Vanadium	40.0	+/-40.0	U	40.0	P	11/19/2024	17:24	LB133524
	Zinc	40.0	+/-40.0	U	40.0	P	11/19/2024	17:24	LB133524
CCB07	Aluminum	100	+/-100	U	100	P	11/19/2024	18:05	LB133524
	Antimony	50.0	+/-50.0	U	50.0	P	11/19/2024	18:05	LB133524
	Arsenic	20.0	+/-20.0	U	20.0	P	11/19/2024	18:05	LB133524
	Barium	25.8	+/-100	J	100	P	11/19/2024	18:05	LB133524
	Beryllium	6.00	+/-6.00	U	6.00	P	11/19/2024	18:05	LB133524
	Cadmium	0.62	+/-6.00	J	6.00	P	11/19/2024	18:05	LB133524
	Calcium	81.8	+/-2000	J	2000	P	11/19/2024	18:05	LB133524
	Chromium	10.0	+/-10.0	U	10.0	P	11/19/2024	18:05	LB133524
	Cobalt	30.0	+/-30.0	U	30.0	P	11/19/2024	18:05	LB133524
	Copper	20.0	+/-20.0	U	20.0	P	11/19/2024	18:05	LB133524
	Iron	100	+/-100	U	100	P	11/19/2024	18:05	LB133524
	Lead	12.0	+/-12.0	U	12.0	P	11/19/2024	18:05	LB133524
	Magnesium	2000	+/-2000	U	2000	P	11/19/2024	18:05	LB133524
	Manganese	7.68	+/-20.0	J	20.0	P	11/19/2024	18:05	LB133524
	Nickel	40.0	+/-40.0	U	40.0	P	11/19/2024	18:05	LB133524
	Potassium	2000	+/-2000	U	2000	P	11/19/2024	18:05	LB133524
	Selenium	20.0	+/-20.0	U	20.0	P	11/19/2024	18:05	LB133524
	Silver	10.0	+/-10.0	U	10.0	P	11/19/2024	18:05	LB133524
	Sodium	2000	+/-2000	U	2000	P	11/19/2024	18:05	LB133524
	Thallium	40.0	+/-40.0	U	40.0	P	11/19/2024	18:05	LB133524
	Vanadium	6.34	+/-40.0	J	40.0	P	11/19/2024	18:05	LB133524
	Zinc	40.0	+/-40.0	U	40.0	P	11/19/2024	18:05	LB133524

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: P4718

Instrument: CV1

Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB164762BL		SOLID		Batch Number:	PB164762		Prep Date:	11/07/2024	
	Mercury	0.013	<0.013	U	0.013	CV	11/07/2024	14:57	LB133334
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB164872BL		WATER		Batch Number:	PB164872		Prep Date:	11/11/2024	
	Mercury	0.20	<0.20	U	0.20	CV	11/11/2024	14:20	LB133392

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: P4718

Instrument: P4

Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB164708BL	SOLID			Batch Number:	PB164708		Prep Date:	11/06/2024	
	Aluminum	4.39	<4.39	U	4.39	P	11/07/2024	23:38	LB133344
	Antimony	2.19	<2.19	U	2.19	P	11/07/2024	23:38	LB133344
	Arsenic	0.88	<0.88	U	0.88	P	11/07/2024	23:38	LB133344
	Barium	4.39	<4.39	U	4.39	P	11/07/2024	23:38	LB133344
	Beryllium	0.26	<0.26	U	0.26	P	11/07/2024	23:38	LB133344
	Cadmium	0.26	<0.26	U	0.26	P	11/07/2024	23:38	LB133344
	Calcium	87.7	<87.7	U	87.7	P	11/07/2024	23:38	LB133344
	Chromium	0.44	<0.44	U	0.44	P	11/07/2024	23:38	LB133344
	Cobalt	1.32	<1.32	U	1.32	P	11/07/2024	23:38	LB133344
	Copper	0.88	<0.88	U	0.88	P	11/07/2024	23:38	LB133344
	Iron	4.39	<4.39	U	4.39	P	11/07/2024	23:38	LB133344
	Lead	0.53	<0.53	U	0.53	P	11/07/2024	23:38	LB133344
	Magnesium	87.7	<87.7	U	87.7	P	11/07/2024	23:38	LB133344
	Manganese	0.88	<0.88	U	0.88	P	11/07/2024	23:38	LB133344
	Nickel	1.75	<1.75	U	1.75	P	11/07/2024	23:38	LB133344
	Potassium	87.7	<87.7	U	87.7	P	11/07/2024	23:38	LB133344
	Selenium	0.88	<0.88	U	0.88	P	11/07/2024	23:38	LB133344
	Silver	0.44	<0.44	U	0.44	P	11/07/2024	23:38	LB133344
	Sodium	87.7	<87.7	U	87.7	P	11/07/2024	23:38	LB133344
	Thallium	1.75	<1.75	U	1.75	P	11/07/2024	23:38	LB133344
	Vanadium	1.75	<1.75	U	1.75	P	11/07/2024	23:38	LB133344
	Zinc	1.75	<1.75	U	1.75	P	11/07/2024	23:38	LB133344
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB164759BL	WATER			Batch Number:	PB164759		Prep Date:	11/07/2024	
	Aluminum	50.0	<50.0	U	50.0	P	11/19/2024	13:00	LB133524
	Antimony	25.0	<25.0	U	25.0	P	11/19/2024	13:00	LB133524
	Arsenic	10.0	<10.0	U	10.0	P	11/19/2024	13:00	LB133524
	Barium	50.0	<50.0	U	50.0	P	11/19/2024	13:00	LB133524
	Beryllium	3.00	<3.00	U	3.00	P	11/19/2024	13:00	LB133524
	Cadmium	3.00	<3.00	U	3.00	P	11/19/2024	13:00	LB133524
	Calcium	1000	<1000	U	1000	P	11/19/2024	13:00	LB133524
	Chromium	5.00	<5.00	U	5.00	P	11/19/2024	13:00	LB133524
	Cobalt	15.0	<15.0	U	15.0	P	11/19/2024	13:00	LB133524
	Copper	10.0	<10.0	U	10.0	P	11/19/2024	13:00	LB133524
	Iron	50.0	<50.0	U	50.0	P	11/19/2024	13:00	LB133524
	Lead	6.00	<6.00	U	6.00	P	11/19/2024	13:00	LB133524
	Magnesium	1000	<1000	U	1000	P	11/19/2024	13:00	LB133524
	Manganese	10.0	<10.0	U	10.0	P	11/19/2024	13:00	LB133524
	Nickel	20.0	<20.0	U	20.0	P	11/19/2024	13:00	LB133524

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: P4718

Instrument: P4

Potassium	1000	<1000	U	1000	P	11/19/2024	13:00	LB133524
Selenium	10.0	<10.0	U	10.0	P	11/19/2024	13:00	LB133524
Silver	5.00	<5.00	U	5.00	P	11/19/2024	13:00	LB133524
Sodium	1000	<1000	U	1000	P	11/19/2024	13:00	LB133524
Thallium	20.0	<20.0	U	20.0	P	11/19/2024	13:00	LB133524
Vanadium	20.0	<20.0	U	20.0	P	11/19/2024	13:00	LB133524
Zinc	20.0	<20.0	U	20.0	P	11/19/2024	13:00	LB133524



METAL CALIBRATION DATA

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV74	Mercury	3.93	4.0	98	90 - 110	CV	11/07/2024	14:41	LB133334

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV47	Mercury	4.81	5.0	96	90 - 110	CV	11/07/2024	14:45	LB133334
CCV48	Mercury	4.90	5.0	98	90 - 110	CV	11/07/2024	15:13	LB133334
CCV49	Mercury	4.92	5.0	98	90 - 110	CV	11/07/2024	15:40	LB133334
CCV50	Mercury	4.92	5.0	98	90 - 110	CV	11/07/2024	16:08	LB133334

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV80	Mercury	4.09	4.0	102	90 - 110	CV	11/11/2024	13:58	LB133392

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV64	Mercury	5.16	5.0	103	90 - 110	CV	11/11/2024	14:03	LB133392
CCV65	Mercury	4.85	5.0	97	90 - 110	CV	11/11/2024	14:36	LB133392
CCV66	Mercury	4.93	5.0	99	90 - 110	CV	11/11/2024	14:54	LB133392

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	2590	2500	104	90 - 110	P	11/07/2024	13:49	LB133344
	Antimony	1010	1000	101	90 - 110	P	11/07/2024	13:49	LB133344
	Arsenic	1030	1000	103	90 - 110	P	11/07/2024	13:49	LB133344
	Barium	521	520	100	90 - 110	P	11/07/2024	13:49	LB133344
	Beryllium	515	510	101	90 - 110	P	11/07/2024	13:49	LB133344
	Cadmium	509	510	100	90 - 110	P	11/07/2024	13:49	LB133344
	Calcium	10200	10000	102	90 - 110	P	11/07/2024	13:49	LB133344
	Chromium	531	520	102	90 - 110	P	11/07/2024	13:49	LB133344
	Cobalt	516	520	99	90 - 110	P	11/07/2024	13:49	LB133344
	Copper	531	510	104	90 - 110	P	11/07/2024	13:49	LB133344
	Iron	9990	10000	100	90 - 110	P	11/07/2024	13:49	LB133344
	Lead	1020	1000	102	90 - 110	P	11/07/2024	13:49	LB133344
	Magnesium	6010	6000	100	90 - 110	P	11/07/2024	13:49	LB133344
	Manganese	526	520	101	90 - 110	P	11/07/2024	13:49	LB133344
	Nickel	518	530	98	90 - 110	P	11/07/2024	13:49	LB133344
	Potassium	9890	9900	100	90 - 110	P	11/07/2024	13:49	LB133344
	Selenium	1040	1000	104	90 - 110	P	11/07/2024	13:49	LB133344
	Silver	256	250	102	90 - 110	P	11/07/2024	13:49	LB133344
	Sodium	9440	10000	94	90 - 110	P	11/07/2024	13:49	LB133344
	Thallium	1100	1000	110	90 - 110	P	11/07/2024	13:49	LB133344
	Vanadium	507	500	101	90 - 110	P	11/07/2024	13:49	LB133344
	Zinc	1050	1000	105	90 - 110	P	11/07/2024	13:49	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	103	100	103	80 - 120	P	11/07/2024	14:23	LB133344
	Antimony	50.8	50.0	102	80 - 120	P	11/07/2024	14:23	LB133344
	Arsenic	18.0	20.0	90	80 - 120	P	11/07/2024	14:23	LB133344
	Barium	100	100	100	80 - 120	P	11/07/2024	14:23	LB133344
	Beryllium	5.93	6.0	99	80 - 120	P	11/07/2024	14:23	LB133344
	Cadmium	6.52	6.0	109	80 - 120	P	11/07/2024	14:23	LB133344
	Calcium	2010	2000	100	80 - 120	P	11/07/2024	14:23	LB133344
	Chromium	9.52	10.0	95	80 - 120	P	11/07/2024	14:23	LB133344
	Cobalt	29.4	30.0	98	80 - 120	P	11/07/2024	14:23	LB133344
	Copper	22.7	20.0	114	80 - 120	P	11/07/2024	14:23	LB133344
	Iron	96.8	100	97	80 - 120	P	11/07/2024	14:23	LB133344
	Lead	11.4	12.0	95	80 - 120	P	11/07/2024	14:23	LB133344
	Magnesium	2020	2000	101	80 - 120	P	11/07/2024	14:23	LB133344
	Manganese	20.2	20.0	101	80 - 120	P	11/07/2024	14:23	LB133344
	Nickel	38.5	40.0	96	80 - 120	P	11/07/2024	14:23	LB133344
	Potassium	2010	2000	101	80 - 120	P	11/07/2024	14:23	LB133344
	Selenium	20.8	20.0	104	80 - 120	P	11/07/2024	14:23	LB133344
	Silver	10.8	10.0	108	80 - 120	P	11/07/2024	14:23	LB133344
	Sodium	1680	2000	84	80 - 120	P	11/07/2024	14:23	LB133344
	Thallium	43.4	40.0	108	80 - 120	P	11/07/2024	14:23	LB133344
	Vanadium	40.8	40.0	102	80 - 120	P	11/07/2024	14:23	LB133344
	Zinc	42.0	40.0	105	80 - 120	P	11/07/2024	14:23	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	10000	10000	100	90 - 110	P	11/07/2024	15:01	LB133344
	Antimony	5120	5000	102	90 - 110	P	11/07/2024	15:01	LB133344
	Arsenic	5040	5000	101	90 - 110	P	11/07/2024	15:01	LB133344
	Barium	9710	10000	97	90 - 110	P	11/07/2024	15:01	LB133344
	Beryllium	239	250	96	90 - 110	P	11/07/2024	15:01	LB133344
	Cadmium	2450	2500	98	90 - 110	P	11/07/2024	15:01	LB133344
	Calcium	24000	25000	96	90 - 110	P	11/07/2024	15:01	LB133344
	Chromium	992	1000	99	90 - 110	P	11/07/2024	15:01	LB133344
	Cobalt	2450	2500	98	90 - 110	P	11/07/2024	15:01	LB133344
	Copper	1270	1250	102	90 - 110	P	11/07/2024	15:01	LB133344
	Iron	4870	5000	97	90 - 110	P	11/07/2024	15:01	LB133344
	Lead	4920	5000	98	90 - 110	P	11/07/2024	15:01	LB133344
	Magnesium	24000	25000	96	90 - 110	P	11/07/2024	15:01	LB133344
	Manganese	2370	2500	95	90 - 110	P	11/07/2024	15:01	LB133344
	Nickel	2450	2500	98	90 - 110	P	11/07/2024	15:01	LB133344
	Potassium	25100	25000	100	90 - 110	P	11/07/2024	15:01	LB133344
	Selenium	5150	5000	103	90 - 110	P	11/07/2024	15:01	LB133344
	Silver	1240	1250	99	90 - 110	P	11/07/2024	15:01	LB133344
	Sodium	25300	25000	101	90 - 110	P	11/07/2024	15:01	LB133344
	Thallium	5140	5000	103	90 - 110	P	11/07/2024	15:01	LB133344
	Vanadium	2450	2500	98	90 - 110	P	11/07/2024	15:01	LB133344
	Zinc	2530	2500	101	90 - 110	P	11/07/2024	15:01	LB133344
CCV02	Aluminum	9620	10000	96	90 - 110	P	11/07/2024	15:53	LB133344
	Antimony	5030	5000	101	90 - 110	P	11/07/2024	15:53	LB133344
	Arsenic	4950	5000	99	90 - 110	P	11/07/2024	15:53	LB133344
	Barium	9390	10000	94	90 - 110	P	11/07/2024	15:53	LB133344
	Beryllium	231	250	92	90 - 110	P	11/07/2024	15:53	LB133344
	Cadmium	2390	2500	96	90 - 110	P	11/07/2024	15:53	LB133344
	Calcium	23100	25000	92	90 - 110	P	11/07/2024	15:53	LB133344
	Chromium	963	1000	96	90 - 110	P	11/07/2024	15:53	LB133344
	Cobalt	2390	2500	96	90 - 110	P	11/07/2024	15:53	LB133344
	Copper	1250	1250	100	90 - 110	P	11/07/2024	15:53	LB133344
	Iron	4810	5000	96	90 - 110	P	11/07/2024	15:53	LB133344
	Lead	4800	5000	96	90 - 110	P	11/07/2024	15:53	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	23100	25000	92	90 - 110	P	11/07/2024	15:53	LB133344
	Manganese	2280	2500	91	90 - 110	P	11/07/2024	15:53	LB133344
	Nickel	2390	2500	96	90 - 110	P	11/07/2024	15:53	LB133344
	Potassium	25000	25000	100	90 - 110	P	11/07/2024	15:53	LB133344
	Selenium	5050	5000	101	90 - 110	P	11/07/2024	15:53	LB133344
	Silver	1220	1250	98	90 - 110	P	11/07/2024	15:53	LB133344
	Sodium	25800	25000	103	90 - 110	P	11/07/2024	15:53	LB133344
	Thallium	4850	5000	97	90 - 110	P	11/07/2024	15:53	LB133344
	Vanadium	2370	2500	95	90 - 110	P	11/07/2024	15:53	LB133344
	Zinc	2480	2500	99	90 - 110	P	11/07/2024	15:53	LB133344
CCV03	Aluminum	10100	10000	101	90 - 110	P	11/07/2024	16:45	LB133344
	Antimony	5250	5000	105	90 - 110	P	11/07/2024	16:45	LB133344
	Arsenic	5190	5000	104	90 - 110	P	11/07/2024	16:45	LB133344
	Barium	9850	10000	98	90 - 110	P	11/07/2024	16:45	LB133344
	Beryllium	235	250	94	90 - 110	P	11/07/2024	16:45	LB133344
	Cadmium	2470	2500	99	90 - 110	P	11/07/2024	16:45	LB133344
	Calcium	24000	25000	96	90 - 110	P	11/07/2024	16:45	LB133344
	Chromium	993	1000	99	90 - 110	P	11/07/2024	16:45	LB133344
	Cobalt	2470	2500	99	90 - 110	P	11/07/2024	16:45	LB133344
	Copper	1300	1250	104	90 - 110	P	11/07/2024	16:45	LB133344
	Iron	4970	5000	99	90 - 110	P	11/07/2024	16:45	LB133344
	Lead	4970	5000	99	90 - 110	P	11/07/2024	16:45	LB133344
	Magnesium	23900	25000	96	90 - 110	P	11/07/2024	16:45	LB133344
	Manganese	2370	2500	95	90 - 110	P	11/07/2024	16:45	LB133344
	Nickel	2480	2500	99	90 - 110	P	11/07/2024	16:45	LB133344
	Potassium	25900	25000	103	90 - 110	P	11/07/2024	16:45	LB133344
	Selenium	5320	5000	106	90 - 110	P	11/07/2024	16:45	LB133344
	Silver	1240	1250	100	90 - 110	P	11/07/2024	16:45	LB133344
	Sodium	25500	25000	102	90 - 110	P	11/07/2024	16:45	LB133344
	Thallium	5400	5000	108	90 - 110	P	11/07/2024	16:45	LB133344
	Vanadium	2460	2500	98	90 - 110	P	11/07/2024	16:45	LB133344
	Zinc	2400	2500	96	90 - 110	P	11/07/2024	16:45	LB133344
CCV04	Aluminum	10200	10000	102	90 - 110	P	11/07/2024	17:56	LB133344
	Antimony	5280	5000	106	90 - 110	P	11/07/2024	17:56	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	5210	5000	104	90 - 110	P	11/07/2024	17:56	LB133344
	Barium	9810	10000	98	90 - 110	P	11/07/2024	17:56	LB133344
	Beryllium	242	250	97	90 - 110	P	11/07/2024	17:56	LB133344
	Cadmium	2500	2500	100	90 - 110	P	11/07/2024	17:56	LB133344
	Calcium	24200	25000	97	90 - 110	P	11/07/2024	17:56	LB133344
	Chromium	1020	1000	102	90 - 110	P	11/07/2024	17:56	LB133344
	Cobalt	2490	2500	100	90 - 110	P	11/07/2024	17:56	LB133344
	Copper	1300	1250	104	90 - 110	P	11/07/2024	17:56	LB133344
	Iron	4980	5000	100	90 - 110	P	11/07/2024	17:56	LB133344
	Lead	5010	5000	100	90 - 110	P	11/07/2024	17:56	LB133344
	Magnesium	24300	25000	97	90 - 110	P	11/07/2024	17:56	LB133344
	Manganese	2370	2500	95	90 - 110	P	11/07/2024	17:56	LB133344
	Nickel	2490	2500	100	90 - 110	P	11/07/2024	17:56	LB133344
	Potassium	26000	25000	104	90 - 110	P	11/07/2024	17:56	LB133344
	Selenium	5390	5000	108	90 - 110	P	11/07/2024	17:56	LB133344
	Silver	1260	1250	101	90 - 110	P	11/07/2024	17:56	LB133344
	Sodium	25400	25000	102	90 - 110	P	11/07/2024	17:56	LB133344
	Thallium	5070	5000	102	90 - 110	P	11/07/2024	17:56	LB133344
	Vanadium	2490	2500	99	90 - 110	P	11/07/2024	17:56	LB133344
	Zinc	2510	2500	100	90 - 110	P	11/07/2024	17:56	LB133344
CCV05	Aluminum	10000	10000	100	90 - 110	P	11/07/2024	18:32	LB133344
	Antimony	5190	5000	104	90 - 110	P	11/07/2024	18:32	LB133344
	Arsenic	5120	5000	102	90 - 110	P	11/07/2024	18:32	LB133344
	Barium	9790	10000	98	90 - 110	P	11/07/2024	18:32	LB133344
	Beryllium	236	250	94	90 - 110	P	11/07/2024	18:32	LB133344
	Cadmium	2430	2500	97	90 - 110	P	11/07/2024	18:32	LB133344
	Calcium	23800	25000	95	90 - 110	P	11/07/2024	18:32	LB133344
	Chromium	988	1000	99	90 - 110	P	11/07/2024	18:32	LB133344
	Cobalt	2430	2500	97	90 - 110	P	11/07/2024	18:32	LB133344
	Copper	1280	1250	102	90 - 110	P	11/07/2024	18:32	LB133344
	Iron	4910	5000	98	90 - 110	P	11/07/2024	18:32	LB133344
	Lead	4880	5000	98	90 - 110	P	11/07/2024	18:32	LB133344
	Magnesium	23800	25000	95	90 - 110	P	11/07/2024	18:32	LB133344
	Manganese	2350	2500	94	90 - 110	P	11/07/2024	18:32	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2430	2500	97	90 - 110	P	11/07/2024	18:32	LB133344
	Potassium	25800	25000	103	90 - 110	P	11/07/2024	18:32	LB133344
	Selenium	5290	5000	106	90 - 110	P	11/07/2024	18:32	LB133344
	Silver	1240	1250	99	90 - 110	P	11/07/2024	18:32	LB133344
	Sodium	25600	25000	102	90 - 110	P	11/07/2024	18:32	LB133344
	Thallium	5000	5000	100	90 - 110	P	11/07/2024	18:32	LB133344
	Vanadium	2440	2500	98	90 - 110	P	11/07/2024	18:32	LB133344
	Zinc	2500	2500	100	90 - 110	P	11/07/2024	18:32	LB133344
CCV06	Aluminum	9950	10000	100	90 - 110	P	11/07/2024	19:08	LB133344
	Antimony	5160	5000	103	90 - 110	P	11/07/2024	19:08	LB133344
	Arsenic	5110	5000	102	90 - 110	P	11/07/2024	19:08	LB133344
	Barium	9640	10000	96	90 - 110	P	11/07/2024	19:08	LB133344
	Beryllium	238	250	95	90 - 110	P	11/07/2024	19:08	LB133344
	Cadmium	2460	2500	99	90 - 110	P	11/07/2024	19:08	LB133344
	Calcium	23800	25000	95	90 - 110	P	11/07/2024	19:08	LB133344
	Chromium	994	1000	99	90 - 110	P	11/07/2024	19:08	LB133344
	Cobalt	2460	2500	98	90 - 110	P	11/07/2024	19:08	LB133344
	Copper	1270	1250	102	90 - 110	P	11/07/2024	19:08	LB133344
	Iron	4870	5000	97	90 - 110	P	11/07/2024	19:08	LB133344
	Lead	4930	5000	99	90 - 110	P	11/07/2024	19:08	LB133344
	Magnesium	23800	25000	95	90 - 110	P	11/07/2024	19:08	LB133344
	Manganese	2330	2500	93	90 - 110	P	11/07/2024	19:08	LB133344
	Nickel	2460	2500	98	90 - 110	P	11/07/2024	19:08	LB133344
	Potassium	25300	25000	101	90 - 110	P	11/07/2024	19:08	LB133344
	Selenium	5260	5000	105	90 - 110	P	11/07/2024	19:08	LB133344
	Silver	1240	1250	99	90 - 110	P	11/07/2024	19:08	LB133344
	Sodium	25300	25000	101	90 - 110	P	11/07/2024	19:08	LB133344
	Thallium	5060	5000	101	90 - 110	P	11/07/2024	19:08	LB133344
	Vanadium	2430	2500	97	90 - 110	P	11/07/2024	19:08	LB133344
	Zinc	2490	2500	100	90 - 110	P	11/07/2024	19:08	LB133344
CCV07	Aluminum	9840	10000	98	90 - 110	P	11/07/2024	19:58	LB133344
	Antimony	5160	5000	103	90 - 110	P	11/07/2024	19:58	LB133344
	Arsenic	5100	5000	102	90 - 110	P	11/07/2024	19:58	LB133344
	Barium	9400	10000	94	90 - 110	P	11/07/2024	19:58	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	232	250	93	90 - 110	P	11/07/2024	19:58	LB133344
	Cadmium	2420	2500	97	90 - 110	P	11/07/2024	19:58	LB133344
	Calcium	23300	25000	93	90 - 110	P	11/07/2024	19:58	LB133344
	Chromium	986	1000	99	90 - 110	P	11/07/2024	19:58	LB133344
	Cobalt	2410	2500	96	90 - 110	P	11/07/2024	19:58	LB133344
	Copper	1270	1250	101	90 - 110	P	11/07/2024	19:58	LB133344
	Iron	4800	5000	96	90 - 110	P	11/07/2024	19:58	LB133344
	Lead	4850	5000	97	90 - 110	P	11/07/2024	19:58	LB133344
	Magnesium	23400	25000	94	90 - 110	P	11/07/2024	19:58	LB133344
	Manganese	2270	2500	91	90 - 110	P	11/07/2024	19:58	LB133344
	Nickel	2410	2500	96	90 - 110	P	11/07/2024	19:58	LB133344
	Potassium	25200	25000	101	90 - 110	P	11/07/2024	19:58	LB133344
	Selenium	5270	5000	105	90 - 110	P	11/07/2024	19:58	LB133344
	Silver	1230	1250	98	90 - 110	P	11/07/2024	19:58	LB133344
	Sodium	24900	25000	100	90 - 110	P	11/07/2024	19:58	LB133344
	Thallium	4910	5000	98	90 - 110	P	11/07/2024	19:58	LB133344
	Vanadium	2400	2500	96	90 - 110	P	11/07/2024	19:58	LB133344
	Zinc	2490	2500	100	90 - 110	P	11/07/2024	19:58	LB133344
CCV08	Aluminum	9850	10000	98	90 - 110	P	11/07/2024	20:58	LB133344
	Antimony	5180	5000	104	90 - 110	P	11/07/2024	20:58	LB133344
	Arsenic	5100	5000	102	90 - 110	P	11/07/2024	20:58	LB133344
	Barium	9480	10000	95	90 - 110	P	11/07/2024	20:58	LB133344
	Beryllium	234	250	94	90 - 110	P	11/07/2024	20:58	LB133344
	Cadmium	2410	2500	96	90 - 110	P	11/07/2024	20:58	LB133344
	Calcium	23200	25000	93	90 - 110	P	11/07/2024	20:58	LB133344
	Chromium	989	1000	99	90 - 110	P	11/07/2024	20:58	LB133344
	Cobalt	2410	2500	96	90 - 110	P	11/07/2024	20:58	LB133344
	Copper	1270	1250	102	90 - 110	P	11/07/2024	20:58	LB133344
	Iron	4890	5000	98	90 - 110	P	11/07/2024	20:58	LB133344
	Lead	4840	5000	97	90 - 110	P	11/07/2024	20:58	LB133344
	Magnesium	23300	25000	93	90 - 110	P	11/07/2024	20:58	LB133344
	Manganese	2280	2500	91	90 - 110	P	11/07/2024	20:58	LB133344
	Nickel	2410	2500	96	90 - 110	P	11/07/2024	20:58	LB133344
	Potassium	25900	25000	104	90 - 110	P	11/07/2024	20:58	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV08	Selenium	5290	5000	106	90 - 110	P	11/07/2024	20:58	LB133344
	Silver	1250	1250	100	90 - 110	P	11/07/2024	20:58	LB133344
	Sodium	26000	25000	104	90 - 110	P	11/07/2024	20:58	LB133344
	Thallium	4970	5000	99	90 - 110	P	11/07/2024	20:58	LB133344
	Vanadium	2390	2500	96	90 - 110	P	11/07/2024	20:58	LB133344
	Zinc	2530	2500	101	90 - 110	P	11/07/2024	20:58	LB133344
CCV09	Aluminum	9720	10000	97	90 - 110	P	11/07/2024	21:51	LB133344
	Antimony	5100	5000	102	90 - 110	P	11/07/2024	21:51	LB133344
	Arsenic	5030	5000	100	90 - 110	P	11/07/2024	21:51	LB133344
	Barium	9510	10000	95	90 - 110	P	11/07/2024	21:51	LB133344
	Beryllium	231	250	92	90 - 110	P	11/07/2024	21:51	LB133344
	Cadmium	2390	2500	95	90 - 110	P	11/07/2024	21:51	LB133344
	Calcium	23100	25000	92	90 - 110	P	11/07/2024	21:51	LB133344
	Chromium	966	1000	97	90 - 110	P	11/07/2024	21:51	LB133344
	Cobalt	2380	2500	95	90 - 110	P	11/07/2024	21:51	LB133344
	Copper	1250	1250	100	90 - 110	P	11/07/2024	21:51	LB133344
	Iron	4750	5000	95	90 - 110	P	11/07/2024	21:51	LB133344
	Lead	4780	5000	96	90 - 110	P	11/07/2024	21:51	LB133344
	Magnesium	22900	25000	92	90 - 110	P	11/07/2024	21:51	LB133344
	Manganese	2250	2500	90	90 - 110	P	11/07/2024	21:51	LB133344
	Nickel	2380	2500	95	90 - 110	P	11/07/2024	21:51	LB133344
	Potassium	25100	25000	100	90 - 110	P	11/07/2024	21:51	LB133344
	Selenium	5170	5000	103	90 - 110	P	11/07/2024	21:51	LB133344
	Silver	1210	1250	97	90 - 110	P	11/07/2024	21:51	LB133344
	Sodium	25300	25000	101	90 - 110	P	11/07/2024	21:51	LB133344
	Thallium	5380	5000	108	90 - 110	P	11/07/2024	21:51	LB133344
	Vanadium	2360	2500	94	90 - 110	P	11/07/2024	21:51	LB133344
	Zinc	2670	2500	107	90 - 110	P	11/07/2024	21:51	LB133344
CCV10	Aluminum	9650	10000	96	90 - 110	P	11/07/2024	22:52	LB133344
	Antimony	5080	5000	102	90 - 110	P	11/07/2024	22:52	LB133344
	Arsenic	5000	5000	100	90 - 110	P	11/07/2024	22:52	LB133344
	Barium	9460	10000	95	90 - 110	P	11/07/2024	22:52	LB133344
	Beryllium	255	250	102	90 - 110	P	11/07/2024	22:52	LB133344
	Cadmium	2340	2500	94	90 - 110	P	11/07/2024	22:52	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV10	Calcium	22600	25000	90	90 - 110	P	11/07/2024	22:52	LB133344
	Chromium	963	1000	96	90 - 110	P	11/07/2024	22:52	LB133344
	Cobalt	2340	2500	94	90 - 110	P	11/07/2024	22:52	LB133344
	Copper	1240	1250	99	90 - 110	P	11/07/2024	22:52	LB133344
	Iron	4790	5000	96	90 - 110	P	11/07/2024	22:52	LB133344
	Lead	4710	5000	94	90 - 110	P	11/07/2024	22:52	LB133344
	Magnesium	26500	25000	106	90 - 110	P	11/07/2024	22:52	LB133344
	Manganese	2500	2500	100	90 - 110	P	11/07/2024	22:52	LB133344
	Nickel	2340	2500	94	90 - 110	P	11/07/2024	22:52	LB133344
	Potassium	26100	25000	104	90 - 110	P	11/07/2024	22:52	LB133344
	Selenium	5160	5000	103	90 - 110	P	11/07/2024	22:52	LB133344
	Silver	1220	1250	98	90 - 110	P	11/07/2024	22:52	LB133344
	Sodium	27200	25000	109	90 - 110	P	11/07/2024	22:52	LB133344
	Thallium	5290	5000	106	90 - 110	P	11/07/2024	22:52	LB133344
	Vanadium	2320	2500	93	90 - 110	P	11/07/2024	22:52	LB133344
	Zinc	2400	2500	96	90 - 110	P	11/07/2024	22:52	LB133344
CCV11	Aluminum	9750	10000	98	90 - 110	P	11/07/2024	23:43	LB133344
	Antimony	5100	5000	102	90 - 110	P	11/07/2024	23:43	LB133344
	Arsenic	5020	5000	100	90 - 110	P	11/07/2024	23:43	LB133344
	Barium	9690	10000	97	90 - 110	P	11/07/2024	23:43	LB133344
	Beryllium	234	250	93	90 - 110	P	11/07/2024	23:43	LB133344
	Cadmium	2390	2500	96	90 - 110	P	11/07/2024	23:43	LB133344
	Calcium	23200	25000	93	90 - 110	P	11/07/2024	23:43	LB133344
	Chromium	962	1000	96	90 - 110	P	11/07/2024	23:43	LB133344
	Cobalt	2390	2500	95	90 - 110	P	11/07/2024	23:43	LB133344
	Copper	1250	1250	100	90 - 110	P	11/07/2024	23:43	LB133344
	Iron	4730	5000	94	90 - 110	P	11/07/2024	23:43	LB133344
	Lead	4800	5000	96	90 - 110	P	11/07/2024	23:43	LB133344
	Magnesium	23000	25000	92	90 - 110	P	11/07/2024	23:43	LB133344
	Manganese	2290	2500	92	90 - 110	P	11/07/2024	23:43	LB133344
	Nickel	2390	2500	95	90 - 110	P	11/07/2024	23:43	LB133344
	Potassium	25300	25000	101	90 - 110	P	11/07/2024	23:43	LB133344
	Selenium	5160	5000	103	90 - 110	P	11/07/2024	23:43	LB133344
	Silver	1210	1250	97	90 - 110	P	11/07/2024	23:43	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV11	Sodium	26300	25000	105	90 - 110	P	11/07/2024	23:43	LB133344
	Thallium	5350	5000	107	90 - 110	P	11/07/2024	23:43	LB133344
	Vanadium	2370	2500	95	90 - 110	P	11/07/2024	23:43	LB133344
	Zinc	2430	2500	97	90 - 110	P	11/07/2024	23:43	LB133344
CCV12	Aluminum	9840	10000	98	90 - 110	P	11/08/2024	00:39	LB133344
	Antimony	5240	5000	105	90 - 110	P	11/08/2024	00:39	LB133344
	Arsenic	5150	5000	103	90 - 110	P	11/08/2024	00:39	LB133344
	Barium	9830	10000	98	90 - 110	P	11/08/2024	00:39	LB133344
	Beryllium	228	250	91	90 - 110	P	11/08/2024	00:39	LB133344
	Cadmium	2390	2500	96	90 - 110	P	11/08/2024	00:39	LB133344
	Calcium	22900	25000	92	90 - 110	P	11/08/2024	00:39	LB133344
	Chromium	965	1000	96	90 - 110	P	11/08/2024	00:39	LB133344
	Cobalt	2390	2500	96	90 - 110	P	11/08/2024	00:39	LB133344
	Copper	1280	1250	102	90 - 110	P	11/08/2024	00:39	LB133344
	Iron	4790	5000	96	90 - 110	P	11/08/2024	00:39	LB133344
	Lead	4820	5000	96	90 - 110	P	11/08/2024	00:39	LB133344
	Magnesium	22700	25000	91	90 - 110	P	11/08/2024	00:39	LB133344
	Manganese	2260	2500	90	90 - 110	P	11/08/2024	00:39	LB133344
	Nickel	2390	2500	96	90 - 110	P	11/08/2024	00:39	LB133344
	Potassium	26400	25000	106	90 - 110	P	11/08/2024	00:39	LB133344
	Selenium	5340	5000	107	90 - 110	P	11/08/2024	00:39	LB133344
	Silver	1230	1250	98	90 - 110	P	11/08/2024	00:39	LB133344
	Sodium	22700	25000	91	90 - 110	P	11/08/2024	00:39	LB133344
	Thallium	5270	5000	105	90 - 110	P	11/08/2024	00:39	LB133344
	Vanadium	2370	2500	95	90 - 110	P	11/08/2024	00:39	LB133344
	Zinc	2380	2500	95	90 - 110	P	11/08/2024	00:39	LB133344
CCV13	Aluminum	9910	10000	99	90 - 110	P	11/08/2024	00:57	LB133344
	Antimony	5250	5000	105	90 - 110	P	11/08/2024	00:57	LB133344
	Arsenic	5160	5000	103	90 - 110	P	11/08/2024	00:57	LB133344
	Barium	9650	10000	96	90 - 110	P	11/08/2024	00:57	LB133344
	Beryllium	233	250	93	90 - 110	P	11/08/2024	00:57	LB133344
	Cadmium	2420	2500	97	90 - 110	P	11/08/2024	00:57	LB133344
	Calcium	23200	25000	93	90 - 110	P	11/08/2024	00:57	LB133344
	Chromium	980	1000	98	90 - 110	P	11/08/2024	00:57	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV13	Cobalt	2420	2500	97	90 - 110	P	11/08/2024	00:57	LB133344
	Copper	1290	1250	103	90 - 110	P	11/08/2024	00:57	LB133344
	Iron	4780	5000	96	90 - 110	P	11/08/2024	00:57	LB133344
	Lead	4870	5000	97	90 - 110	P	11/08/2024	00:57	LB133344
	Magnesium	23100	25000	92	90 - 110	P	11/08/2024	00:57	LB133344
	Manganese	2280	2500	91	90 - 110	P	11/08/2024	00:57	LB133344
	Nickel	2420	2500	97	90 - 110	P	11/08/2024	00:57	LB133344
	Potassium	26000	25000	104	90 - 110	P	11/08/2024	00:57	LB133344
	Selenium	5340	5000	107	90 - 110	P	11/08/2024	00:57	LB133344
	Silver	1230	1250	99	90 - 110	P	11/08/2024	00:57	LB133344
	Sodium	26900	25000	108	90 - 110	P	11/08/2024	00:57	LB133344
	Thallium	5410	5000	108	90 - 110	P	11/08/2024	00:57	LB133344
	Vanadium	2390	2500	96	90 - 110	P	11/08/2024	00:57	LB133344
	Zinc	2470	2500	99	90 - 110	P	11/08/2024	00:57	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	2550	2500	102	90 - 110	P	11/08/2024	13:59	LB133365
	Antimony	979	1000	98	90 - 110	P	11/08/2024	13:59	LB133365
	Arsenic	1000	1000	100	90 - 110	P	11/08/2024	13:59	LB133365
	Barium	513	520	99	90 - 110	P	11/08/2024	13:59	LB133365
	Beryllium	511	510	100	90 - 110	P	11/08/2024	13:59	LB133365
	Cadmium	497	510	97	90 - 110	P	11/08/2024	13:59	LB133365
	Calcium	10100	10000	101	90 - 110	P	11/08/2024	13:59	LB133365
	Chromium	514	520	99	90 - 110	P	11/08/2024	13:59	LB133365
	Cobalt	506	520	97	90 - 110	P	11/08/2024	13:59	LB133365
	Copper	521	510	102	90 - 110	P	11/08/2024	13:59	LB133365
	Iron	9660	10000	97	90 - 110	P	11/08/2024	13:59	LB133365
	Lead	999	1000	100	90 - 110	P	11/08/2024	13:59	LB133365
	Magnesium	5940	6000	99	90 - 110	P	11/08/2024	13:59	LB133365
	Manganese	519	520	100	90 - 110	P	11/08/2024	13:59	LB133365
	Nickel	509	530	96	90 - 110	P	11/08/2024	13:59	LB133365
	Potassium	9550	9900	96	90 - 110	P	11/08/2024	13:59	LB133365
	Selenium	1000	1000	100	90 - 110	P	11/08/2024	13:59	LB133365
	Silver	251	250	100	90 - 110	P	11/08/2024	13:59	LB133365
	Sodium	9170	10000	92	90 - 110	P	11/08/2024	13:59	LB133365
	Thallium	975	1000	98	90 - 110	P	11/08/2024	13:59	LB133365
	Vanadium	496	500	99	90 - 110	P	11/08/2024	13:59	LB133365
	Zinc	1030	1000	103	90 - 110	P	11/08/2024	13:59	LB133365

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	109	100	109	80 - 120	P	11/08/2024	14:11	LB133365
	Antimony	48.4	50.0	97	80 - 120	P	11/08/2024	14:11	LB133365
	Arsenic	19.5	20.0	98	80 - 120	P	11/08/2024	14:11	LB133365
	Barium	95.3	100	95	80 - 120	P	11/08/2024	14:11	LB133365
	Beryllium	5.66	6.0	94	80 - 120	P	11/08/2024	14:11	LB133365
	Cadmium	6.34	6.0	106	80 - 120	P	11/08/2024	14:11	LB133365
	Calcium	1970	2000	98	80 - 120	P	11/08/2024	14:11	LB133365
	Chromium	9.48	10.0	95	80 - 120	P	11/08/2024	14:11	LB133365
	Cobalt	28.3	30.0	94	80 - 120	P	11/08/2024	14:11	LB133365
	Copper	22.8	20.0	114	80 - 120	P	11/08/2024	14:11	LB133365
	Iron	93.6	100	94	80 - 120	P	11/08/2024	14:11	LB133365
	Lead	11.8	12.0	99	80 - 120	P	11/08/2024	14:11	LB133365
	Magnesium	1950	2000	98	80 - 120	P	11/08/2024	14:11	LB133365
	Manganese	19.8	20.0	99	80 - 120	P	11/08/2024	14:11	LB133365
	Nickel	37.8	40.0	94	80 - 120	P	11/08/2024	14:11	LB133365
	Potassium	1870	2000	94	80 - 120	P	11/08/2024	14:11	LB133365
	Selenium	20.6	20.0	103	80 - 120	P	11/08/2024	14:11	LB133365
	Silver	10.6	10.0	106	80 - 120	P	11/08/2024	14:11	LB133365
	Sodium	1630	2000	82	80 - 120	P	11/08/2024	14:11	LB133365
	Thallium	37.6	40.0	94	80 - 120	P	11/08/2024	14:11	LB133365
	Vanadium	38.6	40.0	96	80 - 120	P	11/08/2024	14:11	LB133365
	Zinc	42.1	40.0	105	80 - 120	P	11/08/2024	14:11	LB133365

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9960	10000	100	90 - 110	P	11/08/2024	14:44	LB133365
	Antimony	5040	5000	101	90 - 110	P	11/08/2024	14:44	LB133365
	Arsenic	4950	5000	99	90 - 110	P	11/08/2024	14:44	LB133365
	Barium	9530	10000	95	90 - 110	P	11/08/2024	14:44	LB133365
	Beryllium	241	250	96	90 - 110	P	11/08/2024	14:44	LB133365
	Cadmium	2390	2500	96	90 - 110	P	11/08/2024	14:44	LB133365
	Calcium	24100	25000	97	90 - 110	P	11/08/2024	14:44	LB133365
	Chromium	1050	1000	105	90 - 110	P	11/08/2024	14:44	LB133365
	Cobalt	2420	2500	97	90 - 110	P	11/08/2024	14:44	LB133365
	Copper	1260	1250	101	90 - 110	P	11/08/2024	14:44	LB133365
	Iron	4850	5000	97	90 - 110	P	11/08/2024	14:44	LB133365
	Lead	4870	5000	97	90 - 110	P	11/08/2024	14:44	LB133365
	Magnesium	24200	25000	97	90 - 110	P	11/08/2024	14:44	LB133365
	Manganese	2370	2500	95	90 - 110	P	11/08/2024	14:44	LB133365
	Nickel	2440	2500	97	90 - 110	P	11/08/2024	14:44	LB133365
	Potassium	24600	25000	98	90 - 110	P	11/08/2024	14:44	LB133365
	Selenium	5020	5000	100	90 - 110	P	11/08/2024	14:44	LB133365
	Silver	1230	1250	99	90 - 110	P	11/08/2024	14:44	LB133365
	Sodium	24300	25000	97	90 - 110	P	11/08/2024	14:44	LB133365
	Thallium	4770	5000	95	90 - 110	P	11/08/2024	14:44	LB133365
	Vanadium	2450	2500	98	90 - 110	P	11/08/2024	14:44	LB133365
	Zinc	2530	2500	101	90 - 110	P	11/08/2024	14:44	LB133365
CCV02	Aluminum	9830	10000	98	90 - 110	P	11/08/2024	15:29	LB133365
	Antimony	5060	5000	101	90 - 110	P	11/08/2024	15:29	LB133365
	Arsenic	4950	5000	99	90 - 110	P	11/08/2024	15:29	LB133365
	Barium	9460	10000	95	90 - 110	P	11/08/2024	15:29	LB133365
	Beryllium	231	250	92	90 - 110	P	11/08/2024	15:29	LB133365
	Cadmium	2360	2500	94	90 - 110	P	11/08/2024	15:29	LB133365
	Calcium	23300	25000	93	90 - 110	P	11/08/2024	15:29	LB133365
	Chromium	918	1000	92	90 - 110	P	11/08/2024	15:29	LB133365
	Cobalt	2390	2500	96	90 - 110	P	11/08/2024	15:29	LB133365
	Copper	1260	1250	101	90 - 110	P	11/08/2024	15:29	LB133365
	Iron	4780	5000	96	90 - 110	P	11/08/2024	15:29	LB133365
	Lead	4800	5000	96	90 - 110	P	11/08/2024	15:29	LB133365

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	23400	25000	94	90 - 110	P	11/08/2024	15:29	LB133365
	Manganese	2280	2500	91	90 - 110	P	11/08/2024	15:29	LB133365
	Nickel	2390	2500	96	90 - 110	P	11/08/2024	15:29	LB133365
	Potassium	25100	25000	100	90 - 110	P	11/08/2024	15:29	LB133365
	Selenium	5080	5000	102	90 - 110	P	11/08/2024	15:29	LB133365
	Silver	1230	1250	98	90 - 110	P	11/08/2024	15:29	LB133365
	Sodium	24500	25000	98	90 - 110	P	11/08/2024	15:29	LB133365
	Thallium	5010	5000	100	90 - 110	P	11/08/2024	15:29	LB133365
	Vanadium	2390	2500	96	90 - 110	P	11/08/2024	15:29	LB133365
	Zinc	2520	2500	101	90 - 110	P	11/08/2024	15:29	LB133365
CCV03	Aluminum	9930	10000	99	90 - 110	P	11/08/2024	16:23	LB133365
	Antimony	5110	5000	102	90 - 110	P	11/08/2024	16:23	LB133365
	Arsenic	4980	5000	100	90 - 110	P	11/08/2024	16:23	LB133365
	Barium	9610	10000	96	90 - 110	P	11/08/2024	16:23	LB133365
	Beryllium	234	250	94	90 - 110	P	11/08/2024	16:23	LB133365
	Cadmium	2380	2500	95	90 - 110	P	11/08/2024	16:23	LB133365
	Calcium	23600	25000	94	90 - 110	P	11/08/2024	16:23	LB133365
	Chromium	930	1000	93	90 - 110	P	11/08/2024	16:23	LB133365
	Cobalt	2410	2500	96	90 - 110	P	11/08/2024	16:23	LB133365
	Copper	1270	1250	101	90 - 110	P	11/08/2024	16:23	LB133365
	Iron	4880	5000	98	90 - 110	P	11/08/2024	16:23	LB133365
	Lead	4840	5000	97	90 - 110	P	11/08/2024	16:23	LB133365
	Magnesium	23700	25000	95	90 - 110	P	11/08/2024	16:23	LB133365
	Manganese	2390	2500	96	90 - 110	P	11/08/2024	16:23	LB133365
	Nickel	2410	2500	96	90 - 110	P	11/08/2024	16:23	LB133365
	Potassium	25600	25000	102	90 - 110	P	11/08/2024	16:23	LB133365
	Selenium	5130	5000	103	90 - 110	P	11/08/2024	16:23	LB133365
	Silver	1250	1250	100	90 - 110	P	11/08/2024	16:23	LB133365
	Sodium	25100	25000	100	90 - 110	P	11/08/2024	16:23	LB133365
	Thallium	5140	5000	103	90 - 110	P	11/08/2024	16:23	LB133365
	Vanadium	2420	2500	97	90 - 110	P	11/08/2024	16:23	LB133365
	Zinc	2560	2500	102	90 - 110	P	11/08/2024	16:23	LB133365
CCV04	Aluminum	9690	10000	97	90 - 110	P	11/08/2024	17:16	LB133365
	Antimony	5290	5000	106	90 - 110	P	11/08/2024	17:16	LB133365

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	5190	5000	104	90 - 110	P	11/08/2024	17:16	LB133365
	Barium	9360	10000	94	90 - 110	P	11/08/2024	17:16	LB133365
	Beryllium	227	250	91	90 - 110	P	11/08/2024	17:16	LB133365
	Cadmium	2420	2500	97	90 - 110	P	11/08/2024	17:16	LB133365
	Calcium	22800	25000	91	90 - 110	P	11/08/2024	17:16	LB133365
	Chromium	927	1000	93	90 - 110	P	11/08/2024	17:16	LB133365
	Cobalt	2450	2500	98	90 - 110	P	11/08/2024	17:16	LB133365
	Copper	1310	1250	105	90 - 110	P	11/08/2024	17:16	LB133365
	Iron	4860	5000	97	90 - 110	P	11/08/2024	17:16	LB133365
	Lead	4940	5000	99	90 - 110	P	11/08/2024	17:16	LB133365
	Magnesium	23000	25000	92	90 - 110	P	11/08/2024	17:16	LB133365
	Manganese	2310	2500	93	90 - 110	P	11/08/2024	17:16	LB133365
	Nickel	2460	2500	98	90 - 110	P	11/08/2024	17:16	LB133365
	Potassium	25800	25000	103	90 - 110	P	11/08/2024	17:16	LB133365
	Selenium	5360	5000	107	90 - 110	P	11/08/2024	17:16	LB133365
	Silver	1260	1250	101	90 - 110	P	11/08/2024	17:16	LB133365
	Sodium	24500	25000	98	90 - 110	P	11/08/2024	17:16	LB133365
	Thallium	5230	5000	105	90 - 110	P	11/08/2024	17:16	LB133365
	Vanadium	2350	2500	94	90 - 110	P	11/08/2024	17:16	LB133365
	Zinc	2610	2500	104	90 - 110	P	11/08/2024	17:16	LB133365
CCV05	Aluminum	9880	10000	99	90 - 110	P	11/08/2024	18:04	LB133365
	Antimony	5230	5000	104	90 - 110	P	11/08/2024	18:04	LB133365
	Arsenic	5110	5000	102	90 - 110	P	11/08/2024	18:04	LB133365
	Barium	9290	10000	93	90 - 110	P	11/08/2024	18:04	LB133365
	Beryllium	234	250	94	90 - 110	P	11/08/2024	18:04	LB133365
	Cadmium	2430	2500	97	90 - 110	P	11/08/2024	18:04	LB133365
	Calcium	23500	25000	94	90 - 110	P	11/08/2024	18:04	LB133365
	Chromium	906	1000	91	90 - 110	P	11/08/2024	18:04	LB133365
	Cobalt	2460	2500	98	90 - 110	P	11/08/2024	18:04	LB133365
	Copper	1290	1250	103	90 - 110	P	11/08/2024	18:04	LB133365
	Iron	4750	5000	95	90 - 110	P	11/08/2024	18:04	LB133365
	Lead	4940	5000	99	90 - 110	P	11/08/2024	18:04	LB133365
	Magnesium	23700	25000	95	90 - 110	P	11/08/2024	18:04	LB133365
	Manganese	2280	2500	91	90 - 110	P	11/08/2024	18:04	LB133365

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2460	2500	98	90 - 110	P	11/08/2024	18:04	LB133365
	Potassium	24800	25000	99	90 - 110	P	11/08/2024	18:04	LB133365
	Selenium	5280	5000	106	90 - 110	P	11/08/2024	18:04	LB133365
	Silver	1230	1250	99	90 - 110	P	11/08/2024	18:04	LB133365
	Sodium	23400	25000	94	90 - 110	P	11/08/2024	18:04	LB133365
	Thallium	5280	5000	106	90 - 110	P	11/08/2024	18:04	LB133365
	Vanadium	2410	2500	96	90 - 110	P	11/08/2024	18:04	LB133365
	Zinc	2540	2500	102	90 - 110	P	11/08/2024	18:04	LB133365
CCV06	Aluminum	9790	10000	98	90 - 110	P	11/08/2024	18:55	LB133365
	Antimony	5070	5000	101	90 - 110	P	11/08/2024	18:55	LB133365
	Arsenic	4970	5000	100	90 - 110	P	11/08/2024	18:55	LB133365
	Barium	9460	10000	95	90 - 110	P	11/08/2024	18:55	LB133365
	Beryllium	236	250	95	90 - 110	P	11/08/2024	18:55	LB133365
	Cadmium	2360	2500	95	90 - 110	P	11/08/2024	18:55	LB133365
	Calcium	23500	25000	94	90 - 110	P	11/08/2024	18:55	LB133365
	Chromium	923	1000	92	90 - 110	P	11/08/2024	18:55	LB133365
	Cobalt	2400	2500	96	90 - 110	P	11/08/2024	18:55	LB133365
	Copper	1250	1250	100	90 - 110	P	11/08/2024	18:55	LB133365
	Iron	4700	5000	94	90 - 110	P	11/08/2024	18:55	LB133365
	Lead	4830	5000	97	90 - 110	P	11/08/2024	18:55	LB133365
	Magnesium	23500	25000	94	90 - 110	P	11/08/2024	18:55	LB133365
	Manganese	2290	2500	92	90 - 110	P	11/08/2024	18:55	LB133365
	Nickel	2400	2500	96	90 - 110	P	11/08/2024	18:55	LB133365
	Potassium	24500	25000	98	90 - 110	P	11/08/2024	18:55	LB133365
	Selenium	5100	5000	102	90 - 110	P	11/08/2024	18:55	LB133365
	Silver	1220	1250	98	90 - 110	P	11/08/2024	18:55	LB133365
	Sodium	23900	25000	96	90 - 110	P	11/08/2024	18:55	LB133365
	Thallium	4990	5000	100	90 - 110	P	11/08/2024	18:55	LB133365
	Vanadium	2390	2500	96	90 - 110	P	11/08/2024	18:55	LB133365
	Zinc	2530	2500	101	90 - 110	P	11/08/2024	18:55	LB133365
CCV07	Aluminum	10100	10000	101	90 - 110	P	11/08/2024	19:47	LB133365
	Antimony	5180	5000	104	90 - 110	P	11/08/2024	19:47	LB133365
	Arsenic	5060	5000	101	90 - 110	P	11/08/2024	19:47	LB133365
	Barium	9520	10000	95	90 - 110	P	11/08/2024	19:47	LB133365

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	243	250	97	90 - 110	P	11/08/2024	19:47	LB133365
	Cadmium	2400	2500	96	90 - 110	P	11/08/2024	19:47	LB133365
	Calcium	24000	25000	96	90 - 110	P	11/08/2024	19:47	LB133365
	Chromium	920	1000	92	90 - 110	P	11/08/2024	19:47	LB133365
	Cobalt	2440	2500	98	90 - 110	P	11/08/2024	19:47	LB133365
	Copper	1280	1250	102	90 - 110	P	11/08/2024	19:47	LB133365
	Iron	4710	5000	94	90 - 110	P	11/08/2024	19:47	LB133365
	Lead	4910	5000	98	90 - 110	P	11/08/2024	19:47	LB133365
	Magnesium	24100	25000	97	90 - 110	P	11/08/2024	19:47	LB133365
	Manganese	2310	2500	92	90 - 110	P	11/08/2024	19:47	LB133365
	Nickel	2440	2500	98	90 - 110	P	11/08/2024	19:47	LB133365
	Potassium	25000	25000	100	90 - 110	P	11/08/2024	19:47	LB133365
	Selenium	5210	5000	104	90 - 110	P	11/08/2024	19:47	LB133365
	Silver	1250	1250	100	90 - 110	P	11/08/2024	19:47	LB133365
	Sodium	24600	25000	98	90 - 110	P	11/08/2024	19:47	LB133365
	Thallium	5050	5000	101	90 - 110	P	11/08/2024	19:47	LB133365
	Vanadium	2440	2500	98	90 - 110	P	11/08/2024	19:47	LB133365
	Zinc	2580	2500	103	90 - 110	P	11/08/2024	19:47	LB133365
CCV08	Aluminum	9430	10000	94	90 - 110	P	11/08/2024	20:48	LB133365
	Antimony	4620	5000	92	90 - 110	P	11/08/2024	20:48	LB133365
	Arsenic	4620	5000	92	90 - 110	P	11/08/2024	20:48	LB133365
	Barium	9200	10000	92	90 - 110	P	11/08/2024	20:48	LB133365
	Beryllium	257	250	103	90 - 110	P	11/08/2024	20:48	LB133365
	Cadmium	2370	2500	95	90 - 110	P	11/08/2024	20:48	LB133365
	Calcium	24000	25000	96	90 - 110	P	11/08/2024	20:48	LB133365
	Chromium	920	1000	92	90 - 110	P	11/08/2024	20:48	LB133365
	Cobalt	2370	2500	95	90 - 110	P	11/08/2024	20:48	LB133365
	Copper	1170	1250	94	90 - 110	P	11/08/2024	20:48	LB133365
	Iron	4520	5000	90	90 - 110	P	11/08/2024	20:48	LB133365
	Lead	4790	5000	96	90 - 110	P	11/08/2024	20:48	LB133365
	Magnesium	24600	25000	98	90 - 110	P	11/08/2024	20:48	LB133365
	Manganese	2300	2500	92	90 - 110	P	11/08/2024	20:48	LB133365
	Nickel	2380	2500	95	90 - 110	P	11/08/2024	20:48	LB133365
	Potassium	25700	25000	103	90 - 110	P	11/08/2024	20:48	LB133365

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV08	Selenium	4670	5000	93	90 - 110	P	11/08/2024	20:48	LB133365
	Silver	1190	1250	95	90 - 110	P	11/08/2024	20:48	LB133365
	Sodium	24200	25000	97	90 - 110	P	11/08/2024	20:48	LB133365
	Thallium	4810	5000	96	90 - 110	P	11/08/2024	20:48	LB133365
	Vanadium	2370	2500	95	90 - 110	P	11/08/2024	20:48	LB133365
	Zinc	2380	2500	95	90 - 110	P	11/08/2024	20:48	LB133365
CCV09	Aluminum	9360	10000	94	90 - 110	P	11/08/2024	21:50	LB133365
	Antimony	4620	5000	92	90 - 110	P	11/08/2024	21:50	LB133365
	Arsenic	4610	5000	92	90 - 110	P	11/08/2024	21:50	LB133365
	Barium	9110	10000	91	90 - 110	P	11/08/2024	21:50	LB133365
	Beryllium	251	250	100	90 - 110	P	11/08/2024	21:50	LB133365
	Cadmium	2310	2500	93	90 - 110	P	11/08/2024	21:50	LB133365
	Calcium	23600	25000	95	90 - 110	P	11/08/2024	21:50	LB133365
	Chromium	902	1000	90	90 - 110	P	11/08/2024	21:50	LB133365
	Cobalt	2330	2500	93	90 - 110	P	11/08/2024	21:50	LB133365
	Copper	1160	1250	93	90 - 110	P	11/08/2024	21:50	LB133365
	Iron	5470	5000	109	90 - 110	P	11/08/2024	21:50	LB133365
	Lead	4710	5000	94	90 - 110	P	11/08/2024	21:50	LB133365
	Magnesium	24200	25000	97	90 - 110	P	11/08/2024	21:50	LB133365
	Manganese	2270	2500	91	90 - 110	P	11/08/2024	21:50	LB133365
	Nickel	2330	2500	93	90 - 110	P	11/08/2024	21:50	LB133365
	Potassium	25800	25000	103	90 - 110	P	11/08/2024	21:50	LB133365
	Selenium	4690	5000	94	90 - 110	P	11/08/2024	21:50	LB133365
	Silver	1190	1250	95	90 - 110	P	11/08/2024	21:50	LB133365
	Sodium	23900	25000	96	90 - 110	P	11/08/2024	21:50	LB133365
	Thallium	4710	5000	94	90 - 110	P	11/08/2024	21:50	LB133365
	Vanadium	2350	2500	94	90 - 110	P	11/08/2024	21:50	LB133365
	Zinc	2420	2500	97	90 - 110	P	11/08/2024	21:50	LB133365
CCV10	Aluminum	10100	10000	101	90 - 110	P	11/08/2024	23:24	LB133365
	Antimony	4640	5000	93	90 - 110	P	11/08/2024	23:24	LB133365
	Arsenic	5340	5000	107	90 - 110	P	11/08/2024	23:24	LB133365
	Barium	9550	10000	96	90 - 110	P	11/08/2024	23:24	LB133365
	Beryllium	259	250	104	90 - 110	P	11/08/2024	23:24	LB133365
	Cadmium	2430	2500	97	90 - 110	P	11/08/2024	23:24	LB133365

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture SDG No.: P4718
Contract: PORT06 Lab Code: CHEM Case No.: P4718 SAS No.: P4718
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV10	Calcium	22800	25000	91	90 - 110	P	11/08/2024	23:24	LB133365
	Chromium	1000	1000	100	90 - 110	P	11/08/2024	23:24	LB133365
	Cobalt	2530	2500	101	90 - 110	P	11/08/2024	23:24	LB133365
	Copper	1130	1250	91	90 - 110	P	11/08/2024	23:24	LB133365
	Iron	5150	5000	103	90 - 110	P	11/08/2024	23:24	LB133365
	Lead	4980	5000	100	90 - 110	P	11/08/2024	23:24	LB133365
	Magnesium	24100	25000	96	90 - 110	P	11/08/2024	23:24	LB133365
	Manganese	2380	2500	95	90 - 110	P	11/08/2024	23:24	LB133365
	Nickel	2360	2500	94	90 - 110	P	11/08/2024	23:24	LB133365
	Potassium	26300	25000	105	90 - 110	P	11/08/2024	23:24	LB133365
	Selenium	4900	5000	98	90 - 110	P	11/08/2024	23:24	LB133365
	Silver	1270	1250	102	90 - 110	P	11/08/2024	23:24	LB133365
	Sodium	26100	25000	104	90 - 110	P	11/08/2024	23:24	LB133365
	Thallium	4930	5000	98	90 - 110	P	11/08/2024	23:24	LB133365
	Vanadium	2610	2500	104	90 - 110	P	11/08/2024	23:24	LB133365
	Zinc	2690	2500	108	90 - 110	P	11/08/2024	23:24	LB133365

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	2650	2500	106	90 - 110	P	11/19/2024	11:25	LB133524
	Antimony	1010	1000	101	90 - 110	P	11/19/2024	11:25	LB133524
	Arsenic	1020	1000	102	90 - 110	P	11/19/2024	11:25	LB133524
	Barium	542	520	104	90 - 110	P	11/19/2024	11:25	LB133524
	Beryllium	536	510	105	90 - 110	P	11/19/2024	11:25	LB133524
	Cadmium	526	510	103	90 - 110	P	11/19/2024	11:25	LB133524
	Calcium	10700	10000	107	90 - 110	P	11/19/2024	11:25	LB133524
	Chromium	551	520	106	90 - 110	P	11/19/2024	11:25	LB133524
	Cobalt	531	520	102	90 - 110	P	11/19/2024	11:25	LB133524
	Copper	538	510	105	90 - 110	P	11/19/2024	11:25	LB133524
	Iron	10400	10000	104	90 - 110	P	11/19/2024	11:25	LB133524
	Lead	1040	1000	104	90 - 110	P	11/19/2024	11:25	LB133524
	Magnesium	6280	6000	105	90 - 110	P	11/19/2024	11:25	LB133524
	Manganese	550	520	106	90 - 110	P	11/19/2024	11:25	LB133524
	Nickel	532	530	100	90 - 110	P	11/19/2024	11:25	LB133524
	Potassium	10000	9900	101	90 - 110	P	11/19/2024	11:25	LB133524
	Selenium	1040	1000	104	90 - 110	P	11/19/2024	11:25	LB133524
	Silver	261	250	105	90 - 110	P	11/19/2024	11:25	LB133524
	Sodium	9760	10000	98	90 - 110	P	11/19/2024	11:25	LB133524
	Thallium	1090	1000	109	90 - 110	P	11/19/2024	11:25	LB133524
	Vanadium	524	500	105	90 - 110	P	11/19/2024	11:25	LB133524
	Zinc	1060	1000	106	90 - 110	P	11/19/2024	11:25	LB133524

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	104	100	104	80 - 120	P	11/19/2024	11:34	LB133524
	Antimony	51.1	50.0	102	80 - 120	P	11/19/2024	11:34	LB133524
	Arsenic	20.5	20.0	102	80 - 120	P	11/19/2024	11:34	LB133524
	Barium	104	100	104	80 - 120	P	11/19/2024	11:34	LB133524
	Beryllium	5.85	6.0	98	80 - 120	P	11/19/2024	11:34	LB133524
	Cadmium	5.89	6.0	98	80 - 120	P	11/19/2024	11:34	LB133524
	Calcium	2090	2000	105	80 - 120	P	11/19/2024	11:34	LB133524
	Chromium	10.7	10.0	107	80 - 120	P	11/19/2024	11:34	LB133524
	Cobalt	29.9	30.0	100	80 - 120	P	11/19/2024	11:34	LB133524
	Copper	22.3	20.0	112	80 - 120	P	11/19/2024	11:34	LB133524
	Iron	100	100	100	80 - 120	P	11/19/2024	11:34	LB133524
	Lead	12.4	12.0	103	80 - 120	P	11/19/2024	11:34	LB133524
	Magnesium	2040	2000	102	80 - 120	P	11/19/2024	11:34	LB133524
	Manganese	21.4	20.0	107	80 - 120	P	11/19/2024	11:34	LB133524
	Nickel	39.7	40.0	99	80 - 120	P	11/19/2024	11:34	LB133524
	Potassium	1920	2000	96	80 - 120	P	11/19/2024	11:34	LB133524
	Selenium	20.6	20.0	103	80 - 120	P	11/19/2024	11:34	LB133524
	Silver	10.1	10.0	101	80 - 120	P	11/19/2024	11:34	LB133524
	Sodium	1820	2000	91	80 - 120	P	11/19/2024	11:34	LB133524
	Thallium	39.2	40.0	98	80 - 120	P	11/19/2024	11:34	LB133524
	Vanadium	40.3	40.0	101	80 - 120	P	11/19/2024	11:34	LB133524
	Zinc	45.8	40.0	114	80 - 120	P	11/19/2024	11:34	LB133524

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9940	10000	99	90 - 110	P	11/19/2024	12:43	LB133524
	Antimony	4980	5000	100	90 - 110	P	11/19/2024	12:43	LB133524
	Arsenic	4970	5000	99	90 - 110	P	11/19/2024	12:43	LB133524
	Barium	10000	10000	100	90 - 110	P	11/19/2024	12:43	LB133524
	Beryllium	249	250	100	90 - 110	P	11/19/2024	12:43	LB133524
	Cadmium	2490	2500	100	90 - 110	P	11/19/2024	12:43	LB133524
	Calcium	24800	25000	99	90 - 110	P	11/19/2024	12:43	LB133524
	Chromium	996	1000	100	90 - 110	P	11/19/2024	12:43	LB133524
	Cobalt	2490	2500	100	90 - 110	P	11/19/2024	12:43	LB133524
	Copper	1260	1250	101	90 - 110	P	11/19/2024	12:43	LB133524
	Iron	4880	5000	98	90 - 110	P	11/19/2024	12:43	LB133524
	Lead	4990	5000	100	90 - 110	P	11/19/2024	12:43	LB133524
	Magnesium	24800	25000	99	90 - 110	P	11/19/2024	12:43	LB133524
	Manganese	2490	2500	100	90 - 110	P	11/19/2024	12:43	LB133524
	Nickel	2500	2500	100	90 - 110	P	11/19/2024	12:43	LB133524
	Potassium	24200	25000	97	90 - 110	P	11/19/2024	12:43	LB133524
	Selenium	4950	5000	99	90 - 110	P	11/19/2024	12:43	LB133524
	Silver	1230	1250	98	90 - 110	P	11/19/2024	12:43	LB133524
	Sodium	24200	25000	97	90 - 110	P	11/19/2024	12:43	LB133524
	Thallium	5030	5000	101	90 - 110	P	11/19/2024	12:43	LB133524
	Vanadium	2500	2500	100	90 - 110	P	11/19/2024	12:43	LB133524
	Zinc	2490	2500	100	90 - 110	P	11/19/2024	12:43	LB133524
CCV02	Aluminum	10600	10000	106	90 - 110	P	11/19/2024	13:34	LB133524
	Antimony	5360	5000	107	90 - 110	P	11/19/2024	13:34	LB133524
	Arsenic	5320	5000	106	90 - 110	P	11/19/2024	13:34	LB133524
	Barium	10300	10000	103	90 - 110	P	11/19/2024	13:34	LB133524
	Beryllium	262	250	105	90 - 110	P	11/19/2024	13:34	LB133524
	Cadmium	2610	2500	104	90 - 110	P	11/19/2024	13:34	LB133524
	Calcium	26100	25000	104	90 - 110	P	11/19/2024	13:34	LB133524
	Chromium	1070	1000	107	90 - 110	P	11/19/2024	13:34	LB133524
	Cobalt	2620	2500	105	90 - 110	P	11/19/2024	13:34	LB133524
	Copper	1350	1250	108	90 - 110	P	11/19/2024	13:34	LB133524
	Iron	5270	5000	105	90 - 110	P	11/19/2024	13:34	LB133524
	Lead	5240	5000	105	90 - 110	P	11/19/2024	13:34	LB133524

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	26200	25000	105	90 - 110	P	11/19/2024	13:34	LB133524
	Manganese	2580	2500	103	90 - 110	P	11/19/2024	13:34	LB133524
	Nickel	2620	2500	105	90 - 110	P	11/19/2024	13:34	LB133524
	Potassium	26300	25000	105	90 - 110	P	11/19/2024	13:34	LB133524
	Selenium	5340	5000	107	90 - 110	P	11/19/2024	13:34	LB133524
	Silver	1320	1250	106	90 - 110	P	11/19/2024	13:34	LB133524
	Sodium	25900	25000	103	90 - 110	P	11/19/2024	13:34	LB133524
	Thallium	5230	5000	104	90 - 110	P	11/19/2024	13:34	LB133524
	Vanadium	2640	2500	105	90 - 110	P	11/19/2024	13:34	LB133524
	Zinc	2680	2500	107	90 - 110	P	11/19/2024	13:34	LB133524
CCV03	Aluminum	10400	10000	104	90 - 110	P	11/19/2024	14:24	LB133524
	Antimony	5130	5000	103	90 - 110	P	11/19/2024	14:24	LB133524
	Arsenic	5110	5000	102	90 - 110	P	11/19/2024	14:24	LB133524
	Barium	10300	10000	102	90 - 110	P	11/19/2024	14:24	LB133524
	Beryllium	266	250	106	90 - 110	P	11/19/2024	14:24	LB133524
	Cadmium	2590	2500	104	90 - 110	P	11/19/2024	14:24	LB133524
	Calcium	26000	25000	104	90 - 110	P	11/19/2024	14:24	LB133524
	Chromium	1040	1000	104	90 - 110	P	11/19/2024	14:24	LB133524
	Cobalt	2600	2500	104	90 - 110	P	11/19/2024	14:24	LB133524
	Copper	1300	1250	104	90 - 110	P	11/19/2024	14:24	LB133524
	Iron	4990	5000	100	90 - 110	P	11/19/2024	14:24	LB133524
	Lead	5200	5000	104	90 - 110	P	11/19/2024	14:24	LB133524
	Magnesium	26000	25000	104	90 - 110	P	11/19/2024	14:24	LB133524
	Manganese	2590	2500	104	90 - 110	P	11/19/2024	14:24	LB133524
	Nickel	2600	2500	104	90 - 110	P	11/19/2024	14:24	LB133524
	Potassium	24600	25000	98	90 - 110	P	11/19/2024	14:24	LB133524
	Selenium	5080	5000	102	90 - 110	P	11/19/2024	14:24	LB133524
	Silver	1280	1250	102	90 - 110	P	11/19/2024	14:24	LB133524
	Sodium	24500	25000	98	90 - 110	P	11/19/2024	14:24	LB133524
	Thallium	5160	5000	103	90 - 110	P	11/19/2024	14:24	LB133524
	Vanadium	2600	2500	104	90 - 110	P	11/19/2024	14:24	LB133524
	Zinc	2590	2500	104	90 - 110	P	11/19/2024	14:24	LB133524
CCV04	Aluminum	10400	10000	104	90 - 110	P	11/19/2024	15:31	LB133524
	Antimony	5290	5000	106	90 - 110	P	11/19/2024	15:31	LB133524

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	5260	5000	105	90 - 110	P	11/19/2024	15:31	LB133524
	Barium	10500	10000	105	90 - 110	P	11/19/2024	15:31	LB133524
	Beryllium	261	250	104	90 - 110	P	11/19/2024	15:31	LB133524
	Cadmium	2610	2500	104	90 - 110	P	11/19/2024	15:31	LB133524
	Calcium	25800	25000	103	90 - 110	P	11/19/2024	15:31	LB133524
	Chromium	1050	1000	105	90 - 110	P	11/19/2024	15:31	LB133524
	Cobalt	2610	2500	105	90 - 110	P	11/19/2024	15:31	LB133524
	Copper	1330	1250	106	90 - 110	P	11/19/2024	15:31	LB133524
	Iron	5080	5000	102	90 - 110	P	11/19/2024	15:31	LB133524
	Lead	5230	5000	105	90 - 110	P	11/19/2024	15:31	LB133524
	Magnesium	25800	25000	103	90 - 110	P	11/19/2024	15:31	LB133524
	Manganese	2570	2500	103	90 - 110	P	11/19/2024	15:31	LB133524
	Nickel	2620	2500	105	90 - 110	P	11/19/2024	15:31	LB133524
	Potassium	25500	25000	102	90 - 110	P	11/19/2024	15:31	LB133524
	Selenium	5280	5000	106	90 - 110	P	11/19/2024	15:31	LB133524
	Silver	1290	1250	103	90 - 110	P	11/19/2024	15:31	LB133524
	Sodium	25600	25000	102	90 - 110	P	11/19/2024	15:31	LB133524
	Thallium	5410	5000	108	90 - 110	P	11/19/2024	15:31	LB133524
	Vanadium	2610	2500	104	90 - 110	P	11/19/2024	15:31	LB133524
	Zinc	2620	2500	105	90 - 110	P	11/19/2024	15:31	LB133524
CCV05	Aluminum	10400	10000	104	90 - 110	P	11/19/2024	16:18	LB133524
	Antimony	5220	5000	104	90 - 110	P	11/19/2024	16:18	LB133524
	Arsenic	5210	5000	104	90 - 110	P	11/19/2024	16:18	LB133524
	Barium	10300	10000	103	90 - 110	P	11/19/2024	16:18	LB133524
	Beryllium	257	250	103	90 - 110	P	11/19/2024	16:18	LB133524
	Cadmium	2580	2500	103	90 - 110	P	11/19/2024	16:18	LB133524
	Calcium	25600	25000	102	90 - 110	P	11/19/2024	16:18	LB133524
	Chromium	1050	1000	105	90 - 110	P	11/19/2024	16:18	LB133524
	Cobalt	2580	2500	103	90 - 110	P	11/19/2024	16:18	LB133524
	Copper	1310	1250	105	90 - 110	P	11/19/2024	16:18	LB133524
	Iron	5140	5000	103	90 - 110	P	11/19/2024	16:18	LB133524
	Lead	5170	5000	104	90 - 110	P	11/19/2024	16:18	LB133524
	Magnesium	25600	25000	102	90 - 110	P	11/19/2024	16:18	LB133524
	Manganese	2540	2500	102	90 - 110	P	11/19/2024	16:18	LB133524

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2590	2500	104	90 - 110	P	11/19/2024	16:18	LB133524
	Potassium	25700	25000	103	90 - 110	P	11/19/2024	16:18	LB133524
	Selenium	5240	5000	105	90 - 110	P	11/19/2024	16:18	LB133524
	Silver	1290	1250	103	90 - 110	P	11/19/2024	16:18	LB133524
	Sodium	25300	25000	101	90 - 110	P	11/19/2024	16:18	LB133524
	Thallium	5210	5000	104	90 - 110	P	11/19/2024	16:18	LB133524
	Vanadium	2590	2500	104	90 - 110	P	11/19/2024	16:18	LB133524
	Zinc	2640	2500	106	90 - 110	P	11/19/2024	16:18	LB133524
CCV06	Aluminum	10500	10000	105	90 - 110	P	11/19/2024	17:19	LB133524
	Antimony	5280	5000	106	90 - 110	P	11/19/2024	17:19	LB133524
	Arsenic	5270	5000	105	90 - 110	P	11/19/2024	17:19	LB133524
	Barium	10400	10000	104	90 - 110	P	11/19/2024	17:19	LB133524
	Beryllium	261	250	104	90 - 110	P	11/19/2024	17:19	LB133524
	Cadmium	2610	2500	104	90 - 110	P	11/19/2024	17:19	LB133524
	Calcium	25800	25000	103	90 - 110	P	11/19/2024	17:19	LB133524
	Chromium	1060	1000	106	90 - 110	P	11/19/2024	17:19	LB133524
	Cobalt	2600	2500	104	90 - 110	P	11/19/2024	17:19	LB133524
	Copper	1330	1250	106	90 - 110	P	11/19/2024	17:19	LB133524
	Iron	5190	5000	104	90 - 110	P	11/19/2024	17:19	LB133524
	Lead	5240	5000	105	90 - 110	P	11/19/2024	17:19	LB133524
	Magnesium	25800	25000	103	90 - 110	P	11/19/2024	17:19	LB133524
	Manganese	2540	2500	102	90 - 110	P	11/19/2024	17:19	LB133524
	Nickel	2610	2500	104	90 - 110	P	11/19/2024	17:19	LB133524
	Potassium	25900	25000	104	90 - 110	P	11/19/2024	17:19	LB133524
	Selenium	5310	5000	106	90 - 110	P	11/19/2024	17:19	LB133524
	Silver	1310	1250	105	90 - 110	P	11/19/2024	17:19	LB133524
	Sodium	25900	25000	103	90 - 110	P	11/19/2024	17:19	LB133524
	Thallium	5300	5000	106	90 - 110	P	11/19/2024	17:19	LB133524
	Vanadium	2600	2500	104	90 - 110	P	11/19/2024	17:19	LB133524
	Zinc	2680	2500	107	90 - 110	P	11/19/2024	17:19	LB133524
CCV07	Aluminum	10300	10000	102	90 - 110	P	11/19/2024	18:01	LB133524
	Antimony	5080	5000	102	90 - 110	P	11/19/2024	18:01	LB133524
	Arsenic	5070	5000	101	90 - 110	P	11/19/2024	18:01	LB133524
	Barium	10600	10000	106	90 - 110	P	11/19/2024	18:01	LB133524

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture SDG No.: P4718
Contract: PORT06 Lab Code: CHEM Case No.: P4718 SAS No.: P4718
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	256	250	102	90 - 110	P	11/19/2024	18:01	LB133524
	Cadmium	2490	2500	99	90 - 110	P	11/19/2024	18:01	LB133524
	Calcium	25100	25000	100	90 - 110	P	11/19/2024	18:01	LB133524
	Chromium	1020	1000	102	90 - 110	P	11/19/2024	18:01	LB133524
	Cobalt	2490	2500	100	90 - 110	P	11/19/2024	18:01	LB133524
	Copper	1290	1250	104	90 - 110	P	11/19/2024	18:01	LB133524
	Iron	4990	5000	100	90 - 110	P	11/19/2024	18:01	LB133524
	Lead	5020	5000	100	90 - 110	P	11/19/2024	18:01	LB133524
	Magnesium	25100	25000	100	90 - 110	P	11/19/2024	18:01	LB133524
	Manganese	2600	2500	104	90 - 110	P	11/19/2024	18:01	LB133524
	Nickel	2500	2500	100	90 - 110	P	11/19/2024	18:01	LB133524
	Potassium	25200	25000	101	90 - 110	P	11/19/2024	18:01	LB133524
	Selenium	5100	5000	102	90 - 110	P	11/19/2024	18:01	LB133524
	Silver	1270	1250	101	90 - 110	P	11/19/2024	18:01	LB133524
	Sodium	25600	25000	102	90 - 110	P	11/19/2024	18:01	LB133524
	Thallium	4940	5000	99	90 - 110	P	11/19/2024	18:01	LB133524
	Vanadium	2540	2500	102	90 - 110	P	11/19/2024	18:01	LB133524
	Zinc	2480	2500	99	90 - 110	P	11/19/2024	18:01	LB133524



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Portal Partners Tri-Venture SDG No.: P4718
Contract: PORT06 Lab Code: CHEM Case No.: P4718 SAS No.: P4718
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Aluminum	106	100	106	40 - 160	P	11/07/2024	14:42	LB133344
	Antimony	51.9	50.0	104	40 - 160	P	11/07/2024	14:42	LB133344
	Arsenic	19.9	20.0	99	40 - 160	P	11/07/2024	14:42	LB133344
	Barium	99.8	100	100	40 - 160	P	11/07/2024	14:42	LB133344
	Beryllium	5.88	6.0	98	40 - 160	P	11/07/2024	14:42	LB133344
	Cadmium	6.45	6.0	108	40 - 160	P	11/07/2024	14:42	LB133344
	Calcium	2010	2000	101	40 - 160	P	11/07/2024	14:42	LB133344
	Chromium	9.25	10.0	92	40 - 160	P	11/07/2024	14:42	LB133344
	Cobalt	29.3	30.0	98	40 - 160	P	11/07/2024	14:42	LB133344
	Copper	22.5	20.0	112	40 - 160	P	11/07/2024	14:42	LB133344
	Iron	95.5	100	96	40 - 160	P	11/07/2024	14:42	LB133344
	Lead	11.3	12.0	94	40 - 160	P	11/07/2024	14:42	LB133344
	Magnesium	2000	2000	100	40 - 160	P	11/07/2024	14:42	LB133344
	Manganese	20.5	20.0	102	40 - 160	P	11/07/2024	14:42	LB133344
	Nickel	38.8	40.0	97	40 - 160	P	11/07/2024	14:42	LB133344
	Potassium	1950	2000	98	40 - 160	P	11/07/2024	14:42	LB133344
	Selenium	18.8	20.0	94	40 - 160	P	11/07/2024	14:42	LB133344
	Silver	10.7	10.0	107	40 - 160	P	11/07/2024	14:42	LB133344
	Sodium	1610	2000	81	40 - 160	P	11/07/2024	14:42	LB133344
	Thallium	41.6	40.0	104	40 - 160	P	11/07/2024	14:42	LB133344
	Vanadium	40.4	40.0	101	40 - 160	P	11/07/2024	14:42	LB133344
	Zinc	41.4	40.0	104	40 - 160	P	11/07/2024	14:42	LB133344
CRA	Mercury	0.17	0.2	83	40 - 160	CV	11/07/2024	14:50	LB133334
CRI01	Aluminum	104	100	104	40 - 160	P	11/08/2024	14:20	LB133365
	Antimony	47.8	50.0	96	40 - 160	P	11/08/2024	14:20	LB133365
	Arsenic	20.1	20.0	100	40 - 160	P	11/08/2024	14:20	LB133365
	Barium	95.0	100	95	40 - 160	P	11/08/2024	14:20	LB133365
	Beryllium	5.62	6.0	94	40 - 160	P	11/08/2024	14:20	LB133365
	Cadmium	6.16	6.0	103	40 - 160	P	11/08/2024	14:20	LB133365
	Calcium	1950	2000	98	40 - 160	P	11/08/2024	14:20	LB133365
	Chromium	9.24	10.0	92	40 - 160	P	11/08/2024	14:20	LB133365
	Cobalt	27.8	30.0	92	40 - 160	P	11/08/2024	14:20	LB133365
	Copper	21.4	20.0	107	40 - 160	P	11/08/2024	14:20	LB133365
	Iron	92.9	100	93	40 - 160	P	11/08/2024	14:20	LB133365
	Lead	11.8	12.0	98	40 - 160	P	11/08/2024	14:20	LB133365

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Magnesium	1930	2000	97	40 - 160	P	11/08/2024	14:20	LB133365
	Manganese	19.6	20.0	98	40 - 160	P	11/08/2024	14:20	LB133365
	Nickel	36.8	40.0	92	40 - 160	P	11/08/2024	14:20	LB133365
	Potassium	1840	2000	92	40 - 160	P	11/08/2024	14:20	LB133365
	Selenium	19.8	20.0	99	40 - 160	P	11/08/2024	14:20	LB133365
	Silver	10.3	10.0	103	40 - 160	P	11/08/2024	14:20	LB133365
	Sodium	1600	2000	80	40 - 160	P	11/08/2024	14:20	LB133365
	Thallium	37.3	40.0	93	40 - 160	P	11/08/2024	14:20	LB133365
	Vanadium	38.5	40.0	96	40 - 160	P	11/08/2024	14:20	LB133365
	Zinc	41.9	40.0	105	40 - 160	P	11/08/2024	14:20	LB133365
CRA	Mercury	0.21	0.2	103	40 - 160	CV	11/11/2024	14:10	LB133392
CRI01	Aluminum	107	100	107	40 - 160	P	11/19/2024	11:43	LB133524
	Antimony	49.1	50.0	98	40 - 160	P	11/19/2024	11:43	LB133524
	Arsenic	19.4	20.0	97	40 - 160	P	11/19/2024	11:43	LB133524
	Barium	105	100	105	40 - 160	P	11/19/2024	11:43	LB133524
	Beryllium	5.90	6.0	98	40 - 160	P	11/19/2024	11:43	LB133524
	Cadmium	5.78	6.0	96	40 - 160	P	11/19/2024	11:43	LB133524
	Calcium	2070	2000	104	40 - 160	P	11/19/2024	11:43	LB133524
	Chromium	10.2	10.0	102	40 - 160	P	11/19/2024	11:43	LB133524
	Cobalt	29.6	30.0	99	40 - 160	P	11/19/2024	11:43	LB133524
	Copper	22.2	20.0	111	40 - 160	P	11/19/2024	11:43	LB133524
	Iron	96.0	100	96	40 - 160	P	11/19/2024	11:43	LB133524
	Lead	12.1	12.0	101	40 - 160	P	11/19/2024	11:43	LB133524
	Magnesium	2050	2000	102	40 - 160	P	11/19/2024	11:43	LB133524
	Manganese	21.1	20.0	106	40 - 160	P	11/19/2024	11:43	LB133524
	Nickel	39.1	40.0	98	40 - 160	P	11/19/2024	11:43	LB133524
	Potassium	1920	2000	96	40 - 160	P	11/19/2024	11:43	LB133524
	Selenium	18.1	20.0	90	40 - 160	P	11/19/2024	11:43	LB133524
	Silver	10.1	10.0	101	40 - 160	P	11/19/2024	11:43	LB133524
	Sodium	1810	2000	90	40 - 160	P	11/19/2024	11:43	LB133524
	Thallium	39.1	40.0	98	40 - 160	P	11/19/2024	11:43	LB133524
	Vanadium	39.6	40.0	99	40 - 160	P	11/19/2024	11:43	LB133524
	Zinc	42.4	40.0	106	40 - 160	P	11/19/2024	11:43	LB133524

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client:	Portal Partners Tri-Venture	SDG No.:	P4718
Contract:	PORT06	Lab Code:	CHEM
ICS Source:	EPA	Case No.:	P4718
		SAS No.:	P4718
		Instrument ID:	P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	255000	255000	100	216000	294000	11/07/2024	14:47	LB133344
	Antimony	-0.14			-50	50	11/07/2024	14:47	LB133344
	Arsenic	4.03			-20	20	11/07/2024	14:47	LB133344
	Barium	1.97	6.0	33	-94	106	11/07/2024	14:47	LB133344
	Beryllium	1.25			-6	6	11/07/2024	14:47	LB133344
	Cadmium	6.27	1.0	627	-5	7	11/07/2024	14:47	LB133344
	Calcium	235000	245000	96	208000	282000	11/07/2024	14:47	LB133344
	Chromium	54.7	52.0	105	42	62	11/07/2024	14:47	LB133344
	Cobalt	1.98			-30	30	11/07/2024	14:47	LB133344
	Copper	6.59	2.0	330	-18	22	11/07/2024	14:47	LB133344
	Iron	95800	101000	95	85600	116500	11/07/2024	14:47	LB133344
	Lead	8.91			-12	12	11/07/2024	14:47	LB133344
	Magnesium	257000	255000	101	216000	294000	11/07/2024	14:47	LB133344
	Manganese	4.04	7.0	58	-13	27	11/07/2024	14:47	LB133344
	Nickel	2.14	2.0	107	-38	42	11/07/2024	14:47	LB133344
	Potassium	27.1			0	0	11/07/2024	14:47	LB133344
	Selenium	-18.3			-20	20	11/07/2024	14:47	LB133344
	Silver	0.85			-10	10	11/07/2024	14:47	LB133344
	Sodium	-30.2			0	0	11/07/2024	14:47	LB133344
	Thallium	4.42			-40	40	11/07/2024	14:47	LB133344
	Vanadium	7.40			-40	40	11/07/2024	14:47	LB133344
	Zinc	6.19			-40	40	11/07/2024	14:47	LB133344
ICSAB01	Aluminum	261000	247000	106	209000	285000	11/07/2024	14:51	LB133344
	Antimony	621	618	100	525	711	11/07/2024	14:51	LB133344
	Arsenic	116	104	112	88.4	120	11/07/2024	14:51	LB133344
	Barium	504	537	94	437	637	11/07/2024	14:51	LB133344
	Beryllium	498	495	101	420	570	11/07/2024	14:51	LB133344
	Cadmium	999	972	103	826	1120	11/07/2024	14:51	LB133344
	Calcium	238000	235000	101	199000	271000	11/07/2024	14:51	LB133344
	Chromium	556	542	103	460	624	11/07/2024	14:51	LB133344
	Cobalt	503	476	106	404	548	11/07/2024	14:51	LB133344
	Copper	493	511	96	434	588	11/07/2024	14:51	LB133344
	Iron	98300	99300	99	84400	114500	11/07/2024	14:51	LB133344
	Lead	56.6	49.0	116	37	61	11/07/2024	14:51	LB133344
	Magnesium	260000	248000	105	210000	286000	11/07/2024	14:51	LB133344
	Manganese	484	507	96	430	584	11/07/2024	14:51	LB133344
	Nickel	991	954	104	810	1100	11/07/2024	14:51	LB133344
	Potassium	10.4			0	0	11/07/2024	14:51	LB133344
	Selenium	32.9	46.0	72	26	66	11/07/2024	14:51	LB133344
	Silver	201	201	100	170	232	11/07/2024	14:51	LB133344
	Sodium	-89.4			0	0	11/07/2024	14:51	LB133344
	Thallium	106	108	98	68	148	11/07/2024	14:51	LB133344

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Vanadium	493	491	100	417	565	11/07/2024	14:51	LB133344
	Zinc	835	952	88	809	1095	11/07/2024	14:51	LB133344
ICSA01	Aluminum	264000	255000	104	216000	294000	11/08/2024	14:24	LB133365
	Antimony	-1.65			-50	50	11/08/2024	14:24	LB133365
	Arsenic	5.52			-20	20	11/08/2024	14:24	LB133365
	Barium	1.84	6.0	31	-94	106	11/08/2024	14:24	LB133365
	Beryllium	1.29			-6	6	11/08/2024	14:24	LB133365
	Cadmium	6.11	1.0	611	-5	7	11/08/2024	14:24	LB133365
	Calcium	243000	245000	99	208000	282000	11/08/2024	14:24	LB133365
	Chromium	54.3	52.0	104	42	62	11/08/2024	14:24	LB133365
	Cobalt	2.11			-30	30	11/08/2024	14:24	LB133365
	Copper	14.3	2.0	715	-18	22	11/08/2024	14:24	LB133365
	Iron	98600	101000	98	85600	116500	11/08/2024	14:24	LB133365
	Lead	9.97			-12	12	11/08/2024	14:24	LB133365
	Magnesium	267000	255000	105	216000	294000	11/08/2024	14:24	LB133365
	Manganese	3.65	7.0	52	-13	27	11/08/2024	14:24	LB133365
	Nickel	2.63	2.0	132	-38	42	11/08/2024	14:24	LB133365
	Potassium	17.8			0	0	11/08/2024	14:24	LB133365
	Selenium	-14.4			-20	20	11/08/2024	14:24	LB133365
	Silver	-0.087			-10	10	11/08/2024	14:24	LB133365
	Sodium	15.0			0	0	11/08/2024	14:24	LB133365
	Thallium	-0.89			-40	40	11/08/2024	14:24	LB133365
	Vanadium	8.97			-40	40	11/08/2024	14:24	LB133365
	Zinc	7.04			-40	40	11/08/2024	14:24	LB133365
ICSAB01	Aluminum	260000	247000	105	209000	285000	11/08/2024	14:30	LB133365
	Antimony	622	618	101	525	711	11/08/2024	14:30	LB133365
	Arsenic	115	104	111	88.4	120	11/08/2024	14:30	LB133365
	Barium	496	537	92	437	637	11/08/2024	14:30	LB133365
	Beryllium	493	495	100	420	570	11/08/2024	14:30	LB133365
	Cadmium	979	972	101	826	1120	11/08/2024	14:30	LB133365
	Calcium	237000	235000	101	199000	271000	11/08/2024	14:30	LB133365
	Chromium	521	542	96	460	624	11/08/2024	14:30	LB133365
	Cobalt	498	476	105	404	548	11/08/2024	14:30	LB133365
	Copper	499	511	98	434	588	11/08/2024	14:30	LB133365
	Iron	95700	99300	96	84400	114500	11/08/2024	14:30	LB133365
	Lead	57.3	49.0	117	37	61	11/08/2024	14:30	LB133365
	Magnesium	260000	248000	105	210000	286000	11/08/2024	14:30	LB133365
	Manganese	480	507	95	430	584	11/08/2024	14:30	LB133365
	Nickel	981	954	103	810	1100	11/08/2024	14:30	LB133365
	Potassium	-6.03			0	0	11/08/2024	14:30	LB133365
	Selenium	31.2	46.0	68	26	66	11/08/2024	14:30	LB133365
	Silver	198	201	98	170	232	11/08/2024	14:30	LB133365

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client:	Portal Partners Tri-Venture	SDG No.:	P4718
Contract:	PORT06	Lab Code:	CHEM
ICS Source:	EPA	Case No.:	P4718
		SAS No.:	P4718
		Instrument ID:	P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Sodium	-36.6			0	0	11/08/2024	14:30	LB133365
	Thallium	90.2	108	84	68	148	11/08/2024	14:30	LB133365
	Vanadium	495	491	101	417	565	11/08/2024	14:30	LB133365
	Zinc	1090	952	114	809	1095	11/08/2024	14:30	LB133365
ICSA01	Aluminum	263000	255000	103	216000	294000	11/19/2024	11:47	LB133524
	Antimony	-1.78			-50	50	11/19/2024	11:47	LB133524
	Arsenic	3.30			-20	20	11/19/2024	11:47	LB133524
	Barium	5.39	6.0	90	-94	106	11/19/2024	11:47	LB133524
	Beryllium	1.27			-6	6	11/19/2024	11:47	LB133524
	Cadmium	3.48	1.0	348	-5	7	11/19/2024	11:47	LB133524
	Calcium	251000	245000	102	208000	282000	11/19/2024	11:47	LB133524
	Chromium	57.2	52.0	110	42	62	11/19/2024	11:47	LB133524
	Cobalt	2.14			-30	30	11/19/2024	11:47	LB133524
	Copper	7.48	2.0	374	-18	22	11/19/2024	11:47	LB133524
	Iron	98200	101000	97	85600	116500	11/19/2024	11:47	LB133524
	Lead	8.06			-12	12	11/19/2024	11:47	LB133524
	Magnesium	274000	255000	108	216000	294000	11/19/2024	11:47	LB133524
	Manganese	4.02	7.0	57	-13	27	11/19/2024	11:47	LB133524
	Nickel	2.06	2.0	103	-38	42	11/19/2024	11:47	LB133524
	Potassium	43.8			0	0	11/19/2024	11:47	LB133524
	Selenium	-17.4			-20	20	11/19/2024	11:47	LB133524
	Silver	-0.45			-10	10	11/19/2024	11:47	LB133524
	Sodium	49.8			0	0	11/19/2024	11:47	LB133524
	Thallium	-3.78			-40	40	11/19/2024	11:47	LB133524
	Vanadium	7.04			-40	40	11/19/2024	11:47	LB133524
	Zinc	5.70			-40	40	11/19/2024	11:47	LB133524
ICSAB01	Aluminum	265000	247000	107	209000	285000	11/19/2024	11:51	LB133524
	Antimony	626	618	101	525	711	11/19/2024	11:51	LB133524
	Arsenic	116	104	112	88.4	120	11/19/2024	11:51	LB133524
	Barium	525	537	98	437	637	11/19/2024	11:51	LB133524
	Beryllium	524	495	106	420	570	11/19/2024	11:51	LB133524
	Cadmium	832	972	86	826	1120	11/19/2024	11:51	LB133524
	Calcium	250000	235000	106	199000	271000	11/19/2024	11:51	LB133524
	Chromium	571	542	105	460	624	11/19/2024	11:51	LB133524
	Cobalt	518	476	109	404	548	11/19/2024	11:51	LB133524
	Copper	493	511	96	434	588	11/19/2024	11:51	LB133524
	Iron	99700	99300	100	84400	114500	11/19/2024	11:51	LB133524
	Lead	57.8	49.0	118	37	61	11/19/2024	11:51	LB133524
	Magnesium	273000	248000	110	210000	286000	11/19/2024	11:51	LB133524
	Manganese	509	507	100	430	584	11/19/2024	11:51	LB133524
	Nickel	1020	954	107	810	1100	11/19/2024	11:51	LB133524
	Potassium	62.1			0	0	11/19/2024	11:51	LB133524

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Selenium	31.5	46.0	68	26	66	11/19/2024	11:51	LB133524
	Silver	201	201	100	170	232	11/19/2024	11:51	LB133524
	Sodium	62.8			0	0	11/19/2024	11:51	LB133524
	Thallium	91.5	108	85	68	148	11/19/2024	11:51	LB133524
	Vanadium	509	491	104	417	565	11/19/2024	11:51	LB133524
	Zinc	1090	952	114	809	1095	11/19/2024	11:51	LB133524



METAL QC DATA

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Portal Partners Tri-Venture **level:** low **sdg no.:** P4718
contract: PORT06 **lab code:** CHEM **case no.:** P4718 **sas no.:** P4718
matrix: Solid **sample id:** P4699-01 **client id:** MIXED-DEMOMS
Percent Solids for Sample: 99.5 **Spiked ID:** P4699-01MS **Percent Solids for Spike Sample:** 99.5

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Aluminum	mg/Kg	75 - 125	4160		3900		85.5	297		P
Antimony	mg/Kg	75 - 125	35.5		0.18	J	34.2	103		P
Arsenic	mg/Kg	75 - 125	37.3		0.81	J	34.2	107		P
Barium	mg/Kg	75 - 125	26.5		17.9		8.6	100		P
Beryllium	mg/Kg	75 - 125	8.92		0.17	J	8.6	102		P
Cadmium	mg/Kg	75 - 125	8.39		0.46		8.6	92		P
Calcium	mg/Kg	75 - 125	67400		64800		42.8	6009		P
Chromium	mg/Kg	75 - 125	30.8		16.0		17.1	87		P
Cobalt	mg/Kg	75 - 125	9.10		1.10	J	8.6	93		P
Copper	mg/Kg	75 - 125	16.2		5.06		12.8	87		P
Iron	mg/Kg	75 - 125	3270		3100		130	129		P
Lead	mg/Kg	75 - 125	42.1		4.31		42.8	88		P
Magnesium	mg/Kg	75 - 125	13700		13200		85.5	539		P
Manganese	mg/Kg	75 - 125	152		142		8.6	117		P
Nickel	mg/Kg	75 - 125	23.1		3.50		21.4	92		P
Potassium	mg/Kg	75 - 125	1960		1410		430	128	N	P
Selenium	mg/Kg	75 - 125	86.0		0.82	U	85.5	101		P
Silver	mg/Kg	75 - 125	3.61		0.21	J	3.2	106		P
Sodium	mg/Kg	75 - 125	538		362		130	135	N	P
Thallium	mg/Kg	75 - 125	80.8		1.65	U	85.5	94		P
Vanadium	mg/Kg	75 - 125	16.8		5.40		12.8	89		P
Zinc	mg/Kg	75 - 125	58.9		50.0		8.6	103		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client:	<u>Portal Partners Tri-Venture</u>	level:	<u>low</u>	sdg no.:	<u>P4718</u>
contract:	<u>PORT06</u>	lab code:	<u>CHEM</u>	case no.:	<u>P4718</u>
matrix:	<u>Solid</u>	sample id:	<u>P4699-01</u>	client id:	<u>MIXED-DEMOMSD</u>
Percent Solids for Sample:	99.5	Spiked ID:	P4699-01MSD	Percent Solids for Spike Sample:	99.5

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Aluminum	mg/Kg	75 - 125	4550		3900		91.8	702		P
Antimony	mg/Kg	75 - 125	37.5		0.18	J	36.7	102		P
Arsenic	mg/Kg	75 - 125	39.2		0.81	J	36.7	105		P
Barium	mg/Kg	75 - 125	29.0		17.9		9.2	120		P
Beryllium	mg/Kg	75 - 125	8.59		0.17	J	9.2	91		P
Cadmium	mg/Kg	75 - 125	8.98		0.46		9.2	93		P
Calcium	mg/Kg	75 - 125	74000		64800		45.9	20038		P
Chromium	mg/Kg	75 - 125	33.1		16.0		18.4	93		P
Cobalt	mg/Kg	75 - 125	9.76		1.10	J	9.2	94		P
Copper	mg/Kg	75 - 125	17.3		5.06		13.8	89		P
Iron	mg/Kg	75 - 125	3530		3100		140	306		P
Lead	mg/Kg	75 - 125	45.2		4.31		45.9	89		P
Magnesium	mg/Kg	75 - 125	12800		13200		91.8	-451		P
Manganese	mg/Kg	75 - 125	151		142		9.2	91		P
Nickel	mg/Kg	75 - 125	24.8		3.50		22.9	93		P
Potassium	mg/Kg	75 - 125	2080		1410		460	145	N	P
Selenium	mg/Kg	75 - 125	90.0		0.82	U	91.8	98		P
Silver	mg/Kg	75 - 125	3.84		0.21	J	3.4	107		P
Sodium	mg/Kg	75 - 125	568		362		140	147	N	P
Thallium	mg/Kg	75 - 125	85.5		1.65	U	91.8	93		P
Vanadium	mg/Kg	75 - 125	18.3		5.40		13.8	93		P
Zinc	mg/Kg	75 - 125	63.5		50.0		9.2	146		P

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Portal Partners Tri-Venture **level:** low **sdg no.:** P4718
contract: PORT06 **lab code:** CHEM **case no.:** P4718 **sas no.:** P4718
matrix: Solid **sample id:** P4706-01 **client id:** TR-04-110424MS
Percent Solids for Sample: 93.5 **Spiked ID:** P4706-01MS **Percent Solids for Spike Sample:** 93.5

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	mg/Kg	80 - 120	0.33		0.060		0.27	99		CV

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client:	<u>Portal Partners Tri-Venture</u>	level:	<u>low</u>	sdg no.:	<u>P4718</u>		
contract:	<u>PORT06</u>	lab code:	<u>CHEM</u>	case no.:	<u>P4718</u>	sas no.:	<u>P4718</u>
matrix:	<u>Solid</u>	sample id:	<u>P4706-01</u>	client id:	<u>TR-04-110424MSD</u>		
Percent Solids for Sample:	93.5	Spiked ID:	<u>P4706-01MSD</u>	Percent Solids for Spike Sample:	93.5		

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	mg/Kg	80 - 120	0.33		0.060		0.28	95		CV

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Portal Partners Tri-Venture **level:** low **sdg no.:** P4718
contract: PORT06 **lab code:** CHEM **case no.:** P4718 **sas no.:** P4718
matrix: Water **sample id:** P4718-04 **client id:** WB-307-SWMS
Percent Solids for Sample: NA **Spiked ID:** P4718-04MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Aluminum	ug/L	75 - 125	1670		389		1000	128	N	P
Antimony	ug/L	75 - 125	562		25.0	U	400	140	N	P
Arsenic	ug/L	75 - 125	549		10.0	U	400	137	N	P
Barium	ug/L	75 - 125	134		34.4	J	100	99		P
Beryllium	ug/L	75 - 125	115		3.00	U	100	115		P
Cadmium	ug/L	75 - 125	87.4		0.19	J	100	87		P
Calcium	ug/L	75 - 125	210000		206000		500	779		P
Chromium	ug/L	75 - 125	157		1.78	J	200	78		P
Cobalt	ug/L	75 - 125	92.0		15.0	U	100	92		P
Copper	ug/L	75 - 125	125		10.0	U	150	83		P
Iron	ug/L	75 - 125	2060		521		1500	102		P
Lead	ug/L	75 - 125	436		6.00	U	500	87		P
Magnesium	ug/L	75 - 125	642000		628000		1000	1362		P
Manganese	ug/L	75 - 125	154		69.7		100	84		P
Nickel	ug/L	75 - 125	221		20.0	U	250	88		P
Potassium	ug/L	75 - 125	394000		425000		5000	-622		P
Selenium	ug/L	75 - 125	1370		10.0	U	1000	137	N	P
Silver	ug/L	75 - 125	65.9		5.00	U	37.5	176	N	P
Sodium	ug/L	75 - 125	4920000	D	5110000	D	1500	-12683		P
Thallium	ug/L	75 - 125	814		20.0	U	1000	81		P
Vanadium	ug/L	75 - 125	145		3.65	J	150	94		P
Zinc	ug/L	75 - 125	177		19.8	J	100	157	N	P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: Portal Partners Tri-Venture **level:** low **sdg no.:** P4718
contract: PORT06 **lab code:** CHEM **case no.:** P4718 **sas no.:** P4718
matrix: Water **sample id:** P4718-04 **client id:** WB-307-SWMSD
Percent Solids for Sample: NA **Spiked ID:** P4718-04MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Aluminum	ug/L	75 - 125	1700		389		1000	131	N	P
Antimony	ug/L	75 - 125	591		25.0	U	400	148	N	P
Arsenic	ug/L	75 - 125	570		10.0	U	400	142	N	P
Barium	ug/L	75 - 125	135		34.4	J	100	101		P
Beryllium	ug/L	75 - 125	114		3.00	U	100	114		P
Cadmium	ug/L	75 - 125	87.0		0.19	J	100	87		P
Calcium	ug/L	75 - 125	212000		206000		500	1340		P
Chromium	ug/L	75 - 125	156		1.78	J	200	77		P
Cobalt	ug/L	75 - 125	91.8		15.0	U	100	92		P
Copper	ug/L	75 - 125	127		10.0	U	150	85		P
Iron	ug/L	75 - 125	2120		521		1500	107		P
Lead	ug/L	75 - 125	434		6.00	U	500	87		P
Magnesium	ug/L	75 - 125	654000		628000		1000	2518		P
Manganese	ug/L	75 - 125	156		69.7		100	86		P
Nickel	ug/L	75 - 125	220		20.0	U	250	88		P
Potassium	ug/L	75 - 125	413000		425000		5000	-248		P
Selenium	ug/L	75 - 125	1430		10.0	U	1000	143	N	P
Silver	ug/L	75 - 125	67.9		5.00	U	37.5	181	N	P
Sodium	ug/L	75 - 125	5190000	D	5110000	D	1500	5649		P
Thallium	ug/L	75 - 125	810		20.0	U	1000	81		P
Vanadium	ug/L	75 - 125	145		3.65	J	150	94		P
Zinc	ug/L	75 - 125	182		19.8	J	100	162	N	P

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Portal Partners Tri-Venture **level:** low **sdg no.:** P4718
contract: PORT06 **lab code:** CHEM **case no.:** P4718 **sas no.:** P4718
matrix: Water **sample id:** P4791-02 **client id:** HINCHMAN-OILY-WATERMS
Percent Solids for Sample: NA **Spiked ID:** P4791-02MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	75 - 125	3.06		0.20	U	4.0	76		CV

metals

- 5a -

MATRIX SPIKE DUPLICATE SUMMARY

client: Portal Partners Tri-Venture **level:** low **sdg no.:** P4718
contract: PORT06 **lab code:** CHEM **case no.:** P4718 **sas no.:** P4718
matrix: Water **sample id:** P4791-02 **client id:** HINCHMAN-OILY-WATERMSD
Percent Solids for Sample: NA **Spiked ID:** P4791-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	75 - 125	3.12		0.20	U	4.0	78		CV

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718
Matrix: Solid **Level:** LOW **Client ID:** MIXED-DEMOA
Sample ID: P4699-01 **Spiked ID:** P4699-01A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Potassium	mg/Kg	75 - 125	1890		1410		410	116		P
Sodium	mg/Kg	75 - 125	522		362		120	133		P

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718
Matrix: Water **Level:** LOW **Client ID:** WB-307-SWA
Sample ID: P4718-04 **Spiked ID:** P4718-04A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Aluminum	ug/L	75 - 125	1660		389		10000	13		P
Antimony	ug/L	75 - 125	554		25.0	U	400	138		P
Arsenic	ug/L	75 - 125	539		10.0	U	400	135		P
Selenium	ug/L	75 - 125	1350		10.0	U	1000	135		P
Silver	ug/L	75 - 125	66.6		5.00	U	37.5	178		P
Zinc	ug/L	75 - 125	176		19.8	J	100	156		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Portal Partners Tri-Venture **Level:** LOW **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718
Matrix: Solid **Sample ID:** P4699-01 **Client ID:** MIXED-DEMODUP
Percent Solids for Sample: 99.5 **Duplicate ID** P4699-01DUP **Percent Solids for Spike Sample:** 99.5

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	mg/Kg	20	3900		4320		10		P
Antimony	mg/Kg	20	0.18	J	0.20	J	11		P
Arsenic	mg/Kg	20	0.81	J	0.95		16		P
Barium	mg/Kg	20	17.9		19.9		11		P
Beryllium	mg/Kg	20	0.17	J	0.19	J	12		P
Cadmium	mg/Kg	20	0.46		0.50		8		P
Calcium	mg/Kg	20	64800		72200		11		P
Chromium	mg/Kg	20	16.0		17.6		10		P
Cobalt	mg/Kg	20	1.10	J	1.20	J	9		P
Copper	mg/Kg	20	5.06		5.57		10		P
Iron	mg/Kg	20	3100		3390		9		P
Lead	mg/Kg	20	4.31		4.87		12		P
Magnesium	mg/Kg	20	13200		14600		10		P
Manganese	mg/Kg	20	142		156		9		P
Nickel	mg/Kg	20	3.50		3.86		10		P
Potassium	mg/Kg	20	1410		1540		9		P
Selenium	mg/Kg	20	0.82	U	0.90	U			P
Silver	mg/Kg	20	0.21	J	0.22	J	3		P
Sodium	mg/Kg	20	362		396		9		P
Thallium	mg/Kg	20	1.65	U	1.79	U			P
Vanadium	mg/Kg	20	5.40		6.06		12		P
Zinc	mg/Kg	20	50.0		53.6		7		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Portal Partners Tri-Venture **Level:** LOW **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718
Matrix: Solid **Sample ID:** P4699-01MS **Client ID:** MIXED-DEMOMSD
Percent Solids for Sample: 99.5 **Duplicate ID** P4699-01MSD **Percent Solids for Spike Sample:** 99.5

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	mg/Kg	20	4160		4550		9		P
Antimony	mg/Kg	20	35.5		37.5		5		P
Arsenic	mg/Kg	20	37.3		39.2		5		P
Barium	mg/Kg	20	26.5		29.0		9		P
Beryllium	mg/Kg	20	8.92		8.59		4		P
Cadmium	mg/Kg	20	8.39		8.98		7		P
Calcium	mg/Kg	20	67400		74000		9		P
Chromium	mg/Kg	20	30.8		33.1		7		P
Cobalt	mg/Kg	20	9.10		9.76		7		P
Copper	mg/Kg	20	16.2		17.3		7		P
Iron	mg/Kg	20	3270		3530		8		P
Lead	mg/Kg	20	42.1		45.2		7		P
Magnesium	mg/Kg	20	13700		12800		7		P
Manganese	mg/Kg	20	152		151		1		P
Nickel	mg/Kg	20	23.1		24.8		7		P
Potassium	mg/Kg	20	1960		2080		6		P
Selenium	mg/Kg	20	86.0		90.0		5		P
Silver	mg/Kg	20	3.61		3.84		6		P
Sodium	mg/Kg	20	538		568		5		P
Thallium	mg/Kg	20	80.8		85.5		6		P
Vanadium	mg/Kg	20	16.8		18.3		9		P
Zinc	mg/Kg	20	58.9		63.5		8		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Portal Partners Tri-Venture **Level:** LOW **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718
Matrix: Solid **Sample ID:** P4706-01 **Client ID:** TR-04-110424DUP
Percent Solids for Sample: 93.5 **Duplicate ID** P4706-01DUP **Percent Solids for Spike Sample:** 93.5

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	mg/Kg	20	0.060		0.063		5		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Portal Partners Tri-Venture</u>	Level:	<u>LOW</u>	SDG No.:	<u>P4718</u>
Contract:	<u>PORT06</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>P4718</u>
Matrix:	<u>Solid</u>	Sample ID:	<u>P4706-01MS</u>	Client ID:	<u>TR-04-110424MSD</u>
Percent Solids for Sample:	93.5	Duplicate ID	P4706-01MSD	Percent Solids for Spike Sample:	93.5

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	mg/Kg	20	0.33		0.33		0		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Portal Partners Tri-Venture **Level:** LOW **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718
Matrix: Water **Sample ID:** P4718-04 **Client ID:** WB-307-SWDUP
Percent Solids for Sample: NA **Duplicate ID** P4718-04DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	389		395		2		P
Antimony	ug/L	20	25.0	U	3.27	J	200.0		P
Arsenic	ug/L	20	10.0	U	10.0	U			P
Barium	ug/L	20	34.4	J	33.6	J	2		P
Beryllium	ug/L	20	3.00	U	3.00	U			P
Cadmium	ug/L	20	0.19	J	0.22	J	17		P
Calcium	ug/L	20	206000		201000		2		P
Chromium	ug/L	20	1.78	J	1.15	J	43		P
Cobalt	ug/L	20	15.0	U	15.0	U			P
Copper	ug/L	20	10.0	U	10.0	U			P
Iron	ug/L	20	521		511		2		P
Lead	ug/L	20	6.00	U	6.00	U			P
Magnesium	ug/L	20	628000		615000		2		P
Manganese	ug/L	20	69.7		67.9		3		P
Nickel	ug/L	20	20.0	U	20.0	U			P
Potassium	ug/L	20	425000		422000		1		P
Selenium	ug/L	20	10.0	U	10.0	U			P
Silver	ug/L	20	5.00	U	5.00	U			P
Sodium	ug/L	20	5110000	D	5150000	D	1		P
Thallium	ug/L	20	20.0	U	20.0	U			P
Vanadium	ug/L	20	3.65	J	3.83	J	5		P
Zinc	ug/L	20	19.8	J	19.6	J	1		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Portal Partners Tri-Venture **Level:** LOW **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718
Matrix: Water **Sample ID:** P4718-04MS **Client ID:** WB-307-SWMSD
Percent Solids for Sample: NA **Duplicate ID** P4718-04MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	1670		1700		2		P
Antimony	ug/L	20	562		591		5		P
Arsenic	ug/L	20	549		570		4		P
Barium	ug/L	20	134		135		1		P
Beryllium	ug/L	20	115		114		1		P
Cadmium	ug/L	20	87.4		87.0		0		P
Calcium	ug/L	20	210000		212000		1		P
Chromium	ug/L	20	157		156		1		P
Cobalt	ug/L	20	92.0		91.8		0		P
Copper	ug/L	20	125		127		2		P
Iron	ug/L	20	2060		2120		3		P
Lead	ug/L	20	436		434		0		P
Magnesium	ug/L	20	642000		654000		2		P
Manganese	ug/L	20	154		156		1		P
Nickel	ug/L	20	221		220		0		P
Potassium	ug/L	20	394000		413000		5		P
Selenium	ug/L	20	1370		1430		4		P
Silver	ug/L	20	65.9		67.9		3		P
Sodium	ug/L	20	4920000	D	5190000	D	5		P
Thallium	ug/L	20	814		810		0		P
Vanadium	ug/L	20	145		145		0		P
Zinc	ug/L	20	177		182		3		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Portal Partners Tri-Venture</u>	Level:	<u>LOW</u>	SDG No.:	<u>P4718</u>
Contract:	<u>PORT06</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>P4718</u>
Matrix:	<u>Water</u>	Sample ID:	<u>P4791-02</u>	Client ID:	<u>HINCHMAN-OILY-WATERDUP</u>
Percent Solids for Sample:	<u>NA</u>	Duplicate ID	<u>P4791-02DUP</u>	Percent Solids for Spike Sample:	<u>NA</u>

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	0.20	U	0.20	U			CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Portal Partners Tri-Venture **Level:** LOW **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718
Matrix: Water **Sample ID:** P4791-02MS **Client ID:** HINCHMAN-OILY-WATERMSD
Percent Solids for Sample: NA **Duplicate ID** P4791-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	3.06		3.12		2		CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164708BS							
Aluminum	mg/Kg	87.7	86.1		98	80 - 120	P
Antimony	mg/Kg	35.1	35.0		100	80 - 120	P
Arsenic	mg/Kg	35.1	35.2		100	80 - 120	P
Barium	mg/Kg	8.8	8.35		95	80 - 120	P
Beryllium	mg/Kg	8.8	8.38		95	80 - 120	P
Cadmium	mg/Kg	8.8	8.06		92	80 - 120	P
Calcium	mg/Kg	43.9	42.6	J	97	80 - 120	P
Chromium	mg/Kg	17.5	16.9		97	80 - 120	P
Cobalt	mg/Kg	8.8	8.20		93	80 - 120	P
Copper	mg/Kg	13.2	13.6		103	80 - 120	P
Iron	mg/Kg	130	122		94	80 - 120	P
Lead	mg/Kg	43.9	41.2		94	80 - 120	P
Magnesium	mg/Kg	87.7	80.4	J	92	80 - 120	P
Manganese	mg/Kg	8.8	8.28		94	80 - 120	P
Nickel	mg/Kg	21.9	20.5		94	80 - 120	P
Potassium	mg/Kg	440	430		98	80 - 120	P
Selenium	mg/Kg	87.7	88.9		101	80 - 120	P
Silver	mg/Kg	3.3	3.20		97	80 - 120	P
Sodium	mg/Kg	130	108		83	80 - 120	P
Thallium	mg/Kg	87.7	94.3		108	80 - 120	P
Vanadium	mg/Kg	13.2	12.7		96	80 - 120	P
Zinc	mg/Kg	8.8	8.80		100	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164759BS							
Aluminum	ug/L	1000	1000		100	80 - 120	P
Antimony	ug/L	400	388		97	80 - 120	P
Arsenic	ug/L	400	387		97	80 - 120	P
Barium	ug/L	100	103		103	80 - 120	P
Beryllium	ug/L	100	103		103	80 - 120	P
Cadmium	ug/L	100	97.8		98	80 - 120	P
Calcium	ug/L	500	521	J	104	80 - 120	P
Chromium	ug/L	200	203		102	80 - 120	P
Cobalt	ug/L	100	99.2		99	80 - 120	P
Copper	ug/L	150	155		103	80 - 120	P
Iron	ug/L	1500	1470		98	80 - 120	P
Lead	ug/L	500	489		98	80 - 120	P
Magnesium	ug/L	1000	1000		100	80 - 120	P
Manganese	ug/L	100	105		105	80 - 120	P
Nickel	ug/L	250	248		99	80 - 120	P
Potassium	ug/L	5000	4690		94	80 - 120	P
Selenium	ug/L	1000	964		96	80 - 120	P
Silver	ug/L	37.5	36.8		98	80 - 120	P
Sodium	ug/L	1500	1360		91	80 - 120	P
Thallium	ug/L	1000	1030		103	80 - 120	P
Vanadium	ug/L	150	154		103	80 - 120	P
Zinc	ug/L	100	103		103	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: P4718

Contract: PORT06

Lab Code: CHEM

Case No.: P4718

SAS No.: P4718

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164762BS Mercury	mg/Kg	0.26	0.29		110	80 - 120	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: P4718

Contract: PORT06

Lab Code: CHEM

Case No.: P4718

SAS No.: P4718

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164872BS Mercury	ug/L	4.0	3.62		90	80 - 120	CV

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

MIXED-DEMOL

Lab Name: Chemtech Consulting Group

Contract: PORT06

Lab Code: CHEM **Lb No.:** lb133344

Lab Sample ID : P4699-01L

SDG No.: P4718

Matrix (soil/water): Solid

Level (low/med): LOW

Concentration Units: mg/Kg

Analyte	Initial Sample Result (I)			Serial Dilution Result (S)			% Differ-ence	Q	M
	C			C					
Aluminum	3900			3880			1		P
Antimony	0.18	J		10.3	U		100.0		P
Arsenic	0.81	J		4.12	U		100.0		P
Barium	17.9			17.6	J		2		P
Beryllium	0.17	J		0.19	J		10		P
Cadmium	0.46			0.33	J		29		P
Calcium	64800			71600			11		P
Chromium	16.0			17.6			10		P
Cobalt	1.10	J		1.05	J		4		P
Copper	5.06			6.04			19		P
Iron	3100			3190			3		P
Lead	4.31			4.37			1		P
Magnesium	13200			13600			3		P
Manganese	142			153			7		P
Nickel	3.50			3.50	J		0		P
Potassium	1410			1300			8		P
Selenium	0.82	U		4.12	U				P
Silver	0.21	J		0.31	J		48		P
Sodium	362			248	J		31		P
Thallium	1.65	U		8.24	U				P
Vanadium	5.40			5.87	J		9		P
Zinc	50.0			48.8			2		P

Metals

-9 -

ICP SERIAL DILUTIONS

SAMPLE NO.

TR-04-110424L

Lab Name: Chemtech Consulting Group

Contract: PORT06

Lab Code: CHEM

Lb No.: lb133334

Lab Sample ID : P4706-01L

SDG No.: P4718

Matrix (soil/water): Solid

Level (low/med): LOW

Concentration Units: mg/Kg

Analyte	Initial Sample Result (I) C	Serial Dilution Result (S) C	% Differ- ence	Q	M
Mercury	0.060	0.069 U	100.0		CV

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

WB-307-SWL

Lab Name: Chemtech Consulting Group

Contract: PORT06

Lab Code: CHEM **Lb No.:** lb133365 **Lab Sample ID :** P4718-04L **SDG No.:** P4718

Matrix (soil/water): Water **Level (low/med):** LOW

Concentration Units: ug/L

Analyte	Initial Sample Result (I)		Serial Dilution Result (S)		% Differ-ence	Q	M
		C		C			
Aluminum	389		281		28		P
Antimony	25.0	U	125	U			P
Arsenic	10.0	U	50.0	U			P
Barium	34.4	J	250	U	100.0		P
Beryllium	3.00	U	15.0	U			P
Cadmium	0.19	J	15.0	U	100.0		P
Calcium	206000		231000		12		P
Chromium	1.78	J	25.0	U	100.0		P
Cobalt	15.0	U	75.0	U			P
Copper	10.0	U	50.0	U			P
Iron	521		505		3		P
Lead	6.00	U	30.0	U			P
Magnesium	628000		732000		17		P
Manganese	69.7		75.3		8		P
Nickel	20.0	U	100	U			P
Potassium	425000		279000		34		P
Selenium	10.0	U	50.0	U			P
Silver	5.00	U	25.0	U			P
Sodium	5110000	D	5020000	D	2		P
Thallium	20.0	U	100	U			P
Vanadium	3.65	J	100	U	100.0		P
Zinc	19.8	J	15.6	J	21		P

Metals

-9 -

ICP SERIAL DILUTIONS

SAMPLE NO.

HINCHMAN-OILY-WATERL

Lab Name: Chemtech Consulting Group

Contract: PORT06

Lab Code: CHEM

Lb No.: lb133392

Lab Sample ID : P4791-02L

SDG No.: P4718

Matrix (soil/water): Water

Level (low/med): LOW

Concentration Units: ug/L

Analyte	Initial Sample Result (I) C	Serial Dilution Result (S) C	% Differ- ence	Q	M
Mercury	0.20 U	1.00 U			CV

metals
- 14 -
ANALYSIS RUN LOG

Client: Portal Partners Tri-Venture **Contract:** PORT06

Lab code: CHEM **Case no.:** P4718 **Sas no.:** P4718 **Sdg no.:** P4718

Instrument id number: **Method:** **Run number:** LB133334

Start date: 11/07/2024 **End date:** 11/07/2024

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1417	HG
S0.2	S0.2	1	1419	HG
S2.5	S2.5	1	1421	HG
S5	S5	1	1426	HG
S7.5	S7.5	1	1429	HG
S10	S10	1	1438	HG
ICV74	ICV74	1	1441	HG
ICB74	ICB74	1	1443	HG
CCV47	CCV47	1	1445	HG
CCB47	CCB47	1	1448	HG
CRA	CRA	1	1450	HG
PB164762BL	PB164762BL	1	1457	HG
PB164762BS	PB164762BS	1	1459	HG
P4706-01MS	TR-04-110424MS	1	1508	HG
P4706-01MSD	TR-04-110424MSD	1	1511	HG
CCV48	CCV48	1	1513	HG
CCB48	CCB48	1	1515	HG
P4718-01	WB-307-SB01	1	1522	HG
P4718-02	WB-307-SB02	1	1524	HG
CCV49	CCV49	1	1540	HG
CCB49	CCB49	1	1543	HG
P4706-01DUP	TR-04-110424DUP	1	1601	HG
P4706-01L	TR-04-110424L	5	1603	HG
CCV50	CCV50	1	1608	HG
CCB50	CCB50	1	1610	HG

metals
- 14 -
ANALYSIS RUN LOG

Client: Portal Partners Tri-Venture **Contract:** PORT06

Lab code: CHEM **Case no.:** P4718 **Sas no.:** P4718 **Sdg no.:** P4718

Instrument id number: **Method:** **Run number:** LB133344

Start date: 11/07/2024 **End date:** 11/08/2024

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1323	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S1	S1	1	1328	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S2	S2	1	1332	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S3	S3	1	1336	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S4	S4	1	1340	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S5	S5	1	1345	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICV01	ICV01	1	1349	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
LLICV01	LLICV01	1	1423	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICB01	ICB01	1	1438	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CRI01	CRI01	1	1442	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSA01	ICSA01	1	1447	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSAB01	ICSAB01	1	1451	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV01	CCV01	1	1501	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB01	CCB01	1	1505	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV02	CCV02	1	1553	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB02	CCB02	1	1558	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV03	CCV03	1	1645	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB03	CCB03	1	1649	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV04	CCV04	1	1756	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB04	CCB04	1	1804	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV05	CCV05	1	1832	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB05	CCB05	1	1837	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV06	CCV06	1	1908	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB06	CCB06	1	1912	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV07	CCV07	1	1958	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB07	CCB07	1	2002	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV08	CCV08	1	2058	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB08	CCB08	1	2102	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV09	CCV09	1	2151	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB09	CCB09	1	2155	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
P4699-01DUP	MIXED-DEMODOUP	1	2239	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
P4699-01L	MIXED-DEMOL	5	2243	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
P4699-01MS	MIXED-DEMOMS	1	2247	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV10	CCV10	1	2252	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB10	CCB10	1	2256	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
P4699-01MSD	MIXED-DEMOMSD	1	2300	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
P4699-01A	MIXED-DEMOA	1	2304	K,Na
P4718-01	WB-307-SB01	1	2326	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
P4718-02	WB-307-SB02	1	2330	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
PB164708BL	PB164708BL	1	2338	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV11	CCV11	1	2343	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn

metals
- 14 -
ANALYSIS RUN LOG

Client: Portal Partners Tri-Venture **Contract:** PORT06

Lab code: CHEM **Case no.:** P4718 **Sas no.:** P4718 **Sdg no.:** P4718

Instrument id number: **Method:** **Run number:** LB133344

Start date: 11/07/2024 **End date:** 11/08/2024

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
CCB11	CCB11	1	2347	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
PB164708BS	PB164708BS	1	2351	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV12	CCV12	1	0039	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB12	CCB12	1	0043	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV13	CCV13	1	0057	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB13	CCB13	1	0101	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn

metals
- 14 -
ANALYSIS RUN LOG

Client: Portal Partners Tri-Venture **Contract:** PORT06

Lab code: CHEM **Case no.:** P4718 **Sas no.:** P4718 **Sdg no.:** P4718

Instrument id number: **Method:** **Run number:** LB133365

Start date: 11/08/2024 **End date:** 11/08/2024

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1334	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S1	S1	1	1338	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S2	S2	1	1343	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S3	S3	1	1347	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S4	S4	1	1351	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S5	S5	1	1355	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICV01	ICV01	1	1359	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
LLICV01	LLICV01	1	1411	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICB01	ICB01	1	1416	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CRI01	CRI01	1	1420	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSA01	ICSA01	1	1424	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSAB01	ICSAB01	1	1430	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV01	CCV01	1	1444	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB01	CCB01	1	1450	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV02	CCV02	1	1529	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB02	CCB02	1	1533	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV03	CCV03	1	1623	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB03	CCB03	1	1628	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV04	CCV04	1	1716	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB04	CCB04	1	1722	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV05	CCV05	1	1804	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB05	CCB05	1	1808	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV06	CCV06	1	1855	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB06	CCB06	1	1859	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
P4718-04	WB-307-SW	1	1942	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV07	CCV07	1	1947	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB07	CCB07	1	1951	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
P4718-04DUP	WB-307-SWDUP	1	1955	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
P4718-04L	WB-307-SWL	5	2000	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
P4718-04MS	WB-307-SWMS	1	2004	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
P4718-04MSD	WB-307-SWMSD	1	2009	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
P4718-04A	WB-307-SWA	1	2014	Ag,Al,As,Sb,Se,Zn
CCV08	CCV08	1	2048	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB08	CCB08	1	2057	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
P4718-04	WB-307-SW	10	2118	Na
P4718-04DUP	WB-307-SWDUP	10	2123	Na
P4718-04L	WB-307-SWL	50	2127	Na
P4718-04MS	WB-307-SWMS	10	2131	Na
P4718-04MSD	WB-307-SWMSD	10	2136	Na
CCV09	CCV09	1	2150	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB09	CCB09	1	2155	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn

metals
- 14 -
ANALYSIS RUN LOG

Client: Portal Partners Tri-Venture **Contract:** PORT06
Lab code: CHEM **Case no.:** P4718 **Sas no.:** P4718 **Sdg no.:** P4718
Instrument id number: **Method:** **Run number:** LB133365
Start date: 11/08/2024 **End date:** 11/08/2024

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
CCV10	CCV10	1	2324	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB10	CCB10	1	2335	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn

metals
- 14 -
ANALYSIS RUN LOG

Client: Portal Partners Tri-Venture

Contract: PORT06

Lab code: CHEM **Case no.:** P4718 **Sas no.:** P4718

Sdg no.: P4718

Instrument id number: **Method:**

Run number: LB133392

Start date: 11/11/2024 **End date:** 11/11/2024

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1336	HG
S0.2	S0.2	1	1341	HG
S2.5	S2.5	1	1344	HG
S5	S5	1	1346	HG
S7.5	S7.5	1	1353	HG
S10	S10	1	1355	HG
ICV80	ICV80	1	1358	HG
ICB80	ICB80	1	1401	HG
CCV64	CCV64	1	1403	HG
CCB64	CCB64	1	1405	HG
CRA	CRA	1	1410	HG
PB164872BL	PB164872BL	1	1420	HG
PB164872BS	PB164872BS	1	1422	HG
P4718-04	WB-307-SW	1	1424	HG
P4791-02DUP	HINCHMAN-OILY-WATERDU	1	1431	HG
P4791-02MS	HINCHMAN-OILY-WATERMS	1	1433	HG
CCV65	CCV65	1	1436	HG
CCB65	CCB65	1	1441	HG
P4791-02MSD	HINCHMAN-OILY-WATERMS	1	1443	HG
P4791-02L	HINCHMAN-OILY-WATERL	5	1450	HG
CCV66	CCV66	1	1454	HG
CCB66	CCB66	1	1457	HG

metals
- 14 -
ANALYSIS RUN LOG

Client: Portal Partners Tri-Venture **Contract:** PORT06

Lab code: CHEM **Case no.:** P4718 **Sas no.:** P4718 **Sdg no.:** P4718

Instrument id number: **Method:** **Run number:** LB133524

Start date: 11/19/2024 **End date:** 11/19/2024

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1055	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S1	S1	1	1100	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S2	S2	1	1104	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S3	S3	1	1108	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S4	S4	1	1113	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S5	S5	1	1117	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICV01	ICV01	1	1125	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
LLICV01	LLICV01	1	1134	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICB01	ICB01	1	1138	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CRI01	CRI01	1	1143	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSA01	ICSA01	1	1147	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSAB01	ICSAB01	1	1151	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV01	CCV01	1	1243	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB01	CCB01	1	1248	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
PB164759BL	PB164759BL	1	1300	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
PB164759BS	PB164759BS	1	1305	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV02	CCV02	1	1334	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB02	CCB02	1	1338	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV03	CCV03	1	1424	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB03	CCB03	1	1428	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV04	CCV04	1	1531	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB04	CCB04	1	1536	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV05	CCV05	1	1618	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB05	CCB05	1	1622	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV06	CCV06	1	1719	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB06	CCB06	1	1724	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV07	CCV07	1	1801	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB07	CCB07	1	1805	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn



METAL PREPARATION & INSTRUMENT DATA

Metals

- 11 -

ICP INTERELEMEN CORRECTION FACTORS

Client: Portal Partners Tri-Venture

SDG No.: P4718

Contract: PORT06

Lab Code: CHEM

Case No.: P4718

SAS No.: P4718

Instrument ID:

Date:

Interement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		Al	Ca	Fe	Mg	Ag
Aluminum	396.100	0.0000000	-0.0002060	0.0000000	0.0000000	0.0000000
Antimony	206.833	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	0.0000000	0.0000000	-0.0000440	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000930	0.0000000	0.0000000
Calcium	373.690	0.0000000	0.0000000	-0.0075970	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0000000	0.0007850	0.0000000	0.0000000
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Lead	220.353	-0.0000920	0.0000000	0.0000380	0.0000000	0.0000000
Magnesium	279.079	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	-0.0001440	0.0000000	0.0000000
Silver	328.068	0.0000000	0.0000000	-0.0001490	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Vanadium	292.402	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0001050	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Portal Partners Tri-Venture

SDG No.: P4718

Contract: PORT06

Lab Code: CHEM

Case No.: P4718

SAS No.: P4718

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		As	Ba	Be	Cd	Co
Aluminum	396.100	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Antimony	206.833	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0002870
Calcium	373.690	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0000000	0.0000000	0.0000000	0.0009530
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	-0.0039600
Lead	220.353	0.0000000	0.0003170	0.0000000	0.0000000	0.0000000
Magnesium	279.079	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0000000	-0.0003570
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0000000	0.0054900
Vanadium	292.402	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Portal Partners Tri-Venture

SDG No.: P4718

Contract: PORT06

Lab Code: CHEM

Case No.: P4718

SAS No.: P4718

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Cr	Cu	K	Mn	Mo
Aluminum	396.100	0.0000000	0.0000000	0.0000590	0.0000000	0.0396900
Antimony	206.833	0.0122000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	-0.0029000	0.0000000	0.0000000	0.0000000	0.0004900
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	-0.0000710	-0.0003400
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Calcium	373.690	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000070	0.0002200	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	-0.0007860
Copper	224.700	0.0000000	0.0000000	0.0000000	0.0006510	0.0020500
Iron	240.488	0.0000000	0.0000000	0.0000730	0.0000000	-0.0015250
Lead	220.353	0.0000000	0.0000000	0.0000000	0.0001400	-0.0008600
Magnesium	279.079	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0007460	0.0000000
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	-0.0000120
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0017400	-0.0100400
Vanadium	292.402	-0.0025100	0.0000000	0.0000000	0.0000000	-0.0072000
Zinc	213.800	0.0000000	0.0009010	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Portal Partners Tri-Venture

SDG No.: P4718

Contract: PORT06

Lab Code: CHEM

Case No.: P4718

SAS No.: P4718

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Na	Ni	Pb	Sb	Se
Aluminum	396.100	0.0000000	0.0000000	0.0012800	0.0000000	0.0000000
Antimony	206.833	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Calcium	373.690	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	-0.0047000	0.0036100	0.0000000	0.0000000
Iron	240.488	0.0000000	-0.0017000	0.0000000	0.0000000	0.0000000
Lead	220.353	0.0000000	0.0006580	0.0000000	0.0000000	0.0001290
Magnesium	279.079	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0003330	0.0000000	0.0000000
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Vanadium	292.402	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0067600	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Portal Partners Tri-Venture

SDG No.: P4718

Contract: PORT06

Lab Code: CHEM

Case No.: P4718

SAS No.: P4718

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Sn	Ti	Tl	V	Zn
Aluminum	396.100	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Antimony	206.833	-0.0035600	-0.0007970	0.0000000	-0.0018900	0.0000000
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000630	0.0001280	0.0000000	0.0000000
Calcium	373.690	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0001110	0.0000000
Cobalt	228.616	0.0000000	0.0018800	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0003840	0.0000000	0.0000000	0.0000000
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Lead	220.353	0.0000000	-0.0003610	0.0000000	0.0000000	0.0000000
Magnesium	279.079	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Silver	328.068	0.0000000	-0.0007420	0.0000000	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	-0.0039700	0.0000000	-0.0115600	0.0000000
Vanadium	292.402	0.0000000	0.0005320	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

LAB CHRONICLE

OrderID:	P4718	OrderDate:	11/5/2024 1:31:00 PM
Client:	Portal Partners Tri-Venture	Project:	Amtrak Sawtooth Bridges 2024
Contact:	Joseph Krupansky	Location:	L21,VOA Ref. #2 Soil,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
P4718-01	WB-307-SB01	SOIL			11/04/24			11/05/24
			Mercury	7471B		11/07/24	11/07/24	
			Metals ICP-TAL	6010D		11/06/24	11/07/24	
P4718-02	WB-307-SB02	SOIL			11/04/24			11/05/24
			Mercury	7471B		11/07/24	11/07/24	
			Metals ICP-TAL	6010D		11/06/24	11/07/24	
P4718-03	WB-307-SB02	TCLP			11/04/24			11/05/24
			TCLP ICP Metals	6010D		11/07/24	11/19/24	
			TCLP Mercury	7470A		11/07/24	11/08/24	
P4718-04	WB-307-SW	Water			11/04/24			11/05/24
			Mercury	7470A		11/11/24	11/11/24	
			Metals ICP-TAL	6010D		11/07/24	11/08/24	



METAL PREPARATION & ANALYICAL SUMMARY

Metals

- 13 -

SAMPLE PREPARATION SUMMARY

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Method:**
Case No.: P4718 **SAS No.:** P4718

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(g)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB164708							
P4699-01DUP	MIXED-DEMODUP	DUP	SOLID	11/06/2024	2.24	100.0	99.50
P4699-01MS	MIXED-DEMOMS	MS	SOLID	11/06/2024	2.35	100.0	99.50
P4699-01MSD	MIXED-DEMOMSD	MSD	SOLID	11/06/2024	2.19	100.0	99.50
P4718-01	WB-307-SB01	SAM	SOLID	11/06/2024	2.30	100.0	78.00
P4718-02	WB-307-SB02	SAM	SOLID	11/06/2024	2.44	100.0	77.60
PB164708BL	PB164708BL	MB	SOLID	11/06/2024	2.28	100.0	100.00
PB164708BS	PB164708BS	LCS	SOLID	11/06/2024	2.28	100.0	100.00

Metals

- 13 -

SAMPLE PREPARATION SUMMARY

Client:	<u>Portal Partners Tri-Venture</u>	SDG No.:	<u>P4718</u>
Contract:	<u>PORT06</u>	Lab Code:	<u>CHEM</u>
		Method:	<u></u>
		Case No.:	<u>P4718</u>
		SAS No.:	<u>P4718</u>

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(mL)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB164759							
P4718-04	WB-307-SW	SAM	WATER	11/07/2024	50.0	25.0	
P4718-04DUP	WB-307-SWDUP	DUP	WATER	11/07/2024	50.0	25.0	
P4718-04MS	WB-307-SWMS	MS	WATER	11/07/2024	50.0	25.0	
P4718-04MSD	WB-307-SWMSD	MSD	WATER	11/07/2024	50.0	25.0	
PB164759BL	PB164759BL	MB	WATER	11/07/2024	50.0	25.0	
PB164759BS	PB164759BS	LCS	WATER	11/07/2024	50.0	25.0	

Metals

- 13 -

SAMPLE PREPARATION SUMMARY

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Method:**
Case No.: P4718 **SAS No.:** P4718

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(g)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB164762							
P4706-01DUP	TR-04-110424DUP	DUP	SOLID	11/07/2024	0.58	35.0	93.50
P4706-01MS	TR-04-110424MS	MS	SOLID	11/07/2024	0.55	35.0	93.50
P4706-01MSD	TR-04-110424MSD	MSD	SOLID	11/07/2024	0.53	35.0	93.50
P4718-01	WB-307-SB01	SAM	SOLID	11/07/2024	0.58	35.0	78.00
P4718-02	WB-307-SB02	SAM	SOLID	11/07/2024	0.55	35.0	77.60
PB164762BL	PB164762BL	MB	SOLID	11/07/2024	0.52	35.0	100.00
PB164762BS	PB164762BS	LCS	SOLID	11/07/2024	0.54	35.0	100.00

Metals

- 13 -

SAMPLE PREPARATION SUMMARY

Client:	<u>Portal Partners Tri-Venture</u>	SDG No.:	<u>P4718</u>
Contract:	<u>PORT06</u>	Lab Code:	<u>CHEM</u>
		Method:	<u></u>
		Case No.:	<u>P4718</u>
		SAS No.:	<u>P4718</u>

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(mL)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB164872							
P4718-04	WB-307-SW	SAM	WATER	11/11/2024	30.0	30.0	
P4791-02DUP	HINCHMAN-OILY-WATERDUP	DUP	WATER	11/11/2024	30.0	30.0	
P4791-02MS	HINCHMAN-OILY-WATERMS	MS	WATER	11/11/2024	30.0	30.0	
P4791-02MSD	HINCHMAN-OILY-WATERMSD	MSD	WATER	11/11/2024	30.0	30.0	
PB164872BL	PB164872BL	MB	WATER	11/11/2024	30.0	30.0	
PB164872BS	PB164872BS	LCS	WATER	11/11/2024	30.0	30.0	

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB133334

Review By	jaswal	Review On	11/7/2024 10:00:18 PM
Supervise By	mohan	Supervise On	11/7/2024 10:01:48 PM

STD. NAME	STD REF.#
ICAL Standard	MP83107,MP83108,MP83109,MP83110,MP83111,MP83112
ICV Standard	MP83113
CCV Standard	MP83115
ICSA Standard	
CRI Standard	MP83117
LCS Standard	
Chk Standard	MP83114,MP83116,MP83118,MP83120

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	11/07/24 14:17		Mohan	OK
2	S0.2	S0.2	CAL2	11/07/24 14:19		Mohan	OK
3	S2.5	S2.5	CAL3	11/07/24 14:21		Mohan	OK
4	S5	S5	CAL4	11/07/24 14:26		Mohan	OK
5	S7.5	S7.5	CAL5	11/07/24 14:29		Mohan	OK
6	S10	S10	CAL6	11/07/24 14:38		Mohan	OK
7	ICV74	ICV74	ICV	11/07/24 14:41		Mohan	OK
8	ICB74	ICB74	ICB	11/07/24 14:43		Mohan	OK
9	CCV47	CCV47	CCV	11/07/24 14:45		Mohan	OK
10	CCB47	CCB47	CCB	11/07/24 14:48		Mohan	OK
11	CRA	CRA	CRDL	11/07/24 14:50		Mohan	OK
12	HighStd	HighStd	HIGH STD	11/07/24 14:52		Mohan	OK
13	ChkStd	ChkStd	SAM	11/07/24 14:54		Mohan	OK
14	PB164762BL	PB164762BL	MB	11/07/24 14:57		Mohan	OK
15	PB164762BS	PB164762BS	LCS	11/07/24 14:59		Mohan	OK
16	P4662-06	102524-D	SAM	11/07/24 15:01		Mohan	OK
17	P4706-01	TR-04-110424	SAM	11/07/24 15:04		Mohan	OK
18	P4706-01MS	TR-04-110424MS	MS	11/07/24 15:08		Mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB133334

Review By	jaswal	Review On	11/7/2024 10:00:18 PM
Supervise By	mohan	Supervise On	11/7/2024 10:01:48 PM

STD. NAME	STD REF.#
ICAL Standard	MP83107,MP83108,MP83109,MP83110,MP83111,MP83112
ICV Standard	MP83113
CCV Standard	MP83115
ICSA Standard	
CRI Standard	MP83117
LCS Standard	
Chk Standard	MP83114,MP83116,MP83118,MP83120

19	P4706-01MSD	TR-04-110424MSD	MSD	11/07/24 15:11		Mohan	OK
20	CCV48	CCV48	CCV	11/07/24 15:13		Mohan	OK
21	CCB48	CCB48	CCB	11/07/24 15:15		Mohan	OK
22	P4707-01	HD-02-110424	SAM	11/07/24 15:18		Mohan	OK
23	P4708-01	OR-02-110424	SAM	11/07/24 15:20		Mohan	OK
24	P4718-01	WB-307-SB01	SAM	11/07/24 15:22		Mohan	OK
25	P4718-02	WB-307-SB02	SAM	11/07/24 15:24		Mohan	OK
26	P4719-01	BAYAVE-STOCKPILE	SAM	11/07/24 15:27		Mohan	OK
27	P4720-01	JC-701-COMP-01	SAM	11/07/24 15:29		Mohan	OK
28	P4722-03	WC-1(0-6)	SAM	11/07/24 15:31		Mohan	OK
29	P4722-08	WC-2(0-6)	SAM	11/07/24 15:34		Mohan	OK
30	P4722-13	WC-3(0-6)	SAM	11/07/24 15:36		Mohan	OK
31	P4732-01	PPE-COMP	SAM	11/07/24 15:38		Mohan	OK
32	CCV49	CCV49	CCV	11/07/24 15:40		Mohan	OK
33	CCB49	CCB49	CCB	11/07/24 15:43		Mohan	OK
34	P4734-01	EO-03-11062024	SAM	11/07/24 15:45		Mohan	OK
35	P4737-01	TP2	SAM	11/07/24 15:47		Mohan	OK
36	P4738-01	72-12018	SAM	11/07/24 15:49		Mohan	OK
37	P4739-01	TP-14	SAM	11/07/24 15:52		Mohan	OK
38	P4739-05	BP-G2	SAM	11/07/24 15:54		Mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB133334

Review By	jaswal	Review On	11/7/2024 10:00:18 PM
Supervise By	mohan	Supervise On	11/7/2024 10:01:48 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83107,MP83108,MP83109,MP83110,MP83111,MP83112		
ICV Standard	MP83113		
CCV Standard	MP83115		
ICSA Standard			
CRI Standard	MP83117		
LCS Standard			
Chk Standard	MP83114,MP83116,MP83118,MP83120		

39	P4739-09	BP-B2	SAM	11/07/24 15:56		Mohan	OK
40	P4739-13	TP-11	SAM	11/07/24 15:59		Mohan	OK
41	P4706-01DUP	TR-04-110424DUP	DUP	11/07/24 16:01		Mohan	OK
42	P4706-01L	TR-04-110424L	SD	11/07/24 16:03		Mohan	OK
43	P4706-01A	TR-04-110424A	PS	11/07/24 16:05		Mohan	OK
44	CCV50	CCV50	CCV	11/07/24 16:08		Mohan	OK
45	CCB50	CCB50	CCB	11/07/24 16:10		Mohan	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133344

Review By	mohan	Review On	11/9/2024 12:07:42 AM
Supervise By	kareem	Supervise On	11/13/2024 7:36:39 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83078,MP83079,MP83080,MP83081,MP83082,MP83084		
ICV Standard	MP83085		
CCV Standard	MP83088		
ICSA Standard	MP83086,MP83087		
CRI Standard	MP83084		
LCS Standard			
Chk Standard	MP83091 MP83092		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	11/07/24 13:23		Kareem	OK
2	S1	S1	CAL2	11/07/24 13:28		Kareem	OK
3	S2	S2	CAL3	11/07/24 13:32		Kareem	OK
4	S3	S3	CAL4	11/07/24 13:36		Kareem	OK
5	S4	S4	CAL5	11/07/24 13:40		Kareem	OK
6	S5	S5	CAL6	11/07/24 13:45		Kareem	OK
7	ICV01	ICV01	ICV	11/07/24 13:49		Kareem	OK
8	LLICV01	LLICV01	LLICV	11/07/24 14:23		Kareem	OK
9	ICB01	ICB01	ICB	11/07/24 14:38		Kareem	OK
10	CRI01	CRI01	CRDL	11/07/24 14:42		Kareem	OK
11	ICSA01	ICSA01	ICSA	11/07/24 14:47		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	11/07/24 14:51		Kareem	OK
13	CCV01	CCV01	CCV	11/07/24 15:01		Kareem	OK
14	CCB01	CCB01	CCB	11/07/24 15:05		Kareem	OK
15	P4645-04	Z-02-WC	SAM	11/07/24 15:09		Kareem	OK
16	P4659-04	MH-2	SAM	11/07/24 15:14		Kareem	OK
17	P4660-03	WC-TA2-01-C	SAM	11/07/24 15:18		Kareem	OK
18	P4660-07	WC-WOOD-01-C	SAM	11/07/24 15:22		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133344

Review By	mohan	Review On	11/9/2024 12:07:42 AM
Supervise By	kareem	Supervise On	11/13/2024 7:36:39 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83078,MP83079,MP83080,MP83081,MP83082,MP83084		
ICV Standard	MP83085		
CCV Standard	MP83088		
ICSA Standard	MP83086,MP83087		
CRI Standard	MP83084		
LCS Standard			
Chk Standard	MP83091 MP83092		

19	P4660-11	WC-CONCRETE-01-C	SAM	11/07/24 15:27		Kareem	OK
20	P4667-04	BP-F-6	SAM	11/07/24 15:31		Kareem	OK
21	P4667-08	BP-F-5	SAM	11/07/24 15:36		Kareem	OK
22	P4667-12	TP-10	SAM	11/07/24 15:40		Kareem	OK
23	P4667-16	BP-F-7	SAM	11/07/24 15:45		Kareem	OK
24	P4679-04	MH-1	SAM	11/07/24 15:49		Kareem	OK
25	CCV02	CCV02	CCV	11/07/24 15:53		Kareem	OK
26	CCB02	CCB02	CCB	11/07/24 15:58		Kareem	OK
27	P4680-04	BP-F26	SAM	11/07/24 16:02		Kareem	OK
28	P4680-08	BP-F25	SAM	11/07/24 16:06		Kareem	OK
29	P4684-01	MECHANIC-ST-SWEI	SAM	11/07/24 16:11		Kareem	OK
30	P4684-01DUP	MECHANIC-ST-SWEI	DUP	11/07/24 16:15		Kareem	OK
31	P4684-01L	MECHANIC-ST-SWEI	SD	11/07/24 16:20		Kareem	OK
32	P4684-01MS	MECHANIC-ST-SWEI	MS	11/07/24 16:24	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
33	P4684-01MSD	MECHANIC-ST-SWEI	MSD	11/07/24 16:28	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
34	P4684-01A	MECHANIC-ST-SWEI	PS	11/07/24 16:32	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
35	PB164560TB	PB164560TB	MB	11/07/24 16:36		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133344

Review By	mohan	Review On	11/9/2024 12:07:42 AM
Supervise By	kareem	Supervise On	11/13/2024 7:36:39 PM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP83080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091 MP83092

36	PB164665BL	PB164665BL	MB	11/07/24 16:41		Kareem	OK
37	CCV03	CCV03	CCV	11/07/24 16:45		Kareem	OK
38	CCB03	CCB03	CCB	11/07/24 16:49		Kareem	OK
39	PB164665BS	PB164665BS	LCS	11/07/24 16:54	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
40	P4693-04	BP-G5-WC	SAM	11/07/24 16:58		Kareem	OK
41	P4693-08	BP-G4-WC	SAM	11/07/24 17:02		Kareem	OK
42	P4694-04	Z-03A	SAM	11/07/24 17:06		Kareem	OK
43	P4694-08	Z-04	SAM	11/07/24 17:11		Kareem	OK
44	P4695-04	Z-01	SAM	11/07/24 17:15		Kareem	OK
45	P4700-04	MH-8	SAM	11/07/24 17:20		Kareem	OK
46	P4701-01	BP-F3	SAM	11/07/24 17:24		Kareem	OK
47	P4701-08	BP-F4	SAM	11/07/24 17:29		Kareem	OK
48	P4711-05	CF-613-COMP-16	SAM	11/07/24 17:33		Kareem	OK
49	CCV04	CCV04	CCV	11/07/24 17:56		Kareem	OK
50	CCB04	CCB04	CCB	11/07/24 18:04		Kareem	OK
51	P4711-10	CF-613-COMP-17	SAM	11/07/24 18:08		Kareem	OK
52	P4702-01	TOTE-1	SAM	11/07/24 18:13		Kareem	OK
53	P4662-06DL2	102524-DDL2	SAM	11/07/24 18:17	NOT USE	Kareem	Not Ok
54	P4662-06DL	102524-DDL	SAM	11/07/24 18:22	NOT USE	Kareem	Not Ok

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133344

Review By	mohan	Review On	11/9/2024 12:07:42 AM
Supervise By	kareem	Supervise On	11/13/2024 7:36:39 PM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP83080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091 MP83092

55	P4662-06	102524-D	SAM	11/07/24 18:28		Kareem	OK
56	CCV05	CCV05	CCV	11/07/24 18:32		Kareem	OK
57	CCB05	CCB05	CCB	11/07/24 18:37		Kareem	OK
58	P4702-01DUP	TOTE-1DUP	DUP	11/07/24 18:43		Kareem	OK
59	P4702-01L	TOTE-1L	SD	11/07/24 18:47		Kareem	OK
60	P4706-01	TR-04-110424	SAM	11/07/24 18:51		Kareem	OK
61	P4708-01	OR-02-110424	SAM	11/07/24 18:56		Kareem	OK
62	P4711-01	CF-613-COMP-16	SAM	11/07/24 19:00		Kareem	OK
63	P4711-06	CF-613-COMP-17	SAM	11/07/24 19:04		Kareem	OK
64	CCV06	CCV06	CCV	11/07/24 19:08		Kareem	OK
65	CCB06	CCB06	CCB	11/07/24 19:12		Kareem	OK
66	P4702-01MS	TOTE-1MS	MS	11/07/24 19:16	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
67	P4702-01MSD	TOTE-1MSD	MSD	11/07/24 19:20	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
68	P4702-01A	TOTE-1A	PS	11/07/24 19:24	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
69	PB164662TB	PB164662TB	MB	11/07/24 19:28		Kareem	OK
70	PB164685TB	PB164685TB	MB	11/07/24 19:32		Kareem	OK
71	PB164712BL	PB164712BL	MB	11/07/24 19:37		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133344

Review By	mohan	Review On	11/9/2024 12:07:42 AM
Supervise By	kareem	Supervise On	11/13/2024 7:36:39 PM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP83080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091 MP83092

72	PB164712BS	PB164712BS	LCS	11/07/24 19:41	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
73	P4720-01	JC-701-COMP-01	SAM	11/07/24 19:45		Kareem	OK
74	P4720-01DUP	JC-701-COMP-01DUP	DUP	11/07/24 19:50		Kareem	OK
75	P4720-01L	JC-701-COMP-01L	SD	11/07/24 19:54		Kareem	OK
76	CCV07	CCV07	CCV	11/07/24 19:58		Kareem	OK
77	CCB07	CCB07	CCB	11/07/24 20:02		Kareem	OK
78	P4617-04	CONCRETE-PILE	SAM	11/07/24 20:08		Kareem	OK
79	P4617-04DUP	CONCRETE-PILEDUP	DUP	11/07/24 20:12		Kareem	OK
80	P4617-04L	CONCRETE-PILEL	SD	11/07/24 20:17		Kareem	OK
81	P4617-04MS	CONCRETE-PILEMS	MS	11/07/24 20:21	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
82	P4617-04MSD	CONCRETE-PILEMS	MSD	11/07/24 20:25	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
83	P4549-04	TT-069-IDWGW-2024	SAM	11/07/24 20:34		Kareem	OK
84	P4617-04A	CONCRETE-PILEA	PS	11/07/24 20:41	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
85	P4549-04DUP	TT-069-IDWGW-2024	DUP	11/07/24 20:45		Kareem	OK
86	P4549-04L	TT-069-IDWGW-2024	SD	11/07/24 20:49		Kareem	OK
87	P4549-04MS	TT-069-IDWGW-2024	MS	11/07/24 20:54	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133344

Review By	mohan	Review On	11/9/2024 12:07:42 AM
Supervise By	kareem	Supervise On	11/13/2024 7:36:39 PM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP83080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091 MP83092

88	CCV08	CCV08	CCV	11/07/24 20:58		Kareem	OK
89	CCB08	CCB08	CCB	11/07/24 21:02		Kareem	OK
90	P4549-04MSD	TT-069-IDWGW-2024	MSD	11/07/24 21:08	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
91	P4549-04A	TT-069-IDWGW-2024	PS	11/07/24 21:12	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
92	P4720-01MS	JC-701-COMP-01MS	MS	11/07/24 21:16	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
93	P4720-01MSD	JC-701-COMP-01MS	MSD	11/07/24 21:20	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
94	P4720-01A	JC-701-COMP-01A	PS	11/07/24 21:24	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
95	P4722-03	WC-1(0-6)	SAM	11/07/24 21:29	Fe high	Kareem	Dilution
96	P4722-08	WC-2(0-6)	SAM	11/07/24 21:33		Kareem	OK
97	P4722-13	WC-3(0-6)	SAM	11/07/24 21:38		Kareem	OK
98	PB164723BL	PB164723BL	MB	11/07/24 21:42		Kareem	OK
99	PB164723BS	PB164723BS	LCS	11/07/24 21:47	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
100	CCV09	CCV09	CCV	11/07/24 21:51		Kareem	OK
101	CCB09	CCB09	CCB	11/07/24 21:55		Kareem	OK
102	P4693-01	BP-G5-WC	SAM	11/07/24 22:09		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133344

Review By	mohan	Review On	11/9/2024 12:07:42 AM
Supervise By	kareem	Supervise On	11/13/2024 7:36:39 PM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP83080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091 MP83092

103	P4693-05	BP-G4-WC	SAM	11/07/24 22:13		Kareem	OK
104	P4694-01	Z-03A	SAM	11/07/24 22:17		Kareem	OK
105	P4694-05	Z-04	SAM	11/07/24 22:22		Kareem	OK
106	P4695-01	Z-01	SAM	11/07/24 22:26		Kareem	OK
107	P4697-01	TP-1	SAM	11/07/24 22:30		Kareem	OK
108	P4699-01	MIXED-DEMO	SAM	11/07/24 22:34		Kareem	OK
109	P4699-01DUP	MIXED-DEMODUP	DUP	11/07/24 22:39		Kareem	OK
110	P4699-01L	MIXED-DEMOL	SD	11/07/24 22:43		Kareem	OK
111	P4699-01MS	MIXED-DEMOMS	MS	11/07/24 22:47	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
112	CCV10	CCV10	CCV	11/07/24 22:52		Kareem	OK
113	CCB10	CCB10	CCB	11/07/24 22:56		Kareem	OK
114	P4699-01MSD	MIXED-DEMOMSD	MSD	11/07/24 23:00	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
115	P4699-01A	MIXED-DEMOA	PS	11/07/24 23:04	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
116	P4700-01	MH-8	SAM	11/07/24 23:09		Kareem	OK
117	P4701-01RE	BP-F3RE	SAM	11/07/24 23:13	NOT USE	Kareem	Not Ok
118	P4701-05	BP-F4	SAM	11/07/24 23:17		Kareem	OK
119	P4707-01	HD-02-110424	SAM	11/07/24 23:22		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133344

Review By	mohan	Review On	11/9/2024 12:07:42 AM
Supervise By	kareem	Supervise On	11/13/2024 7:36:39 PM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP83080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091 MP83092

120	P4718-01	WB-307-SB01	SAM	11/07/24 23:26		Kareem	OK
121	P4718-02	WB-307-SB02	SAM	11/07/24 23:30		Kareem	OK
122	P4719-01	BAYAVE-STOCKPILE	SAM	11/07/24 23:34		Kareem	OK
123	PB164708BL	PB164708BL	MB	11/07/24 23:38		Kareem	OK
124	CCV11	CCV11	CCV	11/07/24 23:43		Kareem	OK
125	CCB11	CCB11	CCB	11/07/24 23:47		Kareem	OK
126	PB164708BS	PB164708BS	LCS	11/07/24 23:51	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
127	PB164563BL	PB164563BL	MB	11/07/24 23:55		Kareem	OK
128	PB164563BS	PB164563BS	LCS	11/08/24 00:05	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
129	PB164634BL	PB164634BL	MB	11/08/24 00:08		Kareem	OK
130	PB164634BS	PB164634BS	LCS	11/08/24 00:13	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
131	PB164647BL	PB164647BL	MB	11/08/24 00:17		Kareem	OK
132	PB164647BS	PB164647BS	LCS	11/08/24 00:21	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
133	P4722-03DL	WC-1(0-6)DL	SAM	11/08/24 00:26	5x for Fe	Kareem	Confirms
134	LR1	LR1	HIGH STD	11/08/24 00:30		Kareem	OK
135	LR2	LR2	HIGH STD	11/08/24 00:35		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133344

Review By	mohan	Review On	11/9/2024 12:07:42 AM
Supervise By	kareem	Supervise On	11/13/2024 7:36:39 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83078,MP83079,MP83080,MP83081,MP83082,MP83084		
ICV Standard	MP83085		
CCV Standard	MP83088		
ICSA Standard	MP83086,MP83087		
CRI Standard	MP83084		
LCS Standard			
Chk Standard	MP83091 MP83092		

136	CCV12	CCV12	CCV	11/08/24 00:39		Kareem	OK
137	CCB12	CCB12	CCB	11/08/24 00:43		Kareem	OK
138	P4701-01DL	BP-F3DL	SAM	11/08/24 00:48	NOT USE	Kareem	Not Ok
139	CCV13	CCV13	CCV	11/08/24 00:57		Kareem	OK
140	CCB13	CCB13	CCB	11/08/24 01:01		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133365

Review By	jaswal	Review On	11/12/2024 1:04:50 PM
Supervise By	kareem	Supervise On	11/12/2024 2:08:18 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83078,MP83079,MP83080,MP83081,MP83082,MP83084		
ICV Standard	MP83085		
CCV Standard	MP83088		
ICSA Standard	MP83086,MP83087		
CRI Standard	MP83084		
LCS Standard			
Chk Standard	MP83091,MP83092		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	11/08/24 13:34		Kareem	OK
2	S1	S1	CAL2	11/08/24 13:38		Kareem	OK
3	S2	S2	CAL3	11/08/24 13:43		Kareem	OK
4	S3	S3	CAL4	11/08/24 13:47		Kareem	OK
5	S4	S4	CAL5	11/08/24 13:51		Kareem	OK
6	S5	S5	CAL6	11/08/24 13:55		Kareem	OK
7	ICV01	ICV01	ICV	11/08/24 13:59	ICV fail for Na (200.7)	Kareem	OK
8	LLICV01	LLICV01	LLICV	11/08/24 14:11		Kareem	OK
9	ICB01	ICB01	ICB	11/08/24 14:16		Kareem	OK
10	CRI01	CRI01	CRDL	11/08/24 14:20		Kareem	OK
11	ICSA01	ICSA01	ICSA	11/08/24 14:24		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	11/08/24 14:30		Kareem	OK
13	ICSADL	ICSADL	ICSA	11/08/24 14:35		Kareem	OK
14	ICSABDL	ICSABDL	ICSAB	11/08/24 14:40		Kareem	OK
15	CCV01	CCV01	CCV	11/08/24 14:44		Kareem	OK
16	CCB01	CCB01	CCB	11/08/24 14:50		Kareem	OK
17	P4578-04DL	WB-305-TOP-1DL	SAM	11/08/24 14:55	5x for Cr	Kareem	Confirms
18	P4719-03	BAYAVE-STOCKPILE	SAM	11/08/24 14:59		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133365

Review By	jaswal	Review On	11/12/2024 1:04:50 PM
Supervise By	kareem	Supervise On	11/12/2024 2:08:18 PM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP83080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

19	P4720-05	JC-701-COMP-01	SAM	11/08/24 15:03		Kareem	OK
20	P4734-01	EO-03-11062024	SAM	11/08/24 15:07		Kareem	OK
21	P4757-01	001-WILLETS-PT-BL	SAM	11/08/24 15:12		Kareem	OK
22	P4757-02	002-35TH-AVE(SEP)	SAM	11/08/24 15:16		Kareem	OK
23	P4757-03	001-WILLETS-PT-BL	SAM	11/08/24 15:20		Kareem	OK
24	P4757-04	002-35TH-AVE(OCT)	SAM	11/08/24 15:25		Kareem	OK
25	CCV02	CCV02	CCV	11/08/24 15:29		Kareem	OK
26	CCB02	CCB02	CCB	11/08/24 15:33		Kareem	OK
27	P4718-03	WB-307-SB02	SAM	11/08/24 15:39	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
28	P4718-03DUP	WB-307-SB02DUP	DUP	11/08/24 15:44	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
29	P4718-03L	WB-307-SB02L	SD	11/08/24 15:48	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
30	P4718-03MS	WB-307-SB02MS	MS	11/08/24 15:52	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
31	P4718-03MSD	WB-307-SB02MSD	MSD	11/08/24 15:57	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
32	P4718-03A	WB-307-SB02A	PS	11/08/24 16:01	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
33	P4722-04	WC-1(0-6)	SAM	11/08/24 16:06		Kareem	OK
34	P4722-09	WC-2(0-6)	SAM	11/08/24 16:10		Kareem	OK
35	P4722-14	WC-3(0-6)	SAM	11/08/24 16:15		Kareem	OK
36	P4732-02	PPE-COMP	SAM	11/08/24 16:19		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133365

Review By	jaswal	Review On	11/12/2024 1:04:50 PM
Supervise By	kareem	Supervise On	11/12/2024 2:08:18 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83078,MP83079,MP83080,MP83081,MP83082,MP83084		
ICV Standard	MP83085		
CCV Standard	MP83088		
ICSA Standard	MP83086,MP83087		
CRI Standard	MP83084		
LCS Standard			
Chk Standard	MP83091,MP83092		

37	CCV03	CCV03	CCV	11/08/24 16:23		Kareem	OK
38	CCB03	CCB03	CCB	11/08/24 16:28		Kareem	OK
39	P4738-02	72-12018	SAM	11/08/24 16:32		Kareem	OK
40	P4739-04	TP-14	SAM	11/08/24 16:36		Kareem	OK
41	P4739-08	BP-G2	SAM	11/08/24 16:41		Kareem	OK
42	P4739-12	BP-B2	SAM	11/08/24 16:45		Kareem	OK
43	P4739-16	TP-11	SAM	11/08/24 16:50		Kareem	OK
44	PB164694TB	PB164694TB	MB	11/08/24 16:54		Kareem	OK
45	PB164768BL	PB164768BL	MB	11/08/24 16:59		Kareem	OK
46	PB164768BS	PB164768BS	LCS	11/08/24 17:03	0.1 ML OF M6001 AND M6010 WERE ADDED TO 10 ML OF THE SAMPLE	Kareem	OK
47	P4762-02	EFF-WW	SAM	11/08/24 17:07		Kareem	OK
48	P4762-02DUP	EFF-WWDUP	DUP	11/08/24 17:11		Kareem	OK
49	CCV04	CCV04	CCV	11/08/24 17:16		Kareem	OK
50	CCB04	CCB04	CCB	11/08/24 17:22		Kareem	OK
51	P4762-02L	EFF-WWL	SD	11/08/24 17:26		Kareem	OK
52	P4762-02MS	EFF-WWMS	MS	11/08/24 17:31	0.1 ML OF M6001 AND M6010 WERE ADDED TO 10 ML OF THE SAMPLE	Kareem	OK
53	P4762-02MSD	EFF-WWMSD	MSD	11/08/24 17:35	0.1 ML OF M6001 AND M6010 WERE ADDED TO 10 ML OF THE SAMPLE	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133365

Review By	jaswal	Review On	11/12/2024 1:04:50 PM
Supervise By	kareem	Supervise On	11/12/2024 2:08:18 PM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP83080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

54	P4762-02A	EFF-WWA	PS	11/08/24 17:39	0.1 ML OF M6001 AND M6010 WERE ADDED TO 10 ML OF THE SAMPLE	Kareem	OK
55	PB164782BL	PB164782BL	MB	11/08/24 17:43		Kareem	OK
56	PB164782BS	PB164782BS	LCS	11/08/24 17:47	0.1 ML OF M6001 AND M6010 WERE ADDED TO 10 ML OF THE SAMPLE	Kareem	OK
57	P4732-01	PPE-COMP	SAM	11/08/24 17:51		Kareem	OK
58	P4737-01	TP2	SAM	11/08/24 17:55		Kareem	OK
59	P4738-01	72-12018	SAM	11/08/24 18:00		Kareem	OK
60	CCV05	CCV05	CCV	11/08/24 18:04		Kareem	OK
61	CCB05	CCB05	CCB	11/08/24 18:08		Kareem	OK
62	P4739-01	TP-14	SAM	11/08/24 18:13		Kareem	OK
63	P4739-05	BP-G2	SAM	11/08/24 18:17		Kareem	OK
64	P4739-09	BP-B2	SAM	11/08/24 18:21		Kareem	OK
65	P4739-13	TP-11	SAM	11/08/24 18:26		Kareem	OK
66	P4739-13DUP	TP-11DUP	DUP	11/08/24 18:30		Kareem	OK
67	P4739-13L	TP-11L	SD	11/08/24 18:34		Kareem	OK
68	P4739-13MS	TP-11MS	MS	11/08/24 18:38	0.1 ML OF M6001 AND M6010 WERE ADDED TO 10 ML OF THE SAMPLE	Kareem	OK
69	P4739-13MSD	TP-11MSD	MSD	11/08/24 18:42	0.1 ML OF M6001 AND M6010 WERE ADDED TO 10 ML OF THE SAMPLE	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133365

Review By	jaswal	Review On	11/12/2024 1:04:50 PM
Supervise By	kareem	Supervise On	11/12/2024 2:08:18 PM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP83080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

70	P4739-13A	TP-11A	PS	11/08/24 18:47	0.1 ML OF M6001 AND M6010 WERE ADDED TO 10 ML OF THE SAMPLE	Kareem	OK
71	PB164757BL	PB164757BL	MB	11/08/24 18:51		Kareem	OK
72	CCV06	CCV06	CCV	11/08/24 18:55		Kareem	OK
73	CCB06	CCB06	CCB	11/08/24 18:59		Kareem	OK
74	PB164757BS	PB164757BS	LCS	11/08/24 19:03	0.1 ML OF M6001 AND M6010 WERE ADDED TO 10 ML OF THE SAMPLE	Kareem	OK
75	P4733-01	402	SAM	11/08/24 19:08		Kareem	OK
76	P4733-01DUP	402DUP	DUP	11/08/24 19:12		Kareem	OK
77	P4733-01L	402L	SD	11/08/24 19:17		Kareem	OK
78	P4733-01MS	402MS	MS	11/08/24 19:21	0.1 ML OF M6001 AND M6010 WERE ADDED TO 10 ML OF THE SAMPLE	Kareem	OK
79	P4733-01MSD	402MSD	MSD	11/08/24 19:25	0.1 ML OF M6001 AND M6010 WERE ADDED TO 10 ML OF THE SAMPLE	Kareem	OK
80	P4733-01A	402A	PS	11/08/24 19:29	0.1 ML OF M6001 AND M6010 WERE ADDED TO 10 ML OF THE SAMPLE	Kareem	OK
81	PB164758BL	PB164758BL	MB	11/08/24 19:34		Kareem	OK
82	PB164758BS	PB164758BS	LCS	11/08/24 19:38	0.1 ML OF M6001 AND M6010 WERE ADDED TO 10 ML OF THE SAMPLE	Kareem	OK
83	P4718-04	WB-307-SW	SAM	11/08/24 19:42	Na (Oversaturated)	Kareem	Dilution
84	CCV07	CCV07	CCV	11/08/24 19:47		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133365

Review By	jaswal	Review On	11/12/2024 1:04:50 PM
Supervise By	kareem	Supervise On	11/12/2024 2:08:18 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83078,MP83079,MP83080,MP83081,MP83082,MP83084		
ICV Standard	MP83085		
CCV Standard	MP83088		
ICSA Standard	MP83086,MP83087		
CRI Standard	MP83084		
LCS Standard			
Chk Standard	MP83091,MP83092		

85	CCB07	CCB07	CCB	11/08/24 19:51		Kareem	OK
86	P4718-04DUP	WB-307-SWDUP	DUP	11/08/24 19:55	Na (Oversaturated)	Kareem	Dilution
87	P4718-04L	WB-307-SWL	SD	11/08/24 20:00	Na high	Kareem	Dilution
88	P4718-04MS	WB-307-SWMS	MS	11/08/24 20:04	Na (Oversaturated)	Kareem	Dilution
89	P4718-04MSD	WB-307-SWMSD	MSD	11/08/24 20:09	Na (Oversaturated)	Kareem	Dilution
90	P4718-04A	WB-307-SWA	PS	11/08/24 20:14	Na (Oversaturated)	Kareem	Dilution
91	P4736-02	COMP	SAM	11/08/24 20:18		Kareem	OK
92	PB164759BL	PB164759BL	MB	11/08/24 20:22	NOT USE	Kareem	Not Ok
93	PB164759BS	PB164759BS	LCS	11/08/24 20:27	NOT USE	Kareem	Not Ok
94	LR1	LR1	HIGH STD	11/08/24 20:38		Kareem	OK
95	LR2	LR2	HIGH STD	11/08/24 20:42		Kareem	OK
96	CCV08	CCV08	CCV	11/08/24 20:48		Kareem	OK
97	CCB08	CCB08	CCB	11/08/24 20:57		Kareem	OK
98	PB164646BL	PB164646BL	MB	11/08/24 21:02		Kareem	OK
99	PB164646BS	PB164646BS	LCS	11/08/24 21:06		Kareem	OK
100	PB164645BL	PB164645BL	MB	11/08/24 21:10	NOT USE	Kareem	Not Ok
101	PB164645BS	PB164645BS	LCS	11/08/24 21:14	NOT USE	Kareem	Not Ok
102	P4718-04DL	WB-307-SWDL	SAM	11/08/24 21:18	10x for Na	Kareem	Confirms
103	P4718-04DUPDL	WB-307-SWDUPDL	DUP	11/08/24 21:23	10x for Na	Kareem	Confirms
104	P4718-04LDL	WB-307-SWLDL	SD	11/08/24 21:27	50x for Na	Kareem	Confirms

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133365

Review By	jaswal	Review On	11/12/2024 1:04:50 PM
Supervise By	kareem	Supervise On	11/12/2024 2:08:18 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83078,MP83079,MP83080,MP83081,MP83082,MP83084		
ICV Standard	MP83085		
CCV Standard	MP83088		
ICSA Standard	MP83086,MP83087		
CRI Standard	MP83084		
LCS Standard			
Chk Standard	MP83091,MP83092		

105	P4718-04MSDL	WB-307-SWMSDL	MS	11/08/24 21:31	10x for Na	Kareem	Confirms
106	P4718-04MSDDL	WB-307-SWMSDDL	MSD	11/08/24 21:36	10x for Na	Kareem	Confirms
107	P4718-04ADL	WB-307-SWADL	PS	11/08/24 21:40	10x for Na	Kareem	Confirms
108	CCV09	CCV09	CCV	11/08/24 21:50		Kareem	OK
109	CCB09	CCB09	CCB	11/08/24 21:55		Kareem	OK
110	P4681-01	Q4-WW	SAM	11/08/24 21:59	NOT USE	Kareem	Not Ok
111	P4681-01DL	Q4-WWDL	SAM	11/08/24 22:24	Straight 25x Dilution for all elements	Kareem	OK
112	P4681-01DUPDL	Q4-WWDUPDL	DUP	11/08/24 22:29	Straight 25x Dilution for all elements	Kareem	OK
113	P4681-01LDL	Q4-WWLDL	SD	11/08/24 22:33	Straight 125x Dilution for all elements	Kareem	OK
114	P4681-01MSDL	Q4-WWMSDL	MS	11/08/24 22:44	Straight 25x Dilution for all elements	Kareem	OK
115	P4681-01MSDDL	Q4-WWMSDDL	MSD	11/08/24 22:53	Straight 25x Dilution for all elements	Kareem	OK
116	P4681-01ADL	Q4-WWADL	PS	11/08/24 23:07	Straight 25x Dilution for all elements	Kareem	OK
117	CCV10	CCV10	CCV	11/08/24 23:24		Kareem	OK
118	CCB10	CCB10	CCB	11/08/24 23:35		Kareem	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB133392

Review By	jaswal	Review On	11/11/2024 9:30:43 PM
Supervise By	mohan	Supervise On	11/11/2024 9:32:23 PM

STD. NAME	STD REF.#
ICAL Standard	MP83192,MP83193,MP83194,MP83195,MP83196,MP831MP831MP831MP831MP8317
ICV Standard	MP83198
CCV Standard	MP83200
ICSA Standard	
CRI Standard	MP83202
LCS Standard	
Chk Standard	MP83199,MP83201,MP83203,MP83205

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	11/11/24 13:36		mohan	OK
2	S0.2	S0.2	CAL2	11/11/24 13:41		mohan	OK
3	S2.5	S2.5	CAL3	11/11/24 13:44		mohan	OK
4	S5	S5	CAL4	11/11/24 13:46		mohan	OK
5	S7.5	S7.5	CAL5	11/11/24 13:53		mohan	OK
6	S10	S10	CAL6	11/11/24 13:55		mohan	OK
7	ICV80	ICV80	ICV	11/11/24 13:58		mohan	OK
8	ICB80	ICB80	ICB	11/11/24 14:01		mohan	OK
9	CCV64	CCV64	CCV	11/11/24 14:03		mohan	OK
10	CCB64	CCB64	CCB	11/11/24 14:05		mohan	OK
11	CRA	CRA	CRDL	11/11/24 14:10		mohan	OK
12	HighStd	HighStd	HIGH STD	11/11/24 14:13		mohan	OK
13	ChkStd	ChkStd	SAM	11/11/24 14:15		mohan	OK
14	PB164872BL	PB164872BL	MB	11/11/24 14:20		mohan	OK
15	PB164872BS	PB164872BS	LCS	11/11/24 14:22		mohan	OK
16	P4718-04	WB-307-SW	SAM	11/11/24 14:24		mohan	OK
17	P4736-02	COMP	SAM	11/11/24 14:26		mohan	OK
18	P4791-02	HINCHMAN-OILY-WA	SAM	11/11/24 14:29		mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB133392

Review By	jaswal	Review On	11/11/2024 9:30:43 PM
Supervise By	mohan	Supervise On	11/11/2024 9:32:23 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83192,MP83193,MP83194,MP83195,MP83196,MP831MP831MP831MP831MP831MP8317		
ICV Standard	MP83198		
CCV Standard	MP83200		
ICSA Standard			
CRI Standard	MP83202		
LCS Standard			
Chk Standard	MP83199,MP83201,MP83203,MP83205		

19	P4791-02DUP	HINCHMAN-OILY-WA	DUP	11/11/24 14:31		mohan	OK
20	P4791-02MS	HINCHMAN-OILY-WA	MS	11/11/24 14:33		mohan	OK
21	CCV65	CCV65	CCV	11/11/24 14:36		mohan	OK
22	CCB65	CCB65	CCB	11/11/24 14:41		mohan	OK
23	P4791-02MSD	HINCHMAN-OILY-WA	MSD	11/11/24 14:43		mohan	OK
24	P4792-04	MANHOLE-WASTE-D	SAM	11/11/24 14:45		mohan	OK
25	P4797-01	2885	SAM	11/11/24 14:47		mohan	OK
26	P4791-02L	HINCHMAN-OILY-WA	SD	11/11/24 14:50		mohan	OK
27	P4791-02A	HINCHMAN-OILY-WA	PS	11/11/24 14:52		mohan	OK
28	CCV66	CCV66	CCV	11/11/24 14:54		mohan	OK
29	CCB66	CCB66	CCB	11/11/24 14:57		mohan	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133524

Review By	kareem	Review On	11/20/2024 11:25:14 AM
Supervise By	Jaswal	Supervise On	11/20/2024 11:38:44 AM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	11/19/24 10:55		Kareem	OK
2	S1	S1	CAL2	11/19/24 11:00		Kareem	OK
3	S2	S2	CAL3	11/19/24 11:04		Kareem	OK
4	S3	S3	CAL4	11/19/24 11:08		Kareem	OK
5	S4	S4	CAL5	11/19/24 11:13		Kareem	OK
6	S5	S5	CAL6	11/19/24 11:17		Kareem	OK
7	ICV01	ICV01	ICV	11/19/24 11:25		Kareem	OK
8	LLICV01	LLICV01	LLICV	11/19/24 11:34		Kareem	OK
9	ICB01	ICB01	ICB	11/19/24 11:38		Kareem	OK
10	CRI01	CRI01	CRDL	11/19/24 11:43		Kareem	OK
11	ICSA01	ICSA01	ICSA	11/19/24 11:47		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	11/19/24 11:51		Kareem	OK
13	CCV01	CCV01	CCV	11/19/24 12:43		Kareem	OK
14	CCB01	CCB01	CCB	11/19/24 12:48		Kareem	OK
15	PB165067BL	PB165067BL	MB	11/19/24 12:52		Kareem	OK
16	PB165067BS	PB165067BS	LCS	11/19/24 12:56	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
17	PB164759BL	PB164759BL	MB	11/19/24 13:00		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133524

Review By	kareem	Review On	11/20/2024 11:25:14 AM
Supervise By	Jaswal	Supervise On	11/20/2024 11:38:44 AM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

18	PB164759BS	PB164759BS	LCS	11/19/24 13:05	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
19	PB164645BL	PB164645BL	MB	11/19/24 13:09		Kareem	OK
20	PB164645BS	PB164645BS	LCS	11/19/24 13:13	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
21	PB164944TB	PB164944TB	MB	11/19/24 13:17		Kareem	OK
22	P4892-01	WB-310-TOP	SAM	11/19/24 13:22	Cr high	Kareem	Dilution
23	P4892-02	WB-310-BOT	SAM	11/19/24 13:26		Kareem	OK
24	P4893-01	MH-763	SAM	11/19/24 13:30		Kareem	OK
25	CCV02	CCV02	CCV	11/19/24 13:34		Kareem	OK
26	CCB02	CCB02	CCB	11/19/24 13:38		Kareem	OK
27	P4893-05	MH-762	SAM	11/19/24 13:42		Kareem	OK
28	P4908-01	SP-1	SAM	11/19/24 13:46		Kareem	OK
29	P4908-03	SP-2	SAM	11/19/24 13:51		Kareem	OK
30	P4909-01	BU-02-111824	SAM	11/19/24 13:55		Kareem	OK
31	P4909-01DUP	BU-02-111824DUP	DUP	11/19/24 13:59		Kareem	OK
32	P4909-01L	BU-02-111824L	SD	11/19/24 14:03		Kareem	OK
33	P4909-01MS	BU-02-111824MS	MS	11/19/24 14:08	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
34	P4909-01MSD	BU-02-111824MSD	MSD	11/19/24 14:12	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133524

Review By	kareem	Review On	11/20/2024 11:25:14 AM
Supervise By	Jaswal	Supervise On	11/20/2024 11:38:44 AM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

35	P4909-01A	BU-02-111824A	PS	11/19/24 14:16	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
36	P4890-01	#3711	SAM	11/19/24 14:20		Kareem	OK
37	CCV03	CCV03	CCV	11/19/24 14:24		Kareem	OK
38	CCB03	CCB03	CCB	11/19/24 14:28		Kareem	OK
39	P4890-02	D3617	SAM	11/19/24 14:37		Kareem	OK
40	P4890-03	D3690	SAM	11/19/24 14:41		Kareem	OK
41	P4849-02	RR-1	SAM	11/19/24 14:56		Kareem	OK
42	PB165019TB	PB165019TB	MB	11/19/24 15:00		Kareem	OK
43	P4870-13	TP-1	SAM	11/19/24 15:05		Kareem	OK
44	P4870-14	MH-735	SAM	11/19/24 15:09		Kareem	OK
45	P4870-15	MH-736	SAM	11/19/24 15:13		Kareem	OK
46	P4870-16	TP-15	SAM	11/19/24 15:18		Kareem	OK
47	P4887-02	MH-739	SAM	11/19/24 15:22		Kareem	OK
48	P4887-06	MH-760	SAM	11/19/24 15:27		Kareem	OK
49	CCV04	CCV04	CCV	11/19/24 15:31		Kareem	OK
50	CCB04	CCB04	CCB	11/19/24 15:36		Kareem	OK
51	PB164986TB	PB164986TB	MB	11/19/24 15:40	fail for Cr	Kareem	Not Ok
52	PB165049BL	PB165049BL	MB	11/19/24 15:44		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133524

Review By	kareem	Review On	11/20/2024 11:25:14 AM
Supervise By	Jaswal	Supervise On	11/20/2024 11:38:44 AM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

53	PB165049BS	PB165049BS	LCS	11/19/24 15:49	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
54	P4892-01DL	WB-310-TOPDL	SAM	11/19/24 15:53	5x for Cr	Kareem	Confirms
55	P4341-02	10324-0	SAM	11/19/24 15:57		Kareem	OK
56	P4860-01	DUP-01	SAM	11/19/24 16:01		Kareem	OK
57	P4860-02	PH2-BOT-001	SAM	11/19/24 16:06		Kareem	OK
58	P4860-03	PH2-BOT-002	SAM	11/19/24 16:10		Kareem	OK
59	P4860-04	PH2-BOT-003	SAM	11/19/24 16:14		Kareem	OK
60	CCV05	CCV05	CCV	11/19/24 16:18		Kareem	OK
61	CCB05	CCB05	CCB	11/19/24 16:22		Kareem	OK
62	P4860-05	PH2-BOT-004	SAM	11/19/24 16:27		Kareem	OK
63	P4860-06	PH2-BOT-009	SAM	11/19/24 16:31		Kareem	OK
64	P4860-07	PH2-BOT-008	SAM	11/19/24 16:35		Kareem	OK
65	P4860-08	PH2-BOT-007	SAM	11/19/24 16:39		Kareem	OK
66	P4860-09	PH2-BOT-006	SAM	11/19/24 16:44		Kareem	OK
67	P4860-10	PH2-BOT-005	SAM	11/19/24 16:48		Kareem	OK
68	P4870-01	TP-1	SAM	11/19/24 16:52		Kareem	OK
69	P4870-04	MH-735	SAM	11/19/24 16:57		Kareem	OK
70	P4870-07	MH-736	SAM	11/19/24 17:01		Kareem	OK
71	P4870-10	TP-15	SAM	11/19/24 17:05		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133524

Review By	kareem	Review On	11/20/2024 11:25:14 AM
Supervise By	Jaswal	Supervise On	11/20/2024 11:38:44 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084		
ICV Standard	MP83085		
CCV Standard	MP83088		
ICSA Standard	MP83086,MP83087		
CRI Standard	MP83084		
LCS Standard			
Chk Standard	MP83091,MP83092		

72	CCV06	CCV06	CCV	11/19/24 17:19		Kareem	OK
73	CCB06	CCB06	CCB	11/19/24 17:24		Kareem	OK
74	PB165047BL	PB165047BL	MB	11/19/24 17:28		Kareem	OK
75	PB165047BS	PB165047BS	LCS	11/19/24 17:32		Kareem	OK
76	LR1	LR1	HIGH STD	11/19/24 17:51		Kareem	OK
77	LR2	LR2	HIGH STD	11/19/24 17:56		Kareem	OK
78	CCV07	CCV07	CCV	11/19/24 18:01		Kareem	OK
79	CCB07	CCB07	CCB	11/19/24 18:05		Kareem	OK

SOP ID : M3050B-Digestion-20

SDG No : N/A

Matrix : SOIL

Pipette ID: ICP A

Balance ID : M SC-2

Filter paper ID : N/A

pH Strip ID : N/A

Hood ID : #3

Block ID: 1. HOT BLOCK #2 2. N/A

Start Digest Date: 11/06/2024 Time : 10:00 Temp : 96 °C

End Digest Date: 11/06/2024 Time : 12:10 Temp : 96 °C

Digestion tube ID: M6054

Block thermometer ID: MET-DIG. #2

Dig Technician Signature: *[Signature]*

Supervisor Signature: *[Signature]*

Temp : 1. 96°C 2. N/A

Standard Name	MLS USED	STD REF. # FROM LOG
LFS-1	1.00	M6000
LFS-2	1.00	M6009
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
1:1 HNO3	10.00	MP81119
CONC: HNO3	5.00	M6116
30% H2O2	3.00	M5634
CONC: HCL	10.00	M6095
PTFE Boiling Stones	N/A	M5585
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

HOT BLOCK #2 Cell #34 : 96 C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
11/6/24 12:30	<i>[Signature]</i> Prep Met dig	<i>[Signature]</i> Metals Lab
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	pH	Initial Weight (g)	Final Vol (ml)	Color Before	Color After	Texture	Artifact	Comment	Prep Pos
P4693-01	BP-G5-WC	N/A	2.45	100	Brown	Yellow	Medium	N/A	N/A	1
P4693-05	BP-G4-WC	N/A	2.29	100	Brown	Yellow	Medium	N/A	N/A	2
P4694-01	Z-03A	N/A	2.36	100	Brown	Yellow	Medium	N/A	N/A	3
P4694-05	Z-04	N/A	2.20	100	Brown	Yellow	Medium	N/A	N/A	4
P4695-01	Z-01	N/A	2.22	100	Black	Brown	Medium	N/A	N/A	5
P4697-01	TP-1	N/A	2.26	100	Brown	Yellow	Medium	N/A	N/A	6
P4699-01	MIXED-DEMO	N/A	2.44	100	Light Grey	yellowish Green	Medium	N/A	N/A	7
P4699-01MS	MIXED-DEMOMS	N/A	2.35	100	Light Grey	yellowish Green	Medium	N/A	M6000,M6009	9
P4699-01MSD	MIXED-DEMOMSD	N/A	2.19	100	Light Grey	yellowish Green	Medium	N/A	M6000, M6009	10
P4699-01DUP	MIXED-DEMODUP	N/A	2.24	100	Light Grey	yellowish Green	Medium	N/A	N/A	8
P4700-01	MH-8	N/A	2.30	100	Black	Yellow	Medium	N/A	N/A	11
P4701-01	BP-F3	N/A	2.48	100	Yellow	Yellow	Medium	N/A	N/A	12
P4701-05	BP-F4	N/A	2.21	100	Brown	Yellow	Medium	N/A	N/A	13
P4706-01	TR-04-110424	N/A	2.29	100	Brown	Yellow	Medium	N/A	N/A	14
P4707-01	HD-02-110424	N/A	2.25	100	Brown	Yellow	Medium	N/A	N/A	15
P4708-01	OR-02-110424	N/A	2.46	100	Brown	Yellow	Medium	N/A	N/A	16
P4709-01	HR-02-110424	N/A	2.42	100	Brown	Yellow	Medium	N/A	N/A	17
P4709-03	HR-03-110424	N/A	2.15	100	Brown	Yellow	Medium	N/A	N/A	18
P4711-01	CF-613-COMP-16	N/A	2.23	100	Black	Brown	Medium	N/A	N/A	19
P4711-06	CF-613-COMP-17	N/A	2.12	100	Brown	Yellow	Medium	N/A	N/A	20
P4718-01	WB-307-SB01	N/A	2.30	100	Brown	Yellow	Medium	N/A	N/A	21
P4718-02	WB-307-SB02	N/A	2.44	100	Brown	Yellow	Medium	N/A	N/A	22
P4719-01	BAYAVE-STOCKPILE	N/A	2.21	100	Brown	Yellow	Medium	N/A	N/A	23
PB164708BL	PBS708	N/A	2.28	100	Colorless	Colorless	Fine	N/A	N/A	24
PB164708BS	LCS708	N/A	2.28	100	Colorless	Colorless	Fine	N/A	M6000, M6009	25

SOP ID : M3010A-Digestion-17

SDG No : N/A

Matrix : WATER

Pipette ID: ICP A

Balance ID : M SC-2

Filter paper ID : N/A

pH Strip ID : M6069

Hood ID : #3


Block ID: 1. HOT BLOCK #1 2. N/A

Start Digest Date: 11/07/2024 **Time :** 12:30 **Temp :** 96 °C

End Digest Date: 11/07/2024 **Time :** 15:45 **Temp :** 96 °C

Digestion tube ID: M5595

Block thermometer ID: MET-DIG. #1

Dig Technician Signature: 

Supervisor Signature: 



Temp : 1. 96°C 2. N/A

Standard Name	MLS USED	STD REF. # FROM LOG
LFS-1	0.25	M6000
LFS-2	0.25	M6009
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Conc. HNO3	3.00	M6116
1:1 HCL	5.00	MP82127
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

HOT BLOCK #1 Cell #50 : 96 C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
11/7/24 16:00	 Met dig	
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	pH	Initial Vol (ml)	Final Vol (ml)	Color Before	Color After	Clarity Before	Clarity After	Comment	Prep Pos
P4718-04	WB-307-SW	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	23
P4718-04DUP	WB-307-SWDUP	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	24
P4718-04MS	WB-307-SWMS	<2	50	25	Colorless	Colorless	Clear	Clear	M6000,M6009	25
P4718-04MSD	WB-307-SWMSD	<2	50	25	Colorless	Colorless	Clear	Clear	M6000,M6009	26
P4736-02	COMP	<2	50	25	White	Colorless	Clear	Cloudy	N/A	27
PB164759BL	PBW759	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	28
PB164759BS	LCS759	<2	50	25	Colorless	Colorless	Clear	Clear	M6000,M6009	29

SOP ID : M7471B-Mercury-18

SDG No : NA

Matrix : SOIL

Pipette ID: HG A

Balance ID : M SC-3

Filter paper ID : NA

pH Strip ID : NA

Hood ID : #1

Block ID: 1. HG HOT BLOCK#3 2. N/A

Start Digest Date: 11/07/2024 **Time :** 10:15 **Temp :** 94 °C

End Digest Date: 11/07/2024 **Time :** 10:45 **Temp :** 94 °C

Digestion tube ID: M5595

Block thermometer ID: HG-DIG#3

Dig Technician Signature: 

Supervisor Signature: 

Temp : 1. 94°C 2. N/A

Standardized Name	MLS USED	STD REF. # FROM LOG
ICV	30mL	MP83113
CCV	30mL	MP83115
CRA	30mL	MP83117
Blank Spike	0.48mL	N/A
Matrix Spike	0.48mL	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
AQUA REGIA	1.5mL	MP83119
KMnO ₄ (5%)	4.5mL	MP83108
Hydroxylamine HCL (12%)	2.0mL	MP83110
PTFE Boiling Stones	-----	M4583
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

LAB SAMPLE ID	CLIENT SAMPLE ID	Wt(g)/Vol(ml)	Comment
0.0 ppb	S0	30mL	MP83107
0.05 ppb	S0.05	N/A	N/A
0.2 ppb	S0.2	30mL	MP83108
2.5 ppb	S2.5	30mL	MP83109
5.0 ppb	S5.0	30mL	MP83110
7.5 ppb	S7.5	30mL	MP83111
10.0 ppb	S10.0	30mL	MP83112
ICV	ICV	30mL	MP83113
ICB	ICB	30mL	MP83114
CCV	CCV	30mL	MP83115
CCB	CCB	30mL	MP83116
CRI	CRI	30mL	MP83117
CHK STD	CHK STD	30mL	MP83118

Extraction Conformance/Non-Conformance Comments:

N/A		
Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
11/7/24 @ 11:25	BBG - Prep Lab	BBG - Prep Lab
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	Initial Weight (g)	Final Vol (ml)	pH	Comment	Prep Pos
P4662-06	102524-D	0.51	35	NA	N/A	3-1
P4706-01	TR-04-110424	0.54	35	NA	N/A	2
P4706-01DUP	TR-04-110424DUP	0.58	35	NA	MP83106	3
P4706-01MS	TR-04-110424MS	0.55	35	NA	MP83106	4
P4706-01MSD	TR-04-110424MSD	0.53	35	NA	N/A	5
P4707-01	HD-02-110424	0.51	35	NA	N/A	6
P4708-01	OR-02-110424	0.56	35	NA	N/A	7
P4718-01	WB-307-SB01	0.58	35	NA	N/A	8
P4718-02	WB-307-SB02	0.55	35	NA	N/A	9
P4719-01	BAYAVE-STOCKPILE	0.60	35	NA	N/A	10
P4720-01	JC-701-COMP-01	0.53	35	NA	N/A	11
P4722-03	WC-1(0-6)	0.50	35	NA	N/A	12
P4722-08	WC-2(0-6)	0.53	35	NA	N/A	13
P4722-13	WC-3(0-6)	0.58	35	NA	N/A	14
P4732-01	PPE-COMP	0.50	35	NA	N/A	15
P4734-01	EO-03-11062024	0.50	35	NA	N/A	16
P4737-01	TP2	0.56	35	NA	N/A	17
P4738-01	72-12018	0.59	35	NA	N/A	18
P4739-01	TP-14	0.50	35	NA	N/A	19
P4739-05	BP-G2	0.60	35	NA	N/A	20
P4739-09	BP-B2	0.53	35	NA	N/A	21
P4739-13	TP-11	0.51	35	NA	N/A	22
PB164762BL	PBS762	0.52	35	NA	N/A	23
PB164762BS	LCS762	0.54	35	NA	MP83106	24

SOP ID : M7470A-Mercury-19

SDG No : NA

Matrix : WATER

Pipette ID: HG A

Balance ID : N/A

Filter paper ID : NA

pH Strip ID : M4909

Hood ID : #1

Block ID: 1. HG HOT BLOCK#3 2. N/A

Start Digest Date: 11/11/2024 Time : 09:50 Temp : 94 °C

End Digest Date: 11/11/2024 Time : 11:50 Temp : 94 °C

Digestion tube ID: M5595

Block thermometer ID: HG-DIG#3

Dig Technician Signature:

Supervisor Signature:

Temp : 1. 94°C 2. N/A

Standard Name	MLS USED	STD REF. # FROM LOG
ICV	30mL	MP83198
CCV	30mL	MP83200
CRA	30mL	MP83202
Blank Spike	0.48mL	MP83191
Matrix Spike	0.48mL	MP83191

Chemical Used	ML/SAMPLE USED	Lot Number
HNO3/H2SO4(1:2)	2.5mL	MP82651
KMnO4 (5%)	4.5mL	MP82652
K2S2O8 (5%)	2.5mL	MP82653
Hydroxylamine HCL (12%)	2.0mL	MP82654
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

LAB SAMPLE ID	CLIENT SAMPLE ID	Wt(g)/Vol(ml)	Comment
0.0 ppb	S0	30mL	MP83192
0.05 ppb	S0.05	N/A	N/A
0.2 ppb	S0.2	30mL	MP83193
2.5 ppb	S2.5	30mL	MP83194
5.0 ppb	S5.0	30mL	MP83195
7.5 ppb	S7.5	30mL	MP83196
10.0 ppb	S10.0	30mL	MP83197
ICV	ICV	30mL	MP83198
ICB	ICB	30mL	MP83199
CCV	CCV	30mL	MP83200
CCB	CCB	30mL	MP83201
CRI	CRI	30mL	MP83202
CHK STD	CHK STD	30mL	MP83203

Extraction Conformance/Non-Conformance Comments:

N/A

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
11/11/24 11:15	MB - Dig. Lab	MB - Dig. Lab
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	Initial Vol (ml)	Final Vol (ml)	pH	Comment	Prep Pos
P4718-04	WB-307-SW	30	30	<2	N/A	3-1
P4736-02	COMP	30	30	<2	N/A	2
P4791-02	HINCHMAN-OILY-WATER	30	30	<2	N/A	3
P4791-02DUP	HINCHMAN-OILY-WATERDUP	30	30	<2	N/A	4
P4791-02MS	HINCHMAN-OILY-WATERMS	30	30	<2	MP83191	5
P4791-02MSD	HINCHMAN-OILY-WATERMSD	30	30	<2	MP83191	6
P4792-04	MANHOLE-WASTE-DRUMS	30	30	<2	N/A	7
P4797-01	2885	30	30	<2	N/A	8
PB164872BL	PBW872	30	30	<2	N/A	9
PB164872BS	LCS872	30	30	<2	MP83191	10