

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV74	Mercury	3.93	4.0	98	90 - 110	CV	11/07/2024	14:41	LB133334

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV47	Mercury	4.81	5.0	96	90 - 110	CV	11/07/2024	14:45	LB133334
CCV48	Mercury	4.90	5.0	98	90 - 110	CV	11/07/2024	15:13	LB133334
CCV49	Mercury	4.92	5.0	98	90 - 110	CV	11/07/2024	15:40	LB133334
CCV50	Mercury	4.92	5.0	98	90 - 110	CV	11/07/2024	16:08	LB133334

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV80	Mercury	4.09	4.0	102	90 - 110	CV	11/11/2024	13:58	LB133392

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV64	Mercury	5.16	5.0	103	90 - 110	CV	11/11/2024	14:03	LB133392
CCV65	Mercury	4.85	5.0	97	90 - 110	CV	11/11/2024	14:36	LB133392
CCV66	Mercury	4.93	5.0	99	90 - 110	CV	11/11/2024	14:54	LB133392

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	2590	2500	104	90 - 110	P	11/07/2024	13:49	LB133344
	Antimony	1010	1000	101	90 - 110	P	11/07/2024	13:49	LB133344
	Arsenic	1030	1000	103	90 - 110	P	11/07/2024	13:49	LB133344
	Barium	521	520	100	90 - 110	P	11/07/2024	13:49	LB133344
	Beryllium	515	510	101	90 - 110	P	11/07/2024	13:49	LB133344
	Cadmium	509	510	100	90 - 110	P	11/07/2024	13:49	LB133344
	Calcium	10200	10000	102	90 - 110	P	11/07/2024	13:49	LB133344
	Chromium	531	520	102	90 - 110	P	11/07/2024	13:49	LB133344
	Cobalt	516	520	99	90 - 110	P	11/07/2024	13:49	LB133344
	Copper	531	510	104	90 - 110	P	11/07/2024	13:49	LB133344
	Iron	9990	10000	100	90 - 110	P	11/07/2024	13:49	LB133344
	Lead	1020	1000	102	90 - 110	P	11/07/2024	13:49	LB133344
	Magnesium	6010	6000	100	90 - 110	P	11/07/2024	13:49	LB133344
	Manganese	526	520	101	90 - 110	P	11/07/2024	13:49	LB133344
	Nickel	518	530	98	90 - 110	P	11/07/2024	13:49	LB133344
	Potassium	9890	9900	100	90 - 110	P	11/07/2024	13:49	LB133344
	Selenium	1040	1000	104	90 - 110	P	11/07/2024	13:49	LB133344
	Silver	256	250	102	90 - 110	P	11/07/2024	13:49	LB133344
	Sodium	9440	10000	94	90 - 110	P	11/07/2024	13:49	LB133344
	Thallium	1100	1000	110	90 - 110	P	11/07/2024	13:49	LB133344
	Vanadium	507	500	101	90 - 110	P	11/07/2024	13:49	LB133344
	Zinc	1050	1000	105	90 - 110	P	11/07/2024	13:49	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	103	100	103	80 - 120	P	11/07/2024	14:23	LB133344
	Antimony	50.8	50.0	102	80 - 120	P	11/07/2024	14:23	LB133344
	Arsenic	18.0	20.0	90	80 - 120	P	11/07/2024	14:23	LB133344
	Barium	100	100	100	80 - 120	P	11/07/2024	14:23	LB133344
	Beryllium	5.93	6.0	99	80 - 120	P	11/07/2024	14:23	LB133344
	Cadmium	6.52	6.0	109	80 - 120	P	11/07/2024	14:23	LB133344
	Calcium	2010	2000	100	80 - 120	P	11/07/2024	14:23	LB133344
	Chromium	9.52	10.0	95	80 - 120	P	11/07/2024	14:23	LB133344
	Cobalt	29.4	30.0	98	80 - 120	P	11/07/2024	14:23	LB133344
	Copper	22.7	20.0	114	80 - 120	P	11/07/2024	14:23	LB133344
	Iron	96.8	100	97	80 - 120	P	11/07/2024	14:23	LB133344
	Lead	11.4	12.0	95	80 - 120	P	11/07/2024	14:23	LB133344
	Magnesium	2020	2000	101	80 - 120	P	11/07/2024	14:23	LB133344
	Manganese	20.2	20.0	101	80 - 120	P	11/07/2024	14:23	LB133344
	Nickel	38.5	40.0	96	80 - 120	P	11/07/2024	14:23	LB133344
	Potassium	2010	2000	101	80 - 120	P	11/07/2024	14:23	LB133344
	Selenium	20.8	20.0	104	80 - 120	P	11/07/2024	14:23	LB133344
	Silver	10.8	10.0	108	80 - 120	P	11/07/2024	14:23	LB133344
	Sodium	1680	2000	84	80 - 120	P	11/07/2024	14:23	LB133344
	Thallium	43.4	40.0	108	80 - 120	P	11/07/2024	14:23	LB133344
	Vanadium	40.8	40.0	102	80 - 120	P	11/07/2024	14:23	LB133344
	Zinc	42.0	40.0	105	80 - 120	P	11/07/2024	14:23	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	10000	10000	100	90 - 110	P	11/07/2024	15:01	LB133344
	Antimony	5120	5000	102	90 - 110	P	11/07/2024	15:01	LB133344
	Arsenic	5040	5000	101	90 - 110	P	11/07/2024	15:01	LB133344
	Barium	9710	10000	97	90 - 110	P	11/07/2024	15:01	LB133344
	Beryllium	239	250	96	90 - 110	P	11/07/2024	15:01	LB133344
	Cadmium	2450	2500	98	90 - 110	P	11/07/2024	15:01	LB133344
	Calcium	24000	25000	96	90 - 110	P	11/07/2024	15:01	LB133344
	Chromium	992	1000	99	90 - 110	P	11/07/2024	15:01	LB133344
	Cobalt	2450	2500	98	90 - 110	P	11/07/2024	15:01	LB133344
	Copper	1270	1250	102	90 - 110	P	11/07/2024	15:01	LB133344
	Iron	4870	5000	97	90 - 110	P	11/07/2024	15:01	LB133344
	Lead	4920	5000	98	90 - 110	P	11/07/2024	15:01	LB133344
	Magnesium	24000	25000	96	90 - 110	P	11/07/2024	15:01	LB133344
	Manganese	2370	2500	95	90 - 110	P	11/07/2024	15:01	LB133344
	Nickel	2450	2500	98	90 - 110	P	11/07/2024	15:01	LB133344
	Potassium	25100	25000	100	90 - 110	P	11/07/2024	15:01	LB133344
	Selenium	5150	5000	103	90 - 110	P	11/07/2024	15:01	LB133344
	Silver	1240	1250	99	90 - 110	P	11/07/2024	15:01	LB133344
	Sodium	25300	25000	101	90 - 110	P	11/07/2024	15:01	LB133344
	Thallium	5140	5000	103	90 - 110	P	11/07/2024	15:01	LB133344
	Vanadium	2450	2500	98	90 - 110	P	11/07/2024	15:01	LB133344
	Zinc	2530	2500	101	90 - 110	P	11/07/2024	15:01	LB133344
CCV02	Aluminum	9620	10000	96	90 - 110	P	11/07/2024	15:53	LB133344
	Antimony	5030	5000	101	90 - 110	P	11/07/2024	15:53	LB133344
	Arsenic	4950	5000	99	90 - 110	P	11/07/2024	15:53	LB133344
	Barium	9390	10000	94	90 - 110	P	11/07/2024	15:53	LB133344
	Beryllium	231	250	92	90 - 110	P	11/07/2024	15:53	LB133344
	Cadmium	2390	2500	96	90 - 110	P	11/07/2024	15:53	LB133344
	Calcium	23100	25000	92	90 - 110	P	11/07/2024	15:53	LB133344
	Chromium	963	1000	96	90 - 110	P	11/07/2024	15:53	LB133344
	Cobalt	2390	2500	96	90 - 110	P	11/07/2024	15:53	LB133344
	Copper	1250	1250	100	90 - 110	P	11/07/2024	15:53	LB133344
	Iron	4810	5000	96	90 - 110	P	11/07/2024	15:53	LB133344
	Lead	4800	5000	96	90 - 110	P	11/07/2024	15:53	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	23100	25000	92	90 - 110	P	11/07/2024	15:53	LB133344
	Manganese	2280	2500	91	90 - 110	P	11/07/2024	15:53	LB133344
	Nickel	2390	2500	96	90 - 110	P	11/07/2024	15:53	LB133344
	Potassium	25000	25000	100	90 - 110	P	11/07/2024	15:53	LB133344
	Selenium	5050	5000	101	90 - 110	P	11/07/2024	15:53	LB133344
	Silver	1220	1250	98	90 - 110	P	11/07/2024	15:53	LB133344
	Sodium	25800	25000	103	90 - 110	P	11/07/2024	15:53	LB133344
	Thallium	4850	5000	97	90 - 110	P	11/07/2024	15:53	LB133344
	Vanadium	2370	2500	95	90 - 110	P	11/07/2024	15:53	LB133344
	Zinc	2480	2500	99	90 - 110	P	11/07/2024	15:53	LB133344
CCV03	Aluminum	10100	10000	101	90 - 110	P	11/07/2024	16:45	LB133344
	Antimony	5250	5000	105	90 - 110	P	11/07/2024	16:45	LB133344
	Arsenic	5190	5000	104	90 - 110	P	11/07/2024	16:45	LB133344
	Barium	9850	10000	98	90 - 110	P	11/07/2024	16:45	LB133344
	Beryllium	235	250	94	90 - 110	P	11/07/2024	16:45	LB133344
	Cadmium	2470	2500	99	90 - 110	P	11/07/2024	16:45	LB133344
	Calcium	24000	25000	96	90 - 110	P	11/07/2024	16:45	LB133344
	Chromium	993	1000	99	90 - 110	P	11/07/2024	16:45	LB133344
	Cobalt	2470	2500	99	90 - 110	P	11/07/2024	16:45	LB133344
	Copper	1300	1250	104	90 - 110	P	11/07/2024	16:45	LB133344
	Iron	4970	5000	99	90 - 110	P	11/07/2024	16:45	LB133344
	Lead	4970	5000	99	90 - 110	P	11/07/2024	16:45	LB133344
	Magnesium	23900	25000	96	90 - 110	P	11/07/2024	16:45	LB133344
	Manganese	2370	2500	95	90 - 110	P	11/07/2024	16:45	LB133344
	Nickel	2480	2500	99	90 - 110	P	11/07/2024	16:45	LB133344
	Potassium	25900	25000	103	90 - 110	P	11/07/2024	16:45	LB133344
	Selenium	5320	5000	106	90 - 110	P	11/07/2024	16:45	LB133344
	Silver	1240	1250	100	90 - 110	P	11/07/2024	16:45	LB133344
	Sodium	25500	25000	102	90 - 110	P	11/07/2024	16:45	LB133344
	Thallium	5400	5000	108	90 - 110	P	11/07/2024	16:45	LB133344
	Vanadium	2460	2500	98	90 - 110	P	11/07/2024	16:45	LB133344
	Zinc	2400	2500	96	90 - 110	P	11/07/2024	16:45	LB133344
CCV04	Aluminum	10200	10000	102	90 - 110	P	11/07/2024	17:56	LB133344
	Antimony	5280	5000	106	90 - 110	P	11/07/2024	17:56	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	5210	5000	104	90 - 110	P	11/07/2024	17:56	LB133344
	Barium	9810	10000	98	90 - 110	P	11/07/2024	17:56	LB133344
	Beryllium	242	250	97	90 - 110	P	11/07/2024	17:56	LB133344
	Cadmium	2500	2500	100	90 - 110	P	11/07/2024	17:56	LB133344
	Calcium	24200	25000	97	90 - 110	P	11/07/2024	17:56	LB133344
	Chromium	1020	1000	102	90 - 110	P	11/07/2024	17:56	LB133344
	Cobalt	2490	2500	100	90 - 110	P	11/07/2024	17:56	LB133344
	Copper	1300	1250	104	90 - 110	P	11/07/2024	17:56	LB133344
	Iron	4980	5000	100	90 - 110	P	11/07/2024	17:56	LB133344
	Lead	5010	5000	100	90 - 110	P	11/07/2024	17:56	LB133344
	Magnesium	24300	25000	97	90 - 110	P	11/07/2024	17:56	LB133344
	Manganese	2370	2500	95	90 - 110	P	11/07/2024	17:56	LB133344
	Nickel	2490	2500	100	90 - 110	P	11/07/2024	17:56	LB133344
	Potassium	26000	25000	104	90 - 110	P	11/07/2024	17:56	LB133344
	Selenium	5390	5000	108	90 - 110	P	11/07/2024	17:56	LB133344
	Silver	1260	1250	101	90 - 110	P	11/07/2024	17:56	LB133344
	Sodium	25400	25000	102	90 - 110	P	11/07/2024	17:56	LB133344
	Thallium	5070	5000	102	90 - 110	P	11/07/2024	17:56	LB133344
	Vanadium	2490	2500	99	90 - 110	P	11/07/2024	17:56	LB133344
	Zinc	2510	2500	100	90 - 110	P	11/07/2024	17:56	LB133344
CCV05	Aluminum	10000	10000	100	90 - 110	P	11/07/2024	18:32	LB133344
	Antimony	5190	5000	104	90 - 110	P	11/07/2024	18:32	LB133344
	Arsenic	5120	5000	102	90 - 110	P	11/07/2024	18:32	LB133344
	Barium	9790	10000	98	90 - 110	P	11/07/2024	18:32	LB133344
	Beryllium	236	250	94	90 - 110	P	11/07/2024	18:32	LB133344
	Cadmium	2430	2500	97	90 - 110	P	11/07/2024	18:32	LB133344
	Calcium	23800	25000	95	90 - 110	P	11/07/2024	18:32	LB133344
	Chromium	988	1000	99	90 - 110	P	11/07/2024	18:32	LB133344
	Cobalt	2430	2500	97	90 - 110	P	11/07/2024	18:32	LB133344
	Copper	1280	1250	102	90 - 110	P	11/07/2024	18:32	LB133344
	Iron	4910	5000	98	90 - 110	P	11/07/2024	18:32	LB133344
	Lead	4880	5000	98	90 - 110	P	11/07/2024	18:32	LB133344
	Magnesium	23800	25000	95	90 - 110	P	11/07/2024	18:32	LB133344
	Manganese	2350	2500	94	90 - 110	P	11/07/2024	18:32	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2430	2500	97	90 - 110	P	11/07/2024	18:32	LB133344
	Potassium	25800	25000	103	90 - 110	P	11/07/2024	18:32	LB133344
	Selenium	5290	5000	106	90 - 110	P	11/07/2024	18:32	LB133344
	Silver	1240	1250	99	90 - 110	P	11/07/2024	18:32	LB133344
	Sodium	25600	25000	102	90 - 110	P	11/07/2024	18:32	LB133344
	Thallium	5000	5000	100	90 - 110	P	11/07/2024	18:32	LB133344
	Vanadium	2440	2500	98	90 - 110	P	11/07/2024	18:32	LB133344
	Zinc	2500	2500	100	90 - 110	P	11/07/2024	18:32	LB133344
CCV06	Aluminum	9950	10000	100	90 - 110	P	11/07/2024	19:08	LB133344
	Antimony	5160	5000	103	90 - 110	P	11/07/2024	19:08	LB133344
	Arsenic	5110	5000	102	90 - 110	P	11/07/2024	19:08	LB133344
	Barium	9640	10000	96	90 - 110	P	11/07/2024	19:08	LB133344
	Beryllium	238	250	95	90 - 110	P	11/07/2024	19:08	LB133344
	Cadmium	2460	2500	99	90 - 110	P	11/07/2024	19:08	LB133344
	Calcium	23800	25000	95	90 - 110	P	11/07/2024	19:08	LB133344
	Chromium	994	1000	99	90 - 110	P	11/07/2024	19:08	LB133344
	Cobalt	2460	2500	98	90 - 110	P	11/07/2024	19:08	LB133344
	Copper	1270	1250	102	90 - 110	P	11/07/2024	19:08	LB133344
	Iron	4870	5000	97	90 - 110	P	11/07/2024	19:08	LB133344
	Lead	4930	5000	99	90 - 110	P	11/07/2024	19:08	LB133344
	Magnesium	23800	25000	95	90 - 110	P	11/07/2024	19:08	LB133344
	Manganese	2330	2500	93	90 - 110	P	11/07/2024	19:08	LB133344
	Nickel	2460	2500	98	90 - 110	P	11/07/2024	19:08	LB133344
	Potassium	25300	25000	101	90 - 110	P	11/07/2024	19:08	LB133344
	Selenium	5260	5000	105	90 - 110	P	11/07/2024	19:08	LB133344
	Silver	1240	1250	99	90 - 110	P	11/07/2024	19:08	LB133344
	Sodium	25300	25000	101	90 - 110	P	11/07/2024	19:08	LB133344
	Thallium	5060	5000	101	90 - 110	P	11/07/2024	19:08	LB133344
	Vanadium	2430	2500	97	90 - 110	P	11/07/2024	19:08	LB133344
	Zinc	2490	2500	100	90 - 110	P	11/07/2024	19:08	LB133344
CCV07	Aluminum	9840	10000	98	90 - 110	P	11/07/2024	19:58	LB133344
	Antimony	5160	5000	103	90 - 110	P	11/07/2024	19:58	LB133344
	Arsenic	5100	5000	102	90 - 110	P	11/07/2024	19:58	LB133344
	Barium	9400	10000	94	90 - 110	P	11/07/2024	19:58	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	232	250	93	90 - 110	P	11/07/2024	19:58	LB133344
	Cadmium	2420	2500	97	90 - 110	P	11/07/2024	19:58	LB133344
	Calcium	23300	25000	93	90 - 110	P	11/07/2024	19:58	LB133344
	Chromium	986	1000	99	90 - 110	P	11/07/2024	19:58	LB133344
	Cobalt	2410	2500	96	90 - 110	P	11/07/2024	19:58	LB133344
	Copper	1270	1250	101	90 - 110	P	11/07/2024	19:58	LB133344
	Iron	4800	5000	96	90 - 110	P	11/07/2024	19:58	LB133344
	Lead	4850	5000	97	90 - 110	P	11/07/2024	19:58	LB133344
	Magnesium	23400	25000	94	90 - 110	P	11/07/2024	19:58	LB133344
	Manganese	2270	2500	91	90 - 110	P	11/07/2024	19:58	LB133344
	Nickel	2410	2500	96	90 - 110	P	11/07/2024	19:58	LB133344
	Potassium	25200	25000	101	90 - 110	P	11/07/2024	19:58	LB133344
	Selenium	5270	5000	105	90 - 110	P	11/07/2024	19:58	LB133344
	Silver	1230	1250	98	90 - 110	P	11/07/2024	19:58	LB133344
	Sodium	24900	25000	100	90 - 110	P	11/07/2024	19:58	LB133344
	Thallium	4910	5000	98	90 - 110	P	11/07/2024	19:58	LB133344
	Vanadium	2400	2500	96	90 - 110	P	11/07/2024	19:58	LB133344
	Zinc	2490	2500	100	90 - 110	P	11/07/2024	19:58	LB133344
CCV08	Aluminum	9850	10000	98	90 - 110	P	11/07/2024	20:58	LB133344
	Antimony	5180	5000	104	90 - 110	P	11/07/2024	20:58	LB133344
	Arsenic	5100	5000	102	90 - 110	P	11/07/2024	20:58	LB133344
	Barium	9480	10000	95	90 - 110	P	11/07/2024	20:58	LB133344
	Beryllium	234	250	94	90 - 110	P	11/07/2024	20:58	LB133344
	Cadmium	2410	2500	96	90 - 110	P	11/07/2024	20:58	LB133344
	Calcium	23200	25000	93	90 - 110	P	11/07/2024	20:58	LB133344
	Chromium	989	1000	99	90 - 110	P	11/07/2024	20:58	LB133344
	Cobalt	2410	2500	96	90 - 110	P	11/07/2024	20:58	LB133344
	Copper	1270	1250	102	90 - 110	P	11/07/2024	20:58	LB133344
	Iron	4890	5000	98	90 - 110	P	11/07/2024	20:58	LB133344
	Lead	4840	5000	97	90 - 110	P	11/07/2024	20:58	LB133344
	Magnesium	23300	25000	93	90 - 110	P	11/07/2024	20:58	LB133344
	Manganese	2280	2500	91	90 - 110	P	11/07/2024	20:58	LB133344
	Nickel	2410	2500	96	90 - 110	P	11/07/2024	20:58	LB133344
	Potassium	25900	25000	104	90 - 110	P	11/07/2024	20:58	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV08	Selenium	5290	5000	106	90 - 110	P	11/07/2024	20:58	LB133344
	Silver	1250	1250	100	90 - 110	P	11/07/2024	20:58	LB133344
	Sodium	26000	25000	104	90 - 110	P	11/07/2024	20:58	LB133344
	Thallium	4970	5000	99	90 - 110	P	11/07/2024	20:58	LB133344
	Vanadium	2390	2500	96	90 - 110	P	11/07/2024	20:58	LB133344
	Zinc	2530	2500	101	90 - 110	P	11/07/2024	20:58	LB133344
CCV09	Aluminum	9720	10000	97	90 - 110	P	11/07/2024	21:51	LB133344
	Antimony	5100	5000	102	90 - 110	P	11/07/2024	21:51	LB133344
	Arsenic	5030	5000	100	90 - 110	P	11/07/2024	21:51	LB133344
	Barium	9510	10000	95	90 - 110	P	11/07/2024	21:51	LB133344
	Beryllium	231	250	92	90 - 110	P	11/07/2024	21:51	LB133344
	Cadmium	2390	2500	95	90 - 110	P	11/07/2024	21:51	LB133344
	Calcium	23100	25000	92	90 - 110	P	11/07/2024	21:51	LB133344
	Chromium	966	1000	97	90 - 110	P	11/07/2024	21:51	LB133344
	Cobalt	2380	2500	95	90 - 110	P	11/07/2024	21:51	LB133344
	Copper	1250	1250	100	90 - 110	P	11/07/2024	21:51	LB133344
	Iron	4750	5000	95	90 - 110	P	11/07/2024	21:51	LB133344
	Lead	4780	5000	96	90 - 110	P	11/07/2024	21:51	LB133344
	Magnesium	22900	25000	92	90 - 110	P	11/07/2024	21:51	LB133344
	Manganese	2250	2500	90	90 - 110	P	11/07/2024	21:51	LB133344
	Nickel	2380	2500	95	90 - 110	P	11/07/2024	21:51	LB133344
	Potassium	25100	25000	100	90 - 110	P	11/07/2024	21:51	LB133344
	Selenium	5170	5000	103	90 - 110	P	11/07/2024	21:51	LB133344
	Silver	1210	1250	97	90 - 110	P	11/07/2024	21:51	LB133344
	Sodium	25300	25000	101	90 - 110	P	11/07/2024	21:51	LB133344
	Thallium	5380	5000	108	90 - 110	P	11/07/2024	21:51	LB133344
	Vanadium	2360	2500	94	90 - 110	P	11/07/2024	21:51	LB133344
	Zinc	2670	2500	107	90 - 110	P	11/07/2024	21:51	LB133344
CCV10	Aluminum	9650	10000	96	90 - 110	P	11/07/2024	22:52	LB133344
	Antimony	5080	5000	102	90 - 110	P	11/07/2024	22:52	LB133344
	Arsenic	5000	5000	100	90 - 110	P	11/07/2024	22:52	LB133344
	Barium	9460	10000	95	90 - 110	P	11/07/2024	22:52	LB133344
	Beryllium	255	250	102	90 - 110	P	11/07/2024	22:52	LB133344
	Cadmium	2340	2500	94	90 - 110	P	11/07/2024	22:52	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV10	Calcium	22600	25000	90	90 - 110	P	11/07/2024	22:52	LB133344
	Chromium	963	1000	96	90 - 110	P	11/07/2024	22:52	LB133344
	Cobalt	2340	2500	94	90 - 110	P	11/07/2024	22:52	LB133344
	Copper	1240	1250	99	90 - 110	P	11/07/2024	22:52	LB133344
	Iron	4790	5000	96	90 - 110	P	11/07/2024	22:52	LB133344
	Lead	4710	5000	94	90 - 110	P	11/07/2024	22:52	LB133344
	Magnesium	26500	25000	106	90 - 110	P	11/07/2024	22:52	LB133344
	Manganese	2500	2500	100	90 - 110	P	11/07/2024	22:52	LB133344
	Nickel	2340	2500	94	90 - 110	P	11/07/2024	22:52	LB133344
	Potassium	26100	25000	104	90 - 110	P	11/07/2024	22:52	LB133344
	Selenium	5160	5000	103	90 - 110	P	11/07/2024	22:52	LB133344
	Silver	1220	1250	98	90 - 110	P	11/07/2024	22:52	LB133344
	Sodium	27200	25000	109	90 - 110	P	11/07/2024	22:52	LB133344
	Thallium	5290	5000	106	90 - 110	P	11/07/2024	22:52	LB133344
	Vanadium	2320	2500	93	90 - 110	P	11/07/2024	22:52	LB133344
	Zinc	2400	2500	96	90 - 110	P	11/07/2024	22:52	LB133344
CCV11	Aluminum	9750	10000	98	90 - 110	P	11/07/2024	23:43	LB133344
	Antimony	5100	5000	102	90 - 110	P	11/07/2024	23:43	LB133344
	Arsenic	5020	5000	100	90 - 110	P	11/07/2024	23:43	LB133344
	Barium	9690	10000	97	90 - 110	P	11/07/2024	23:43	LB133344
	Beryllium	234	250	93	90 - 110	P	11/07/2024	23:43	LB133344
	Cadmium	2390	2500	96	90 - 110	P	11/07/2024	23:43	LB133344
	Calcium	23200	25000	93	90 - 110	P	11/07/2024	23:43	LB133344
	Chromium	962	1000	96	90 - 110	P	11/07/2024	23:43	LB133344
	Cobalt	2390	2500	95	90 - 110	P	11/07/2024	23:43	LB133344
	Copper	1250	1250	100	90 - 110	P	11/07/2024	23:43	LB133344
	Iron	4730	5000	94	90 - 110	P	11/07/2024	23:43	LB133344
	Lead	4800	5000	96	90 - 110	P	11/07/2024	23:43	LB133344
	Magnesium	23000	25000	92	90 - 110	P	11/07/2024	23:43	LB133344
	Manganese	2290	2500	92	90 - 110	P	11/07/2024	23:43	LB133344
	Nickel	2390	2500	95	90 - 110	P	11/07/2024	23:43	LB133344
	Potassium	25300	25000	101	90 - 110	P	11/07/2024	23:43	LB133344
	Selenium	5160	5000	103	90 - 110	P	11/07/2024	23:43	LB133344
	Silver	1210	1250	97	90 - 110	P	11/07/2024	23:43	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV11	Sodium	26300	25000	105	90 - 110	P	11/07/2024	23:43	LB133344
	Thallium	5350	5000	107	90 - 110	P	11/07/2024	23:43	LB133344
	Vanadium	2370	2500	95	90 - 110	P	11/07/2024	23:43	LB133344
	Zinc	2430	2500	97	90 - 110	P	11/07/2024	23:43	LB133344
CCV12	Aluminum	9840	10000	98	90 - 110	P	11/08/2024	00:39	LB133344
	Antimony	5240	5000	105	90 - 110	P	11/08/2024	00:39	LB133344
	Arsenic	5150	5000	103	90 - 110	P	11/08/2024	00:39	LB133344
	Barium	9830	10000	98	90 - 110	P	11/08/2024	00:39	LB133344
	Beryllium	228	250	91	90 - 110	P	11/08/2024	00:39	LB133344
	Cadmium	2390	2500	96	90 - 110	P	11/08/2024	00:39	LB133344
	Calcium	22900	25000	92	90 - 110	P	11/08/2024	00:39	LB133344
	Chromium	965	1000	96	90 - 110	P	11/08/2024	00:39	LB133344
	Cobalt	2390	2500	96	90 - 110	P	11/08/2024	00:39	LB133344
	Copper	1280	1250	102	90 - 110	P	11/08/2024	00:39	LB133344
	Iron	4790	5000	96	90 - 110	P	11/08/2024	00:39	LB133344
	Lead	4820	5000	96	90 - 110	P	11/08/2024	00:39	LB133344
	Magnesium	22700	25000	91	90 - 110	P	11/08/2024	00:39	LB133344
	Manganese	2260	2500	90	90 - 110	P	11/08/2024	00:39	LB133344
	Nickel	2390	2500	96	90 - 110	P	11/08/2024	00:39	LB133344
	Potassium	26400	25000	106	90 - 110	P	11/08/2024	00:39	LB133344
	Selenium	5340	5000	107	90 - 110	P	11/08/2024	00:39	LB133344
	Silver	1230	1250	98	90 - 110	P	11/08/2024	00:39	LB133344
	Sodium	22700	25000	91	90 - 110	P	11/08/2024	00:39	LB133344
	Thallium	5270	5000	105	90 - 110	P	11/08/2024	00:39	LB133344
	Vanadium	2370	2500	95	90 - 110	P	11/08/2024	00:39	LB133344
	Zinc	2380	2500	95	90 - 110	P	11/08/2024	00:39	LB133344
CCV13	Aluminum	9910	10000	99	90 - 110	P	11/08/2024	00:57	LB133344
	Antimony	5250	5000	105	90 - 110	P	11/08/2024	00:57	LB133344
	Arsenic	5160	5000	103	90 - 110	P	11/08/2024	00:57	LB133344
	Barium	9650	10000	96	90 - 110	P	11/08/2024	00:57	LB133344
	Beryllium	233	250	93	90 - 110	P	11/08/2024	00:57	LB133344
	Cadmium	2420	2500	97	90 - 110	P	11/08/2024	00:57	LB133344
	Calcium	23200	25000	93	90 - 110	P	11/08/2024	00:57	LB133344
	Chromium	980	1000	98	90 - 110	P	11/08/2024	00:57	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV13	Cobalt	2420	2500	97	90 - 110	P	11/08/2024	00:57	LB133344
	Copper	1290	1250	103	90 - 110	P	11/08/2024	00:57	LB133344
	Iron	4780	5000	96	90 - 110	P	11/08/2024	00:57	LB133344
	Lead	4870	5000	97	90 - 110	P	11/08/2024	00:57	LB133344
	Magnesium	23100	25000	92	90 - 110	P	11/08/2024	00:57	LB133344
	Manganese	2280	2500	91	90 - 110	P	11/08/2024	00:57	LB133344
	Nickel	2420	2500	97	90 - 110	P	11/08/2024	00:57	LB133344
	Potassium	26000	25000	104	90 - 110	P	11/08/2024	00:57	LB133344
	Selenium	5340	5000	107	90 - 110	P	11/08/2024	00:57	LB133344
	Silver	1230	1250	99	90 - 110	P	11/08/2024	00:57	LB133344
	Sodium	26900	25000	108	90 - 110	P	11/08/2024	00:57	LB133344
	Thallium	5410	5000	108	90 - 110	P	11/08/2024	00:57	LB133344
	Vanadium	2390	2500	96	90 - 110	P	11/08/2024	00:57	LB133344
	Zinc	2470	2500	99	90 - 110	P	11/08/2024	00:57	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	2550	2500	102	90 - 110	P	11/08/2024	13:59	LB133365
	Antimony	979	1000	98	90 - 110	P	11/08/2024	13:59	LB133365
	Arsenic	1000	1000	100	90 - 110	P	11/08/2024	13:59	LB133365
	Barium	513	520	99	90 - 110	P	11/08/2024	13:59	LB133365
	Beryllium	511	510	100	90 - 110	P	11/08/2024	13:59	LB133365
	Cadmium	497	510	97	90 - 110	P	11/08/2024	13:59	LB133365
	Calcium	10100	10000	101	90 - 110	P	11/08/2024	13:59	LB133365
	Chromium	514	520	99	90 - 110	P	11/08/2024	13:59	LB133365
	Cobalt	506	520	97	90 - 110	P	11/08/2024	13:59	LB133365
	Copper	521	510	102	90 - 110	P	11/08/2024	13:59	LB133365
	Iron	9660	10000	97	90 - 110	P	11/08/2024	13:59	LB133365
	Lead	999	1000	100	90 - 110	P	11/08/2024	13:59	LB133365
	Magnesium	5940	6000	99	90 - 110	P	11/08/2024	13:59	LB133365
	Manganese	519	520	100	90 - 110	P	11/08/2024	13:59	LB133365
	Nickel	509	530	96	90 - 110	P	11/08/2024	13:59	LB133365
	Potassium	9550	9900	96	90 - 110	P	11/08/2024	13:59	LB133365
	Selenium	1000	1000	100	90 - 110	P	11/08/2024	13:59	LB133365
	Silver	251	250	100	90 - 110	P	11/08/2024	13:59	LB133365
	Sodium	9170	10000	92	90 - 110	P	11/08/2024	13:59	LB133365
	Thallium	975	1000	98	90 - 110	P	11/08/2024	13:59	LB133365
	Vanadium	496	500	99	90 - 110	P	11/08/2024	13:59	LB133365
	Zinc	1030	1000	103	90 - 110	P	11/08/2024	13:59	LB133365

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	109	100	109	80 - 120	P	11/08/2024	14:11	LB133365
	Antimony	48.4	50.0	97	80 - 120	P	11/08/2024	14:11	LB133365
	Arsenic	19.5	20.0	98	80 - 120	P	11/08/2024	14:11	LB133365
	Barium	95.3	100	95	80 - 120	P	11/08/2024	14:11	LB133365
	Beryllium	5.66	6.0	94	80 - 120	P	11/08/2024	14:11	LB133365
	Cadmium	6.34	6.0	106	80 - 120	P	11/08/2024	14:11	LB133365
	Calcium	1970	2000	98	80 - 120	P	11/08/2024	14:11	LB133365
	Chromium	9.48	10.0	95	80 - 120	P	11/08/2024	14:11	LB133365
	Cobalt	28.3	30.0	94	80 - 120	P	11/08/2024	14:11	LB133365
	Copper	22.8	20.0	114	80 - 120	P	11/08/2024	14:11	LB133365
	Iron	93.6	100	94	80 - 120	P	11/08/2024	14:11	LB133365
	Lead	11.8	12.0	99	80 - 120	P	11/08/2024	14:11	LB133365
	Magnesium	1950	2000	98	80 - 120	P	11/08/2024	14:11	LB133365
	Manganese	19.8	20.0	99	80 - 120	P	11/08/2024	14:11	LB133365
	Nickel	37.8	40.0	94	80 - 120	P	11/08/2024	14:11	LB133365
	Potassium	1870	2000	94	80 - 120	P	11/08/2024	14:11	LB133365
	Selenium	20.6	20.0	103	80 - 120	P	11/08/2024	14:11	LB133365
	Silver	10.6	10.0	106	80 - 120	P	11/08/2024	14:11	LB133365
	Sodium	1630	2000	82	80 - 120	P	11/08/2024	14:11	LB133365
	Thallium	37.6	40.0	94	80 - 120	P	11/08/2024	14:11	LB133365
	Vanadium	38.6	40.0	96	80 - 120	P	11/08/2024	14:11	LB133365
	Zinc	42.1	40.0	105	80 - 120	P	11/08/2024	14:11	LB133365

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9960	10000	100	90 - 110	P	11/08/2024	14:44	LB133365
	Antimony	5040	5000	101	90 - 110	P	11/08/2024	14:44	LB133365
	Arsenic	4950	5000	99	90 - 110	P	11/08/2024	14:44	LB133365
	Barium	9530	10000	95	90 - 110	P	11/08/2024	14:44	LB133365
	Beryllium	241	250	96	90 - 110	P	11/08/2024	14:44	LB133365
	Cadmium	2390	2500	96	90 - 110	P	11/08/2024	14:44	LB133365
	Calcium	24100	25000	97	90 - 110	P	11/08/2024	14:44	LB133365
	Chromium	1050	1000	105	90 - 110	P	11/08/2024	14:44	LB133365
	Cobalt	2420	2500	97	90 - 110	P	11/08/2024	14:44	LB133365
	Copper	1260	1250	101	90 - 110	P	11/08/2024	14:44	LB133365
	Iron	4850	5000	97	90 - 110	P	11/08/2024	14:44	LB133365
	Lead	4870	5000	97	90 - 110	P	11/08/2024	14:44	LB133365
	Magnesium	24200	25000	97	90 - 110	P	11/08/2024	14:44	LB133365
	Manganese	2370	2500	95	90 - 110	P	11/08/2024	14:44	LB133365
	Nickel	2440	2500	97	90 - 110	P	11/08/2024	14:44	LB133365
	Potassium	24600	25000	98	90 - 110	P	11/08/2024	14:44	LB133365
	Selenium	5020	5000	100	90 - 110	P	11/08/2024	14:44	LB133365
	Silver	1230	1250	99	90 - 110	P	11/08/2024	14:44	LB133365
	Sodium	24300	25000	97	90 - 110	P	11/08/2024	14:44	LB133365
	Thallium	4770	5000	95	90 - 110	P	11/08/2024	14:44	LB133365
	Vanadium	2450	2500	98	90 - 110	P	11/08/2024	14:44	LB133365
	Zinc	2530	2500	101	90 - 110	P	11/08/2024	14:44	LB133365
CCV02	Aluminum	9830	10000	98	90 - 110	P	11/08/2024	15:29	LB133365
	Antimony	5060	5000	101	90 - 110	P	11/08/2024	15:29	LB133365
	Arsenic	4950	5000	99	90 - 110	P	11/08/2024	15:29	LB133365
	Barium	9460	10000	95	90 - 110	P	11/08/2024	15:29	LB133365
	Beryllium	231	250	92	90 - 110	P	11/08/2024	15:29	LB133365
	Cadmium	2360	2500	94	90 - 110	P	11/08/2024	15:29	LB133365
	Calcium	23300	25000	93	90 - 110	P	11/08/2024	15:29	LB133365
	Chromium	918	1000	92	90 - 110	P	11/08/2024	15:29	LB133365
	Cobalt	2390	2500	96	90 - 110	P	11/08/2024	15:29	LB133365
	Copper	1260	1250	101	90 - 110	P	11/08/2024	15:29	LB133365
	Iron	4780	5000	96	90 - 110	P	11/08/2024	15:29	LB133365
	Lead	4800	5000	96	90 - 110	P	11/08/2024	15:29	LB133365

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	23400	25000	94	90 - 110	P	11/08/2024	15:29	LB133365
	Manganese	2280	2500	91	90 - 110	P	11/08/2024	15:29	LB133365
	Nickel	2390	2500	96	90 - 110	P	11/08/2024	15:29	LB133365
	Potassium	25100	25000	100	90 - 110	P	11/08/2024	15:29	LB133365
	Selenium	5080	5000	102	90 - 110	P	11/08/2024	15:29	LB133365
	Silver	1230	1250	98	90 - 110	P	11/08/2024	15:29	LB133365
	Sodium	24500	25000	98	90 - 110	P	11/08/2024	15:29	LB133365
	Thallium	5010	5000	100	90 - 110	P	11/08/2024	15:29	LB133365
	Vanadium	2390	2500	96	90 - 110	P	11/08/2024	15:29	LB133365
	Zinc	2520	2500	101	90 - 110	P	11/08/2024	15:29	LB133365
CCV03	Aluminum	9930	10000	99	90 - 110	P	11/08/2024	16:23	LB133365
	Antimony	5110	5000	102	90 - 110	P	11/08/2024	16:23	LB133365
	Arsenic	4980	5000	100	90 - 110	P	11/08/2024	16:23	LB133365
	Barium	9610	10000	96	90 - 110	P	11/08/2024	16:23	LB133365
	Beryllium	234	250	94	90 - 110	P	11/08/2024	16:23	LB133365
	Cadmium	2380	2500	95	90 - 110	P	11/08/2024	16:23	LB133365
	Calcium	23600	25000	94	90 - 110	P	11/08/2024	16:23	LB133365
	Chromium	930	1000	93	90 - 110	P	11/08/2024	16:23	LB133365
	Cobalt	2410	2500	96	90 - 110	P	11/08/2024	16:23	LB133365
	Copper	1270	1250	101	90 - 110	P	11/08/2024	16:23	LB133365
	Iron	4880	5000	98	90 - 110	P	11/08/2024	16:23	LB133365
	Lead	4840	5000	97	90 - 110	P	11/08/2024	16:23	LB133365
	Magnesium	23700	25000	95	90 - 110	P	11/08/2024	16:23	LB133365
	Manganese	2390	2500	96	90 - 110	P	11/08/2024	16:23	LB133365
	Nickel	2410	2500	96	90 - 110	P	11/08/2024	16:23	LB133365
	Potassium	25600	25000	102	90 - 110	P	11/08/2024	16:23	LB133365
	Selenium	5130	5000	103	90 - 110	P	11/08/2024	16:23	LB133365
	Silver	1250	1250	100	90 - 110	P	11/08/2024	16:23	LB133365
	Sodium	25100	25000	100	90 - 110	P	11/08/2024	16:23	LB133365
	Thallium	5140	5000	103	90 - 110	P	11/08/2024	16:23	LB133365
	Vanadium	2420	2500	97	90 - 110	P	11/08/2024	16:23	LB133365
	Zinc	2560	2500	102	90 - 110	P	11/08/2024	16:23	LB133365
CCV04	Aluminum	9690	10000	97	90 - 110	P	11/08/2024	17:16	LB133365
	Antimony	5290	5000	106	90 - 110	P	11/08/2024	17:16	LB133365

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	5190	5000	104	90 - 110	P	11/08/2024	17:16	LB133365
	Barium	9360	10000	94	90 - 110	P	11/08/2024	17:16	LB133365
	Beryllium	227	250	91	90 - 110	P	11/08/2024	17:16	LB133365
	Cadmium	2420	2500	97	90 - 110	P	11/08/2024	17:16	LB133365
	Calcium	22800	25000	91	90 - 110	P	11/08/2024	17:16	LB133365
	Chromium	927	1000	93	90 - 110	P	11/08/2024	17:16	LB133365
	Cobalt	2450	2500	98	90 - 110	P	11/08/2024	17:16	LB133365
	Copper	1310	1250	105	90 - 110	P	11/08/2024	17:16	LB133365
	Iron	4860	5000	97	90 - 110	P	11/08/2024	17:16	LB133365
	Lead	4940	5000	99	90 - 110	P	11/08/2024	17:16	LB133365
	Magnesium	23000	25000	92	90 - 110	P	11/08/2024	17:16	LB133365
	Manganese	2310	2500	93	90 - 110	P	11/08/2024	17:16	LB133365
	Nickel	2460	2500	98	90 - 110	P	11/08/2024	17:16	LB133365
	Potassium	25800	25000	103	90 - 110	P	11/08/2024	17:16	LB133365
	Selenium	5360	5000	107	90 - 110	P	11/08/2024	17:16	LB133365
	Silver	1260	1250	101	90 - 110	P	11/08/2024	17:16	LB133365
	Sodium	24500	25000	98	90 - 110	P	11/08/2024	17:16	LB133365
	Thallium	5230	5000	105	90 - 110	P	11/08/2024	17:16	LB133365
	Vanadium	2350	2500	94	90 - 110	P	11/08/2024	17:16	LB133365
	Zinc	2610	2500	104	90 - 110	P	11/08/2024	17:16	LB133365
CCV05	Aluminum	9880	10000	99	90 - 110	P	11/08/2024	18:04	LB133365
	Antimony	5230	5000	104	90 - 110	P	11/08/2024	18:04	LB133365
	Arsenic	5110	5000	102	90 - 110	P	11/08/2024	18:04	LB133365
	Barium	9290	10000	93	90 - 110	P	11/08/2024	18:04	LB133365
	Beryllium	234	250	94	90 - 110	P	11/08/2024	18:04	LB133365
	Cadmium	2430	2500	97	90 - 110	P	11/08/2024	18:04	LB133365
	Calcium	23500	25000	94	90 - 110	P	11/08/2024	18:04	LB133365
	Chromium	906	1000	91	90 - 110	P	11/08/2024	18:04	LB133365
	Cobalt	2460	2500	98	90 - 110	P	11/08/2024	18:04	LB133365
	Copper	1290	1250	103	90 - 110	P	11/08/2024	18:04	LB133365
	Iron	4750	5000	95	90 - 110	P	11/08/2024	18:04	LB133365
	Lead	4940	5000	99	90 - 110	P	11/08/2024	18:04	LB133365
	Magnesium	23700	25000	95	90 - 110	P	11/08/2024	18:04	LB133365
	Manganese	2280	2500	91	90 - 110	P	11/08/2024	18:04	LB133365

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2460	2500	98	90 - 110	P	11/08/2024	18:04	LB133365
	Potassium	24800	25000	99	90 - 110	P	11/08/2024	18:04	LB133365
	Selenium	5280	5000	106	90 - 110	P	11/08/2024	18:04	LB133365
	Silver	1230	1250	99	90 - 110	P	11/08/2024	18:04	LB133365
	Sodium	23400	25000	94	90 - 110	P	11/08/2024	18:04	LB133365
	Thallium	5280	5000	106	90 - 110	P	11/08/2024	18:04	LB133365
	Vanadium	2410	2500	96	90 - 110	P	11/08/2024	18:04	LB133365
	Zinc	2540	2500	102	90 - 110	P	11/08/2024	18:04	LB133365
CCV06	Aluminum	9790	10000	98	90 - 110	P	11/08/2024	18:55	LB133365
	Antimony	5070	5000	101	90 - 110	P	11/08/2024	18:55	LB133365
	Arsenic	4970	5000	100	90 - 110	P	11/08/2024	18:55	LB133365
	Barium	9460	10000	95	90 - 110	P	11/08/2024	18:55	LB133365
	Beryllium	236	250	95	90 - 110	P	11/08/2024	18:55	LB133365
	Cadmium	2360	2500	95	90 - 110	P	11/08/2024	18:55	LB133365
	Calcium	23500	25000	94	90 - 110	P	11/08/2024	18:55	LB133365
	Chromium	923	1000	92	90 - 110	P	11/08/2024	18:55	LB133365
	Cobalt	2400	2500	96	90 - 110	P	11/08/2024	18:55	LB133365
	Copper	1250	1250	100	90 - 110	P	11/08/2024	18:55	LB133365
	Iron	4700	5000	94	90 - 110	P	11/08/2024	18:55	LB133365
	Lead	4830	5000	97	90 - 110	P	11/08/2024	18:55	LB133365
	Magnesium	23500	25000	94	90 - 110	P	11/08/2024	18:55	LB133365
	Manganese	2290	2500	92	90 - 110	P	11/08/2024	18:55	LB133365
	Nickel	2400	2500	96	90 - 110	P	11/08/2024	18:55	LB133365
	Potassium	24500	25000	98	90 - 110	P	11/08/2024	18:55	LB133365
	Selenium	5100	5000	102	90 - 110	P	11/08/2024	18:55	LB133365
	Silver	1220	1250	98	90 - 110	P	11/08/2024	18:55	LB133365
	Sodium	23900	25000	96	90 - 110	P	11/08/2024	18:55	LB133365
	Thallium	4990	5000	100	90 - 110	P	11/08/2024	18:55	LB133365
	Vanadium	2390	2500	96	90 - 110	P	11/08/2024	18:55	LB133365
	Zinc	2530	2500	101	90 - 110	P	11/08/2024	18:55	LB133365
CCV07	Aluminum	10100	10000	101	90 - 110	P	11/08/2024	19:47	LB133365
	Antimony	5180	5000	104	90 - 110	P	11/08/2024	19:47	LB133365
	Arsenic	5060	5000	101	90 - 110	P	11/08/2024	19:47	LB133365
	Barium	9520	10000	95	90 - 110	P	11/08/2024	19:47	LB133365

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	243	250	97	90 - 110	P	11/08/2024	19:47	LB133365
	Cadmium	2400	2500	96	90 - 110	P	11/08/2024	19:47	LB133365
	Calcium	24000	25000	96	90 - 110	P	11/08/2024	19:47	LB133365
	Chromium	920	1000	92	90 - 110	P	11/08/2024	19:47	LB133365
	Cobalt	2440	2500	98	90 - 110	P	11/08/2024	19:47	LB133365
	Copper	1280	1250	102	90 - 110	P	11/08/2024	19:47	LB133365
	Iron	4710	5000	94	90 - 110	P	11/08/2024	19:47	LB133365
	Lead	4910	5000	98	90 - 110	P	11/08/2024	19:47	LB133365
	Magnesium	24100	25000	97	90 - 110	P	11/08/2024	19:47	LB133365
	Manganese	2310	2500	92	90 - 110	P	11/08/2024	19:47	LB133365
	Nickel	2440	2500	98	90 - 110	P	11/08/2024	19:47	LB133365
	Potassium	25000	25000	100	90 - 110	P	11/08/2024	19:47	LB133365
	Selenium	5210	5000	104	90 - 110	P	11/08/2024	19:47	LB133365
	Silver	1250	1250	100	90 - 110	P	11/08/2024	19:47	LB133365
	Sodium	24600	25000	98	90 - 110	P	11/08/2024	19:47	LB133365
	Thallium	5050	5000	101	90 - 110	P	11/08/2024	19:47	LB133365
	Vanadium	2440	2500	98	90 - 110	P	11/08/2024	19:47	LB133365
	Zinc	2580	2500	103	90 - 110	P	11/08/2024	19:47	LB133365
CCV08	Aluminum	9430	10000	94	90 - 110	P	11/08/2024	20:48	LB133365
	Antimony	4620	5000	92	90 - 110	P	11/08/2024	20:48	LB133365
	Arsenic	4620	5000	92	90 - 110	P	11/08/2024	20:48	LB133365
	Barium	9200	10000	92	90 - 110	P	11/08/2024	20:48	LB133365
	Beryllium	257	250	103	90 - 110	P	11/08/2024	20:48	LB133365
	Cadmium	2370	2500	95	90 - 110	P	11/08/2024	20:48	LB133365
	Calcium	24000	25000	96	90 - 110	P	11/08/2024	20:48	LB133365
	Chromium	920	1000	92	90 - 110	P	11/08/2024	20:48	LB133365
	Cobalt	2370	2500	95	90 - 110	P	11/08/2024	20:48	LB133365
	Copper	1170	1250	94	90 - 110	P	11/08/2024	20:48	LB133365
	Iron	4520	5000	90	90 - 110	P	11/08/2024	20:48	LB133365
	Lead	4790	5000	96	90 - 110	P	11/08/2024	20:48	LB133365
	Magnesium	24600	25000	98	90 - 110	P	11/08/2024	20:48	LB133365
	Manganese	2300	2500	92	90 - 110	P	11/08/2024	20:48	LB133365
	Nickel	2380	2500	95	90 - 110	P	11/08/2024	20:48	LB133365
	Potassium	25700	25000	103	90 - 110	P	11/08/2024	20:48	LB133365

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV08	Selenium	4670	5000	93	90 - 110	P	11/08/2024	20:48	LB133365
	Silver	1190	1250	95	90 - 110	P	11/08/2024	20:48	LB133365
	Sodium	24200	25000	97	90 - 110	P	11/08/2024	20:48	LB133365
	Thallium	4810	5000	96	90 - 110	P	11/08/2024	20:48	LB133365
	Vanadium	2370	2500	95	90 - 110	P	11/08/2024	20:48	LB133365
	Zinc	2380	2500	95	90 - 110	P	11/08/2024	20:48	LB133365
CCV09	Aluminum	9360	10000	94	90 - 110	P	11/08/2024	21:50	LB133365
	Antimony	4620	5000	92	90 - 110	P	11/08/2024	21:50	LB133365
	Arsenic	4610	5000	92	90 - 110	P	11/08/2024	21:50	LB133365
	Barium	9110	10000	91	90 - 110	P	11/08/2024	21:50	LB133365
	Beryllium	251	250	100	90 - 110	P	11/08/2024	21:50	LB133365
	Cadmium	2310	2500	93	90 - 110	P	11/08/2024	21:50	LB133365
	Calcium	23600	25000	95	90 - 110	P	11/08/2024	21:50	LB133365
	Chromium	902	1000	90	90 - 110	P	11/08/2024	21:50	LB133365
	Cobalt	2330	2500	93	90 - 110	P	11/08/2024	21:50	LB133365
	Copper	1160	1250	93	90 - 110	P	11/08/2024	21:50	LB133365
	Iron	5470	5000	109	90 - 110	P	11/08/2024	21:50	LB133365
	Lead	4710	5000	94	90 - 110	P	11/08/2024	21:50	LB133365
	Magnesium	24200	25000	97	90 - 110	P	11/08/2024	21:50	LB133365
	Manganese	2270	2500	91	90 - 110	P	11/08/2024	21:50	LB133365
	Nickel	2330	2500	93	90 - 110	P	11/08/2024	21:50	LB133365
	Potassium	25800	25000	103	90 - 110	P	11/08/2024	21:50	LB133365
	Selenium	4690	5000	94	90 - 110	P	11/08/2024	21:50	LB133365
	Silver	1190	1250	95	90 - 110	P	11/08/2024	21:50	LB133365
	Sodium	23900	25000	96	90 - 110	P	11/08/2024	21:50	LB133365
	Thallium	4710	5000	94	90 - 110	P	11/08/2024	21:50	LB133365
	Vanadium	2350	2500	94	90 - 110	P	11/08/2024	21:50	LB133365
	Zinc	2420	2500	97	90 - 110	P	11/08/2024	21:50	LB133365
CCV10	Aluminum	10100	10000	101	90 - 110	P	11/08/2024	23:24	LB133365
	Antimony	4640	5000	93	90 - 110	P	11/08/2024	23:24	LB133365
	Arsenic	5340	5000	107	90 - 110	P	11/08/2024	23:24	LB133365
	Barium	9550	10000	96	90 - 110	P	11/08/2024	23:24	LB133365
	Beryllium	259	250	104	90 - 110	P	11/08/2024	23:24	LB133365
	Cadmium	2430	2500	97	90 - 110	P	11/08/2024	23:24	LB133365

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV10	Calcium	22800	25000	91	90 - 110	P	11/08/2024	23:24	LB133365
	Chromium	1000	1000	100	90 - 110	P	11/08/2024	23:24	LB133365
	Cobalt	2530	2500	101	90 - 110	P	11/08/2024	23:24	LB133365
	Copper	1130	1250	91	90 - 110	P	11/08/2024	23:24	LB133365
	Iron	5150	5000	103	90 - 110	P	11/08/2024	23:24	LB133365
	Lead	4980	5000	100	90 - 110	P	11/08/2024	23:24	LB133365
	Magnesium	24100	25000	96	90 - 110	P	11/08/2024	23:24	LB133365
	Manganese	2380	2500	95	90 - 110	P	11/08/2024	23:24	LB133365
	Nickel	2360	2500	94	90 - 110	P	11/08/2024	23:24	LB133365
	Potassium	26300	25000	105	90 - 110	P	11/08/2024	23:24	LB133365
	Selenium	4900	5000	98	90 - 110	P	11/08/2024	23:24	LB133365
	Silver	1270	1250	102	90 - 110	P	11/08/2024	23:24	LB133365
	Sodium	26100	25000	104	90 - 110	P	11/08/2024	23:24	LB133365
	Thallium	4930	5000	98	90 - 110	P	11/08/2024	23:24	LB133365
	Vanadium	2610	2500	104	90 - 110	P	11/08/2024	23:24	LB133365
	Zinc	2690	2500	108	90 - 110	P	11/08/2024	23:24	LB133365

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	2650	2500	106	90 - 110	P	11/19/2024	11:25	LB133524
	Antimony	1010	1000	101	90 - 110	P	11/19/2024	11:25	LB133524
	Arsenic	1020	1000	102	90 - 110	P	11/19/2024	11:25	LB133524
	Barium	542	520	104	90 - 110	P	11/19/2024	11:25	LB133524
	Beryllium	536	510	105	90 - 110	P	11/19/2024	11:25	LB133524
	Cadmium	526	510	103	90 - 110	P	11/19/2024	11:25	LB133524
	Calcium	10700	10000	107	90 - 110	P	11/19/2024	11:25	LB133524
	Chromium	551	520	106	90 - 110	P	11/19/2024	11:25	LB133524
	Cobalt	531	520	102	90 - 110	P	11/19/2024	11:25	LB133524
	Copper	538	510	105	90 - 110	P	11/19/2024	11:25	LB133524
	Iron	10400	10000	104	90 - 110	P	11/19/2024	11:25	LB133524
	Lead	1040	1000	104	90 - 110	P	11/19/2024	11:25	LB133524
	Magnesium	6280	6000	105	90 - 110	P	11/19/2024	11:25	LB133524
	Manganese	550	520	106	90 - 110	P	11/19/2024	11:25	LB133524
	Nickel	532	530	100	90 - 110	P	11/19/2024	11:25	LB133524
	Potassium	10000	9900	101	90 - 110	P	11/19/2024	11:25	LB133524
	Selenium	1040	1000	104	90 - 110	P	11/19/2024	11:25	LB133524
	Silver	261	250	105	90 - 110	P	11/19/2024	11:25	LB133524
	Sodium	9760	10000	98	90 - 110	P	11/19/2024	11:25	LB133524
	Thallium	1090	1000	109	90 - 110	P	11/19/2024	11:25	LB133524
	Vanadium	524	500	105	90 - 110	P	11/19/2024	11:25	LB133524
	Zinc	1060	1000	106	90 - 110	P	11/19/2024	11:25	LB133524

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	104	100	104	80 - 120	P	11/19/2024	11:34	LB133524
	Antimony	51.1	50.0	102	80 - 120	P	11/19/2024	11:34	LB133524
	Arsenic	20.5	20.0	102	80 - 120	P	11/19/2024	11:34	LB133524
	Barium	104	100	104	80 - 120	P	11/19/2024	11:34	LB133524
	Beryllium	5.85	6.0	98	80 - 120	P	11/19/2024	11:34	LB133524
	Cadmium	5.89	6.0	98	80 - 120	P	11/19/2024	11:34	LB133524
	Calcium	2090	2000	105	80 - 120	P	11/19/2024	11:34	LB133524
	Chromium	10.7	10.0	107	80 - 120	P	11/19/2024	11:34	LB133524
	Cobalt	29.9	30.0	100	80 - 120	P	11/19/2024	11:34	LB133524
	Copper	22.3	20.0	112	80 - 120	P	11/19/2024	11:34	LB133524
	Iron	100	100	100	80 - 120	P	11/19/2024	11:34	LB133524
	Lead	12.4	12.0	103	80 - 120	P	11/19/2024	11:34	LB133524
	Magnesium	2040	2000	102	80 - 120	P	11/19/2024	11:34	LB133524
	Manganese	21.4	20.0	107	80 - 120	P	11/19/2024	11:34	LB133524
	Nickel	39.7	40.0	99	80 - 120	P	11/19/2024	11:34	LB133524
	Potassium	1920	2000	96	80 - 120	P	11/19/2024	11:34	LB133524
	Selenium	20.6	20.0	103	80 - 120	P	11/19/2024	11:34	LB133524
	Silver	10.1	10.0	101	80 - 120	P	11/19/2024	11:34	LB133524
	Sodium	1820	2000	91	80 - 120	P	11/19/2024	11:34	LB133524
	Thallium	39.2	40.0	98	80 - 120	P	11/19/2024	11:34	LB133524
	Vanadium	40.3	40.0	101	80 - 120	P	11/19/2024	11:34	LB133524
	Zinc	45.8	40.0	114	80 - 120	P	11/19/2024	11:34	LB133524

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9940	10000	99	90 - 110	P	11/19/2024	12:43	LB133524
	Antimony	4980	5000	100	90 - 110	P	11/19/2024	12:43	LB133524
	Arsenic	4970	5000	99	90 - 110	P	11/19/2024	12:43	LB133524
	Barium	10000	10000	100	90 - 110	P	11/19/2024	12:43	LB133524
	Beryllium	249	250	100	90 - 110	P	11/19/2024	12:43	LB133524
	Cadmium	2490	2500	100	90 - 110	P	11/19/2024	12:43	LB133524
	Calcium	24800	25000	99	90 - 110	P	11/19/2024	12:43	LB133524
	Chromium	996	1000	100	90 - 110	P	11/19/2024	12:43	LB133524
	Cobalt	2490	2500	100	90 - 110	P	11/19/2024	12:43	LB133524
	Copper	1260	1250	101	90 - 110	P	11/19/2024	12:43	LB133524
	Iron	4880	5000	98	90 - 110	P	11/19/2024	12:43	LB133524
	Lead	4990	5000	100	90 - 110	P	11/19/2024	12:43	LB133524
	Magnesium	24800	25000	99	90 - 110	P	11/19/2024	12:43	LB133524
	Manganese	2490	2500	100	90 - 110	P	11/19/2024	12:43	LB133524
	Nickel	2500	2500	100	90 - 110	P	11/19/2024	12:43	LB133524
	Potassium	24200	25000	97	90 - 110	P	11/19/2024	12:43	LB133524
	Selenium	4950	5000	99	90 - 110	P	11/19/2024	12:43	LB133524
	Silver	1230	1250	98	90 - 110	P	11/19/2024	12:43	LB133524
	Sodium	24200	25000	97	90 - 110	P	11/19/2024	12:43	LB133524
	Thallium	5030	5000	101	90 - 110	P	11/19/2024	12:43	LB133524
	Vanadium	2500	2500	100	90 - 110	P	11/19/2024	12:43	LB133524
	Zinc	2490	2500	100	90 - 110	P	11/19/2024	12:43	LB133524
CCV02	Aluminum	10600	10000	106	90 - 110	P	11/19/2024	13:34	LB133524
	Antimony	5360	5000	107	90 - 110	P	11/19/2024	13:34	LB133524
	Arsenic	5320	5000	106	90 - 110	P	11/19/2024	13:34	LB133524
	Barium	10300	10000	103	90 - 110	P	11/19/2024	13:34	LB133524
	Beryllium	262	250	105	90 - 110	P	11/19/2024	13:34	LB133524
	Cadmium	2610	2500	104	90 - 110	P	11/19/2024	13:34	LB133524
	Calcium	26100	25000	104	90 - 110	P	11/19/2024	13:34	LB133524
	Chromium	1070	1000	107	90 - 110	P	11/19/2024	13:34	LB133524
	Cobalt	2620	2500	105	90 - 110	P	11/19/2024	13:34	LB133524
	Copper	1350	1250	108	90 - 110	P	11/19/2024	13:34	LB133524
	Iron	5270	5000	105	90 - 110	P	11/19/2024	13:34	LB133524
	Lead	5240	5000	105	90 - 110	P	11/19/2024	13:34	LB133524

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	26200	25000	105	90 - 110	P	11/19/2024	13:34	LB133524
	Manganese	2580	2500	103	90 - 110	P	11/19/2024	13:34	LB133524
	Nickel	2620	2500	105	90 - 110	P	11/19/2024	13:34	LB133524
	Potassium	26300	25000	105	90 - 110	P	11/19/2024	13:34	LB133524
	Selenium	5340	5000	107	90 - 110	P	11/19/2024	13:34	LB133524
	Silver	1320	1250	106	90 - 110	P	11/19/2024	13:34	LB133524
	Sodium	25900	25000	103	90 - 110	P	11/19/2024	13:34	LB133524
	Thallium	5230	5000	104	90 - 110	P	11/19/2024	13:34	LB133524
	Vanadium	2640	2500	105	90 - 110	P	11/19/2024	13:34	LB133524
	Zinc	2680	2500	107	90 - 110	P	11/19/2024	13:34	LB133524
CCV03	Aluminum	10400	10000	104	90 - 110	P	11/19/2024	14:24	LB133524
	Antimony	5130	5000	103	90 - 110	P	11/19/2024	14:24	LB133524
	Arsenic	5110	5000	102	90 - 110	P	11/19/2024	14:24	LB133524
	Barium	10300	10000	102	90 - 110	P	11/19/2024	14:24	LB133524
	Beryllium	266	250	106	90 - 110	P	11/19/2024	14:24	LB133524
	Cadmium	2590	2500	104	90 - 110	P	11/19/2024	14:24	LB133524
	Calcium	26000	25000	104	90 - 110	P	11/19/2024	14:24	LB133524
	Chromium	1040	1000	104	90 - 110	P	11/19/2024	14:24	LB133524
	Cobalt	2600	2500	104	90 - 110	P	11/19/2024	14:24	LB133524
	Copper	1300	1250	104	90 - 110	P	11/19/2024	14:24	LB133524
	Iron	4990	5000	100	90 - 110	P	11/19/2024	14:24	LB133524
	Lead	5200	5000	104	90 - 110	P	11/19/2024	14:24	LB133524
	Magnesium	26000	25000	104	90 - 110	P	11/19/2024	14:24	LB133524
	Manganese	2590	2500	104	90 - 110	P	11/19/2024	14:24	LB133524
	Nickel	2600	2500	104	90 - 110	P	11/19/2024	14:24	LB133524
	Potassium	24600	25000	98	90 - 110	P	11/19/2024	14:24	LB133524
	Selenium	5080	5000	102	90 - 110	P	11/19/2024	14:24	LB133524
	Silver	1280	1250	102	90 - 110	P	11/19/2024	14:24	LB133524
	Sodium	24500	25000	98	90 - 110	P	11/19/2024	14:24	LB133524
	Thallium	5160	5000	103	90 - 110	P	11/19/2024	14:24	LB133524
	Vanadium	2600	2500	104	90 - 110	P	11/19/2024	14:24	LB133524
	Zinc	2590	2500	104	90 - 110	P	11/19/2024	14:24	LB133524
CCV04	Aluminum	10400	10000	104	90 - 110	P	11/19/2024	15:31	LB133524
	Antimony	5290	5000	106	90 - 110	P	11/19/2024	15:31	LB133524

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	5260	5000	105	90 - 110	P	11/19/2024	15:31	LB133524
	Barium	10500	10000	105	90 - 110	P	11/19/2024	15:31	LB133524
	Beryllium	261	250	104	90 - 110	P	11/19/2024	15:31	LB133524
	Cadmium	2610	2500	104	90 - 110	P	11/19/2024	15:31	LB133524
	Calcium	25800	25000	103	90 - 110	P	11/19/2024	15:31	LB133524
	Chromium	1050	1000	105	90 - 110	P	11/19/2024	15:31	LB133524
	Cobalt	2610	2500	105	90 - 110	P	11/19/2024	15:31	LB133524
	Copper	1330	1250	106	90 - 110	P	11/19/2024	15:31	LB133524
	Iron	5080	5000	102	90 - 110	P	11/19/2024	15:31	LB133524
	Lead	5230	5000	105	90 - 110	P	11/19/2024	15:31	LB133524
	Magnesium	25800	25000	103	90 - 110	P	11/19/2024	15:31	LB133524
	Manganese	2570	2500	103	90 - 110	P	11/19/2024	15:31	LB133524
	Nickel	2620	2500	105	90 - 110	P	11/19/2024	15:31	LB133524
	Potassium	25500	25000	102	90 - 110	P	11/19/2024	15:31	LB133524
	Selenium	5280	5000	106	90 - 110	P	11/19/2024	15:31	LB133524
	Silver	1290	1250	103	90 - 110	P	11/19/2024	15:31	LB133524
	Sodium	25600	25000	102	90 - 110	P	11/19/2024	15:31	LB133524
	Thallium	5410	5000	108	90 - 110	P	11/19/2024	15:31	LB133524
	Vanadium	2610	2500	104	90 - 110	P	11/19/2024	15:31	LB133524
	Zinc	2620	2500	105	90 - 110	P	11/19/2024	15:31	LB133524
CCV05	Aluminum	10400	10000	104	90 - 110	P	11/19/2024	16:18	LB133524
	Antimony	5220	5000	104	90 - 110	P	11/19/2024	16:18	LB133524
	Arsenic	5210	5000	104	90 - 110	P	11/19/2024	16:18	LB133524
	Barium	10300	10000	103	90 - 110	P	11/19/2024	16:18	LB133524
	Beryllium	257	250	103	90 - 110	P	11/19/2024	16:18	LB133524
	Cadmium	2580	2500	103	90 - 110	P	11/19/2024	16:18	LB133524
	Calcium	25600	25000	102	90 - 110	P	11/19/2024	16:18	LB133524
	Chromium	1050	1000	105	90 - 110	P	11/19/2024	16:18	LB133524
	Cobalt	2580	2500	103	90 - 110	P	11/19/2024	16:18	LB133524
	Copper	1310	1250	105	90 - 110	P	11/19/2024	16:18	LB133524
	Iron	5140	5000	103	90 - 110	P	11/19/2024	16:18	LB133524
	Lead	5170	5000	104	90 - 110	P	11/19/2024	16:18	LB133524
	Magnesium	25600	25000	102	90 - 110	P	11/19/2024	16:18	LB133524
	Manganese	2540	2500	102	90 - 110	P	11/19/2024	16:18	LB133524

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2590	2500	104	90 - 110	P	11/19/2024	16:18	LB133524
	Potassium	25700	25000	103	90 - 110	P	11/19/2024	16:18	LB133524
	Selenium	5240	5000	105	90 - 110	P	11/19/2024	16:18	LB133524
	Silver	1290	1250	103	90 - 110	P	11/19/2024	16:18	LB133524
	Sodium	25300	25000	101	90 - 110	P	11/19/2024	16:18	LB133524
	Thallium	5210	5000	104	90 - 110	P	11/19/2024	16:18	LB133524
	Vanadium	2590	2500	104	90 - 110	P	11/19/2024	16:18	LB133524
	Zinc	2640	2500	106	90 - 110	P	11/19/2024	16:18	LB133524
CCV06	Aluminum	10500	10000	105	90 - 110	P	11/19/2024	17:19	LB133524
	Antimony	5280	5000	106	90 - 110	P	11/19/2024	17:19	LB133524
	Arsenic	5270	5000	105	90 - 110	P	11/19/2024	17:19	LB133524
	Barium	10400	10000	104	90 - 110	P	11/19/2024	17:19	LB133524
	Beryllium	261	250	104	90 - 110	P	11/19/2024	17:19	LB133524
	Cadmium	2610	2500	104	90 - 110	P	11/19/2024	17:19	LB133524
	Calcium	25800	25000	103	90 - 110	P	11/19/2024	17:19	LB133524
	Chromium	1060	1000	106	90 - 110	P	11/19/2024	17:19	LB133524
	Cobalt	2600	2500	104	90 - 110	P	11/19/2024	17:19	LB133524
	Copper	1330	1250	106	90 - 110	P	11/19/2024	17:19	LB133524
	Iron	5190	5000	104	90 - 110	P	11/19/2024	17:19	LB133524
	Lead	5240	5000	105	90 - 110	P	11/19/2024	17:19	LB133524
	Magnesium	25800	25000	103	90 - 110	P	11/19/2024	17:19	LB133524
	Manganese	2540	2500	102	90 - 110	P	11/19/2024	17:19	LB133524
	Nickel	2610	2500	104	90 - 110	P	11/19/2024	17:19	LB133524
	Potassium	25900	25000	104	90 - 110	P	11/19/2024	17:19	LB133524
	Selenium	5310	5000	106	90 - 110	P	11/19/2024	17:19	LB133524
	Silver	1310	1250	105	90 - 110	P	11/19/2024	17:19	LB133524
	Sodium	25900	25000	103	90 - 110	P	11/19/2024	17:19	LB133524
	Thallium	5300	5000	106	90 - 110	P	11/19/2024	17:19	LB133524
	Vanadium	2600	2500	104	90 - 110	P	11/19/2024	17:19	LB133524
	Zinc	2680	2500	107	90 - 110	P	11/19/2024	17:19	LB133524
CCV07	Aluminum	10300	10000	102	90 - 110	P	11/19/2024	18:01	LB133524
	Antimony	5080	5000	102	90 - 110	P	11/19/2024	18:01	LB133524
	Arsenic	5070	5000	101	90 - 110	P	11/19/2024	18:01	LB133524
	Barium	10600	10000	106	90 - 110	P	11/19/2024	18:01	LB133524

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	256	250	102	90 - 110	P	11/19/2024	18:01	LB133524
	Cadmium	2490	2500	99	90 - 110	P	11/19/2024	18:01	LB133524
	Calcium	25100	25000	100	90 - 110	P	11/19/2024	18:01	LB133524
	Chromium	1020	1000	102	90 - 110	P	11/19/2024	18:01	LB133524
	Cobalt	2490	2500	100	90 - 110	P	11/19/2024	18:01	LB133524
	Copper	1290	1250	104	90 - 110	P	11/19/2024	18:01	LB133524
	Iron	4990	5000	100	90 - 110	P	11/19/2024	18:01	LB133524
	Lead	5020	5000	100	90 - 110	P	11/19/2024	18:01	LB133524
	Magnesium	25100	25000	100	90 - 110	P	11/19/2024	18:01	LB133524
	Manganese	2600	2500	104	90 - 110	P	11/19/2024	18:01	LB133524
	Nickel	2500	2500	100	90 - 110	P	11/19/2024	18:01	LB133524
	Potassium	25200	25000	101	90 - 110	P	11/19/2024	18:01	LB133524
	Selenium	5100	5000	102	90 - 110	P	11/19/2024	18:01	LB133524
	Silver	1270	1250	101	90 - 110	P	11/19/2024	18:01	LB133524
	Sodium	25600	25000	102	90 - 110	P	11/19/2024	18:01	LB133524
	Thallium	4940	5000	99	90 - 110	P	11/19/2024	18:01	LB133524
	Vanadium	2540	2500	102	90 - 110	P	11/19/2024	18:01	LB133524
	Zinc	2480	2500	99	90 - 110	P	11/19/2024	18:01	LB133524