



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Portal Partners Tri-Venture				SDG No.:	P4718				
Contract:	PORT06	Lab Code:	CHEM		Case No.:	P4718	SAS No.:	P4718		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB74	Mercury	0.20	+/-0.20	U	0.20	CV	11/07/2024	14:43	LB133334	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: P4718

Contract: PORT06

Lab Code: CHEM

Case No.: P4718

SAS No.: P4718

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB47	Mercury	0.20	+/-0.20	U	0.20	CV	11/07/2024	14:48	LB133334
CCB48	Mercury	0.20	+/-0.20	U	0.20	CV	11/07/2024	15:15	LB133334
CCB49	Mercury	0.20	+/-0.20	U	0.20	CV	11/07/2024	15:43	LB133334
CCB50	Mercury	0.20	+/-0.20	U	0.20	CV	11/07/2024	16:10	LB133334

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Portal Partners Tri-Venture				SDG No.:	P4718			
Contract:	PORT06	Lab Code:	CHEM		Case No.:	P4718	SAS No.:	P4718	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB80	Mercury	0.20	+/-0.20	U	0.20	CV	11/11/2024	14:01	LB133392

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: P4718

Contract: PORT06

Lab Code: CHEM

Case No.: P4718

SAS No.: P4718

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB64	Mercury	0.20	+/-0.20	U	0.20	CV	11/11/2024	14:05	LB133392
CCB65	Mercury	0.20	+/-0.20	U	0.20	CV	11/11/2024	14:41	LB133392
CCB66	Mercury	0.20	+/-0.20	U	0.20	CV	11/11/2024	14:57	LB133392

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Portal Partners Tri-Venture				SDG No.:	P4718			
Contract:	PORT06	Lab Code:	CHEM	Case No.:	P4718	SAS No.:	P4718		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	11/07/2024	14:38	LB133344
	Antimony	50.0	+/-50.0	U	50.0	P	11/07/2024	14:38	LB133344
	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	14:38	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	14:38	LB133344
	Beryllium	6.00	+/-6.00	U	6.00	P	11/07/2024	14:38	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	14:38	LB133344
	Calcium	2000	+/-2000	U	2000	P	11/07/2024	14:38	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	14:38	LB133344
	Cobalt	30.0	+/-30.0	U	30.0	P	11/07/2024	14:38	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	14:38	LB133344
	Iron	100	+/-100	U	100	P	11/07/2024	14:38	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	14:38	LB133344
	Magnesium	2000	+/-2000	U	2000	P	11/07/2024	14:38	LB133344
	Manganese	20.0	+/-20.0	U	20.0	P	11/07/2024	14:38	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	14:38	LB133344
	Potassium	2000	+/-2000	U	2000	P	11/07/2024	14:38	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	14:38	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	14:38	LB133344
	Sodium	2000	+/-2000	U	2000	P	11/07/2024	14:38	LB133344
	Thallium	40.0	+/-40.0	U	40.0	P	11/07/2024	14:38	LB133344
	Vanadium	40.0	+/-40.0	U	40.0	P	11/07/2024	14:38	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	14:38	LB133344

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	11/07/2024	15:05	LB133344
	Antimony	50.0	+/-50.0	U	50.0	P	11/07/2024	15:05	LB133344
	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	15:05	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	15:05	LB133344
	Beryllium	6.00	+/-6.00	U	6.00	P	11/07/2024	15:05	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	15:05	LB133344
	Calcium	2000	+/-2000	U	2000	P	11/07/2024	15:05	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	15:05	LB133344
	Cobalt	30.0	+/-30.0	U	30.0	P	11/07/2024	15:05	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	15:05	LB133344
	Iron	100	+/-100	U	100	P	11/07/2024	15:05	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	15:05	LB133344
	Magnesium	2000	+/-2000	U	2000	P	11/07/2024	15:05	LB133344
	Manganese	20.0	+/-20.0	U	20.0	P	11/07/2024	15:05	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	15:05	LB133344
	Potassium	2000	+/-2000	U	2000	P	11/07/2024	15:05	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	15:05	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	15:05	LB133344
	Sodium	2000	+/-2000	U	2000	P	11/07/2024	15:05	LB133344
	Thallium	40.0	+/-40.0	U	40.0	P	11/07/2024	15:05	LB133344
	Vanadium	40.0	+/-40.0	U	40.0	P	11/07/2024	15:05	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	15:05	LB133344
CCB02	Aluminum	100	+/-100	U	100	P	11/07/2024	15:58	LB133344
	Antimony	50.0	+/-50.0	U	50.0	P	11/07/2024	15:58	LB133344
	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	15:58	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	15:58	LB133344
	Beryllium	6.00	+/-6.00	U	6.00	P	11/07/2024	15:58	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	15:58	LB133344
	Calcium	2000	+/-2000	U	2000	P	11/07/2024	15:58	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	15:58	LB133344
	Cobalt	30.0	+/-30.0	U	30.0	P	11/07/2024	15:58	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	15:58	LB133344
	Iron	100	+/-100	U	100	P	11/07/2024	15:58	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	15:58	LB133344
	Magnesium	2000	+/-2000	U	2000	P	11/07/2024	15:58	LB133344
	Manganese	20.0	+/-20.0	U	20.0	P	11/07/2024	15:58	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	15:58	LB133344
	Potassium	2000	+/-2000	U	2000	P	11/07/2024	15:58	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	15:58	LB133344

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	15:58	LB133344
	Sodium	2000	+/-2000	U	2000	P	11/07/2024	15:58	LB133344
	Thallium	40.0	+/-40.0	U	40.0	P	11/07/2024	15:58	LB133344
	Vanadium	40.0	+/-40.0	U	40.0	P	11/07/2024	15:58	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	15:58	LB133344
CCB03	Aluminum	100	+/-100	U	100	P	11/07/2024	16:49	LB133344
	Antimony	50.0	+/-50.0	U	50.0	P	11/07/2024	16:49	LB133344
	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	16:49	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	16:49	LB133344
	Beryllium	6.00	+/-6.00	U	6.00	P	11/07/2024	16:49	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	16:49	LB133344
	Calcium	2000	+/-2000	U	2000	P	11/07/2024	16:49	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	16:49	LB133344
	Cobalt	30.0	+/-30.0	U	30.0	P	11/07/2024	16:49	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	16:49	LB133344
	Iron	100	+/-100	U	100	P	11/07/2024	16:49	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	16:49	LB133344
	Magnesium	2000	+/-2000	U	2000	P	11/07/2024	16:49	LB133344
	Manganese	20.0	+/-20.0	U	20.0	P	11/07/2024	16:49	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	16:49	LB133344
	Potassium	2000	+/-2000	U	2000	P	11/07/2024	16:49	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	16:49	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	16:49	LB133344
	Sodium	2000	+/-2000	U	2000	P	11/07/2024	16:49	LB133344
	Thallium	40.0	+/-40.0	U	40.0	P	11/07/2024	16:49	LB133344
	Vanadium	40.0	+/-40.0	U	40.0	P	11/07/2024	16:49	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	16:49	LB133344
CCB04	Aluminum	100	+/-100	U	100	P	11/07/2024	18:04	LB133344
	Antimony	50.0	+/-50.0	U	50.0	P	11/07/2024	18:04	LB133344
	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	18:04	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	18:04	LB133344
	Beryllium	6.00	+/-6.00	U	6.00	P	11/07/2024	18:04	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	18:04	LB133344
	Calcium	2000	+/-2000	U	2000	P	11/07/2024	18:04	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	18:04	LB133344
	Cobalt	30.0	+/-30.0	U	30.0	P	11/07/2024	18:04	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	18:04	LB133344
	Iron	100	+/-100	U	100	P	11/07/2024	18:04	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	18:04	LB133344

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	11/07/2024	18:04	LB133344
	Manganese	20.0	+/-20.0	U	20.0	P	11/07/2024	18:04	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	18:04	LB133344
	Potassium	2000	+/-2000	U	2000	P	11/07/2024	18:04	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	18:04	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	18:04	LB133344
	Sodium	2000	+/-2000	U	2000	P	11/07/2024	18:04	LB133344
	Thallium	40.0	+/-40.0	U	40.0	P	11/07/2024	18:04	LB133344
	Vanadium	40.0	+/-40.0	U	40.0	P	11/07/2024	18:04	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	18:04	LB133344
CCB05	Aluminum	100	+/-100	U	100	P	11/07/2024	18:37	LB133344
	Antimony	50.0	+/-50.0	U	50.0	P	11/07/2024	18:37	LB133344
	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	18:37	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	18:37	LB133344
	Beryllium	6.00	+/-6.00	U	6.00	P	11/07/2024	18:37	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	18:37	LB133344
	Calcium	2000	+/-2000	U	2000	P	11/07/2024	18:37	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	18:37	LB133344
	Cobalt	30.0	+/-30.0	U	30.0	P	11/07/2024	18:37	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	18:37	LB133344
	Iron	100	+/-100	U	100	P	11/07/2024	18:37	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	18:37	LB133344
	Magnesium	2000	+/-2000	U	2000	P	11/07/2024	18:37	LB133344
	Manganese	20.0	+/-20.0	U	20.0	P	11/07/2024	18:37	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	18:37	LB133344
	Potassium	2000	+/-2000	U	2000	P	11/07/2024	18:37	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	18:37	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	18:37	LB133344
	Sodium	2000	+/-2000	U	2000	P	11/07/2024	18:37	LB133344
	Thallium	40.0	+/-40.0	U	40.0	P	11/07/2024	18:37	LB133344
	Vanadium	40.0	+/-40.0	U	40.0	P	11/07/2024	18:37	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	18:37	LB133344
CCB06	Aluminum	100	+/-100	U	100	P	11/07/2024	19:12	LB133344
	Antimony	50.0	+/-50.0	U	50.0	P	11/07/2024	19:12	LB133344
	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	19:12	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	19:12	LB133344
	Beryllium	6.00	+/-6.00	U	6.00	P	11/07/2024	19:12	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	19:12	LB133344
	Calcium	2000	+/-2000	U	2000	P	11/07/2024	19:12	LB133344

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	19:12	LB133344
	Cobalt	30.0	+/-30.0	U	30.0	P	11/07/2024	19:12	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	19:12	LB133344
	Iron	100	+/-100	U	100	P	11/07/2024	19:12	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	19:12	LB133344
	Magnesium	2000	+/-2000	U	2000	P	11/07/2024	19:12	LB133344
	Manganese	20.0	+/-20.0	U	20.0	P	11/07/2024	19:12	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	19:12	LB133344
	Potassium	2000	+/-2000	U	2000	P	11/07/2024	19:12	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	19:12	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	19:12	LB133344
	Sodium	2000	+/-2000	U	2000	P	11/07/2024	19:12	LB133344
	Thallium	40.0	+/-40.0	U	40.0	P	11/07/2024	19:12	LB133344
	Vanadium	40.0	+/-40.0	U	40.0	P	11/07/2024	19:12	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	19:12	LB133344
CCB07	Aluminum	100	+/-100	U	100	P	11/07/2024	20:02	LB133344
	Antimony	50.0	+/-50.0	U	50.0	P	11/07/2024	20:02	LB133344
	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	20:02	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	20:02	LB133344
	Beryllium	6.00	+/-6.00	U	6.00	P	11/07/2024	20:02	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	20:02	LB133344
	Calcium	2000	+/-2000	U	2000	P	11/07/2024	20:02	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	20:02	LB133344
	Cobalt	30.0	+/-30.0	U	30.0	P	11/07/2024	20:02	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	20:02	LB133344
	Iron	100	+/-100	U	100	P	11/07/2024	20:02	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	20:02	LB133344
	Magnesium	2000	+/-2000	U	2000	P	11/07/2024	20:02	LB133344
	Manganese	20.0	+/-20.0	U	20.0	P	11/07/2024	20:02	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	20:02	LB133344
	Potassium	2000	+/-2000	U	2000	P	11/07/2024	20:02	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	20:02	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	20:02	LB133344
	Sodium	2000	+/-2000	U	2000	P	11/07/2024	20:02	LB133344
	Thallium	40.0	+/-40.0	U	40.0	P	11/07/2024	20:02	LB133344
	Vanadium	40.0	+/-40.0	U	40.0	P	11/07/2024	20:02	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	20:02	LB133344
CCB08	Aluminum	100	+/-100	U	100	P	11/07/2024	21:02	LB133344
	Antimony	50.0	+/-50.0	U	50.0	P	11/07/2024	21:02	LB133344

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	21:02	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	21:02	LB133344
	Beryllium	6.00	+/-6.00	U	6.00	P	11/07/2024	21:02	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	21:02	LB133344
	Calcium	2000	+/-2000	U	2000	P	11/07/2024	21:02	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	21:02	LB133344
	Cobalt	30.0	+/-30.0	U	30.0	P	11/07/2024	21:02	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	21:02	LB133344
	Iron	100	+/-100	U	100	P	11/07/2024	21:02	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	21:02	LB133344
	Magnesium	2000	+/-2000	U	2000	P	11/07/2024	21:02	LB133344
	Manganese	20.0	+/-20.0	U	20.0	P	11/07/2024	21:02	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	21:02	LB133344
	Potassium	2000	+/-2000	U	2000	P	11/07/2024	21:02	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	21:02	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	21:02	LB133344
	Sodium	2000	+/-2000	U	2000	P	11/07/2024	21:02	LB133344
	Thallium	40.0	+/-40.0	U	40.0	P	11/07/2024	21:02	LB133344
	Vanadium	40.0	+/-40.0	U	40.0	P	11/07/2024	21:02	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	21:02	LB133344
CCB09	Aluminum	100	+/-100	U	100	P	11/07/2024	21:55	LB133344
	Antimony	50.0	+/-50.0	U	50.0	P	11/07/2024	21:55	LB133344
	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	21:55	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	21:55	LB133344
	Beryllium	6.00	+/-6.00	U	6.00	P	11/07/2024	21:55	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	21:55	LB133344
	Calcium	2000	+/-2000	U	2000	P	11/07/2024	21:55	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	21:55	LB133344
	Cobalt	30.0	+/-30.0	U	30.0	P	11/07/2024	21:55	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	21:55	LB133344
	Iron	100	+/-100	U	100	P	11/07/2024	21:55	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	21:55	LB133344
	Magnesium	2000	+/-2000	U	2000	P	11/07/2024	21:55	LB133344
	Manganese	20.0	+/-20.0	U	20.0	P	11/07/2024	21:55	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	21:55	LB133344
	Potassium	2000	+/-2000	U	2000	P	11/07/2024	21:55	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	21:55	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	21:55	LB133344
	Sodium	2000	+/-2000	U	2000	P	11/07/2024	21:55	LB133344

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Thallium	40.0	+/-40.0	U	40.0	P	11/07/2024	21:55	LB133344
	Vanadium	40.0	+/-40.0	U	40.0	P	11/07/2024	21:55	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	21:55	LB133344
CCB10	Aluminum	100	+/-100	U	100	P	11/07/2024	22:56	LB133344
	Antimony	50.0	+/-50.0	U	50.0	P	11/07/2024	22:56	LB133344
	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	22:56	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	22:56	LB133344
	Beryllium	6.00	+/-6.00	U	6.00	P	11/07/2024	22:56	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	22:56	LB133344
	Calcium	2000	+/-2000	U	2000	P	11/07/2024	22:56	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	22:56	LB133344
	Cobalt	30.0	+/-30.0	U	30.0	P	11/07/2024	22:56	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	22:56	LB133344
	Iron	100	+/-100	U	100	P	11/07/2024	22:56	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	22:56	LB133344
	Magnesium	2000	+/-2000	U	2000	P	11/07/2024	22:56	LB133344
	Manganese	20.0	+/-20.0	U	20.0	P	11/07/2024	22:56	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	22:56	LB133344
	Potassium	2000	+/-2000	U	2000	P	11/07/2024	22:56	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	22:56	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	22:56	LB133344
	Sodium	2000	+/-2000	U	2000	P	11/07/2024	22:56	LB133344
	Thallium	40.0	+/-40.0	U	40.0	P	11/07/2024	22:56	LB133344
	Vanadium	40.0	+/-40.0	U	40.0	P	11/07/2024	22:56	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	22:56	LB133344
CCB11	Aluminum	100	+/-100	U	100	P	11/07/2024	23:47	LB133344
	Antimony	50.0	+/-50.0	U	50.0	P	11/07/2024	23:47	LB133344
	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	23:47	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	23:47	LB133344
	Beryllium	6.00	+/-6.00	U	6.00	P	11/07/2024	23:47	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	23:47	LB133344
	Calcium	2000	+/-2000	U	2000	P	11/07/2024	23:47	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	23:47	LB133344
	Cobalt	30.0	+/-30.0	U	30.0	P	11/07/2024	23:47	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	23:47	LB133344
	Iron	100	+/-100	U	100	P	11/07/2024	23:47	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	23:47	LB133344
	Magnesium	2000	+/-2000	U	2000	P	11/07/2024	23:47	LB133344
	Manganese	20.0	+/-20.0	U	20.0	P	11/07/2024	23:47	LB133344

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB11	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	23:47	LB133344
	Potassium	2000	+/-2000	U	2000	P	11/07/2024	23:47	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	23:47	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	23:47	LB133344
	Sodium	2000	+/-2000	U	2000	P	11/07/2024	23:47	LB133344
	Thallium	40.0	+/-40.0	U	40.0	P	11/07/2024	23:47	LB133344
	Vanadium	40.0	+/-40.0	U	40.0	P	11/07/2024	23:47	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	23:47	LB133344
CCB12	Aluminum	100	+/-100	U	100	P	11/08/2024	00:43	LB133344
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	00:43	LB133344
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	00:43	LB133344
	Barium	100	+/-100	U	100	P	11/08/2024	00:43	LB133344
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	00:43	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	00:43	LB133344
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	00:43	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	00:43	LB133344
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	00:43	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	00:43	LB133344
	Iron	100	+/-100	U	100	P	11/08/2024	00:43	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	00:43	LB133344
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	00:43	LB133344
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	00:43	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	00:43	LB133344
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	00:43	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	00:43	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	00:43	LB133344
	Sodium	2000	+/-2000	U	2000	P	11/08/2024	00:43	LB133344
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	00:43	LB133344
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	00:43	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	00:43	LB133344
CCB13	Aluminum	100	+/-100	U	100	P	11/08/2024	01:01	LB133344
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	01:01	LB133344
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	01:01	LB133344
	Barium	100	+/-100	U	100	P	11/08/2024	01:01	LB133344
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	01:01	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	01:01	LB133344
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	01:01	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	01:01	LB133344
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	01:01	LB133344

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB13	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	01:01	LB133344
	Iron	100	+/-100	U	100	P	11/08/2024	01:01	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	01:01	LB133344
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	01:01	LB133344
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	01:01	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	01:01	LB133344
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	01:01	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	01:01	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	01:01	LB133344
	Sodium	2000	+/-2000	U	2000	P	11/08/2024	01:01	LB133344
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	01:01	LB133344
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	01:01	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	01:01	LB133344

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	11/08/2024	14:16	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	14:16	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	14:16	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	14:16	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	14:16	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	14:16	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	14:16	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	14:16	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	14:16	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	14:16	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	14:16	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	14:16	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	14:16	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	14:16	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	14:16	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	14:16	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	14:16	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	14:16	LB133365
	Sodium	2000	+/-2000	U	2000	P	11/08/2024	14:16	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	14:16	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	14:16	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	14:16	LB133365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	11/08/2024	14:50	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	14:50	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	14:50	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	14:50	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	14:50	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	14:50	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	14:50	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	14:50	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	14:50	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	14:50	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	14:50	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	14:50	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	14:50	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	14:50	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	14:50	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	14:50	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	14:50	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	14:50	LB133365
	Sodium	2000	+/-2000	U	2000	P	11/08/2024	14:50	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	14:50	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	14:50	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	14:50	LB133365
CCB02	Aluminum	100	+/-100	U	100	P	11/08/2024	15:33	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	15:33	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	15:33	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	15:33	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	15:33	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	15:33	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	15:33	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	15:33	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	15:33	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	15:33	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	15:33	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	15:33	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	15:33	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	15:33	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	15:33	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	15:33	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	15:33	LB133365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	15:33	LB133365
	Sodium	2000	+/-2000	U	2000	P	11/08/2024	15:33	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	15:33	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	15:33	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	15:33	LB133365
CCB03	Aluminum	100	+/-100	U	100	P	11/08/2024	16:28	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	16:28	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	16:28	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	16:28	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	16:28	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	16:28	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	16:28	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	16:28	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	16:28	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	16:28	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	16:28	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	16:28	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	16:28	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	16:28	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	16:28	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	16:28	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	16:28	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	16:28	LB133365
	Sodium	2000	+/-2000	U	2000	P	11/08/2024	16:28	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	16:28	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	16:28	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	16:28	LB133365
CCB04	Aluminum	100	+/-100	U	100	P	11/08/2024	17:22	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	17:22	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	17:22	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	17:22	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	17:22	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	17:22	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	17:22	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	17:22	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	17:22	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	17:22	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	17:22	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	17:22	LB133365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	17:22	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	17:22	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	17:22	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	17:22	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	17:22	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	17:22	LB133365
	Sodium	2000	+/-2000	U	2000	P	11/08/2024	17:22	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	17:22	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	17:22	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	17:22	LB133365
CCB05	Aluminum	100	+/-100	U	100	P	11/08/2024	18:08	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	18:08	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	18:08	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	18:08	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	18:08	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	18:08	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	18:08	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	18:08	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	18:08	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	18:08	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	18:08	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	18:08	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	18:08	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	18:08	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	18:08	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	18:08	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	18:08	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	18:08	LB133365
	Sodium	2000	+/-2000	U	2000	P	11/08/2024	18:08	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	18:08	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	18:08	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	18:08	LB133365
CCB06	Aluminum	100	+/-100	U	100	P	11/08/2024	18:59	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	18:59	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	18:59	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	18:59	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	18:59	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	18:59	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	18:59	LB133365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	18:59	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	18:59	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	18:59	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	18:59	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	18:59	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	18:59	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	18:59	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	18:59	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	18:59	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	18:59	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	18:59	LB133365
	Sodium	2000	+/-2000	U	2000	P	11/08/2024	18:59	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	18:59	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	18:59	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	18:59	LB133365
CCB07	Aluminum	100	+/-100	U	100	P	11/08/2024	19:51	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	19:51	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	19:51	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	19:51	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	19:51	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	19:51	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	19:51	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	19:51	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	19:51	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	19:51	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	19:51	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	19:51	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	19:51	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	19:51	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	19:51	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	19:51	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	19:51	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	19:51	LB133365
	Sodium	941	+/-2000	J	2000	P	11/08/2024	19:51	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	19:51	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	19:51	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	19:51	LB133365
CCB08	Aluminum	100	+/-100	U	100	P	11/08/2024	20:57	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	20:57	LB133365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	20:57	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	20:57	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	20:57	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	20:57	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	20:57	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	20:57	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	20:57	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	20:57	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	20:57	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	20:57	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	20:57	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	20:57	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	20:57	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	20:57	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	20:57	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	20:57	LB133365
	Sodium	720	+/-2000	J	2000	P	11/08/2024	20:57	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	20:57	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	20:57	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	20:57	LB133365
CCB09	Aluminum	100	+/-100	U	100	P	11/08/2024	21:55	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	21:55	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	21:55	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	21:55	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	21:55	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	21:55	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	21:55	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	21:55	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	21:55	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	21:55	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	21:55	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	21:55	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	21:55	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	21:55	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	21:55	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	21:55	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	21:55	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	21:55	LB133365
	Sodium	2000	+/-2000	U	2000	P	11/08/2024	21:55	LB133365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	21:55	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	21:55	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	21:55	LB133365
CCB10	Aluminum	100	+/-100	U	100	P	11/08/2024	23:35	LB133365
	Antimony	50.0	+/-50.0	U	50.0	P	11/08/2024	23:35	LB133365
	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	23:35	LB133365
	Barium	100	+/-100	U	100	P	11/08/2024	23:35	LB133365
	Beryllium	6.00	+/-6.00	U	6.00	P	11/08/2024	23:35	LB133365
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	23:35	LB133365
	Calcium	2000	+/-2000	U	2000	P	11/08/2024	23:35	LB133365
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	23:35	LB133365
	Cobalt	30.0	+/-30.0	U	30.0	P	11/08/2024	23:35	LB133365
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	23:35	LB133365
	Iron	100	+/-100	U	100	P	11/08/2024	23:35	LB133365
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	23:35	LB133365
	Magnesium	2000	+/-2000	U	2000	P	11/08/2024	23:35	LB133365
	Manganese	20.0	+/-20.0	U	20.0	P	11/08/2024	23:35	LB133365
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	23:35	LB133365
	Potassium	2000	+/-2000	U	2000	P	11/08/2024	23:35	LB133365
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	23:35	LB133365
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	23:35	LB133365
	Sodium	742	+/-2000	J	2000	P	11/08/2024	23:35	LB133365
	Thallium	40.0	+/-40.0	U	40.0	P	11/08/2024	23:35	LB133365
	Vanadium	40.0	+/-40.0	U	40.0	P	11/08/2024	23:35	LB133365
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	23:35	LB133365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	11/19/2024	11:38	LB133524
	Antimony	50.0	+/-50.0	U	50.0	P	11/19/2024	11:38	LB133524
	Arsenic	20.0	+/-20.0	U	20.0	P	11/19/2024	11:38	LB133524
	Barium	100	+/-100	U	100	P	11/19/2024	11:38	LB133524
	Beryllium	6.00	+/-6.00	U	6.00	P	11/19/2024	11:38	LB133524
	Cadmium	6.00	+/-6.00	U	6.00	P	11/19/2024	11:38	LB133524
	Calcium	2000	+/-2000	U	2000	P	11/19/2024	11:38	LB133524
	Chromium	10.0	+/-10.0	U	10.0	P	11/19/2024	11:38	LB133524
	Cobalt	30.0	+/-30.0	U	30.0	P	11/19/2024	11:38	LB133524
	Copper	20.0	+/-20.0	U	20.0	P	11/19/2024	11:38	LB133524
	Iron	100	+/-100	U	100	P	11/19/2024	11:38	LB133524
	Lead	12.0	+/-12.0	U	12.0	P	11/19/2024	11:38	LB133524
	Magnesium	2000	+/-2000	U	2000	P	11/19/2024	11:38	LB133524
	Manganese	20.0	+/-20.0	U	20.0	P	11/19/2024	11:38	LB133524
	Nickel	40.0	+/-40.0	U	40.0	P	11/19/2024	11:38	LB133524
	Potassium	2000	+/-2000	U	2000	P	11/19/2024	11:38	LB133524
	Selenium	20.0	+/-20.0	U	20.0	P	11/19/2024	11:38	LB133524
	Silver	10.0	+/-10.0	U	10.0	P	11/19/2024	11:38	LB133524
	Sodium	2000	+/-2000	U	2000	P	11/19/2024	11:38	LB133524
	Thallium	40.0	+/-40.0	U	40.0	P	11/19/2024	11:38	LB133524
	Vanadium	40.0	+/-40.0	U	40.0	P	11/19/2024	11:38	LB133524
	Zinc	40.0	+/-40.0	U	40.0	P	11/19/2024	11:38	LB133524

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	11/19/2024	12:48	LB133524
	Antimony	50.0	+/-50.0	U	50.0	P	11/19/2024	12:48	LB133524
	Arsenic	20.0	+/-20.0	U	20.0	P	11/19/2024	12:48	LB133524
	Barium	100	+/-100	U	100	P	11/19/2024	12:48	LB133524
	Beryllium	6.00	+/-6.00	U	6.00	P	11/19/2024	12:48	LB133524
	Cadmium	6.00	+/-6.00	U	6.00	P	11/19/2024	12:48	LB133524
	Calcium	2000	+/-2000	U	2000	P	11/19/2024	12:48	LB133524
	Chromium	10.0	+/-10.0	U	10.0	P	11/19/2024	12:48	LB133524
	Cobalt	30.0	+/-30.0	U	30.0	P	11/19/2024	12:48	LB133524
	Copper	20.0	+/-20.0	U	20.0	P	11/19/2024	12:48	LB133524
	Iron	100	+/-100	U	100	P	11/19/2024	12:48	LB133524
	Lead	12.0	+/-12.0	U	12.0	P	11/19/2024	12:48	LB133524
	Magnesium	2000	+/-2000	U	2000	P	11/19/2024	12:48	LB133524
	Manganese	20.0	+/-20.0	U	20.0	P	11/19/2024	12:48	LB133524
	Nickel	40.0	+/-40.0	U	40.0	P	11/19/2024	12:48	LB133524
	Potassium	2000	+/-2000	U	2000	P	11/19/2024	12:48	LB133524
	Selenium	20.0	+/-20.0	U	20.0	P	11/19/2024	12:48	LB133524
	Silver	10.0	+/-10.0	U	10.0	P	11/19/2024	12:48	LB133524
	Sodium	2000	+/-2000	U	2000	P	11/19/2024	12:48	LB133524
	Thallium	40.0	+/-40.0	U	40.0	P	11/19/2024	12:48	LB133524
	Vanadium	40.0	+/-40.0	U	40.0	P	11/19/2024	12:48	LB133524
	Zinc	40.0	+/-40.0	U	40.0	P	11/19/2024	12:48	LB133524
CCB02	Aluminum	100	+/-100	U	100	P	11/19/2024	13:38	LB133524
	Antimony	50.0	+/-50.0	U	50.0	P	11/19/2024	13:38	LB133524
	Arsenic	20.0	+/-20.0	U	20.0	P	11/19/2024	13:38	LB133524
	Barium	100	+/-100	U	100	P	11/19/2024	13:38	LB133524
	Beryllium	6.00	+/-6.00	U	6.00	P	11/19/2024	13:38	LB133524
	Cadmium	6.00	+/-6.00	U	6.00	P	11/19/2024	13:38	LB133524
	Calcium	2000	+/-2000	U	2000	P	11/19/2024	13:38	LB133524
	Chromium	10.0	+/-10.0	U	10.0	P	11/19/2024	13:38	LB133524
	Cobalt	30.0	+/-30.0	U	30.0	P	11/19/2024	13:38	LB133524
	Copper	20.0	+/-20.0	U	20.0	P	11/19/2024	13:38	LB133524
	Iron	100	+/-100	U	100	P	11/19/2024	13:38	LB133524
	Lead	12.0	+/-12.0	U	12.0	P	11/19/2024	13:38	LB133524
	Magnesium	2000	+/-2000	U	2000	P	11/19/2024	13:38	LB133524
	Manganese	20.0	+/-20.0	U	20.0	P	11/19/2024	13:38	LB133524
	Nickel	40.0	+/-40.0	U	40.0	P	11/19/2024	13:38	LB133524
	Potassium	2000	+/-2000	U	2000	P	11/19/2024	13:38	LB133524
	Selenium	20.0	+/-20.0	U	20.0	P	11/19/2024	13:38	LB133524

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	11/19/2024	13:38	LB133524
	Sodium	2000	+/-2000	U	2000	P	11/19/2024	13:38	LB133524
	Thallium	40.0	+/-40.0	U	40.0	P	11/19/2024	13:38	LB133524
	Vanadium	40.0	+/-40.0	U	40.0	P	11/19/2024	13:38	LB133524
	Zinc	40.0	+/-40.0	U	40.0	P	11/19/2024	13:38	LB133524
CCB03	Aluminum	100	+/-100	U	100	P	11/19/2024	14:28	LB133524
	Antimony	50.0	+/-50.0	U	50.0	P	11/19/2024	14:28	LB133524
	Arsenic	20.0	+/-20.0	U	20.0	P	11/19/2024	14:28	LB133524
	Barium	100	+/-100	U	100	P	11/19/2024	14:28	LB133524
	Beryllium	6.00	+/-6.00	U	6.00	P	11/19/2024	14:28	LB133524
	Cadmium	6.00	+/-6.00	U	6.00	P	11/19/2024	14:28	LB133524
	Calcium	2000	+/-2000	U	2000	P	11/19/2024	14:28	LB133524
	Chromium	10.0	+/-10.0	U	10.0	P	11/19/2024	14:28	LB133524
	Cobalt	30.0	+/-30.0	U	30.0	P	11/19/2024	14:28	LB133524
	Copper	20.0	+/-20.0	U	20.0	P	11/19/2024	14:28	LB133524
	Iron	100	+/-100	U	100	P	11/19/2024	14:28	LB133524
	Lead	12.0	+/-12.0	U	12.0	P	11/19/2024	14:28	LB133524
	Magnesium	2000	+/-2000	U	2000	P	11/19/2024	14:28	LB133524
	Manganese	20.0	+/-20.0	U	20.0	P	11/19/2024	14:28	LB133524
	Nickel	40.0	+/-40.0	U	40.0	P	11/19/2024	14:28	LB133524
	Potassium	2000	+/-2000	U	2000	P	11/19/2024	14:28	LB133524
	Selenium	20.0	+/-20.0	U	20.0	P	11/19/2024	14:28	LB133524
	Silver	10.0	+/-10.0	U	10.0	P	11/19/2024	14:28	LB133524
	Sodium	2000	+/-2000	U	2000	P	11/19/2024	14:28	LB133524
	Thallium	40.0	+/-40.0	U	40.0	P	11/19/2024	14:28	LB133524
	Vanadium	40.0	+/-40.0	U	40.0	P	11/19/2024	14:28	LB133524
	Zinc	40.0	+/-40.0	U	40.0	P	11/19/2024	14:28	LB133524
CCB04	Aluminum	100	+/-100	U	100	P	11/19/2024	15:36	LB133524
	Antimony	50.0	+/-50.0	U	50.0	P	11/19/2024	15:36	LB133524
	Arsenic	20.0	+/-20.0	U	20.0	P	11/19/2024	15:36	LB133524
	Barium	100	+/-100	U	100	P	11/19/2024	15:36	LB133524
	Beryllium	6.00	+/-6.00	U	6.00	P	11/19/2024	15:36	LB133524
	Cadmium	6.00	+/-6.00	U	6.00	P	11/19/2024	15:36	LB133524
	Calcium	2000	+/-2000	U	2000	P	11/19/2024	15:36	LB133524
	Chromium	10.0	+/-10.0	U	10.0	P	11/19/2024	15:36	LB133524
	Cobalt	30.0	+/-30.0	U	30.0	P	11/19/2024	15:36	LB133524
	Copper	20.0	+/-20.0	U	20.0	P	11/19/2024	15:36	LB133524
	Iron	100	+/-100	U	100	P	11/19/2024	15:36	LB133524
	Lead	12.0	+/-12.0	U	12.0	P	11/19/2024	15:36	LB133524

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4718							
Contract: PORT06	Lab Code: CHEM	Case No.: P4718	SAS No.: P4718						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	11/19/2024	15:36	LB133524
	Manganese	20.0	+/-20.0	U	20.0	P	11/19/2024	15:36	LB133524
	Nickel	40.0	+/-40.0	U	40.0	P	11/19/2024	15:36	LB133524
	Potassium	2000	+/-2000	U	2000	P	11/19/2024	15:36	LB133524
	Selenium	20.0	+/-20.0	U	20.0	P	11/19/2024	15:36	LB133524
	Silver	10.0	+/-10.0	U	10.0	P	11/19/2024	15:36	LB133524
	Sodium	2000	+/-2000	U	2000	P	11/19/2024	15:36	LB133524
	Thallium	40.0	+/-40.0	U	40.0	P	11/19/2024	15:36	LB133524
	Vanadium	40.0	+/-40.0	U	40.0	P	11/19/2024	15:36	LB133524
	Zinc	40.0	+/-40.0	U	40.0	P	11/19/2024	15:36	LB133524
CCB05	Aluminum	100	+/-100	U	100	P	11/19/2024	16:22	LB133524
	Antimony	50.0	+/-50.0	U	50.0	P	11/19/2024	16:22	LB133524
	Arsenic	20.0	+/-20.0	U	20.0	P	11/19/2024	16:22	LB133524
	Barium	100	+/-100	U	100	P	11/19/2024	16:22	LB133524
	Beryllium	6.00	+/-6.00	U	6.00	P	11/19/2024	16:22	LB133524
	Cadmium	6.00	+/-6.00	U	6.00	P	11/19/2024	16:22	LB133524
	Calcium	2000	+/-2000	U	2000	P	11/19/2024	16:22	LB133524
	Chromium	10.0	+/-10.0	U	10.0	P	11/19/2024	16:22	LB133524
	Cobalt	30.0	+/-30.0	U	30.0	P	11/19/2024	16:22	LB133524
	Copper	20.0	+/-20.0	U	20.0	P	11/19/2024	16:22	LB133524
	Iron	100	+/-100	U	100	P	11/19/2024	16:22	LB133524
	Lead	12.0	+/-12.0	U	12.0	P	11/19/2024	16:22	LB133524
	Magnesium	2000	+/-2000	U	2000	P	11/19/2024	16:22	LB133524
	Manganese	20.0	+/-20.0	U	20.0	P	11/19/2024	16:22	LB133524
	Nickel	40.0	+/-40.0	U	40.0	P	11/19/2024	16:22	LB133524
	Potassium	2000	+/-2000	U	2000	P	11/19/2024	16:22	LB133524
	Selenium	20.0	+/-20.0	U	20.0	P	11/19/2024	16:22	LB133524
	Silver	10.0	+/-10.0	U	10.0	P	11/19/2024	16:22	LB133524
	Sodium	2000	+/-2000	U	2000	P	11/19/2024	16:22	LB133524
	Thallium	40.0	+/-40.0	U	40.0	P	11/19/2024	16:22	LB133524
	Vanadium	40.0	+/-40.0	U	40.0	P	11/19/2024	16:22	LB133524
	Zinc	40.0	+/-40.0	U	40.0	P	11/19/2024	16:22	LB133524
CCB06	Aluminum	100	+/-100	U	100	P	11/19/2024	17:24	LB133524
	Antimony	50.0	+/-50.0	U	50.0	P	11/19/2024	17:24	LB133524
	Arsenic	20.0	+/-20.0	U	20.0	P	11/19/2024	17:24	LB133524
	Barium	22.0	+/-100	J	100	P	11/19/2024	17:24	LB133524
	Beryllium	6.00	+/-6.00	U	6.00	P	11/19/2024	17:24	LB133524
	Cadmium	6.00	+/-6.00	U	6.00	P	11/19/2024	17:24	LB133524
	Calcium	2000	+/-2000	U	2000	P	11/19/2024	17:24	LB133524

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: P4718

Contract: PORT06

Lab Code: CHEM

Case No.: P4718

SAS No.: P4718

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	11/19/2024	17:24	LB133524
	Cobalt	30.0	+/-30.0	U	30.0	P	11/19/2024	17:24	LB133524
	Copper	20.0	+/-20.0	U	20.0	P	11/19/2024	17:24	LB133524
	Iron	100	+/-100	U	100	P	11/19/2024	17:24	LB133524
	Lead	12.0	+/-12.0	U	12.0	P	11/19/2024	17:24	LB133524
	Magnesium	2000	+/-2000	U	2000	P	11/19/2024	17:24	LB133524
	Manganese	6.51	+/-20.0	J	20.0	P	11/19/2024	17:24	LB133524
	Nickel	40.0	+/-40.0	U	40.0	P	11/19/2024	17:24	LB133524
	Potassium	2000	+/-2000	U	2000	P	11/19/2024	17:24	LB133524
	Selenium	20.0	+/-20.0	U	20.0	P	11/19/2024	17:24	LB133524
	Silver	10.0	+/-10.0	U	10.0	P	11/19/2024	17:24	LB133524
	Sodium	2000	+/-2000	U	2000	P	11/19/2024	17:24	LB133524
	Thallium	40.0	+/-40.0	U	40.0	P	11/19/2024	17:24	LB133524
	Vanadium	40.0	+/-40.0	U	40.0	P	11/19/2024	17:24	LB133524
	Zinc	40.0	+/-40.0	U	40.0	P	11/19/2024	17:24	LB133524
CCB07	Aluminum	100	+/-100	U	100	P	11/19/2024	18:05	LB133524
	Antimony	50.0	+/-50.0	U	50.0	P	11/19/2024	18:05	LB133524
	Arsenic	20.0	+/-20.0	U	20.0	P	11/19/2024	18:05	LB133524
	Barium	25.8	+/-100	J	100	P	11/19/2024	18:05	LB133524
	Beryllium	6.00	+/-6.00	U	6.00	P	11/19/2024	18:05	LB133524
	Cadmium	0.62	+/-6.00	J	6.00	P	11/19/2024	18:05	LB133524
	Calcium	81.8	+/-2000	J	2000	P	11/19/2024	18:05	LB133524
	Chromium	10.0	+/-10.0	U	10.0	P	11/19/2024	18:05	LB133524
	Cobalt	30.0	+/-30.0	U	30.0	P	11/19/2024	18:05	LB133524
	Copper	20.0	+/-20.0	U	20.0	P	11/19/2024	18:05	LB133524
	Iron	100	+/-100	U	100	P	11/19/2024	18:05	LB133524
	Lead	12.0	+/-12.0	U	12.0	P	11/19/2024	18:05	LB133524
	Magnesium	2000	+/-2000	U	2000	P	11/19/2024	18:05	LB133524
	Manganese	7.68	+/-20.0	J	20.0	P	11/19/2024	18:05	LB133524
	Nickel	40.0	+/-40.0	U	40.0	P	11/19/2024	18:05	LB133524
	Potassium	2000	+/-2000	U	2000	P	11/19/2024	18:05	LB133524
	Selenium	20.0	+/-20.0	U	20.0	P	11/19/2024	18:05	LB133524
	Silver	10.0	+/-10.0	U	10.0	P	11/19/2024	18:05	LB133524
	Sodium	2000	+/-2000	U	2000	P	11/19/2024	18:05	LB133524
	Thallium	40.0	+/-40.0	U	40.0	P	11/19/2024	18:05	LB133524
	Vanadium	6.34	+/-40.0	J	40.0	P	11/19/2024	18:05	LB133524
	Zinc	40.0	+/-40.0	U	40.0	P	11/19/2024	18:05	LB133524