


LB133352

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000
B= 2.6139e-004
C= 4.2402e-002
Rho= 0.9998343
Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.067	0.067	93	0.000	93	0				1
0.2	0.200	0.203	0.003	614	0.0 %	614					2
2.5	2.500	2.382	-0.118	8950	0.0 %	8950					5
5.0	5.000	5.070	0.070	19234	0.0 %	19234					1
7.5	7.500	7.447	-0.053	28329	0.0 %	28329					1
10.0	10.000	10.031	0.031	38215	0.0 %	38215					0

LB133352 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 Nov 2024 09:40:21

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	93	-	0.0000	7470A	PPB	08 Nov 2024 10:04:53	S	Std
0.2 - 1	50.2	614	-	0.2000	7470A	PPB	08 Nov 2024 10:07:10	S	Std
2.5 - 1	52.7	8950	-	2.5000	7470A	PPB	08 Nov 2024 10:09:27	S	Std
5.0 - 1	55	19234	-	5.0000	7470A	PPB	08 Nov 2024 10:11:43	S	Std
7.5 - 1	57.5	28329	-	7.5000	7470A	PPB	08 Nov 2024 10:13:59	S	Std
10.0 - 1	510	38215	-	10.0000	7470A	PPB	08 Nov 2024 10:16:17	S	Std
ICV75 - 1	ICV75	13850	3.6626	-	7470A	PPB	08 Nov 2024 10:19:04	U	SMPL
ICB75 - 1	ICB75	-223	-0.0159	-	7470A	PPB	08 Nov 2024 10:21:18	U	SMPL
CCV51 - 1	CCV51	19260	5.0767	-	7470A	PPB	08 Nov 2024 10:23:35	U	SMPL
CCB51 - 1	CCB51	-209	-0.0122	-	7470A	PPB	08 Nov 2024 10:25:49	U	SMPL
CRA - 1	CRA	564	0.1898	-	7470A	PPB	08 Nov 2024 10:28:06	U	SMPL
HighStd - 1	HighStd	36747	9.6476	-	7470A	PPB	08 Nov 2024 10:30:21	U	SMPL
ChkStd - 1	ChkStd	25382	6.6770	-	7470A	PPB	08 Nov 2024 10:35:46	U	SMPL
PB164791BL - 1	PBW	-362	-0.0522	-	7470A	PPB	08 Nov 2024 10:38:01	U	SMPL
PB164791BS - 1	LCSW	15875	4.1919	-	7470A	PPB	08 Nov 2024 10:40:19	U	SMPL
P4718-03 - 1	WB-307-SB02	-255	-0.0243	-	7470A	PPB	08 Nov 2024 10:42:34	U	SMPL
P4718-03DUP - 1	WB-307-SB02DUP	-92	0.0184	-	7470A	PPB	08 Nov 2024 10:44:52	U	SMPL
P4718-03MS - 1	WB-307-SB02MS	12687	3.3586	-	7470A	PPB	08 Nov 2024 10:47:09	U	SMPL
P4718-03MSD - 1	WB-307-SB02MSD	15411	4.0707	-	7470A	PPB	08 Nov 2024 10:49:26	U	SMPL
P4719-03 - 1	BAYAVE-STOCKPILE	-195	-0.0086	-	7470A	PPB	08 Nov 2024 10:51:42	U	SMPL
CCV52 - 1	CCV52	18837	4.9662	-	7470A	PPB	08 Nov 2024 10:53:59	U	SMPL
CCB52 - 1	CCB52	-279	-0.0305	-	7470A	PPB	08 Nov 2024 10:56:15	U	SMPL
P4720-05 - 1	JC-701-COMP-01	-108	0.0142	-	7470A	PPB	08 Nov 2024 10:58:31	U	SMPL
P4722-04 - 1	WC-1(0-6)	-61	0.0265	-	7470A	PPB	08 Nov 2024 11:00:46	U	SMPL
P4722-09 - 1	WC-2(0-6)	-98	0.0168	-	7470A	PPB	08 Nov 2024 11:03:02	U	SMPL
P4722-14 - 1	WC-3(0-6)	-42	0.0314	-	7470A	PPB	08 Nov 2024 11:05:18	U	SMPL
P4732-02 - 1	PPE-COMP	94	0.0670	-	7470A	PPB	08 Nov 2024 11:07:33	U	SMPL
P4738-02 - 1	72-12018	-10	0.0398	-	7470A	PPB	08 Nov 2024 11:09:50	U	SMPL
P4739-04 - 1	TP-14	65	0.0594	-	7470A	PPB	08 Nov 2024 11:12:06	U	SMPL
P4739-08 - 1	BP-G2	157	0.0834	-	7470A	PPB	08 Nov 2024 11:14:23	U	SMPL
P4739-12 - 1	BP-B2	-254	-0.0240	-	7470A	PPB	08 Nov 2024 11:16:40	U	SMPL
P4739-16 - 1	TP-11	9	0.0448	-	7470A	PPB	08 Nov 2024 11:18:57	U	SMPL
CCV53 - 1	CCV53	17701	4.6692	-	7470A	PPB	08 Nov 2024 11:21:12	U	SMPL
CCB53 - 1	CCB53	-253	-0.0237	-	7470A	PPB	08 Nov 2024 11:23:27	U	SMPL
PB164694TB - 1	PB164694TB	-114	0.0126	-	7470A	PPB	08 Nov 2024 11:25:44	U	SMPL
P4718-03LX5 - 1		-90	0.0189	-	7470A	PPB	08 Nov 2024 11:28:00	U	SMPL
P4718-03A - 1		14079	3.7225	-	7470A	PPB	08 Nov 2024 11:30:15	U	SMPL
CCV54 - 1	CCV54	17291	4.5621	-	7470A	PPB	08 Nov 2024 11:32:30	U	SMPL
CCB54 - 1	CCB54	-218	-0.0146	-	7470A	PPB	08 Nov 2024 11:34:48	U	SMPL