



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.19	0.2	95	40 - 160	CV	11/08/2024	10:28	LB133352
CRI01	Arsenic	20.1	20.0	100	40 - 160	P	11/08/2024	14:20	LB133365
	Barium	95.0	100	95	40 - 160	P	11/08/2024	14:20	LB133365
	Cadmium	6.16	6.0	103	40 - 160	P	11/08/2024	14:20	LB133365
	Chromium	9.24	10.0	92	40 - 160	P	11/08/2024	14:20	LB133365
	Lead	11.8	12.0	98	40 - 160	P	11/08/2024	14:20	LB133365
	Selenium	19.8	20.0	99	40 - 160	P	11/08/2024	14:20	LB133365
	Silver	10.3	10.0	103	40 - 160	P	11/08/2024	14:20	LB133365
CRI01	Arsenic	21.8	20.0	109	40 - 160	P	11/19/2024	20:19	LB133525
	Barium	99.9	100	100	40 - 160	P	11/19/2024	20:19	LB133525
	Cadmium	7.16	6.0	119	40 - 160	P	11/19/2024	20:19	LB133525
	Chromium	10.5	10.0	105	40 - 160	P	11/19/2024	20:19	LB133525
	Lead	11.3	12.0	94	40 - 160	P	11/19/2024	20:19	LB133525
	Selenium	17.2	20.0	86	40 - 160	P	11/19/2024	20:19	LB133525
	Silver	10.3	10.0	103	40 - 160	P	11/19/2024	20:19	LB133525