

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Portal Partners Tri-Venture **SDG No.:** P4718
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4718 **SAS No.:** P4718

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164768BS							
Arsenic	ug/L	4000	3890		97	80 - 120	P
Barium	ug/L	1000	956		96	80 - 120	P
Cadmium	ug/L	1000	901		90	80 - 120	P
Chromium	ug/L	2000	1800		90	80 - 120	P
Lead	ug/L	5000	4650		93	80 - 120	P
Selenium	ug/L	10000	9760		98	80 - 120	P
Silver	ug/L	380	363		96	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: P4718

Contract: PORT06

Lab Code: CHEM

Case No.: P4718

SAS No.: P4718

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164791BS Mercury	ug/L	4.0	4.19		105	80 - 120	CV