

Continuing Calibration Report for SequenceID : FC111124AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067701.D

Aliphatic C9-C12	8328348.000	60.000	3.170	6.450	138805.800	135469.641	-2.463
Aliphatic C12-C16	5483001.000	40.000	6.451	9.840	137075.025	131681.730	-4.096
Aliphatic C16-C21	7943258.000	60.000	9.841	13.198	132387.633	127235.092	-4.050
Aliphatic C21-C28	10046123.000	80.000	13.199	16.853	125576.538	120090.398	-4.568
Aliphatic C28-C40	13170565.000	120.000	16.854	21.695	109754.708	101093.195	-8.568
Aliphatic EPH	44971295.000	360.000	3.170	21.695	124920.264	118799.912	-5.152