

Continuing Calibration Report for SequenceID : FD110824AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048686.D

Aromatic C10-C12	7558477.000	40.000	4.092	5.810	188961.925	176202.694	-7.241
Aromatic C12-C16	11312555.000	60.000	5.811	8.415	188542.583	177273.493	-6.357
Aromatic C16-C21	13328267.000	80.000	8.416	12.678	166603.338	163832.550	-1.691
Aromatic C21-C36	30539691.000	180.000	12.679	18.083	169664.950	152713.153	-11.100
Aromatic EPH	62738990.000	360.000	4.092	18.083	174274.972	161887.469	-7.652