

Continuing Calibration Report for SequenceID : FD111124AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048705.D

Aromatic C10-C12	7212813.000	40.000	4.090	5.808	180320.325	176202.694	-2.337
Aromatic C12-C16	10865815.000	60.000	5.809	8.414	181096.917	177273.493	-2.157
Aromatic C16-C21	13319448.000	80.000	8.415	12.677	166493.100	163832.550	-1.624
Aromatic C21-C36	29488863.000	180.000	12.678	18.084	163827.017	152713.153	-7.278
Aromatic EPH	60886939.000	360.000	4.090	18.084	169130.386	161887.469	-4.474