

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** P4718 **SAS No :** P4718 **SDG No:** P4718
Sample No : PB164790BSD **Datafile:** FC067672.D
Client ID : PB164790BSD

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	RPD QC LIMITS	QC Limit Of RPD
Aliphatic C9-C12	10	5.31	53		0.66 (40-140)	50
Aliphatic C12-C16	6.7	4.55	68		0.76 (40-140)	50
Aliphatic C16-C21	10	8.11	81		0.84 (40-140)	50
Aliphatic C21-C28	13.3	11.1	83		1 (40-140)	50
Aliphatic C28-C40	20.0	18.9	95		0.09 (40-140)	50