

**WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY**

**Lab Name:** Chemtech **Client:** Portal Partners Tri-Venture  
**Lab Code:** CHEM **Cas No:** P4718 **SAS No :** P4718 **SDG No:** P4718  
**Sample No :** PB164847BS **Datafile:** FC067695.D  
**Client ID :** PB164847BS

| COMPOUND          | SPIKE<br>ADDED<br>ug/l | LCS/LCSD<br>CONCENTRATION<br>ug/l | % REC | Qual | QC LIMITS |
|-------------------|------------------------|-----------------------------------|-------|------|-----------|
| Aliphatic C9-C12  | 300                    | 164                               | 55    |      | (40-140)  |
| Aliphatic C12-C16 | 200                    | 143                               | 72    |      | (40-140)  |
| Aliphatic C16-C21 | 300                    | 256                               | 85    |      | (40-140)  |
| Aliphatic C21-C28 | 400                    | 360                               | 90    |      | (40-140)  |
| Aliphatic C28-C40 | 600                    | 602                               | 100   |      | (40-140)  |