

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** P4718 **SAS No :** P4718 **SDG No:** P4718
Sample No : PB164847BSD **Datafile:** FC067696.D
Client ID : PB164847BSD

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	RPD QC LIMITS	QC Limit Of RPD
Aliphatic C9-C12	300	179	60		8.7 (40-140)	50
Aliphatic C12-C16	200	155	78		8.1 (40-140)	50
Aliphatic C16-C21	300	279	93		8.6 (40-140)	50
Aliphatic C21-C28	400	395	99		9.3 (40-140)	50
Aliphatic C28-C40	600	660	110		9.5 (40-140)	50