

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB164700BS

Contract: PORT06

Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718

Lab Sample ID: PB164700BS Date(s) Analyzed: 11/06/2024 11/06/2024

Instrument ID (1): ECD_O Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO107732.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.524	5.474	5.574	166	163	3.12
COLUMN 1	2	5.546	5.496	5.596	160		
	3	5.609	5.559	5.659	159		
	4	5.706	5.656	5.756	167		
	5	6.001	5.951	6.051	163		
	1	4.724	4.674	4.774	162		
COLUMN 2	2	4.743	4.693	4.793	168	158	
	3	4.919	4.869	4.969	161		
	4	4.96	4.91	5.01	146		
	5	5.174	5.124	5.224	155		
Aroclor-1260	1	7.131	7.081	7.181	171	168	5.5
COLUMN 1	2	7.388	7.338	7.438	178		
	3	7.75	7.7	7.8	157		
	4	7.975	7.925	8.025	174		
	5	8.29	8.24	8.34	160		
	1	6.205	6.155	6.255	162	159	
COLUMN 2	2	6.393	6.343	6.443	170		
	3	6.545	6.495	6.595	166		
	4	7.016	6.966	7.066	144		
	5	7.258	7.208	7.308	154		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

TR-04-110424MS

Contract: PORT06

Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718

Lab Sample ID: P4706-01MS Date(s) Analyzed: 11/06/2024 11/06/2024

Instrument ID (1): ECD_O Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO107736.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.523	5.473	5.573	145	139	2.84
COLUMN 1	2	5.546	5.496	5.596	133		
	3	5.608	5.558	5.658	136		
	4	5.705	5.655	5.755	148		
	5	5.999	5.949	6.049	131		
COLUMN 2	1	4.723	4.673	4.773	150	143	
	2	4.742	4.692	4.792	152		
	3	4.918	4.868	4.968	144		
	4	4.96	4.91	5.01	132		
	5	5.172	5.122	5.222	134		
Aroclor-1260	1	7.13	7.08	7.18	139	142	15.15
COLUMN 1	2	7.387	7.337	7.437	155		
	3	7.749	7.699	7.799	123		
	4	7.973	7.923	8.023	150		
	5	8.288	8.238	8.338	144		
COLUMN 2	1	6.204	6.154	6.254	128	122	
	2	6.391	6.341	6.441	136		
	3	6.544	6.494	6.594	128		
	4	7.016	6.966	7.066	106		
	5	7.257	7.207	7.307	113		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

TR-04-110424MSD

Contract: PORT06

Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718

Lab Sample ID: P4706-01MSD Date(s) Analyzed: 11/06/2024 11/06/2024

Instrument ID (1): ECD_O Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO107737.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.523	5.473	5.573	136	131	5.93
COLUMN 1	2	5.545	5.495	5.595	126		
	3	5.608	5.558	5.658	128		
	4	5.705	5.655	5.755	139		
	5	6	5.95	6.05	127		
	1	4.724	4.674	4.774	146		
COLUMN 2	2	4.743	4.693	4.793	149	139	
	3	4.919	4.869	4.969	141		
	4	4.96	4.91	5.01	129		
	5	5.173	5.123	5.223	131		
Aroclor-1260	1	7.13	7.08	7.18	131	135	14.29
COLUMN 1	2	7.387	7.337	7.437	146		
	3	7.749	7.699	7.799	120		
	4	7.974	7.924	8.024	140		
	5	8.289	8.239	8.339	139		
	1	6.204	6.154	6.254	123	117	
COLUMN 2	2	6.391	6.341	6.441	130		
	3	6.545	6.495	6.595	123		
	4	7.016	6.966	7.066	101		
	5	7.257	7.207	7.307	108		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB164704BS

Contract: PORT06

Lab Code: CHEM

Case No.: P4718

SAS No.: P4718

SDG NO.: P4718

Lab Sample ID: PB164704BS

Date(s) Analyzed: 11/06/2024

11/06/2024

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO107747.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.523	5.473	5.573	4.34	4.20	0
	2	5.545	5.495	5.595	4.25		
	3	5.608	5.558	5.658	4.18		
	4	5.705	5.655	5.755	4.34		
	5	6	5.95	6.05	4.00		
COLUMN 1	1	4.723	4.673	4.773	4.43	4.20	
	2	4.742	4.692	4.792	4.53		
	3	4.918	4.868	4.968	4.26		
	4	4.959	4.909	5.009	3.86		
	5	5.173	5.123	5.223	4.08		
COLUMN 2	1	7.129	7.079	7.179	4.04	4.20	2.35
	2	7.386	7.336	7.436	4.48		
	3	7.748	7.698	7.798	3.76		
	4	7.974	7.924	8.024	4.65		
	5	8.289	8.239	8.339	3.93		
Aroclor-1260	1	6.204	6.154	6.254	4.39	4.30	
	2	6.391	6.341	6.441	4.63		
	3	6.545	6.495	6.595	4.51		
	4	7.015	6.965	7.065	3.90		
	5	7.256	7.206	7.306	4.15		
COLUMN 1	1						
	2						
	3						
	4						
	5						
COLUMN 2	1						
	2						
	3						
	4						
	5						

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB164704BSD

Contract: PORT06

Lab Code: CHEM

Case No.: P4718

SAS No.: P4718

SDG NO.: P4718

Lab Sample ID: PB164704BSD

Date(s) Analyzed: 11/06/2024

11/06/2024

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO107748.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.522	5.472	5.572	4.50	4.40	2.25
COLUMN 1	2	5.545	5.495	5.595	4.39		
	3	5.607	5.557	5.657	4.39		
	4	5.705	5.655	5.755	4.53		
	5	6	5.95	6.05	4.32		
	1	4.723	4.673	4.773	4.61		
COLUMN 2	2	4.742	4.692	4.792	4.70	4.50	
	3	4.918	4.868	4.968	4.52		
	4	4.96	4.91	5.01	4.10		
	5	5.173	5.123	5.223	4.33		
Aroclor-1260	1	7.129	7.079	7.179	4.10	4.20	6.9
COLUMN 1	2	7.386	7.336	7.436	4.66		
	3	7.749	7.699	7.799	3.99		
	4	7.974	7.924	8.024	4.34		
	5	8.289	8.239	8.339	4.11		
	1	6.204	6.154	6.254	4.53		
COLUMN 2	2	6.391	6.341	6.441	4.80	4.50	
	3	6.544	6.494	6.594	4.68		
	4	7.015	6.965	7.065	4.06		
	5	7.257	7.207	7.307	4.32		