

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718

Continuing Calib Date: 11/06/2024 Initial Calibration Date(s): 10/15/2024 10/16/2024

Continuing Calib Time: 22:14 Initial Calibration Time(s): 18:27 02:36

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	5.52	5.52	5.42	5.62	0.00
Aroclor-1248-2 (2)	5.80	5.80	5.70	5.90	0.00
Aroclor-1248-3 (3)	6.00	6.00	5.90	6.10	0.00
Aroclor-1248-4 (4)	6.40	6.40	6.30	6.50	0.00
Aroclor-1248-5 (5)	6.44	6.44	6.34	6.54	0.00
Tetrachloro-m-xylene	4.37	4.37	4.27	4.47	0.00
Decachlorobiphenyl	10.08	10.06	9.96	10.16	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718

Continuing Calib Date: 11/06/2024 Initial Calibration Date(s): 10/15/2024 10/16/2024

Continuing Calib Time: 22:14 Initial Calibration Time(s): 18:27 02:36

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.72	4.73	4.63	4.83	0.01
Aroclor-1248-2 (2)	4.96	4.96	4.86	5.06	0.00
Aroclor-1248-3 (3)	5.00	5.00	4.90	5.10	0.00
Aroclor-1248-4 (4)	5.17	5.18	5.08	5.28	0.01
Aroclor-1248-5 (5)	5.56	5.57	5.47	5.67	0.01
Tetrachloro-m-xylene	3.64	3.64	3.54	3.74	0.00
Decachlorobiphenyl	8.64	8.64	8.54	8.74	0.00